

Data Path : Z:\HPCHEM1\BNA_G\DATA\BG051817\
 Data File : BG027044.D
 Acq On : 19 May 2017 17:41
 Operator : SJ/MA
 Sample : I3137-09
 Misc :
 ALS Vial : 42 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 DAC38

Quant Time: May 19 18:21:17 2017
 Quant Method : Z:\HPCHEM1\BNA_G\METHODS\SOM-EPA-BG051817MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Fri May 19 17:37:52 2017
 Response via : Initial Calibration


Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	8.01	152	209877	20.00	ng/ul	0.00
2) Naphthalene-d8	10.81	136	943143	20.00	ng/ul	0.00
6) Acenaphthene-d10	14.62	164	462489	20.00	ng/ul	0.00
12) Phenanthrene-d10	17.36	188	833998	20.00	ng/ul	0.00
17) Chrysene-d12	21.61	240	757603	20.00	ng/ul	0.01
22) Perylene-d12	24.80	264	636988	20.00	ng/ul	0.04
System Monitoring Compounds						
7) Acenaphthylene-d8	14.32	160	1041837	21.96	ng/ul	0.00
10) Fluorene-d10	15.61	176	657871	20.79	ng/ul	0.00
14) Anthracene-d10	17.36	188	833998	20.51	ng/ul	-0.09
18) Pyrene-d10	19.74	212	879974	22.25	ng/ul	0.00
25) Benzo(a)pyrene-d12	24.58	264	730899	23.09	ng/ul	0.04
Target Compounds						
23) Benzo(b)fluoranthene	23.79	252	39341	1.04	ng/ul#	Qvalue 20


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA_G\DATA\BG051817\
Data File : BG027044.D
Acq On : 19 May 2017 17:41
Operator : SJ/MA
Sample : I3137-09
Misc :
ALS Vial : 42 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
DAC38

Quant Time: May 19 18:21:17 2017
Quant Method : Z:\HPCHEM1\BNA_G\METHODS\SOM-EPA-BG051817MA.M
Quant Title : SVOA CALIBRATION
QLast Update : Fri May 19 17:37:52 2017
Response via : Initial Calibration





#23

Benzo(b)fluoranthene

Concen: 1.04 ng/ul

RT: 23.79 min Scan# 3523

Delta R.T. 0.03 min

Lab File: BG027044.D


Acq: 19 May 2017 17:41

Instrument :

BNA_G

ClientSampleId :

DAC38



Tgt Ion:	252	Resp:	39341
Ion	Ratio	Lower	Upper
252	100		
253	31.0	18.0	27.0#
125	91.9	8.0	12.0#

