

Data Path : Z:\SVOASRV\HPCHEM1\BNA G\DATA\BG051820\
 Data File : BG045413.D
 Acq On : 18 May 2020 14:54
 Operator : CG/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_G
 LabSampleId :
 SSTDCCC040

Quant Time: May 18 16:21:49 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA G\METHODS\8270-BG051520.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon May 18 15:47:17 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	125	0.00
2	1,4-Dioxane	0.507	0.418	17.6	103	0.00
3	Pyridine	1.413	1.352	4.3	115	0.00
4	n-Nitrosodimethylamine	0.462	0.426	7.8	113	0.00
5 S	2-Fluorophenol	1.023	1.001	2.2	119	0.00
6	Aniline	1.901	1.931	-1.6	126	0.00
7 S	Phenol-d6	1.457	1.531	-5.1	128	0.00
8	2-Chlorophenol	1.122	1.123	-0.1	123	0.00
9	Benzaldehyde	0.949	0.934	1.6	118	0.00
10 C	Phenol	1.501	1.544	-2.9	124	0.00
11	bis(2-Chloroethyl)ether	1.171	1.187	-1.4	122	0.00
12	1,3-Dichlorobenzene	1.349	1.322	2.0	121	0.00
13 C	1,4-Dichlorobenzene	1.331	1.308	1.7	121	0.00
14	1,2-Dichlorobenzene	1.280	1.270	0.8	124	0.00
15	Benzyl Alcohol	1.203	1.289	-7.1	131	0.00
16	2,2'-oxybis(1-Chloropropane	1.111	1.159	-4.3	130	0.00
17	2-Methylphenol	1.124	1.202	-6.9	129	0.00
18	Hexachloroethane	0.510	0.489	4.1	117	0.00
19 P	n-Nitroso-di-n-propylamine	1.007	1.138	-13.0	138	0.00
20	3+4-Methylphenols	1.530	1.580	-3.3	125	0.00
21 I	Naphthalene-d8	1.000	1.000	0.0	131	0.00
22	Acetophenone	0.474	0.478	-0.8	131	0.00
23 S	Nitrobenzene-d5	0.367	0.363	1.1	128	0.00
24	Nitrobenzene	0.393	0.385	2.0	130	0.00
25	Isophorone	0.722	0.748	-3.6	133	0.00
26 C	2-Nitrophenol	0.168	0.174	-3.6	135	0.00
27	2,4-Dimethylphenol	0.249	0.260	-4.4	134	0.00
28	bis(2-Chloroethoxy)methane	0.379	0.390	-2.9	135	0.00
29 C	2,4-Dichlorophenol	0.294	0.301	-2.4	135	0.00
30	1,2,4-Trichlorobenzene	0.348	0.339	2.6	128	0.00
31	Naphthalene	0.932	0.940	-0.9	132	0.00
32	Benzoic acid	0.150	0.215	-43.3#	198#	0.00
33	4-Chloroaniline	0.390	0.419	-7.4	137	0.00
34 C	Hexachlorobutadiene	0.246	0.236	4.1	126	0.00
35	Caprolactam	0.108	0.114	-5.6	133	0.00
36 C	4-Chloro-3-methylphenol	0.332	0.358	-7.8	140	0.00
37	2-Methylnaphthalene	0.695	0.723	-4.0	135	0.00
38	1-Methylnaphthalene	0.656	0.676	-3.0	133	0.00
39 I	Acenaphthene-d10	1.000	1.000	0.0	137	0.00
40	1,2,4,5-Tetrachlorobenzene	0.559	0.546	2.3	135	0.00
41 P	Hexachlorocyclopentadiene	0.322	0.317	1.6	134	0.00
42 S	2,4,6-Tribromophenol	0.256	0.244	4.7	128	0.00
43 C	2,4,6-Trichlorophenol	0.388	0.385	0.8	134	0.00
44	2,4,5-Trichlorophenol	0.443	0.444	-0.2	135	0.00

Data Path : Z:\SVOASRV\HPCHEM1\BNA G\DATA\BG051820\
 Data File : BG045413.D
 Acq On : 18 May 2020 14:54
 Operator : CG/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_G
 LabSampleId :
 SSTDCCC040

Quant Time: May 18 16:21:49 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA G\METHODS\8270-BG051520.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon May 18 15:47:17 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
45 S	2-Fluorobiphenyl	1.179	1.167	1.0	132	0.00
46	1,1'-Biphenyl	1.229	1.232	-0.2	136	0.00
47	2-Chloronaphthalene	0.998	0.995	0.3	137	0.00
48	2-Nitroaniline	0.328	0.331	-0.9	133	0.00
49	Acenaphthylene	1.603	1.616	-0.8	137	0.00
50	Dimethylphthalate	1.372	1.364	0.6	132	0.00
51	2,6-Dinitrotoluene	0.310	0.313	-1.0	134	0.00
52 C	Acenaphthene	1.073	0.982	8.5	122	0.00
53	3-Nitroaniline	0.315	0.325	-3.2	140	0.00
54 P	2,4-Dinitrophenol	0.180	0.148	17.8	107	0.00
55	Dibenzofuran	1.593	1.590	0.2	133	0.00
56 P	4-Nitrophenol	0.238	0.226	5.0	122	0.00
57	2,4-Dinitrotoluene	0.448	0.442	1.3	133	0.00
58	Fluorene	1.309	1.332	-1.8	136	0.00
59	2,3,4,6-Tetrachlorophenol	0.390	0.359	7.9	123	0.00
60	Diethylphthalate	1.428	1.408	1.4	131	0.00
61	4-Chlorophenyl-phenylether	0.749	0.755	-0.8	135	0.00
62	4-Nitroaniline	0.329	0.332	-0.9	133	0.00
63	Azobenzene	1.332	1.323	0.7	133	0.00
64 I	Phenanthrene-d10	1.000	1.000	0.0	128	0.00
65	4,6-Dinitro-2-methylphenol	0.116	0.110	5.2	117	0.00
66 c	n-Nitrosodiphenylamine	0.480	0.496	-3.3	133	0.00
67	4-Bromophenyl-phenylether	0.217	0.225	-3.7	133	0.00
68	Hexachlorobenzene	0.227	0.234	-3.1	132	0.00
69	Atrazine	0.198	0.198	0.0	126	0.00
70 C	Pentachlorophenol	0.118	0.137	-16.1	145	0.00
71	Phenanthrene	0.873	0.886	-1.5	128	0.00
72	Anthracene	0.877	0.881	-0.5	127	0.00
73	Carbazole	0.855	0.853	0.2	125	0.00
74	Di-n-butylphthalate	1.026	1.004	2.1	120	0.00
75 C	Fluoranthene	1.111	1.069	3.8	118	0.00
76 I	Chrysene-d12	1.000	1.000	0.0	111	0.00
77	Benzidine	0.562	0.566	-0.7	107	0.00
78	Pyrene	1.140	1.227	-7.6	116	0.00
79 S	Terphenyl-d14	0.858	0.906	-5.6	112	0.00
80	Butylbenzylphthalate	0.492	0.507	-3.0	110	0.00
81	Benzo(a)anthracene	1.114	1.135	-1.9	112	0.00
82	3,3'-Dichlorobenzidine	0.437	0.439	-0.5	109	0.00
83	Chrysene	1.071	1.077	-0.6	108	0.00
84	Bis(2-ethylhexyl)phthalate	0.676	0.684	-1.2	110	0.00
85 c	Di-n-octyl phthalate	1.162	1.150	1.0	109	0.00
86	Indeno(1,2,3-cd)pyrene	1.401	1.325	5.4	106	0.00
87 I	Perylene-d12	1.000	1.000	0.0	108	0.00

Data Path : Z:\SVOASRV\HPCHEM1\BNA G\DATA\BG051820\
Data File : BG045413.D
Acq On : 18 May 2020 14:54
Operator : CG/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_G
LabSampleId :
SSTDCCC040

Quant Time: May 18 16:21:49 2020
Quant Method : Z:\SVOASRV\HPCHEM1\BNA G\METHODS\8270-BG051520.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon May 18 15:47:17 2020
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
88	Benzo(b)fluoranthene	1.073	1.076	-0.3	106	0.00
89	Benzo(k)fluoranthene	1.008	1.056	-4.8	112	0.00
90 C	Benzo(a)pyrene	0.999	1.008	-0.9	108	0.00
91	Dibenzo(a,h)anthracene	1.004	1.007	-0.3	109	0.00
92	Benzo(g,h,i)perylene	1.004	1.000	0.4	107	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0