

Data Path : Z:\HPCHEM1\BNA_G\DATA\BG051916\
 Data File : BG021957.D
 Acq On : 19 May 2016 8:58
 Operator : UM/SJ
 Sample : SSTDC0020
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_G
 LabSampleId :
 SSTD02003

Quant Time: May 20 05:44:15 2016
 Quant Method : Z:\HPCHEM1\BNA_G\METHODS\SOM02.2-EPA-BG051816.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu May 19 03:59:04 2016
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	133	0.00
2 S	1,4-Dioxane-d8	0.436	0.409	6.2	120	0.00
3 S	Phenol-d5	1.766	1.604	9.2	123	0.00
4 S	Bis-(2-Chloroethyl)ether-d8	0.941	0.898	4.6	123	0.00
5 S	2-Chlorophenol-d4	1.450	1.428	1.5	124	0.00
6 S	4-Methylphenol-d8	1.573	1.378	12.4	112	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	121	0.00
8 S	Nitrobenzene-d5	0.140	0.145	-3.6	119	0.00
9 S	2-Nitrophenol-d4	0.150	0.147	2.0	115	0.00
10 S	2,4-Dichlorophenol-d3	0.265	0.273	-3.0	121	0.00
11	Naphthalene	0.973	0.987	-1.4	120	0.00
12 S	4-Chloroaniline-d4	0.208	0.196	5.8	122	0.00
13	2-Methylnaphthalene	0.705	0.695	1.4	117	0.00
14	1-Methylnaphthalene	0.721	0.696	3.5	112	0.00
15 I	Acenaphthene-d10	1.000	1.000	0.0	113	0.00
16 S	Dimethylphthalate-d6	1.542	1.450	6.0	105	0.00
17 S	Acenaphthylene-d8	1.986	1.970	0.8	109	0.00
18	Acenaphthylene	2.068	2.068	0.0	110	0.00
19 C	Acenaphthene	1.420	1.421	-0.1	111	0.00
20 S	4-Nitrophenol-d4	0.271	0.223	17.7	96	0.00
21 S	Fluorene-d10	1.388	1.341	3.4	108	0.00
22	Fluorene	1.573	1.536	2.4	108	0.00
23 I	Phenanthrene-d10	1.000	1.000	0.0	109	0.00
24 S	4,6-Dinitro-2-methylphenol-	0.099	0.087	12.1	103	0.00
25	Phenanthrene	1.069	1.075	-0.6	108	0.00
26 S	Anthracene-d10	0.921	0.885	3.9	104	0.00
27	Anthracene	1.074	1.065	0.8	105	0.00
28 C	Fluoranthene	1.111	1.168	-5.1	112	0.00
29 I	Chrysene-d12	1.000	1.000	0.0	115	0.00
30 S	Pyrene-d10	1.055	1.037	1.7	111	0.00
31	Pyrene	1.377	1.324	3.8	108	0.00
32	Benzo(a)anthracene	1.109	1.149	-3.6	117	0.00
33	Chrysene	1.079	1.065	1.3	113	0.00
34 I	Perylene-d12	1.000	1.000	0.0	114	0.00
35	Benzo(b)fluoranthene	1.150	1.176	-2.3	116	0.00
36	Benzo(k)fluoranthene	1.131	1.167	-3.2	118	0.00
37 S	Benzo(a)pyrene-d12	0.888	0.906	-2.0	116	0.00
38 C	Benzo(a)pyrene	1.099	1.119	-1.8	115	0.00
39	Indeno(1,2,3-cd)pyrene	1.250	1.293	-3.4	119	-0.03
40	Dibenzo(a,h)anthracene	1.035	1.038	-0.3	113	-0.03
41	Benzo(g,h,i)perylene	1.057	1.070	-1.2	114	-0.02

Data Path : Z:\HPCHEM1\BNA_G\DATA\BG051916\
Data File : BG021957.D
Acq On : 19 May 2016 8:58
Operator : UM/SJ
Sample : SSTDCCC020
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_G
LabSampleId :
SSTD02003

Quant Time: May 20 05:44:15 2016
Quant Method : Z:\HPCHEM1\BNA_G\METHODS\SOM02.2-EPA-BG051816.M
Quant Title : SVOA CALIBRATION
QLast Update : Thu May 19 03:59:04 2016
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
Max. RRF Dev : 20% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)
----------	-------	------	-----------	------------

(#) = Out of Range

SPCC's out = 0 CCC's out = 0