

Data Path : Z:\HPCHEM1\BNA_G\DATA\BG051916\
 Data File : BG021970.D
 Acq On : 19 May 2016 19:44
 Operator : UM/SJ
 Sample : SSTDC0020
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_G
 LabSampleId :
 SSTD02004

Quant Time: May 20 07:11:18 2016
 Quant Method : Z:\HPCHEM1\BNA_G\METHODS\SOM02.2-EPA-BG051816.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Fri May 20 05:44:47 2016
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	172#	0.00
2 S	1,4-Dioxane-d8	0.436	0.374	14.2	141	0.00
3 S	Phenol-d5	1.766	1.719	2.7	170#	0.00
4 S	Bis-(2-Chloroethyl)ether-d8	0.941	0.895	4.9	159#	0.00
5 S	2-Chlorophenol-d4	1.450	1.445	0.3	163#	0.00
6 S	4-Methylphenol-d8	1.573	1.456	7.4	153#	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	159#	0.00
8 S	Nitrobenzene-d5	0.140	0.140	0.0	152#	0.00
9 S	2-Nitrophenol-d4	0.150	0.153	-2.0	158#	0.00
10 S	2,4-Dichlorophenol-d3	0.265	0.275	-3.8	160#	0.00
11	Naphthalene	0.973	0.983	-1.0	157#	0.00
12 S	4-Chloroaniline-d4	0.208	0.212	-1.9	174#	0.00
13	2-Methylnaphthalene	0.705	0.699	0.9	154#	0.00
14	1-Methylnaphthalene	0.721	0.694	3.7	147	0.00
15 I	Acenaphthene-d10	1.000	1.000	0.0	147	0.00
16 S	Dimethylphthalate-d6	1.542	1.415	8.2	133	0.00
17 S	Acenaphthylene-d8	1.986	1.959	1.4	140	0.00
18	Acenaphthylene	2.068	2.057	0.5	143	0.00
19 C	Acenaphthene	1.420	1.387	2.3	141	0.00
20 S	4-Nitrophenol-d4	0.271	0.249	8.1	140	0.00
21 S	Fluorene-d10	1.388	1.306	5.9	137	0.00
22	Fluorene	1.573	1.478	6.0	135	0.00
23 I	Phenanthrene-d10	1.000	1.000	0.0	133	0.00
24 S	4,6-Dinitro-2-methylphenol-	0.099	0.093	6.1	134	0.00
25	Phenanthrene	1.069	1.077	-0.7	132	0.00
26 S	Anthracene-d10	0.921	0.910	1.2	130	0.00
27	Anthracene	1.074	1.074	0.0	129	0.00
28 C	Fluoranthene	1.111	1.111	0.0	130	0.00
29 I	Chrysene-d12	1.000	1.000	0.0	131	0.00
30 S	Pyrene-d10	1.055	1.065	-0.9	130	0.00
31	Pyrene	1.377	1.355	1.6	126	0.00
32	Benzo(a)anthracene	1.109	1.127	-1.6	131	0.00
33	Chrysene	1.079	1.111	-3.0	135	0.00
34 I	Perylene-d12	1.000	1.000	0.0	135	0.00
35	Benzo(b)fluoranthene	1.150	1.171	-1.8	135	0.00
36	Benzo(k)fluoranthene	1.131	1.111	1.8	132	0.00
37 S	Benzo(a)pyrene-d12	0.888	0.902	-1.6	136	0.00
38 C	Benzo(a)pyrene	1.099	1.121	-2.0	136	0.00
39	Indeno(1,2,3-cd)pyrene	1.250	1.255	-0.4	136	0.00
40	Dibenzo(a,h)anthracene	1.035	1.012	2.2	130	0.00
41	Benzo(g,h,i)perylene	1.057	0.987	6.6	124	0.00

Data Path : Z:\HPCHEM1\BNA_G\DATA\BG051916\
Data File : BG021970.D
Acq On : 19 May 2016 19:44
Operator : UM/SJ
Sample : SSTDCCC020
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_G
LabSampleId :
SSTD02004

Quant Time: May 20 07:11:18 2016
Quant Method : Z:\HPCHEM1\BNA_G\METHODS\SOM02.2-EPA-BG051816.M
Quant Title : SVOA CALIBRATION
QLast Update : Fri May 20 05:44:47 2016
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
Max. RRF Dev : 20% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)
----------	-------	------	-----------	------------

(#) = Out of Range

SPCC's out = 0 CCC's out = 0