

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG051923\
 Data File : BG057480.D
 Acq On : 19 May 2023 23:41
 Operator : CG/JU
 Sample : 02849-01
 Misc :
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 TP7

Quant Time: May 20 04:31:25 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG042723.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Sat May 20 04:19:55 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.355	152	17396	20.000	ng	0.00
21) Naphthalene-d8	11.193	136	67451	20.000	ng	0.00
39) Acenaphthene-d10	14.970	164	52735	20.000	ng	0.00
64) Phenanthrene-d10	17.707	188	134205	20.000	ng	0.00
76) Chrysene-d12	22.013	240	127704	20.000	ng	0.00
86) Perylene-d12	25.509	264	134777	20.000	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.870	112	60713	59.701	ng	0.00
7) Phenol-d6	7.498	99	88276	57.163	ng	0.00
23) Nitrobenzene-d5	9.530	82	58376	36.324	ng	0.00
42) 2,4,6-Tribromophenol	16.450	330	44545	59.532	ng	0.00
45) 2-Fluorobiphenyl	13.589	172	149271	38.122	ng	-0.01
79) Terphenyl-d14	20.274	244	274843	35.122	ng	0.00

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG051923\
Data File : BG057480.D
Acq On : 19 May 2023 23:41
Operator : CG/JU
Sample : 02849-01
Misc :
ALS Vial : 12 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
TP7

Quant Time: May 20 04:31:25 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG042723.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Sat May 20 04:19:55 2023
Response via : Initial Calibration

