

Data Path : Z:\svoasrv\HPCHEM1\BNA\_G\Data\BG052022\  
 Data File : BG053643.D  
 Acq On : 20 May 2022 14:27  
 Operator : CG/JU  
 Sample : PB144936BL  
 Misc :  
 ALS Vial : 3 Sample Multiplier: 1

Instrument :  
 BNA\_G  
 ClientSampleId :  
 PB144936BL

Quant Time: May 23 01:30:03 2022  
 Quant Method : Z:\svoasrv\HPCHEM1\BNA\_G\Methods\8270-BG050522.M  
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
 QLast Update : Fri May 20 03:55:28 2022  
 Response via : Initial Calibration

| Compound                    | R.T.   | QIon | Response | Conc    | Units | Dev(Min) |
|-----------------------------|--------|------|----------|---------|-------|----------|
| -----                       |        |      |          |         |       |          |
| Internal Standards          |        |      |          |         |       |          |
| 1) 1,4-Dichlorobenzene-d4   | 8.069  | 152  | 24579    | 20.000  | ng    | 0.00     |
| 21) Naphthalene-d8          | 10.877 | 136  | 104552   | 20.000  | ng    | # 0.00   |
| 39) Acenaphthene-d10        | 14.696 | 164  | 85525    | 20.000  | ng    | 0.00     |
| 64) Phenanthrene-d10        | 17.445 | 188  | 215008   | 20.000  | ng    | 0.00     |
| 76) Chrysene-d12            | 21.721 | 240  | 209654   | 20.000  | ng    | #-0.01   |
| 86) Perylene-d12            | 24.988 | 264  | 204340   | 20.000  | ng    | -0.02    |
|                             |        |      |          |         |       |          |
| System Monitoring Compounds |        |      |          |         |       |          |
| 5) 2-Fluorophenol           | 5.637  | 112  | 215061   | 148.496 | ng    | 0.00     |
| 7) Phenol-d6                | 7.235  | 99   | 318112   | 144.730 | ng    | 0.00     |
| 23) Nitrobenzene-d5         | 9.232  | 82   | 215209   | 87.623  | ng    | 0.00     |
| 42) 2,4,6-Tribromophenol    | 16.188 | 330  | 161939   | 128.381 | ng    | 0.00     |
| 45) 2-Fluorobiphenyl        | 13.321 | 172  | 523994   | 89.530  | ng    | 0.00     |
| 79) Terphenyl-d14           | 20.047 | 244  | 1081751  | 96.098  | ng    | 0.00     |

|                  |        |
|------------------|--------|
| Target Compounds | Qvalue |
| -----            |        |

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA\_G\Data\BG052022\  
Data File : BG053643.D  
Acq On : 20 May 2022 14:27  
Operator : CG/JU  
Sample : PB144936BL  
Misc :  
ALS Vial : 3 Sample Multiplier: 1

Instrument :  
BNA\_G  
ClientSampleId :  
PB144936BL

Quant Time: May 23 01:30:03 2022  
Quant Method : Z:\svoasrv\HPCHEM1\BNA\_G\Methods\8270-BG050522.M  
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
QLast Update : Fri May 20 03:55:28 2022  
Response via : Initial Calibration

