

Data Path : Z:\HPCHEM1\BNA_G\DATA\BG052116\
 Data File : BG021995.D
 Acq On : 22 May 2016 1:58
 Operator : UM/SJ
 Sample : SSTDC0020
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_G
 LabSampleId :
 SSTD02008

Quant Time: May 23 18:15:07 2016
 Quant Method : Z:\HPCHEM1\BNA_G\METHODS\SOM02.2-EPA-BG052116.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Mon May 23 17:16:55 2016
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	84	0.00
2	1,4-Dioxane	0.696	0.811	-16.5	101	0.00
3 S	1,4-Dioxane-d8	0.384	0.440	-14.6	99	0.00
4	Benzaldehyde	0.996	1.099	-10.3	83	0.00
5 S	Phenol-d5	1.807	1.815	-0.4	87	0.00
6	Phenol	1.820	1.819	0.1	84	0.00
7 S	Bis-(2-Chloroethyl)ether-d8	0.947	0.949	-0.2	85	0.00
8	Bis(2-Chloroethyl)ether	1.330	1.385	-4.1	88	0.00
9 S	2-Chlorophenol-d4	1.511	1.526	-1.0	85	0.00
10	2-Chlorophenol	1.489	1.537	-3.2	86	0.00
11	2-Methylphenol	1.454	1.487	-2.3	86	0.00
12	2,2'-oxybis(1-Chloropropane	1.589	1.639	-3.1	85	0.00
13 S	4-Methylphenol-d8	1.533	1.577	-2.9	86	0.00
14	Acetophenone	2.144	2.255	-5.2	87	0.00
15 P	N-Nitroso-di-n-propylamine	1.107	1.164	-5.1	87	0.00
16	4-Methylphenol	1.567	1.623	-3.6	86	0.00
17	Hexachloroethane	0.594	0.613	-3.2	86	0.00
18 I	Naphthalene-d8	1.000	1.000	0.0	87	0.00
19 S	Nitrobenzene-d5	0.148	0.153	-3.4	89	0.00
20	Nitrobenzene	0.299	0.315	-5.4	90	0.00
21	Isophorone	0.636	0.668	-5.0	88	0.00
22 S	2-Nitrophenol-d4	0.160	0.157	1.9	83	0.00
23 C	2-Nitrophenol	0.167	0.176	-5.4	88	0.00
24	2,4-Dimethylphenol	0.322	0.328	-1.9	86	0.00
25	Bis(2-Chloroethoxy)methane	0.365	0.378	-3.6	86	0.00
26 S	2,4-Dichlorophenol-d3	0.282	0.295	-4.6	86	0.00
27 C	2,4-Dichlorophenol	0.277	0.285	-2.9	88	0.00
28	Naphthalene	1.014	1.047	-3.3	88	0.00
29 S	4-Chloroaniline-d4	0.307	0.339	-10.4	92	0.00
30	4-Chloroaniline	0.309	0.336	-8.7	89	0.00
31 C	Hexachlorobutadiene	0.163	0.163	0.0	84	0.00
32	Caprolactam	0.111	0.115	-3.6	89	0.00
33 C	4-Chloro-3-methylphenol	0.295	0.314	-6.4	90	0.00
34	2-Methylnaphthalene	0.725	0.743	-2.5	87	0.00
35 I	Acenaphthene-d10	1.000	1.000	0.0	88	0.00
36	1,2,4,5-Tetrachlorobenzene	0.593	0.588	0.8	86	0.00
37	Hexachlorocyclopentadiene	0.256	0.247	3.5	96	0.00
38 C	2,4,6-Trichlorophenol	0.401	0.395	1.5	85	0.00
39	2,4,5-Trichlorophenol	0.425	0.445	-4.7	94	0.00
40	1,1'-Biphenyl	1.647	1.711	-3.9	89	0.00
41	2-Chloronaphthalene	1.265	1.310	-3.6	89	0.00
42	2-Nitroaniline	0.285	0.282	1.1	84	0.00
43 S	Dimethylphthalate-d6	1.502	1.592	-6.0	91	0.00
44	Dimethylphthalate	1.510	1.605	-6.3	92	0.00

Data Path : Z:\HPCHEM1\BNA_G\DATA\BG052116\
 Data File : BG021995.D
 Acq On : 22 May 2016 1:58
 Operator : UM/SJ
 Sample : SSTDC0020
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_G
 LabSampleId :
 SSTD02008

Quant Time: May 23 18:15:07 2016
 Quant Method : Z:\HPCHEM1\BNA_G\METHODS\SOM02.2-EPA-BG052116.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Mon May 23 17:16:55 2016
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
45	2,6-Dinitrotoluene	0.324	0.339	-4.6	89	0.00
46 S	Acenaphthylene-d8	2.098	2.186	-4.2	89	0.00
47	Acenaphthylene	2.137	2.242	-4.9	91	0.00
48	3-Nitroaniline	0.310	0.281	9.4	85	0.00
49 C	Acenaphthene	1.476	1.517	-2.8	88	0.00
50	2,4-Dinitrophenol	0.125	0.087	30.4#	78	0.00
51 S	4-Nitrophenol-d4	0.259	0.260	-0.4	95	0.00
52	4-Nitrophenol	0.146	0.127	13.0	81	0.00
53	Dibenzofuran	1.825	1.897	-3.9	89	0.00
54	2,4-Dinitrotoluene	0.437	0.467	-6.9	90	0.00
55	2,3,4,6-Tetrachlorophenol	0.336	0.342	-1.8	87	0.00
56	Diethylphthalate	1.566	1.682	-7.4	92	0.00
57 S	Fluorene-d10	1.407	1.483	-5.4	91	0.00
58	Fluorene	1.541	1.644	-6.7	92	0.00
59	4-Chlorophenyl-phenylether	0.694	0.732	-5.5	90	0.00
60	4-Nitroaniline	0.341	0.317	7.0	78	0.00
61 I	Phenanthrene-d10	1.000	1.000	0.0	94	0.00
62 S	4,6-Dinitro-2-methylphenol-	0.105	0.094	10.5	87	0.00
63	4,6-Dinitro-2-methylphenol	0.108	0.100	7.4	96	0.00
64	N-Nitrosodiphenylamine	0.632	0.626	0.9	89	0.00
65	4-Bromophenyl-phenylether	0.202	0.204	-1.0	92	0.00
66	Hexachlorobenzene	0.235	0.227	3.4	94	0.00
67	Atrazine	0.218	0.223	-2.3	92	0.00
68 C	Pentachlorophenol	0.122	0.106	13.1	84	0.00
69	Phenanthrene	1.112	1.157	-4.0	95	0.00
70 S	Anthracene-d10	0.960	1.010	-5.2	97	0.00
71	Anthracene	1.127	1.148	-1.9	92	0.00
72	Carbazole	0.977	1.000	-2.4	95	0.00
73	Di-n-butylphthalate	1.309	1.375	-5.0	95	0.00
74 C	Fluoranthene	1.137	1.250	-9.9	100	0.00
75 I	Chrysene-d12	1.000	1.000	0.0	114	0.00
76 S	Pyrene-d10	1.128	1.091	3.3	103	0.00
77	Pyrene	1.407	1.371	2.6	103	0.00
78	Butylbenzylphthalate	0.667	0.657	1.5	106	0.00
79	3,3'-Dichlorobenzidine	0.376	0.388	-3.2	110	0.00
80	Benzo(a)anthracene	1.173	1.175	-0.2	108	0.00
81	Bis(2-ethylhexyl)phthalate	0.949	0.954	-0.5	109	0.00
82	Chrysene	1.125	1.152	-2.4	114	0.00
83 I	Perylene-d12	1.000	1.000	0.0	104	0.00
84	Di-n-octyl phthalate	1.630	1.764	-8.2	108	0.00
85	Benzo(b)fluoranthene	1.170	1.208	-3.2	105	0.00
86	Benzo(k)fluoranthene	1.139	1.222	-7.3	113	0.00
87 S	Benzo(a)pyrene-d12	0.928	1.001	-7.9	110	0.00

Data Path : Z:\HPCHEM1\BNA_G\DATA\BG052116\
Data File : BG021995.D
Acq On : 22 May 2016 1:58
Operator : UM/SJ
Sample : SSTDCCC020
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_G
LabSampleId :
SSTD02008

Quant Time: May 23 18:15:07 2016
Quant Method : Z:\HPCHEM1\BNA_G\METHODS\SOM02.2-EPA-BG052116.M
Quant Title : SVOA CALIBRATION
QLast Update : Mon May 23 17:16:55 2016
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
88 C	Benzo(a)pyrene	1.129	1.175	-4.1	108	0.00
89	Indeno(1,2,3-cd)pyrene	1.295	1.292	0.2	105	-0.02
90	Dibenzo(a,h)anthracene	1.083	1.086	-0.3	104	0.00
91	Benzo(g,h,i)perylene	1.037	1.001	3.5	105	0.01

(#) = Out of Range

SPCC's out = 0 CCC's out = 0