

Data Path : Z:\HPCHEM1\BNA_G\DATA\BG052215\
 Data File : BG017273.D
 Acq On : 23 May 2015 9:06
 Operator : TP/IZ
 Sample : G2359-05 2X
 Misc :
 ALS Vial : 30 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 SB-29AS

Quant Time: May 26 00:59:23 2015
 Quant Method : Z:\HPCHEM1\BNA_G\METHODS\8270-BG051315.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon May 25 08:11:45 2015
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.68	152	34861	20.00	ng	0.00
21) Naphthalene-d8	10.47	136	143396	20.00	ng	0.00
38) Acenaphthene-d10	14.33	164	81823	20.00	ng	0.00
63) Phenanthrene-d10	17.08	188	170569	20.00	ng	0.00
75) Chrysene-d12	21.26	240	185550	20.00	ng	0.00
86) Perylene-d12	23.52	264	185149	20.00	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.29	112	90319	45.97	ng	0.00
7) Phenol-d6	6.89	99	125196	45.40	ng	0.00
23) Nitrobenzene-d5	8.84	82	86464	28.15	ng	0.00
41) 2,4,6-Tribromophenol	15.83	330	43675	44.16	ng	0.00
44) 2-Fluorobiphenyl	12.95	172	166628	29.88	ng	0.00
78) Terphenyl-d14	19.71	244	222171	24.95	ng	0.00
Target Compounds						Qvalue
70) Phenanthrene	17.12	178	54315	6.18	ng	94
74) Fluoranthene	19.14	202	107238	10.15	ng	98
77) Pyrene	19.51	202	99833	8.77	ng	100
80) Benzo(a)anthracene	21.25	228	54008	5.08	ng	95
82) Chrysene	21.30	228	51157	5.17	ng	98
83) Bis(2-ethylhexyl)phthalate	21.18	149	39724	5.40	ng	98
85) Indeno(1,2,3-cd)pyrene	25.80	276	37054	3.13	ng	97
87) Benzo(b)fluoranthene	22.84	252	74773	6.93	ng	98
88) Benzo(k)fluoranthene	22.88	252	23677m	2.31	ng	
89) Benzo(a)pyrene	23.41	252	53563	5.40	ng	97
91) Benzo(g,h,i)perylene	26.50	276	39809	4.16	ng	# 91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA_G\DATA\BG052215\
Data File : BG017273.D
Acq On : 23 May 2015 9:06
Operator : TP/IZ
Sample : G2359-05 2X
Misc :
ALS Vial : 30 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
SB-29AS

Quant Time: May 26 00:59:23 2015
Quant Method : Z:\HPCHEM1\BNA_G\METHODS\8270-BG051315.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon May 25 08:11:45 2015
Response via : Initial Calibration

