


Data Path : Z:\HPCHEM1\BNA_G\DATA\BG052216\
 Data File : BG022009.D
 Acq On : 22 May 2016 12:41
 Operator : UM/SJ
 Sample : DFTPP11
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 DFTPP11

Integration File: LSCINT.P

Method : Z:\HPCHEM1\BNA_G\METHODS\SOM02.2-EPA-BG051816.M
 Title : SVOA CALIBRATION
 Last Update : Fri May 20 08:11:42 2016



AutoFind: Scans 2568, 2569, 2570; Background Corrected with Scan 2562

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	29.6	15173	PASS
68	69	0.00	2	1.5	282	PASS
69	198	0.00	100	37.2	19061	PASS
70	69	0.00	2	0.0	0	PASS
127	198	10	80	55.4	28429	PASS
197	198	0.00	2	0.0	0	PASS
198	198	100	100	100.0	51290	PASS
199	198	5	9	6.9	3525	PASS
275	198	10	60	24.5	12565	PASS
365	198	1	100	2.6	1347	PASS
441	442	0.01	100	16.0	5863	PASS
442	198	50	100	71.3	36562	PASS
443	442	15	24	19.1	6987	PASS

m/z	Abundance
39.90	1932.0
43.90	444.0
206.90	183.0

Instrument :
BNA_G
ClientSampleId :
DFTPP11

m/z	Abundance
39.00	973.0
39.90	1823.0
44.00	825.0
49.00	178.0
50.00	3170.0
51.00	13327.0
52.00	603.0
55.00	161.0
55.90	568.0
56.90	994.0
62.00	198.0
62.90	726.0
65.00	397.0
68.00	434.0
68.90	16017.0
73.90	1218.0
75.00	2356.0
76.00	1141.0
77.00	20560.0
78.00	1405.0
78.90	1241.0
79.90	881.0
80.90	1476.0
81.90	450.0
82.90	332.0
84.90	263.0
85.90	366.0
90.90	339.0
91.90	468.0
92.90	1635.0
97.90	1581.0
98.90	1153.0
99.90	188.0
100.90	894.0
102.90	319.0
103.80	539.0
104.90	446.0
105.80	207.0
106.90	5853.0
107.90	897.0
108.90	335.0
109.90	11184.0
110.90	1798.0
111.90	257.0
115.90	341.0
116.90	3895.0
117.90	400.0
121.90	333.0
122.90	693.0
123.90	307.0
124.90	302.0
126.90	23936.0
127.90	1999.0
128.90	8486.0
129.90	747.0
130.90	160.0
133.90	344.0
134.90	694.0
135.90	258.0
137.00	459.0
139.90	171.0
140.90	1093.0
141.90	228.0
142.80	307.0
145.80	193.0
146.90	678.0
147.90	974.0
148.90	219.0
152.90	343.0
153.80	313.0
154.90	468.0
155.90	659.0
156.90	203.0
157.90	226.0
159.90	364.0
160.90	510.0
161.80	263.0
164.70	197.0
166.00	257.0
166.90	1909.0
168.00	1074.0
168.90	180.0
172.80	299.0
173.90	414.0
174.90	1005.0
175.90	248.0
176.90	354.0
178.20	174.0
178.90	1580.0
179.90	1054.0
180.90	538.0

Instrument :
BNA_G
ClientSampleId :
DFTPP11

183.80	211.0
184.90	927.0
185.90	5854.0
186.90	1547.0
187.80	220.0
188.80	458.0
191.80	450.0
192.90	583.0
195.80	1157.0
196.50	455.0
197.80	38952.0
198.80	2804.0
199.80	197.0
201.30	328.0
202.90	298.0
203.90	1588.0
204.90	2746.0
205.90	10402.0
206.90	1519.0
207.90	406.0
208.90	163.0
209.70	195.0
210.40	187.0
211.00	402.0
215.80	238.0
216.80	2569.0
217.80	381.0
219.90	193.0
220.90	1962.0
221.90	351.0
222.80	744.0
223.90	5922.0
224.90	1405.0
225.90	214.0
226.90	2174.0
227.90	158.0
228.90	425.0
230.80	190.0
235.90	187.0
236.90	191.0
240.90	184.0
241.90	282.0
242.90	467.0
243.90	4231.0
244.90	717.0
245.90	725.0
246.90	195.0
252.90	202.0
254.90	21400.0
255.90	3091.0
256.90	193.0
257.90	1179.0
258.80	228.0
264.90	495.0
272.80	790.0
273.80	1521.0
274.90	9406.0
275.90	1195.0
276.80	815.0
292.80	189.0
295.90	2492.0
296.90	503.0
302.80	278.0
313.90	156.0
314.80	258.0
315.90	183.0
322.90	972.0
323.80	176.0
326.80	165.0
333.90	342.0
345.80	183.0
351.80	280.0
353.80	260.0
364.80	863.0
371.90	513.0
401.80	161.0
402.90	423.0
420.80	268.0
421.80	194.0
422.80	1288.0
423.80	211.0
440.90	3847.0
441.90	22848.0
442.90	4802.0
443.90	504.0

Instrument :
BNA_G
ClientSampleId :
DFTPP11

m/z	Abundance
39.00	1020.0
40.00	1884.0
44.00	924.0
50.00	3839.0
51.00	16880.0
52.00	886.0
56.00	567.0
57.00	1347.0
60.90	187.0
61.90	232.0
63.00	657.0
64.90	456.0
68.90	21376.0
73.10	247.0
74.00	2424.0
75.00	3430.0
76.00	1177.0
77.00	27608.0
78.00	1758.0
78.90	1492.0
79.90	1175.0
80.90	1959.0
81.90	623.0
83.00	451.0
83.90	164.0
84.90	402.0
85.90	462.0
86.90	320.0
91.00	422.0
91.90	462.0
92.90	2331.0
93.90	179.0
95.90	160.0
97.90	1867.0
98.90	1549.0
100.00	215.0
100.90	1081.0
102.90	292.0
103.90	584.0
104.90	547.0
105.90	219.0
106.90	7946.0
107.90	1292.0
109.90	15612.0
110.90	2798.0
112.00	264.0
113.00	154.0
116.00	635.0
116.90	5293.0
117.90	428.0
121.90	489.0
122.90	744.0
123.90	466.0
124.90	385.0
126.90	32288.0
127.90	2779.0
128.90	12137.0
129.90	1055.0
131.00	187.0
133.90	308.0
134.90	893.0
135.90	268.0
136.90	553.0
137.90	165.0
139.90	199.0
140.90	1411.0
141.90	501.0
142.90	418.0
145.80	342.0
146.90	698.0
147.90	1597.0
148.70	255.0
151.00	168.0
152.80	537.0
153.90	486.0
154.90	931.0
155.90	1441.0
156.90	303.0
157.90	307.0
158.90	287.0
159.90	451.0
160.90	662.0
161.90	190.0
164.90	669.0
165.80	576.0
166.90	2619.0
167.90	1067.0
168.90	344.0
171.90	317.0
172.90	340.0
173.90	558.0

Instrument :
BNA_G
ClientSampleId :
DFTPP11

174.90	1490.0
175.90	309.0
176.80	408.0
177.90	167.0
178.90	2203.0
179.90	1401.0
180.90	935.0
182.00	201.0
183.90	190.0
184.90	1042.0
185.90	7790.0
186.90	2196.0
187.90	218.0
188.90	711.0
190.80	276.0
191.90	829.0
192.90	776.0
193.90	228.0
195.80	1701.0
197.80	58712.0
198.80	3844.0
199.90	347.0
201.30	281.0
202.80	459.0
203.90	2188.0
204.90	3624.0
205.90	16163.0
206.90	2318.0
207.90	699.0
209.90	259.0
210.90	453.0
214.90	205.0
215.80	450.0
216.90	3708.0
217.90	577.0
220.90	2285.0
221.90	486.0
222.90	1050.0
223.90	8536.0
224.90	1715.0
225.90	201.0
226.90	3399.0
227.90	432.0
228.90	743.0
230.80	243.0
233.90	304.0
234.80	426.0
235.70	174.0
236.90	198.0
241.80	580.0
242.90	515.0
243.90	6512.0
244.80	1096.0
245.90	1337.0
246.90	323.0
248.90	180.0
252.90	341.0
254.90	32848.0
255.90	4940.0
256.80	401.0
257.90	1484.0
258.80	360.0
264.90	959.0
270.70	186.0
272.00	170.0
272.90	1197.0
273.90	2431.0
274.90	14528.0
275.90	2085.0
276.90	1103.0
277.70	168.0
285.00	185.0
292.80	237.0
295.90	4065.0
296.80	530.0
302.80	535.0
303.90	185.0
313.80	201.0
314.90	497.0
315.80	352.0
320.80	159.0
322.90	1283.0
323.90	181.0
326.80	208.0
333.90	742.0
340.90	185.0
345.90	214.0
351.80	422.0
352.70	193.0
353.90	374.0
364.80	1519.0
365.80	225.0
371.90	709.0
372.80	152.0
382.70	203.0

Instrument :

BNA_G

ClientSampleId :

DFTPP11

401.80	221.0
402.80	385.0
420.90	433.0
421.90	384.0
422.90	2021.0
423.90	503.0
440.90	6709.0
441.90	40552.0
442.90	7551.0
443.90	670.0

Instrument :
BNA_G
ClientSampleId :
DFTPP11

m/z	Abundance
38.00	166.0
39.00	1082.0
39.90	1787.0
43.90	977.0
50.00	3578.0
51.00	15314.0
52.00	943.0
55.90	573.0
56.90	1019.0
60.90	252.0
61.90	214.0
62.90	715.0
65.00	385.0
67.90	413.0
68.90	19792.0
73.00	180.0
74.00	1936.0
74.90	3165.0
76.00	1061.0
77.00	23208.0
78.00	1627.0
78.90	1295.0
79.90	1215.0
80.90	1612.0
81.90	445.0
82.90	555.0
85.00	327.0
86.00	414.0
86.90	202.0
91.00	297.0
91.90	366.0
92.90	2673.0
93.90	294.0
97.90	1801.0
98.90	1680.0
100.90	927.0
102.90	361.0
103.90	964.0
105.00	588.0
106.00	303.0
106.90	7187.0
107.90	1135.0
109.90	14819.0
110.90	2425.0
111.90	209.0
116.00	360.0
116.90	5038.0
117.90	314.0
119.90	164.0
122.00	409.0
122.90	943.0
123.90	373.0
124.90	489.0
126.90	29064.0
127.90	2467.0
128.90	10675.0
129.90	880.0
130.80	250.0
131.90	211.0
133.80	355.0
134.90	822.0
135.80	206.0
136.90	471.0
139.80	151.0
140.90	1547.0
141.90	469.0
142.90	401.0
146.00	269.0
146.90	856.0
147.90	1416.0
149.00	233.0
149.90	162.0
152.90	502.0
153.90	407.0
154.90	714.0
155.90	1281.0
156.90	276.0
157.90	246.0
158.80	178.0
159.90	510.0
160.90	895.0
161.90	263.0
164.90	599.0
165.90	452.0
166.90	3121.0
167.90	939.0
168.90	258.0
170.90	200.0
171.90	299.0
172.90	315.0
173.90	586.0

Instrument :
BNA_G
ClientSampleId :
DFTPP11

174.90	1225.0
175.90	424.0
176.90	565.0
177.90	220.0
178.80	1871.0
179.90	1671.0
180.90	733.0
181.90	238.0
184.90	1309.0
185.90	7687.0
186.90	2290.0
187.90	305.0
188.80	379.0
190.80	228.0
191.90	636.0
192.90	855.0
195.90	1975.0
197.80	56208.0
198.80	3929.0
199.80	388.0
201.30	280.0
202.90	565.0
203.90	2605.0
204.90	4026.0
205.90	14614.0
206.90	2308.0
207.90	459.0
208.90	194.0
209.90	198.0
210.90	565.0
215.90	345.0
216.90	4079.0
217.80	554.0
219.90	213.0
220.90	2714.0
222.90	935.0
223.90	8609.0
224.90	2143.0
225.90	177.0
226.90	3609.0
227.90	605.0
228.90	665.0
230.80	294.0
233.90	211.0
234.80	342.0
236.80	274.0
238.80	189.0
240.90	205.0
241.90	458.0
242.90	383.0
243.90	6893.0
244.90	836.0
245.80	993.0
246.80	241.0
248.80	247.0
250.80	167.0
253.00	192.0
253.90	286.0
254.90	33368.0
255.90	4912.0
256.90	513.0
257.90	1753.0
258.80	314.0
264.80	736.0
270.90	214.0
272.10	270.0
272.90	1377.0
273.90	1990.0
274.90	13763.0
275.90	1939.0
276.90	1058.0
278.00	174.0
282.80	170.0
284.80	164.0
292.80	310.0
295.90	4031.0
296.80	480.0
302.80	446.0
303.80	173.0
313.90	253.0
314.80	420.0
315.80	215.0
320.90	176.0
322.90	1326.0
323.90	214.0
326.70	254.0
331.70	191.0
333.90	747.0
334.80	178.0
340.80	211.0
345.70	256.0
351.90	384.0
352.80	274.0
353.90	394.0
364.80	1661.0

Instrument :
BNA_G
ClientSampleId :
DFTPP11

365.70	214.0
371.90	794.0
372.90	232.0
382.80	249.0
401.80	331.0
402.80	418.0
403.90	163.0
420.80	326.0
421.90	380.0
422.90	2170.0
423.80	430.0
440.90	7034.0
441.90	46288.0
442.90	8609.0
443.80	712.0

Instrument :
BNA_G
ClientSampleId :
DFTPP11