

Data Path : Z:\HPCHEM1\BNA_G\DATA\BG052316\
 Data File : BG022049.D
 Acq On : 23 May 2016 16:36
 Operator : UM/SJ
 Sample : SSTDC0020
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_G
 LabSampleId :
 SSTD02014

Quant Time: May 24 07:27:44 2016
 Quant Method : Z:\HPCHEM1\BNA_G\METHODS\SOM02.2-EPA-BG052116.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Tue May 24 05:44:46 2016
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	114	0.00
2	1,4-Dioxane	0.696	0.718	-3.2	122	0.00
3 S	1,4-Dioxane-d8	0.384	0.366	4.7	112	0.00
4	Benzaldehyde	0.996	1.070	-7.4	110	0.00
5 S	Phenol-d5	1.807	1.760	2.6	115	0.00
6	Phenol	1.820	1.763	3.1	110	0.00
7 S	Bis-(2-Chloroethyl)ether-d8	0.947	0.919	3.0	111	0.00
8	Bis(2-Chloroethyl)ether	1.330	1.364	-2.6	117	0.00
9 S	2-Chlorophenol-d4	1.511	1.532	-1.4	116	0.00
10	2-Chlorophenol	1.489	1.520	-2.1	115	0.00
11	2-Methylphenol	1.454	1.401	3.6	110	0.00
12	2,2'-oxybis(1-Chloropropane	1.589	1.577	0.8	110	0.00
13 S	4-Methylphenol-d8	1.533	1.519	0.9	112	0.00
14	Acetophenone	2.144	2.164	-0.9	113	0.00
15 P	N-Nitroso-di-n-propylamine	1.107	1.097	0.9	111	0.00
16	4-Methylphenol	1.567	1.560	0.4	113	0.00
17	Hexachloroethane	0.594	0.596	-0.3	114	0.00
18 I	Naphthalene-d8	1.000	1.000	0.0	112	0.00
19 S	Nitrobenzene-d5	0.148	0.154	-4.1	115	0.00
20	Nitrobenzene	0.299	0.316	-5.7	116	0.00
21	Isophorone	0.636	0.642	-0.9	109	0.00
22 S	2-Nitrophenol-d4	0.160	0.163	-1.9	111	0.00
23 C	2-Nitrophenol	0.167	0.176	-5.4	114	0.00
24	2,4-Dimethylphenol	0.322	0.326	-1.2	110	0.00
25	Bis(2-Chloroethoxy)methane	0.365	0.382	-4.7	112	0.00
26 S	2,4-Dichlorophenol-d3	0.282	0.296	-5.0	112	0.00
27 C	2,4-Dichlorophenol	0.277	0.287	-3.6	114	0.00
28	Naphthalene	1.014	1.052	-3.7	114	0.00
29 S	4-Chloroaniline-d4	0.307	0.363	-18.2	127	0.00
30	4-Chloroaniline	0.309	0.374	-21.0#	127	0.00
31 C	Hexachlorobutadiene	0.163	0.165	-1.2	110	0.00
32	Caprolactam	0.111	0.106	4.5	105	0.00
33 C	4-Chloro-3-methylphenol	0.295	0.300	-1.7	111	0.00
34	2-Methylnaphthalene	0.725	0.734	-1.2	111	0.00
35 I	Acenaphthene-d10	1.000	1.000	0.0	108	0.00
36	1,2,4,5-Tetrachlorobenzene	0.593	0.618	-4.2	111	0.00
37	Hexachlorocyclopentadiene	0.256	0.171	33.2#	81	0.00
38 C	2,4,6-Trichlorophenol	0.401	0.413	-3.0	109	0.00
39	2,4,5-Trichlorophenol	0.425	0.418	1.6	108	0.00
40	1,1'-Biphenyl	1.647	1.728	-4.9	110	0.00
41	2-Chloronaphthalene	1.265	1.330	-5.1	111	0.00
42	2-Nitroaniline	0.285	0.293	-2.8	106	0.00
43 S	Dimethylphthalate-d6	1.502	1.543	-2.7	108	0.00
44	Dimethylphthalate	1.510	1.559	-3.2	109	0.00

Data Path : Z:\HPCHEM1\BNA_G\DATA\BG052316\
 Data File : BG022049.D
 Acq On : 23 May 2016 16:36
 Operator : UM/SJ
 Sample : SSTDC0020
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_G
 LabSampleId :
 SSTD02014

Quant Time: May 24 07:27:44 2016
 Quant Method : Z:\HPCHEM1\BNA_G\METHODS\SOM02.2-EPA-BG052116.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Tue May 24 05:44:46 2016
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
45	2,6-Dinitrotoluene	0.324	0.331	-2.2	107	0.00
46 S	Acenaphthylene-d8	2.098	2.165	-3.2	108	0.00
47	Acenaphthylene	2.137	2.193	-2.6	110	0.00
48	3-Nitroaniline	0.310	0.313	-1.0	115	0.00
49 C	Acenaphthene	1.476	1.507	-2.1	108	0.00
50	2,4-Dinitrophenol	0.125	0.093	25.6#	101	0.00
51 S	4-Nitrophenol-d4	0.259	0.241	6.9	108	0.00
52	4-Nitrophenol	0.146	0.141	3.4	111	0.00
53	Dibenzofuran	1.825	1.883	-3.2	108	0.00
54	2,4-Dinitrotoluene	0.437	0.450	-3.0	106	0.00
55	2,3,4,6-Tetrachlorophenol	0.336	0.338	-0.6	105	0.00
56	Diethylphthalate	1.566	1.600	-2.2	108	0.00
57 S	Fluorene-d10	1.407	1.402	0.4	105	0.00
58	Fluorene	1.541	1.560	-1.2	107	0.00
59	4-Chlorophenyl-phenylether	0.694	0.706	-1.7	106	0.00
60	4-Nitroaniline	0.341	0.316	7.3	95	0.00
61 I	Phenanthrene-d10	1.000	1.000	0.0	106	0.00
62 S	4,6-Dinitro-2-methylphenol-	0.105	0.094	10.5	98	0.00
63	4,6-Dinitro-2-methylphenol	0.108	0.097	10.2	106	0.00
64	N-Nitrosodiphenylamine	0.632	0.662	-4.7	106	0.00
65	4-Bromophenyl-phenylether	0.202	0.208	-3.0	106	0.00
66	Hexachlorobenzene	0.235	0.239	-1.7	112	0.00
67	Atrazine	0.218	0.231	-6.0	107	0.00
68 C	Pentachlorophenol	0.122	0.115	5.7	103	0.00
69	Phenanthrene	1.112	1.158	-4.1	107	0.00
70 S	Anthracene-d10	0.960	0.964	-0.4	104	0.00
71	Anthracene	1.127	1.175	-4.3	106	0.00
72	Carbazole	0.977	1.001	-2.5	107	0.00
73	Di-n-butylphthalate	1.309	1.333	-1.8	104	0.00
74 C	Fluoranthene	1.137	1.112	2.2	101	0.00
75 I	Chrysene-d12	1.000	1.000	0.0	107	0.00
76 S	Pyrene-d10	1.128	1.149	-1.9	101	0.00
77	Pyrene	1.407	1.402	0.4	99	0.00
78	Butylbenzylphthalate	0.667	0.678	-1.6	102	0.00
79	3,3'-Dichlorobenzidine	0.376	0.420	-11.7	112	0.00
80	Benzo(a)anthracene	1.173	1.194	-1.8	103	0.00
81	Bis(2-ethylhexyl)phthalate	0.949	0.964	-1.6	103	0.00
82	Chrysene	1.125	1.152	-2.4	107	0.00
83 I	Perylene-d12	1.000	1.000	0.0	101	0.00
84	Di-n-octyl phthalate	1.630	1.794	-10.1	107	0.00
85	Benzo(b)fluoranthene	1.170	1.262	-7.9	107	0.01
86	Benzo(k)fluoranthene	1.139	1.173	-3.0	105	0.01
87 S	Benzo(a)pyrene-d12	0.928	0.990	-6.7	106	0.01

Data Path : Z:\HPCHEM1\BNA_G\DATA\BG052316\
Data File : BG022049.D
Acq On : 23 May 2016 16:36
Operator : UM/SJ
Sample : SSTDCCC020
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_G
LabSampleId :
SSTD02014

Quant Time: May 24 07:27:44 2016
Quant Method : Z:\HPCHEM1\BNA_G\METHODS\SOM02.2-EPA-BG052116.M
Quant Title : SVOA CALIBRATION
QLast Update : Tue May 24 05:44:46 2016
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
88 C	Benzo(a)pyrene	1.129	1.184	-4.9	105	0.00
89	Indeno(1,2,3-cd)pyrene	1.295	1.277	1.4	101	0.01
90	Dibenzo(a,h)anthracene	1.083	0.991	8.5	92	0.02
91	Benzo(g,h,i)perylene	1.037	0.935	9.8	95	0.02

(#) = Out of Range

SPCC's out = 0 CCC's out = 0