


Data Path : Z:\HPCHEM1\BNA_G\DATA\BG052316\
Data File : BG022041.D
Acq On : 23 May 2016 10:15
Operator : UM/SJ
Sample : DFTPP12
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
DFTPP12

Integration File: LSCINT.P

Method : Z:\HPCHEM1\BNA_G\METHODS\SOM02.2-EPA-BG052116.M
Title : SVOA CALIBRATION
Last Update : Tue May 24 05:44:46 2016



AutoFind: Scans 2569, 2570, 2571; Background Corrected with Scan 2563

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	29.9	16090	PASS
68	69	0.00	2	0.5	101	PASS
69	198	0.00	100	36.4	19584	PASS
70	69	0.00	2	0.0	0	PASS
127	198	10	80	56.4	30298	PASS
197	198	0.00	2	0.3	182	PASS
198	198	100	100	100.0	53762	PASS
199	198	5	9	6.9	3718	PASS
275	198	10	60	25.2	13537	PASS
365	198	1	100	3.3	1765	PASS
441	442	0.01	100	16.8	6460	PASS
442	198	50	100	71.4	38360	PASS
443	442	15	24	19.7	7574	PASS

Scan 2563 (17.897 min): BG022041.D
DFTPP12

m/z	Abundance
39.90	1868.0
44.00	501.0
48.90	262.0
83.80	252.0
206.90	186.0

Instrument :
BNA_G
ClientSampleId :
DFTPP12

m/z	Abundance
39.00	944.0
39.90	2020.0
40.90	159.0
43.90	781.0
49.00	335.0
50.00	3375.0
51.00	14079.0
52.00	618.0
55.90	494.0
57.00	1078.0
61.90	232.0
63.00	722.0
65.00	376.0
67.90	303.0
68.90	15761.0
74.00	1628.0
75.00	2318.0
76.00	908.0
77.00	20144.0
78.00	1473.0
78.90	1260.0
79.90	943.0
80.90	1439.0
82.00	392.0
82.90	315.0
83.90	156.0
85.00	219.0
85.90	539.0
86.90	165.0
90.90	323.0
91.90	330.0
92.90	1748.0
93.90	176.0
97.90	1438.0
98.90	1128.0
100.90	888.0
102.90	274.0
103.90	430.0
104.90	345.0
106.00	170.0
106.90	5464.0
107.90	916.0
109.90	11273.0
110.90	1733.0
112.00	229.0
116.00	318.0
116.90	3763.0
117.90	225.0
121.90	324.0
122.90	456.0
123.90	334.0
124.80	266.0
126.90	23632.0
128.00	1887.0
128.90	9252.0
129.90	562.0
133.80	311.0
134.90	810.0
135.80	366.0
136.90	275.0
140.90	909.0
141.80	331.0
142.80	163.0
145.90	158.0
146.90	387.0
147.90	811.0
148.90	382.0
150.90	177.0
152.90	361.0
153.90	262.0
154.90	521.0
155.90	828.0
156.90	165.0
157.90	191.0
159.90	345.0
160.90	577.0
164.90	521.0
165.80	446.0
166.90	2072.0
167.90	811.0
172.80	312.0
173.90	456.0
174.90	1031.0
175.90	265.0
176.90	458.0
178.80	1474.0
179.90	1191.0
180.80	447.0
183.90	201.0
184.90	852.0
185.90	5537.0

Instrument :
BNA_G
ClientSampleId :
DFTPP12

186.90	1887.0
187.90	209.0
188.90	433.0
190.80	230.0
191.90	456.0
192.90	517.0
195.90	1427.0
196.70	547.0
197.90	39512.0
198.80	2780.0
199.80	271.0
202.90	512.0
203.90	1720.0
204.90	2825.0
205.90	11280.0
206.90	1503.0
207.80	321.0
209.90	204.0
210.90	358.0
214.80	262.0
215.80	398.0
216.80	2842.0
217.90	378.0
220.90	1336.0
221.80	278.0
222.80	567.0
223.90	5506.0
224.90	1626.0
226.80	2395.0
227.80	367.0
228.80	407.0
230.80	213.0
236.80	158.0
241.80	234.0
242.90	397.0
243.90	4445.0
244.90	471.0
245.90	770.0
246.70	165.0
248.80	190.0
252.80	232.0
253.90	364.0
254.90	21656.0
255.90	3778.0
256.80	344.0
257.90	1002.0
258.80	190.0
264.80	415.0
270.80	176.0
272.90	626.0
273.90	1592.0
274.90	8756.0
275.90	1289.0
276.90	891.0
295.90	2629.0
296.80	361.0
302.90	288.0
303.90	151.0
313.80	173.0
314.90	280.0
315.90	250.0
322.90	706.0
323.90	215.0
326.70	199.0
327.80	173.0
333.80	502.0
345.80	289.0
351.90	303.0
353.90	248.0
364.80	1111.0
365.80	215.0
371.90	421.0
402.80	227.0
420.90	195.0
421.90	161.0
422.80	1301.0
423.90	279.0
440.90	3728.0
441.90	22000.0
442.90	4318.0
443.80	339.0

Instrument :
BNA_G
ClientSampleId :
DFTPP12

m/z	Abundance
38.00	186.0
39.00	1067.0
39.90	1914.0
43.90	808.0
49.00	398.0
50.00	4411.0
51.00	17640.0
52.00	887.0
54.90	156.0
56.00	528.0
57.00	1207.0
61.00	300.0
62.00	314.0
63.00	810.0
64.90	502.0
68.90	23512.0
73.90	2166.0
75.00	3788.0
76.00	1384.0
77.00	28528.0
78.00	2124.0
78.90	1511.0
79.90	1440.0
80.90	1695.0
81.90	544.0
82.90	595.0
83.90	288.0
84.90	391.0
85.90	515.0
86.90	164.0
90.90	457.0
91.90	363.0
92.90	2598.0
96.00	231.0
97.90	2163.0
98.90	1533.0
100.90	1423.0
102.90	351.0
103.90	629.0
105.00	624.0
106.00	299.0
106.90	8243.0
107.90	1287.0
109.90	16302.0
110.90	2558.0
111.90	339.0
115.90	589.0
116.90	5693.0
117.90	378.0
121.90	438.0
122.90	934.0
123.90	524.0
124.90	397.0
126.90	32872.0
127.90	2441.0
128.90	12896.0
129.90	1115.0
130.90	341.0
132.00	153.0
133.90	476.0
134.90	1272.0
135.90	545.0
136.90	450.0
137.50	165.0
139.80	234.0
140.90	1530.0
141.90	423.0
142.90	323.0
145.90	241.0
146.90	760.0
147.90	1616.0
148.90	394.0
149.90	177.0
150.90	157.0
152.90	537.0
153.90	461.0
154.90	951.0
155.90	1477.0
156.90	221.0
157.80	364.0
158.90	262.0
159.90	533.0
160.90	762.0
161.90	315.0
164.90	682.0
165.90	617.0
166.90	2787.0
167.90	1291.0
168.90	306.0
171.80	270.0
172.90	445.0

Instrument :
BNA_G
ClientSampleId :
DFTPP12

173.90	640.0
174.90	1185.0
175.80	437.0
176.90	679.0
178.00	224.0
178.80	2084.0
179.90	1842.0
180.90	911.0
183.80	182.0
184.90	1318.0
185.90	8286.0
186.90	2330.0
187.90	283.0
188.90	591.0
190.90	303.0
191.90	728.0
192.90	996.0
193.90	245.0
195.90	2010.0
197.90	61048.0
198.80	3999.0
199.80	401.0
201.30	417.0
202.90	405.0
203.90	2664.0
204.90	4198.0
205.90	17592.0
206.90	2551.0
207.90	602.0
208.90	235.0
210.90	609.0
214.90	170.0
215.90	372.0
216.90	4350.0
217.80	629.0
220.10	197.0
220.90	2368.0
222.90	1134.0
223.90	9186.0
224.90	2234.0
226.00	244.0
226.90	3576.0
227.90	497.0
228.80	825.0
230.90	327.0
233.80	287.0
235.00	249.0
235.90	211.0
236.80	312.0
239.90	177.0
240.80	212.0
241.90	413.0
242.90	424.0
243.90	6870.0
244.90	1169.0
245.80	1165.0
246.80	374.0
248.80	306.0
251.90	180.0
252.90	230.0
254.90	34544.0
255.90	5702.0
256.90	334.0
257.90	1759.0
258.90	438.0
264.90	599.0
271.00	200.0
272.00	232.0
272.90	1160.0
273.90	2595.0
274.90	15192.0
275.90	1903.0
276.80	1297.0
284.90	293.0
292.80	289.0
295.80	4028.0
296.80	668.0
302.90	367.0
303.90	191.0
313.80	203.0
314.80	503.0
315.90	345.0
320.80	189.0
322.90	1437.0
323.80	192.0
326.80	234.0
333.90	838.0
334.90	236.0
340.80	202.0
345.80	298.0
351.90	542.0
352.80	263.0
353.90	464.0
364.80	1704.0
365.90	229.0

Instrument :
BNA_G
ClientSampleId :
DFTPP12

371.80	647.0
372.90	157.0
382.80	200.0
401.90	361.0
402.80	291.0
420.80	283.0
421.90	288.0
422.90	2163.0
423.80	426.0
440.90	6832.0
441.90	41240.0
442.90	8127.0
443.90	694.0

Instrument :
BNA_G
ClientSampleId :
DFTPP12

m/z	Abundance
39.00	1051.0
39.90	2167.0
44.00	886.0
49.00	278.0
50.00	4249.0
51.00	16552.0
52.00	766.0
55.00	198.0
56.00	593.0
57.00	1256.0
57.90	158.0
60.90	252.0
62.00	332.0
63.00	644.0
65.00	481.0
68.90	19480.0
72.90	189.0
74.00	2531.0
74.90	3267.0
76.00	1282.0
77.00	27416.0
78.00	1774.0
78.90	1370.0
79.90	1109.0
80.90	1642.0
81.90	383.0
82.90	553.0
83.80	211.0
84.90	335.0
85.90	635.0
86.90	220.0
90.90	464.0
91.90	464.0
92.90	2433.0
93.90	313.0
98.00	1845.0
98.90	1673.0
100.00	170.0
100.90	1113.0
102.90	458.0
104.00	624.0
104.90	644.0
106.90	8409.0
108.00	1224.0
109.90	16568.0
110.90	2237.0
111.90	261.0
115.90	646.0
116.90	5342.0
117.90	424.0
121.90	474.0
122.90	852.0
123.90	332.0
124.80	394.0
126.90	34392.0
127.90	2845.0
128.90	13176.0
129.90	936.0
130.90	279.0
133.80	351.0
134.90	1064.0
135.90	403.0
136.90	360.0
139.90	173.0
140.90	1451.0
141.80	551.0
142.90	488.0
143.90	204.0
144.80	151.0
145.90	294.0
146.90	965.0
147.90	1432.0
148.90	317.0
151.00	284.0
151.60	187.0
152.80	425.0
153.90	411.0
154.90	1039.0
155.90	1142.0
157.00	351.0
157.80	307.0
158.80	325.0
159.90	420.0
160.90	722.0
161.80	227.0
164.90	578.0
165.90	523.0
166.90	2728.0
167.90	907.0
168.90	154.0
171.80	215.0

Instrument :
BNA_G
ClientSampleId :
DFTPP12

172.90	351.0
173.90	531.0
174.90	1383.0
175.90	484.0
176.80	623.0
177.90	220.0
178.90	2377.0
179.90	1913.0
180.90	751.0
183.90	212.0
184.90	1266.0
185.90	8969.0
186.90	2294.0
187.90	436.0
188.90	562.0
191.00	236.0
191.90	881.0
192.90	795.0
193.90	151.0
194.80	188.0
195.80	1844.0
197.80	60728.0
198.90	4375.0
199.90	452.0
201.40	364.0
203.00	478.0
203.90	2477.0
204.90	4423.0
205.90	16888.0
206.90	2618.0
207.90	523.0
208.80	196.0
209.90	361.0
210.20	385.0
210.90	598.0
214.90	286.0
215.90	399.0
216.80	4458.0
217.90	496.0
220.80	2680.0
222.90	1340.0
223.90	8685.0
224.90	2250.0
225.80	257.0
226.90	4171.0
227.90	607.0
228.80	897.0
230.80	379.0
233.80	232.0
234.90	295.0
235.80	199.0
236.90	310.0
238.90	183.0
241.90	490.0
242.90	631.0
243.90	7330.0
244.90	929.0
245.90	1404.0
246.80	268.0
248.90	455.0
252.80	268.0
254.90	37960.0
255.90	5408.0
256.80	493.0
257.90	2172.0
258.80	507.0
264.90	916.0
270.90	223.0
272.90	1180.0
273.90	2783.0
274.90	16664.0
275.90	2223.0
276.90	1223.0
282.90	254.0
284.90	208.0
292.80	267.0
295.90	5057.0
296.80	595.0
302.90	497.0
304.00	163.0
313.90	292.0
314.90	401.0
315.90	264.0
320.80	256.0
322.90	1584.0
323.90	375.0
326.80	275.0
327.90	164.0
331.90	209.0
333.80	1062.0
334.80	205.0
340.90	182.0
345.90	334.0
351.80	437.0
352.80	262.0

Instrument :
BNA_G
ClientSampleId :
DFTPP12

353.90	585.0
364.80	2482.0
365.80	211.0
370.80	152.0
371.90	725.0
372.90	173.0
382.80	172.0
401.80	285.0
402.80	498.0
420.80	426.0
421.90	379.0
422.90	2879.0
423.90	524.0
440.90	8820.0
441.90	51840.0
442.90	10279.0
443.90	888.0

Instrument :
BNA_G
ClientSampleId :
DFTPP12