


Data Path : Z:\HPCHEM1\BNA_G\Data\BG052317\
 Data File : BG027061.D
 Acq On : 22 May 2017 16:59
 Operator : SJ/MA
 Sample : DFTPP66
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 DFTPP66

Integration File: LSCINT.P

Method : Z:\HPCHEM1\BNA_G\METHODS\SOM-EPA-BG051817MA.M
 Title : SVOA CALIBRATION
 Last Update : Tue May 23 17:47:51 2017



AutoFind: Scans 2500, 2501, 2502; Background Corrected with Scan 2494

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	32.3	56522	PASS
68	69	0.00	2	1.0	597	PASS
69	198	0.00	100	35.4	61978	PASS
70	69	0.00	2	0.7	426	PASS
127	198	10	80	52.9	92581	PASS
197	198	0.00	2	0.0	0	PASS
198	198	100	100	100.0	175061	PASS
199	198	5	9	7.4	12898	PASS
275	198	10	60	25.6	44808	PASS
365	198	1	100	3.3	5825	PASS
441	442	0.01	100	17.1	21408	PASS
442	198	50	100	71.5	125178	PASS
443	442	15	24	19.9	24938	PASS

m/z	Abundance
36.90	200.0
38.00	654.0
39.00	3956.0
40.00	499.0
41.00	280.0
42.10	152.0
43.00	263.0
44.00	1714.0
45.00	210.0
48.90	591.0
50.00	15649.0
51.00	65392.0
52.00	3295.0
52.90	381.0
55.00	843.0
55.90	2277.0
57.00	5202.0
58.00	327.0
59.00	296.0
60.30	208.0
61.00	942.0
62.00	1535.0
63.00	3130.0
64.00	593.0
65.00	1965.0
66.90	229.0
68.00	1793.0
68.90	73360.0
70.00	505.0
71.00	292.0
73.10	794.0
74.00	7984.0
75.00	12459.0
76.00	4253.0
77.00	88312.0
78.00	6218.0
79.00	4690.0
80.00	4274.0
81.00	6291.0
82.00	1657.0
83.00	1676.0
84.10	319.0
85.00	1045.0
86.00	2196.0
86.90	513.0
88.00	748.0
90.10	274.0
90.90	1848.0
92.00	1486.0
93.00	9351.0
94.00	736.0
95.00	291.0
96.10	459.0
98.00	6208.0
99.00	6081.0
99.90	480.0
101.00	3789.0
102.00	366.0
102.80	1097.0
104.00	2512.0
104.90	2167.0
107.00	27528.0
108.00	4363.0
110.00	50712.0
111.00	8036.0
112.00	850.0
113.00	325.0
114.90	166.0
115.90	1576.0
117.00	18144.0
117.90	1418.0
118.70	277.0
120.00	398.0
122.00	1757.0
123.00	2931.0
123.90	1243.0
125.00	827.0
127.00	102872.0
128.00	8482.0
128.90	36952.0
130.00	3430.0
131.00	964.0
132.00	309.0
132.90	182.0
134.10	1352.0
134.90	3222.0
135.90	1476.0
137.00	2219.0
137.90	397.0
140.00	669.0
141.00	4594.0

Instrument :
BNA_G
ClientSampleId :
DFTPP66

142.00	1408.0
143.00	1009.0
144.00	264.0
144.90	370.0
145.90	1503.0
147.00	2440.0
147.90	4667.0
148.90	1357.0
149.90	173.0
150.80	376.0
151.80	856.0
153.00	1706.0
153.90	1364.0
155.00	2656.0
156.00	3943.0
157.00	731.0
158.00	1136.0
158.90	769.0
160.00	1790.0
161.00	2306.0
161.70	725.0
163.10	176.0
164.10	244.0
165.00	1781.0
165.90	1867.0
167.00	7592.0
168.00	4885.0
168.90	740.0
169.80	166.0
170.80	606.0
171.90	920.0
172.90	994.0
174.00	2070.0
175.00	3613.0
176.00	977.0
177.00	1537.0
177.90	516.0
178.90	6594.0
180.00	4368.0
181.00	1709.0
182.00	664.0
182.60	225.0
182.90	187.0
183.80	441.0
185.00	3176.0
186.00	24672.0
187.00	6869.0
188.10	739.0
189.00	1599.0
190.00	250.0
191.00	644.0
192.00	2428.0
193.00	2183.0
193.90	337.0
194.60	233.0
196.00	5825.0
197.90	179648.0
198.90	12850.0
199.90	1058.0
201.40	842.0
203.00	1064.0
204.00	6084.0
205.00	11702.0
206.00	42640.0
207.00	6120.0
208.00	1441.0
209.00	547.0
210.00	751.0
210.90	1422.0
211.70	432.0
214.90	485.0
215.80	682.0
216.90	11449.0
217.90	1336.0
218.90	269.0
219.20	193.0
219.90	544.0
221.00	7148.0
223.00	2419.0
224.00	22432.0
225.00	5712.0
225.90	481.0
227.00	8972.0
227.90	1265.0
228.90	1771.0
231.00	948.0
231.80	280.0
232.80	153.0
234.00	880.0
234.90	388.0
236.00	301.0
236.90	813.0
238.90	437.0
239.80	271.0
240.90	423.0

Instrument :
BNA_G
ClientSampleId :
DFTPP66

242.00	1439.0
243.00	1782.0
244.00	18048.0
245.00	2651.0
245.90	2890.0
247.00	878.0
248.00	434.0
249.00	668.0
249.90	197.0
250.90	260.0
251.80	384.0
253.00	761.0
255.00	89288.0
256.00	13568.0
257.00	853.0
258.00	4348.0
259.00	1103.0
263.10	182.0
264.90	2412.0
266.00	284.0
269.90	171.0
270.90	201.0
271.90	382.0
273.00	3225.0
273.90	7333.0
275.00	43648.0
276.00	6819.0
276.90	3778.0
278.00	671.0
282.90	278.0
284.00	472.0
285.00	612.0
291.90	204.0
292.80	690.0
293.70	216.0
294.10	172.0
294.90	462.0
295.90	11481.0
296.90	1449.0
298.00	219.0
302.00	248.0
303.00	1189.0
304.10	472.0
309.90	152.0
313.90	435.0
315.00	1166.0
316.00	829.0
316.90	371.0
321.00	361.0
323.00	3905.0
324.00	809.0
325.10	178.0
326.90	686.0
327.90	296.0
331.90	304.0
332.80	382.0
334.00	2680.0
335.00	862.0
341.00	353.0
345.90	819.0
352.00	981.0
352.90	734.0
354.00	1275.0
355.10	238.0
364.90	4121.0
366.00	744.0
370.80	171.0
372.00	1768.0
372.90	533.0
382.90	615.0
392.00	251.0
402.00	864.0
403.00	783.0
404.00	366.0
415.20	173.0
420.90	999.0
422.20	559.0
423.00	5004.0
424.00	1356.0
437.70	178.0
438.50	208.0
441.00	16768.0
442.00	99808.0
443.00	20552.0
444.00	1284.0
444.80	175.0

Instrument :
BNA_G
ClientSampleId :
DFTPP66

m/z	Abundance
37.90	536.0
39.00	4620.0
39.90	613.0
41.00	347.0
44.00	1568.0
45.00	175.0
49.00	567.0
50.00	15310.0
51.00	63464.0
52.00	3782.0
55.00	491.0
55.90	2493.0
57.00	5599.0
57.90	463.0
60.90	985.0
61.90	1295.0
63.00	3487.0
64.00	366.0
65.00	1810.0
66.00	237.0
67.10	323.0
68.90	68696.0
70.00	448.0
73.00	770.0
74.00	6813.0
75.00	11295.0
76.00	4893.0
77.00	89064.0
78.00	5372.0
79.00	5624.0
80.00	3609.0
81.00	5779.0
82.00	1642.0
83.00	1029.0
83.90	552.0
84.90	736.0
85.90	1486.0
87.00	673.0
87.80	296.0
89.00	258.0
89.90	154.0
91.00	1322.0
92.00	1730.0
92.90	9074.0
94.00	902.0
95.00	216.0
95.90	731.0
98.00	6262.0
98.90	6236.0
100.00	457.0
101.00	4012.0
101.90	294.0
102.90	1042.0
103.90	2273.0
104.90	2055.0
107.00	28616.0
108.00	4498.0
109.90	56080.0
110.90	8354.0
112.00	998.0
112.90	269.0
113.40	203.0
114.70	346.0
115.90	1590.0
117.00	16888.0
118.00	1550.0
118.90	200.0
120.00	614.0
122.00	1946.0
123.00	3409.0
123.90	1101.0
124.90	1546.0
127.00	105224.0
128.00	9220.0
128.90	38528.0
129.90	3710.0
131.00	1122.0
132.00	477.0
132.90	164.0
133.90	1400.0
134.90	3624.0
135.90	1516.0
136.90	2158.0
137.90	503.0
139.00	372.0
140.00	818.0
140.90	4963.0
141.90	1843.0
142.90	1190.0
144.00	494.0
146.00	1176.0

Instrument :
BNA_G
ClientSampleId :
DFTPP66

147.00	2801.0
147.90	5717.0
148.90	1223.0
150.00	348.0
150.90	720.0
151.60	489.0
153.00	2190.0
154.00	1346.0
155.00	3534.0
156.00	4297.0
157.00	806.0
157.90	1025.0
158.80	427.0
159.90	1508.0
160.90	3092.0
162.00	860.0
162.90	403.0
164.00	843.0
164.90	2251.0
166.00	1488.0
167.00	8477.0
167.90	4286.0
169.00	697.0
169.90	403.0
171.00	507.0
171.90	1134.0
172.90	1191.0
174.00	2489.0
175.00	4486.0
175.90	928.0
177.00	1606.0
177.90	790.0
178.90	7599.0
179.90	5049.0
180.90	2181.0
181.90	457.0
183.10	159.0
183.90	608.0
185.00	4252.0
186.00	27816.0
187.00	7414.0
188.00	849.0
188.90	1772.0
189.90	255.0
191.00	984.0
192.00	2021.0
193.00	2577.0
194.00	696.0
195.00	903.0
195.90	5432.0
197.90	202048.0
198.90	15420.0
199.90	1135.0
200.80	507.0
201.60	922.0
202.90	1296.0
204.00	6806.0
205.00	12852.0
206.00	53136.0
207.00	7426.0
207.90	2152.0
208.90	719.0
210.00	797.0
211.00	2076.0
211.90	505.0
213.20	237.0
214.80	510.0
216.00	1100.0
216.90	12687.0
217.90	1814.0
219.80	434.0
221.00	7696.0
222.90	2770.0
224.00	26288.0
225.00	7125.0
226.00	939.0
227.00	9830.0
227.90	1444.0
228.90	2680.0
230.00	375.0
231.00	956.0
234.00	648.0
234.90	944.0
235.90	385.0
236.90	1077.0
237.90	159.0
239.10	502.0
240.00	409.0
241.00	903.0
242.00	1278.0
243.00	2094.0
244.00	22424.0
245.00	3233.0
245.90	3708.0
246.90	1006.0

Instrument :
BNA_G
ClientSampleId :
DFTPP66

247.90	243.0
248.90	1192.0
249.80	435.0
251.00	250.0
251.90	361.0
253.00	930.0
255.00	108104.0
256.00	16600.0
257.00	1330.0
257.90	6379.0
259.00	1169.0
259.90	216.0
261.00	169.0
263.80	282.0
265.00	2179.0
266.00	360.0
269.90	200.0
271.10	489.0
272.00	490.0
272.90	3764.0
274.00	10398.0
275.00	50800.0
276.00	7092.0
277.00	4005.0
277.90	662.0
278.90	296.0
282.20	235.0
282.90	795.0
284.00	248.0
284.90	874.0
290.10	187.0
291.10	177.0
292.10	358.0
293.00	954.0
294.10	569.0
295.00	517.0
296.00	13682.0
297.00	2023.0
297.90	270.0
301.00	278.0
302.00	398.0
303.00	1628.0
304.00	433.0
308.00	286.0
310.00	238.0
313.00	167.0
314.00	493.0
315.00	1607.0
316.00	679.0
317.10	151.0
320.90	405.0
321.90	541.0
323.00	4537.0
323.90	602.0
327.00	722.0
328.00	455.0
332.00	420.0
332.90	339.0
334.00	3220.0
335.00	773.0
340.90	937.0
343.10	165.0
345.90	900.0
352.00	1720.0
353.00	1162.0
354.00	1890.0
355.00	534.0
359.00	192.0
364.10	196.0
364.90	7550.0
366.10	969.0
369.90	244.0
371.10	383.0
372.00	2637.0
372.90	609.0
377.70	189.0
383.00	748.0
384.00	335.0
385.00	255.0
389.40	173.0
390.00	415.0
391.10	181.0
400.70	185.0
402.00	1254.0
403.00	1412.0
403.90	782.0
405.10	325.0
421.00	1181.0
422.00	1479.0
423.00	8665.0
424.00	1792.0
425.10	198.0
437.30	225.0
439.50	185.0
441.00	24928.0

Instrument :
BNA_G
ClientSampleId :
DFTPP66

442.00	150208.0
443.00	30808.0
444.00	2794.0
445.00	163.0

Instrument :
BNA_G
ClientSampleId :
DFTPP66

m/z	Abundance
38.00	273.0
39.00	2743.0
41.00	170.0
42.70	158.0
43.10	248.0
44.00	1252.0
50.00	10903.0
51.00	40712.0
52.00	1723.0
55.00	416.0
56.00	1338.0
57.00	3522.0
57.90	254.0
59.00	276.0
61.00	640.0
62.00	611.0
63.00	2614.0
63.90	247.0
65.00	837.0
65.90	167.0
67.10	374.0
68.90	43880.0
69.90	325.0
71.00	295.0
73.00	728.0
74.00	4861.0
75.00	7888.0
76.00	2744.0
77.00	56120.0
78.00	3794.0
79.00	3368.0
79.90	2930.0
80.90	3247.0
81.90	986.0
83.00	772.0
83.70	396.0
83.90	392.0
84.90	651.0
86.00	1147.0
86.90	404.0
87.90	170.0
91.00	761.0
92.00	1053.0
93.00	5485.0
94.00	568.0
95.10	250.0
95.90	328.0
96.70	181.0
98.00	4492.0
99.00	3657.0
101.00	2568.0
102.00	199.0
102.90	745.0
104.00	1488.0
105.00	1809.0
105.90	295.0
107.00	16680.0
108.00	2784.0
109.00	1044.0
109.90	35128.0
110.90	4942.0
111.90	667.0
113.00	186.0
114.20	151.0
115.00	220.0
115.90	681.0
116.90	11013.0
117.90	966.0
118.90	242.0
119.90	170.0
122.00	979.0
123.00	1765.0
123.90	897.0
125.00	1019.0
127.00	69648.0
128.00	5852.0
128.90	25752.0
129.90	2177.0
131.00	503.0
131.80	355.0
133.00	217.0
133.90	917.0
135.00	2597.0
136.00	701.0
136.90	1460.0
137.60	634.0
139.00	263.0
140.90	3106.0
141.90	1398.0
143.10	818.0
144.10	219.0

Instrument :
BNA_G
ClientSampleId :
DFTPP66

146.00	781.0
147.00	2068.0
148.00	2780.0
148.90	1006.0
150.10	256.0
151.30	534.0
153.00	1289.0
154.00	666.0
155.00	1640.0
156.00	3264.0
156.90	655.0
157.90	1238.0
159.00	437.0
160.00	913.0
161.00	1575.0
161.90	412.0
163.00	326.0
163.90	191.0
165.00	1384.0
166.00	1137.0
167.00	6809.0
168.00	3683.0
169.00	567.0
170.00	373.0
171.00	357.0
171.90	505.0
172.90	1011.0
174.00	1494.0
175.00	3424.0
175.90	1048.0
176.90	1179.0
178.00	607.0
178.90	5089.0
180.00	3857.0
181.00	1952.0
181.90	409.0
183.00	182.0
184.00	540.0
185.00	2787.0
186.00	18864.0
187.00	5146.0
188.00	706.0
189.00	816.0
189.90	450.0
191.00	922.0
191.90	1684.0
193.00	1617.0
194.00	529.0
195.00	648.0
196.00	4408.0
197.90	143488.0
198.90	10425.0
200.00	582.0
201.20	712.0
201.90	514.0
203.00	1029.0
204.00	5468.0
205.00	9070.0
206.00	37472.0
207.00	5228.0
208.00	1305.0
208.80	424.0
210.00	750.0
211.00	1577.0
211.70	308.0
214.90	492.0
215.90	979.0
216.90	9097.0
218.00	1555.0
219.00	271.0
221.00	5762.0
221.70	1361.0
222.90	2327.0
224.00	18208.0
225.00	5012.0
226.00	469.0
227.00	8179.0
228.00	1135.0
228.90	1968.0
229.90	183.0
231.00	1020.0
233.10	223.0
234.00	804.0
235.00	776.0
235.80	643.0
237.00	998.0
238.90	461.0
239.90	178.0
240.90	338.0
242.00	1077.0
243.00	1061.0
244.00	16672.0
245.00	1954.0
245.90	2887.0
246.90	714.0

Instrument :
BNA_G
ClientSampleId :
DFTPP66

248.00	204.0
248.90	595.0
250.90	296.0
252.00	256.0
252.90	480.0
255.00	83288.0
256.00	11230.0
257.00	982.0
257.90	5012.0
259.00	1022.0
259.90	259.0
265.00	1832.0
265.90	256.0
270.10	208.0
272.10	341.0
273.00	3201.0
274.00	6118.0
275.00	39976.0
276.00	4976.0
277.00	3913.0
277.90	613.0
281.10	234.0
281.90	157.0
282.90	317.0
285.00	517.0
286.10	176.0
291.10	150.0
291.90	268.0
292.90	696.0
294.90	195.0
296.00	11110.0
297.00	1648.0
301.00	175.0
303.00	1480.0
304.00	468.0
305.10	170.0
307.90	229.0
309.90	206.0
313.10	332.0
314.00	442.0
314.90	1132.0
316.00	769.0
320.90	304.0
323.00	3422.0
323.90	509.0
326.90	703.0
327.90	368.0
332.00	337.0
333.00	528.0
334.00	3059.0
335.00	497.0
339.00	191.0
341.00	524.0
342.00	247.0
345.90	877.0
346.70	220.0
352.00	1231.0
353.00	905.0
354.00	1007.0
355.00	321.0
359.00	215.0
364.90	5806.0
365.90	1073.0
367.00	187.0
371.00	277.0
372.00	2371.0
373.00	750.0
382.90	557.0
384.00	304.0
390.10	231.0
390.90	277.0
394.80	168.0
402.00	1200.0
403.00	1121.0
404.00	478.0
421.00	963.0
422.00	1068.0
423.00	6923.0
424.00	1499.0
426.80	158.0
434.90	153.0
438.00	229.0
438.60	226.0
439.90	240.0
441.00	22528.0
442.00	125520.0
443.00	23456.0
444.00	2711.0
445.00	280.0
483.70	153.0

Instrument :
BNA_G
ClientSampleId :
DFTPP66

m/z	Abundance
39.90	246.0
43.90	1268.0
49.00	228.0
55.10	267.0
56.90	177.0
67.00	211.0
70.90	194.0
73.10	239.0
81.10	152.0
91.00	280.0
97.00	217.0
126.10	172.0
135.10	190.0
136.90	256.0
147.10	161.0
206.10	173.0
207.00	643.0
276.10	273.0

Instrument :
BNA_G
ClientSampleId :
DFTPP66