

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG052322\
 Data File : BG053660.D
 Acq On : 23 May 2022 12:01
 Operator : CG/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 SSTDCCC040

Manual Integrations
 APPROVED

Reviewed By : Christian
 Giraldo

Quant Time: May 24 05:08:48 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG050522.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue May 24 05:05:32 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)	
Internal Standards							05/24/2022
1) 1,4-Dichlorobenzene-d4	8.058	152	32181	20.000	ng	0.00	Supervised By : Jagrut Padhyay
21) Naphthalene-d8	10.872	136	123885	20.000	ng	0.00	
39) Acenaphthene-d10	14.690	164	93024	20.000	ng	0.00	
64) Phenanthrene-d10	17.439	188	216891	20.000	ng	0.00	
76) Chrysene-d12	21.716	240	187606	20.000	ng	0.00	
86) Perylene-d12	24.982	264	187219	20.000	ng	0.00	05/24/2022
System Monitoring Compounds							
5) 2-Fluorophenol	5.626	112	149070	78.615	ng	0.00	
7) Phenol-d6	7.229	99	227060	78.901	ng	0.00	
23) Nitrobenzene-d5	9.227	82	242636	83.373	ng	0.00	
42) 2,4,6-Tribromophenol	16.182	330	111145	81.010	ng	0.00	
45) 2-Fluorobiphenyl	13.315	172	533005	83.728	ng	0.00	
79) Terphenyl-d14	20.041	244	868925	86.263	ng	0.00	
Target Compounds							Qvalue
2) 1,4-Dioxane	3.493	88	33546	33.718	ng		95
3) Pyridine	3.899	79	91546	37.678	ng		93
4) n-Nitrosodimethylamine	3.810	42	44616	37.696	ng	#	98
6) Aniline	7.382	93	126467	39.678	ng		99
8) 2-Chlorophenol	7.629	128	78751	39.727	ng		96
9) Benzaldehyde	7.194	77	62783m	34.341	ng		
10) Phenol	7.253	94	111880	39.751	ng		95
11) bis(2-Chloroethyl)ether	7.470	93	82976	39.529	ng		98
12) 1,3-Dichlorobenzene	7.952	146	92109	39.388	ng		95
13) 1,4-Dichlorobenzene	8.093	146	91636	38.832	ng		96
14) 1,2-Dichlorobenzene	8.416	146	90109	40.663	ng		96
15) Benzyl Alcohol	8.299	79	93121	38.705	ng		97
16) 2,2'-oxybis(1-Chloropr...	8.586	45	133314	39.022	ng		99
17) 2-Methylphenol	8.510	107	75902	39.900	ng		96
18) Hexachloroethane	9.144	117	34404	38.391	ng		95
19) n-Nitroso-di-n-propyla...	8.862	70	80473	38.974	ng		98
20) 3+4-Methylphenols	8.839	107	103961	38.695	ng		94
22) Acetophenone	8.880	105	140844	41.540	ng	#	99
24) Nitrobenzene	9.268	77	114835	40.768	ng		98
25) Isophorone	9.791	82	202579	41.361	ng		100
26) 2-Nitrophenol	9.984	139	48531	42.540	ng		95
27) 2,4-Dimethylphenol	10.043	122	69612	40.545	ng		94
28) bis(2-Chloroethoxy)met...	10.266	93	112620	39.980	ng		98
29) 2,4-Dichlorophenol	10.531	162	87634	42.292	ng		91
30) 1,2,4-Trichlorobenzene	10.736	180	100198	42.559	ng		98
31) Naphthalene	10.924	128	263435	41.804	ng		100
32) Benzoic acid	10.225	122	34834m	35.205	ng		
33) 4-Chloroaniline	11.030	127	110409	41.836	ng		98
34) Hexachlorobutadiene	11.206	225	76972	44.090	ng		95
35) Caprolactam	11.806	113	29298	38.908	ng		92
36) 4-Chloro-3-methylphenol	12.158	107	98493	39.737	ng		98
37) 2-Methylnaphthalene	12.522	142	190617	40.532	ng		99
38) 1-Methylnaphthalene	12.740	142	185583	40.446	ng		98
40) 1,2,4,5-Tetrachloroben...	12.892	216	125000	42.630	ng		100
41) Hexachlorocyclopentadiene	12.869	237	30525	31.571	ng		98
43) 2,4,6-Trichlorophenol	13.133	196	81491	42.576	ng		97

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG052322\
 Data File : BG053660.D
 Acq On : 23 May 2022 12:01
 Operator : CG/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 SSTDCCC040

Manual Integrations
 APPROVED

Reviewed By : Christian
 Giraldo

Quant Time: May 24 05:08:48 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG050522.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue May 24 05:05:32 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)	
44) 2,4,5-Trichlorophenol	13.215	196	86259	42.344	ng	98	05/24/2022
46) 1,1'-Biphenyl	13.521	154	264942	41.783	ng	98	Supervised By :Jagrut Upadhyay
47) 2-Chloronaphthalene	13.574	162	205630	41.392	ng	99	
48) 2-Nitroaniline	13.773	65	77990	42.131	ng	97	
49) Acenaphthylene	14.414	152	321769	40.589	ng	99	
50) Dimethylphthalate	14.144	163	276096	39.058	ng	99	
51) 2,6-Dinitrotoluene	14.261	165	57663	40.039	ng	95	05/24/2022
52) Acenaphthene	14.755	154	190825	40.392	ng	99	
53) 3-Nitroaniline	14.596	138	62457	40.137	ng	96	
54) 2,4-Dinitrophenol	14.819	184	30865	37.372	ng	92	
55) Dibenzofuran	15.089	168	338096	39.842	ng	98	
56) 4-Nitrophenol	14.925	139	41164	37.037	ng	90	
57) 2,4-Dinitrotoluene	15.054	165	84353	40.408	ng	92	
58) Fluorene	15.736	166	269526	39.579	ng	97	
59) 2,3,4,6-Tetrachlorophenol	15.324	232	79514	38.471	ng	98	
60) Diethylphthalate	15.495	149	285050	39.037	ng	99	
61) 4-Chlorophenyl-phenyle...	15.724	204	150536	39.615	ng	98	
62) 4-Nitroaniline	15.759	138	64249	39.808	ng	99	
63) Azobenzene	16.017	77	295684	39.544	ng	98	
65) 4,6-Dinitro-2-methylph...	15.830	198	57447	44.450	ng	94	
66) n-Nitrosodiphenylamine	15.941	169	239118	41.991	ng	99	
67) 4-Bromophenyl-phenylether	16.617	248	107297	42.359	ng	98	
68) Hexachlorobenzene	16.752	284	113976	41.085	ng	97	
69) Atrazine	16.881	200	92439	38.560	ng	97	
70) Pentachlorophenol	17.098	266	50493	36.484	ng	96	
71) Phenanthrene	17.480	178	445936	40.435	ng	97	
72) Anthracene	17.568	178	441062	40.688	ng	99	
73) Carbazole	17.839	167	425822	40.082	ng	100	
74) Di-n-butylphthalate	18.385	149	479044	40.093	ng	99	
75) Fluoranthene	19.489	202	540541	37.853	ng	98	
77) Benzidine	19.660	184	202009	49.958	ng	98	
78) Pyrene	19.853	202	541799	44.612	ng	99	
80) Butylbenzylphthalate	20.723	149	193613	43.475	ng	96	
81) Benzo(a)anthracene	21.698	228	484044	40.452	ng	99	
82) 3,3'-Dichlorobenzidine	21.604	252	131205	43.239	ng	96	
83) Chrysene	21.763	228	462899	41.420	ng	98	
84) Bis(2-ethylhexyl)phtha...	21.581	149	258748	42.095	ng	98	
85) Di-n-octyl phthalate	22.802	149	430803	41.447	ng	99	
87) Indeno(1,2,3-cd)pyrene	28.730	276	539848	41.401	ng	# 95	
88) Benzo(b)fluoranthene	23.942	252	457116	41.249	ng	99	
89) Benzo(k)fluoranthene	24.013	252	457822	42.042	ng	96	
90) Benzo(a)pyrene	24.829	252	389430	41.287	ng	99	
91) Dibenzo(a,h)anthracene	28.783	278	446821	41.586	ng	98	
92) Benzo(g,h,i)perylene	29.905	276	439220	41.240	ng	95	

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG052322\
Data File : BG053660.D
Acq On : 23 May 2022 12:01
Operator : CG/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1


Instrument :
BNA_G
ClientSampleId :
SSTDCCC040

Manual Integrations
APPROVED

Reviewed By :Christian
Giraldo

Quant Time: May 24 05:08:48 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG050522.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue May 24 05:05:32 2022
Response via : Initial Calibration






#1
1,4-Dichlorobenzene-d4
Concen: 20.000 ng
RT: 8.058 min Scan# 79
Delta R.T. -0.000 min
Lab File: BG053660.D
Acq: 23 May 2022 12:01

Instrument :

BNA_G

ClientSampleId :


SSTDCCC040



Tgt Ion:152 Resp: 3218
Ion Ratio Lower Upper
152 100
150 152.9 131.6 197.4
115 57.3 50.9 76.3


Manual Integrations
APPROVED

Reviewed By :Christian
Giraldo




05/24/2022
Supervised By :Jagrut
Upadhyay



05/24/2022




#2
1,4-Dioxane
Concen: 33.718 ng
RT: 3.493 min Scan# 18
Delta R.T. 0.000 min
Lab File: BG053660.D
Acq: 23 May 2022 12:01



Tgt Ion: 88 Resp: 33546
Ion Ratio Lower Upper
88 100
58 78.4 59.2 88.8
43 33.1 24.3 36.5







#3
Pyridine
Concen: 37.678 ng
RT: 3.899 min Scan# 81
Delta R.T. 0.000 min
Lab File: BG053660.D
Acq: 23 May 2022 12:01

Instrument :
BNA_G
ClientSampleId :
SSTDCCC040

Tgt Ion: 79 Resp: 91540
Ion Ratio Lower Upper
79 100
52 77.2 54.6 81.8
51 41.4 32.5 48.7



Manual Integrations
APPROVED

Reviewed By :Christian
Giraldo






05/24/2022
Supervised By :Jagrut
Upadhyay


05/24/2022



#4
n-Nitrosodimethylamine
Concen: 37.696 ng
RT: 3.810 min Scan# 72
Delta R.T. 0.000 min
Lab File: BG053660.D
Acq: 23 May 2022 12:01

Tgt Ion: 42 Resp: 44616
Ion Ratio Lower Upper
42 100
74 119.4 97.0 145.4
44 7.0 7.0 10.4






#5
2-Fluorophenol
Concen: 78.615 ng
RT: 5.626 min Scan# 388
Delta R.T. 0.000 min
Lab File: BG053660.D
Acq: 23 May 2022 12:01

Instrument :

BNA_G

ClientSampleId :



SSTDCCC040



Tgt Ion: 112 Resp: 149070
Ion Ratio Lower Upper
112 100
64 69.5 55.1 82.7
63 40.8 30.4 45.6


Manual Integrations
APPROVED

Reviewed By :Christian
Giraldo




05/24/2022
Supervised By :Jagrut
Upadhyay



05/24/2022




#6
Aniline
Concen: 39.678 ng
RT: 7.382 min Scan# 680
Delta R.T. 0.000 min
Lab File: BG053660.D
Acq: 23 May 2022 12:01




Tgt Ion: 93 Resp: 126467
Ion Ratio Lower Upper
93 100
66 41.7 32.9 49.3
65 21.8 17.8 26.8





#7
Phenol-d6
Concen: 78.901 ng
RT: 7.229 min Scan# 611
Delta R.T. 0.000 min
Lab File: BG053660.D
Acq: 23 May 2022 12:01



Instrument :
BNA_G
ClientSampleId :
SSTDCCC040



Tgt Ion: 99 Resp: 227060
Ion Ratio Lower Upper
99 100
42 20.8 17.0 25.6
71 43.4 33.8 50.8


Manual Integrations
APPROVED

Reviewed By :Christian
Giraldo




05/24/2022
Supervised By :Jagrut
Upadhyay



05/24/2022




#8
2-Chlorophenol
Concen: 39.727 ng
RT: 7.629 min Scan# 722
Delta R.T. 0.000 min
Lab File: BG053660.D
Acq: 23 May 2022 12:01



Tgt Ion:128 Resp: 78751
Ion Ratio Lower Upper
128 100
130 33.5 10.5 50.5
64 57.4 35.1 75.1





#9
Benzaldehyde
Concen: 34.341 ng m
RT: 7.194 min Scan# 648
Delta R.T. 0.000 min
Lab File: BG053660.D
Acq: 23 May 2022 12:01

Instrument :

BNA_G



ClientSampleId :

SSTDCCC040

Tgt Ion: 77 Resp: 6278
Ion Ratio Lower Upper
77 100
105 93.4 59.6 99.6
106 85.9 56.8 96.8



Manual Integrations
APPROVED

Reviewed By :Christian
Giraldo






05/24/2022
Supervised By :Jagrut
Upadhyay


05/24/2022



#10
Phenol
Concen: 39.751 ng
RT: 7.253 min Scan# 658
Delta R.T. 0.000 min
Lab File: BG053660.D
Acq: 23 May 2022 12:01

Tgt Ion: 94 Resp: 111880
Ion Ratio Lower Upper
94 100
65 30.9 14.8 54.8
66 47.8 25.2 65.2






#11
bis(2-Chloroethyl)ether
Concen: 39.529 ng
RT: 7.470 min Scan# 695
Delta R.T. 0.000 min
Lab File: BG053660.D
Acq: 23 May 2022 12:01

Instrument :

BNA_G

ClientSampleId :



SSTDCCC040



Tgt Ion: 93 Resp: 82970
Ion Ratio Lower Upper
93 100
63 82.8 63.6 103.6
95 32.8 10.1 50.1


Manual Integrations
APPROVED

Reviewed By :Christian
Giraldo




05/24/2022
Supervised By :Jagrut
Upadhyay



05/24/2022




#12
1,3-Dichlorobenzene
Concen: 39.388 ng
RT: 7.952 min Scan# 777
Delta R.T. 0.000 min
Lab File: BG053660.D
Acq: 23 May 2022 12:01



Tgt Ion:146 Resp: 92109
Ion Ratio Lower Upper
146 100
148 61.0 46.1 69.1
75 38.3 27.6 41.4






#13
1,4-Dichlorobenzene
Concen: 38.832 ng
RT: 8.093 min Scan# 80
Delta R.T. 0.000 min
Lab File: BG053660.D
Acq: 23 May 2022 12:01

Instrument :

BNA_G

ClientSampleId :


SSTDCCC040



Tgt Ion:146 Resp: 91630
Ion Ratio Lower Upper
146 100
148 59.1 51.5 77.3
111 42.0 33.9 50.9



Manual Integrations
APPROVED

Reviewed By :Christian
Giraldo






05/24/2022
Supervised By :Jagrut
Upadhyay


05/24/2022



#14
1,2-Dichlorobenzene
Concen: 40.663 ng
RT: 8.416 min Scan# 856
Delta R.T. 0.000 min
Lab File: BG053660.D
Acq: 23 May 2022 12:01

Tgt Ion:146 Resp: 90109
Ion Ratio Lower Upper
146 100
148 61.8 52.6 78.8
111 44.6 34.5 51.7





#15
Benzyl Alcohol
Concen: 38.705 ng
RT: 8.299 min Scan# 81
Delta R.T. 0.000 min
Lab File: BG053660.D
Acq: 23 May 2022 12:01

Instrument :

BNA_G



ClientSampleId :

SSTDCCC040

Tgt Ion: 79 Resp: 9312
Ion Ratio Lower Upper
79 100
108 63.9 53.7 80.5
77 67.1 52.2 78.4



Manual Integrations
APPROVED

Reviewed By :Christian
Giraldo






05/24/2022
Supervised By :Jagrut
Upadhyay


05/24/2022



#16
2,2'-oxybis(1-Chloropropane)
Concen: 39.022 ng
RT: 8.586 min Scan# 885
Delta R.T. 0.000 min
Lab File: BG053660.D
Acq: 23 May 2022 12:01

Tgt Ion: 45 Resp: 133314
Ion Ratio Lower Upper
45 100
77 13.5 0.0 32.7
79 9.7 0.0 29.7






#17
2-Methylphenol
Concen: 39.900 ng
RT: 8.510 min Scan# 81
Delta R.T. 0.000 min
Lab File: BG053660.D
Acq: 23 May 2022 12:01

Instrument :

BNA_G

ClientSampleId :



SSTDCCC040



Tgt Ion:107 Resp: 7590
Ion Ratio Lower Upper
107 100
108 111.5 85.4 128.0
77 60.2 46.2 69.2
79 56.2 46.1 69.1


Manual Integrations
APPROVED

Reviewed By :Christian
Giraldo




05/24/2022
Supervised By :Jagrut
Upadhyay



05/24/2022




#18
Hexachloroethane
Concen: 38.391 ng
RT: 9.144 min Scan# 980
Delta R.T. 0.000 min
Lab File: BG053660.D
Acq: 23 May 2022 12:01



Tgt Ion:117 Resp: 34404
Ion Ratio Lower Upper
117 100
119 97.7 75.0 112.6
201 118.2 89.0 133.6





#19
n-Nitroso-di-n-propylamine
Concen: 38.974 ng
RT: 8.862 min Scan# 911
Delta R.T. 0.000 min
Lab File: BG053660.D
Acq: 23 May 2022 12:01

Instrument :

BNA_G



ClientSampleId :

SSTDCCC040

Tgt Ion: 70 Resp: 8047
Ion Ratio Lower Upper
70 100
42 61.0 47.3 70.9
101 8.9 7.0 10.4
130 18.0 14.6 22.0



Manual Integrations
APPROVED

Reviewed By :Christian
Giraldo






05/24/2022
Supervised By :Jagrut
Upadhyay


05/24/2022



#20
3+4-Methylphenols
Concen: 38.695 ng
RT: 8.839 min Scan# 928
Delta R.T. 0.000 min
Lab File: BG053660.D
Acq: 23 May 2022 12:01

Tgt Ion:107 Resp: 103961
Ion Ratio Lower Upper
107 100
108 96.2 67.5 107.5
77 41.0 20.3 60.3
79 35.6 16.7 56.7






#21
Naphthalene-d8
Concen: 20.000 ng
RT: 10.872 min Scan# 12388
Delta R.T. 0.000 min
Lab File: BG053660.D
Acq: 23 May 2022 12:01

Instrument :

BNA_G

ClientSampleId :



SSTDCCC040



Tgt Ion:136 Resp: 12388
Ion Ratio Lower Upper
136 100
137 11.5 8.5 12.7
54 11.0 7.4 11.2
68 6.2 4.9 7.3


Manual Integrations
APPROVED

Reviewed By :Christian
Giraldo




05/24/2022
Supervised By :Jagrut
Upadhyay



05/24/2022




#22
Acetophenone
Concen: 41.540 ng
RT: 8.880 min Scan# 935
Delta R.T. 0.000 min
Lab File: BG053660.D
Acq: 23 May 2022 12:01



Tgt Ion:105 Resp: 140844
Ion Ratio Lower Upper
105 100
71 2.9 1.1 1.7#
51 31.9 25.4 38.0
120 20.2 17.3 25.9






#23
Nitrobenzene-d5
Concen: 83.373 ng
RT: 9.227 min Scan# 994
Delta R.T. 0.000 min
Lab File: BG053660.D
Acq: 23 May 2022 12:01

Instrument :

BNA_G

ClientSampleId :



SSTDCCC040



Tgt Ion: 82 Resp: 242630
Ion Ratio Lower Upper
82 100
128 37.2 27.9 41.9
54 51.4 41.2 61.8


Manual Integrations
APPROVED

Reviewed By :Christian
Giraldo




05/24/2022
Supervised By :Jagrut
Upadhyay



05/24/2022




#24
Nitrobenzene
Concen: 40.768 ng
RT: 9.268 min Scan# 1001
Delta R.T. 0.000 min
Lab File: BG053660.D
Acq: 23 May 2022 12:01



Tgt Ion: 77 Resp: 114835
Ion Ratio Lower Upper
77 100
123 35.4 27.4 41.0
65 13.8 11.4 17.0





#25
Isophorone
Concen: 41.361 ng
RT: 9.791 min Scan# 1096
Delta R.T. 0.000 min
Lab File: BG053660.D
Acq: 23 May 2022 12:01

Instrument :

BNA_G



ClientSampleId :

SSTDCCC040

Tgt Ion: 82 Resp: 202579
Ion Ratio Lower Upper
82 100
95 7.6 6.2 9.4
138 14.6 11.6 17.4



Manual Integrations
APPROVED

Reviewed By :Christian
Giraldo






05/24/2022
Supervised By :Jagrut
Upadhyay


05/24/2022



#26
2-Nitrophenol
Concen: 42.540 ng
RT: 9.984 min Scan# 1123
Delta R.T. 0.000 min
Lab File: BG053660.D
Acq: 23 May 2022 12:01

Tgt Ion:139 Resp: 48531
Ion Ratio Lower Upper
139 100
109 40.4 36.1 54.1
65 66.7 55.8 83.6





#27
2,4-Dimethylphenol
Concen: 40.545 ng
RT: 10.043 min Scan# 1133
Delta R.T. 0.000 min
Lab File: BG053660.D
Acq: 23 May 2022 12:01

Instrument :

BNA_G



ClientSampleId :

SSTDCCC040

Tgt Ion:122 Resp: 6961
Ion Ratio Lower Upper
122 100
107 127.4 109.4 164.0
121 58.7 48.7 73.1



Manual Integrations
APPROVED

Reviewed By :Christian
Giraldo






05/24/2022
Supervised By :Jagrut
Upadhyay


05/24/2022



#28
bis(2-Chloroethoxy)methane
Concen: 39.980 ng
RT: 10.266 min Scan# 1171
Delta R.T. 0.000 min
Lab File: BG053660.D
Acq: 23 May 2022 12:01

Tgt Ion: 93 Resp: 112620
Ion Ratio Lower Upper
93 100
95 34.9 26.9 40.3
123 10.3 7.7 11.5






#29
2,4-Dichlorophenol
Concen: 42.292 ng
RT: 10.531 min Scan# 11
Delta R.T. 0.000 min
Lab File: BG053660.D
Acq: 23 May 2022 12:01

Instrument :

BNA_G

ClientSampleId :


SSTDCCC040



Tgt Ion:162 Resp: 87634
Ion Ratio Lower Upper
162 100
164 60.7 46.6 86.6
98 34.8 22.8 62.8


Manual Integrations
APPROVED

Reviewed By :Christian
Giraldo




05/24/2022
Supervised By :Jagrut
Upadhyay



05/24/2022




#30
1,2,4-Trichlorobenzene
Concen: 42.559 ng
RT: 10.736 min Scan# 1251
Delta R.T. 0.000 min
Lab File: BG053660.D
Acq: 23 May 2022 12:01



Tgt Ion:180 Resp: 100198
Ion Ratio Lower Upper
180 100
182 94.2 76.6 115.0
145 28.7 25.0 37.4






#31
Naphthalene
Concen: 41.804 ng
RT: 10.924 min Scan# 1164
Delta R.T. 0.000 min
Lab File: BG053660.D
Acq: 23 May 2022 12:01

Instrument :

BNA_G

ClientSampleId :


SSTDCCC040



Tgt Ion:128 Resp: 26343
Ion Ratio Lower Upper
128 100
129 11.5 9.0 13.6
127 13.9 11.1 16.7


Manual Integrations
APPROVED

Reviewed By :Christian
Giraldo






05/24/2022
Supervised By :Jagrut
Upadhyay


05/24/2022



#32
Benzoic acid
Concen: 35.205 ng m
RT: 10.225 min Scan# 1164
Delta R.T. 0.000 min
Lab File: BG053660.D
Acq: 23 May 2022 12:01

Tgt Ion:122 Resp: 34834
Ion Ratio Lower Upper
122 100
105 142.3 111.8 151.8
77 127.9 109.4 149.4





#33
4-Chloroaniline
Concen: 41.836 ng
RT: 11.030 min Scan# 11040
Delta R.T. 0.000 min
Lab File: BG053660.D
Acq: 23 May 2022 12:01

Instrument :

BNA_G



ClientSampleId :

SSTDCCC040

Tgt Ion:127 Resp: 11040
Ion Ratio Lower Upper
127 100
129 31.9 26.3 39.5
65 41.0 31.1 46.7
92 23.5 19.1 28.7



Manual Integrations
APPROVED

Reviewed By :Christian
Giraldo






05/24/2022
Supervised By :Jagrut
Upadhyay


05/24/2022



#34
Hexachlorobutadiene
Concen: 44.090 ng
RT: 11.206 min Scan# 1331
Delta R.T. 0.000 min
Lab File: BG053660.D
Acq: 23 May 2022 12:01


Tgt Ion:225 Resp: 76972
Ion Ratio Lower Upper
225 100
223 63.2 48.9 73.3
227 60.7 52.8 79.2





#35
Caprolactam
Concen: 38.908 ng
RT: 11.806 min Scan# 1433
Delta R.T. 0.000 min
Lab File: BG053660.D
Acq: 23 May 2022 12:01



Instrument :
BNA_G
ClientSampleId :
SSTDCCC040



Tgt Ion:113 Resp: 29298
Ion Ratio Lower Upper
113 100
55 203.8 169.8 209.8
56 153.6 141.2 181.2


Manual Integrations
APPROVED

Reviewed By :Christian
Giraldo




05/24/2022
Supervised By :Jagrut
Upadhyay



05/24/2022




#36
4-Chloro-3-methylphenol
Concen: 39.737 ng
RT: 12.158 min Scan# 1493
Delta R.T. 0.000 min
Lab File: BG053660.D
Acq: 23 May 2022 12:01



Tgt Ion:107 Resp: 98493
Ion Ratio Lower Upper
107 100
144 23.9 18.5 27.7
142 71.0 55.0 82.4





#37

2-Methylnaphthalene

Concen: 40.532 ng

RT: 12.522 min Scan# 1555

Delta R.T. 0.000 min

Lab File: BG053660.D

Acq: 23 May 2022 12:01

Instrument :

BNA_G

ClientSampleId :

SSTDCCC040

Tgt Ion:142 Resp: 19061

Ion Ratio Lower Upper

142 100



141 90.7 72.3 108.5

115 42.7 33.7 50.5

Manual Integrations



APPROVED

Reviewed By :Christian
Giraldo



05/24/2022
Supervised By :Jagrut
Upadhyay

05/24/2022



#38

1-Methylnaphthalene

Concen: 40.446 ng

RT: 12.740 min Scan# 1592

Delta R.T. 0.000 min

Lab File: BG053660.D

Acq: 23 May 2022 12:01




Tgt Ion:142 Resp: 185583


Ion Ratio Lower Upper

142 100

141 93.5 76.4 114.6

116 4.7 3.3 4.9






#39
Acenaphthene-d10
Concen: 20.000 ng
RT: 14.690 min Scan# 1924
Delta R.T. 0.000 min
Lab File: BG053660.D
Acq: 23 May 2022 12:01

Instrument :

BNA_G

ClientSampleId :


SSTDCCC040



Tgt Ion:164 Resp: 93024
Ion Ratio Lower Upper
164 100
162 105.8 82.3 123.5
160 44.8 34.2 51.2


Manual Integrations
APPROVED

Reviewed By :Christian
Giraldo




05/24/2022
Supervised By :Jagrut
Upadhyay



05/24/2022




#40
1,2,4,5-Tetrachlorobenzene
Concen: 42.630 ng
RT: 12.892 min Scan# 1618
Delta R.T. 0.000 min
Lab File: BG053660.D
Acq: 23 May 2022 12:01




Tgt Ion:216 Resp: 125000
Ion Ratio Lower Upper
216 100
214 78.5 62.9 94.3
179 20.0 16.1 24.1
108 15.1 12.2 18.2





#41
Hexachlorocyclopentadiene
Concen: 31.571 ng
RT: 12.869 min Scan# 10
Delta R.T. 0.000 min
Lab File: BG053660.D
Acq: 23 May 2022 12:01


Instrument :
BNA_G
ClientSampleId :
SSTDCCC040



Tgt Ion:237 Resp: 3052
Ion Ratio Lower Upper
237 100
235 64.2 43.5 83.5
272 13.8 0.0 30.1


Manual Integrations
APPROVED

Reviewed By :Christian
Giraldo




05/24/2022
Supervised By :Jagrut
Upadhyay



05/24/2022




#42
2,4,6-Tribromophenol
Concen: 81.010 ng
RT: 16.182 min Scan# 2178
Delta R.T. 0.000 min
Lab File: BG053660.D
Acq: 23 May 2022 12:01




Tgt Ion:330 Resp: 111145
Ion Ratio Lower Upper
330 100
332 96.6 76.0 114.0
141 32.7 27.1 40.7





#43
2,4,6-Trichlorophenol
Concen: 42.576 ng
RT: 13.133 min Scan# 1659
Delta R.T. 0.000 min
Lab File: BG053660.D
Acq: 23 May 2022 12:01



Instrument :
BNA_G
ClientSampleId :
SSTDCCC040



Tgt Ion:196 Resp: 8149
Ion Ratio Lower Upper
196 100
198 96.0 74.1 111.1
200 29.6 24.7 37.1


Manual Integrations
APPROVED

Reviewed By :Christian
Giraldo






05/24/2022
Supervised By :Jagrut
Upadhyay


05/24/2022



#44
2,4,5-Trichlorophenol
Concen: 42.344 ng
RT: 13.215 min Scan# 1673
Delta R.T. 0.000 min
Lab File: BG053660.D
Acq: 23 May 2022 12:01

Tgt Ion:196 Resp: 86259
Ion Ratio Lower Upper
196 100
198 98.1 76.8 115.2
97 51.9 39.8 59.6
132 23.1 18.4 27.6





#45
2-Fluorobiphenyl
Concen: 83.728 ng
RT: 13.315 min Scan# 1690
Delta R.T. 0.000 min
Lab File: BG053660.D
Acq: 23 May 2022 12:01

Instrument :

BNA_G



ClientSampleId :

SSTDCCC040

Tgt Ion:172 Resp: 533000
Ion Ratio Lower Upper
172 100
171 34.6 28.5 42.7
170 23.5 18.9 28.3



Manual Integrations
APPROVED

Reviewed By :Christian
Giraldo






05/24/2022
Supervised By :Jagrut
Upadhyay


05/24/2022



#46
1,1'-Biphenyl
Concen: 41.783 ng
RT: 13.521 min Scan# 1725
Delta R.T. 0.000 min
Lab File: BG053660.D
Acq: 23 May 2022 12:01

Tgt Ion:154 Resp: 264942
Ion Ratio Lower Upper
154 100
153 40.1 21.6 61.6
76 13.4 0.0 32.6






#47
2-Chloronaphthalene
Concen: 41.392 ng
RT: 13.574 min Scan# 1734
Delta R.T. 0.000 min
Lab File: BG053660.D
Acq: 23 May 2022 12:01

Instrument :

BNA_G

ClientSampleId :


SSTDCCC040



Tgt Ion:162 Resp: 205630
Ion Ratio Lower Upper
162 100
127 38.5 30.6 45.8
164 34.5 26.5 39.7


Manual Integrations
APPROVED

Reviewed By :Christian
Giraldo




05/24/2022
Supervised By :Jagrut
Upadhyay



05/24/2022




#48
2-Nitroaniline
Concen: 42.131 ng
RT: 13.773 min Scan# 1768
Delta R.T. 0.000 min
Lab File: BG053660.D
Acq: 23 May 2022 12:01



Tgt Ion: 65 Resp: 77990
Ion Ratio Lower Upper
65 100
92 61.1 51.5 77.3
138 86.9 71.3 106.9






#49
Acenaphthylene
Concen: 40.589 ng
RT: 14.414 min Scan# 1877
Delta R.T. 0.000 min
Lab File: BG053660.D
Acq: 23 May 2022 12:01

Instrument :

BNA_G

ClientSampleId :



SSTDCCC040



Tgt Ion:152 Resp: 321769
Ion Ratio Lower Upper
152 100
151 20.1 16.4 24.6
153 13.1 10.8 16.2


Manual Integrations
APPROVED

Reviewed By :Christian
Giraldo




05/24/2022
Supervised By :Jagrut
Upadhyay



05/24/2022




#50
Dimethylphthalate
Concen: 39.058 ng
RT: 14.144 min Scan# 1831
Delta R.T. 0.000 min
Lab File: BG053660.D
Acq: 23 May 2022 12:01



Tgt Ion:163 Resp: 276096
Ion Ratio Lower Upper
163 100
194 4.0 3.0 4.6
164 10.4 8.0 12.0






#51
2,6-Dinitrotoluene
Concen: 40.039 ng
RT: 14.261 min Scan# 11
Delta R.T. 0.000 min
Lab File: BG053660.D
Acq: 23 May 2022 12:01

Instrument :

BNA_G

ClientSampleId :


SSTDCCC040



Tgt Ion:165 Resp: 5766
Ion Ratio Lower Upper
165 100
63 63.5 54.1 81.1
89 72.1 54.8 82.2


Manual Integrations
APPROVED

Reviewed By :Christian
Giraldo




05/24/2022
Supervised By :Jagrut
Upadhyay



05/24/2022




#52
Acenaphthene
Concen: 40.392 ng
RT: 14.755 min Scan# 1935
Delta R.T. 0.000 min
Lab File: BG053660.D
Acq: 23 May 2022 12:01




Tgt Ion:154 Resp: 190825
Ion Ratio Lower Upper
154 100
153 115.9 92.7 139.1
152 55.2 42.6 64.0





#53
3-Nitroaniline
Concen: 40.137 ng
RT: 14.596 min Scan# 1908
Delta R.T. 0.000 min
Lab File: BG053660.D
Acq: 23 May 2022 12:01



Instrument :
BNA_G
ClientSampleId :
SSTDCCC040



Tgt Ion:138 Resp: 6245
Ion Ratio Lower Upper
138 100
108 10.7 9.2 13.8
92 123.0 101.9 152.9


Manual Integrations
APPROVED

Reviewed By :Christian
Giraldo




05/24/2022
Supervised By :Jagrut
Upadhyay



05/24/2022




#54
2,4-Dinitrophenol
Concen: 37.372 ng
RT: 14.819 min Scan# 1946
Delta R.T. 0.000 min
Lab File: BG053660.D
Acq: 23 May 2022 12:01



Tgt Ion:184 Resp: 30865
Ion Ratio Lower Upper
184 100
63 68.5 61.1 91.7
154 70.4 60.6 91.0





#55
Dibenzenofuran
Concen: 39.842 ng
RT: 15.089 min Scan# 1997
Delta R.T. 0.000 min
Lab File: BG053660.D
Acq: 23 May 2022 12:01

Instrument :

BNA_G



ClientSampleId :

SSTDCCC040

Tgt Ion:168 Resp: 338090
Ion Ratio Lower Upper
168 100
139 39.8 30.5 45.7
169 13.6 10.3 15.5



Manual Integrations
APPROVED

Reviewed By :Christian
Giraldo






05/24/2022
Supervised By :Jagrut
Upadhyay


05/24/2022



#56
4-Nitrophenol
Concen: 37.037 ng
RT: 14.925 min Scan# 1964
Delta R.T. 0.000 min
Lab File: BG053660.D
Acq: 23 May 2022 12:01

Tgt Ion:139 Resp: 41164
Ion Ratio Lower Upper
139 100
109 114.5 88.5 128.5
65 128.8 93.9 133.9





#57
2,4-Dinitrotoluene
Concen: 40.408 ng
RT: 15.054 min Scan# 1991
Delta R.T. 0.000 min
Lab File: BG053660.D
Acq: 23 May 2022 12:01

Instrument :

BNA_G



ClientSampleId :

SSTDCCC040

Tgt Ion:165 Resp: 8435
Ion Ratio Lower Upper
165 100
63 48.4 34.6 52.0
89 82.4 60.3 90.5
182 2.6 2.2 3.4



Manual Integrations
APPROVED

Reviewed By :Christian
Giraldo






05/24/2022
Supervised By :Jagrut
Upadhyay


05/24/2022



#58
Fluorene
Concen: 39.579 ng
RT: 15.736 min Scan# 2102
Delta R.T. 0.000 min
Lab File: BG053660.D
Acq: 23 May 2022 12:01


Tgt Ion:166 Resp: 269526
Ion Ratio Lower Upper
166 100
165 97.0 75.0 112.6
167 13.9 10.8 16.2





#59
2,3,4,6-Tetrachlorophenol
Concen: 38.471 ng
RT: 15.324 min Scan# 2032
Delta R.T. 0.000 min
Lab File: BG053660.D
Acq: 23 May 2022 12:01



Instrument :
BNA_G
ClientSampleId :
SSTDC040



Tgt Ion:232 Resp: 79514
Ion Ratio Lower Upper
232 100
131 38.9 30.6 46.0
130 2.1 1.8 2.8
166 28.0 21.1 31.7


Manual Integrations
APPROVED

Reviewed By :Christian
Giraldo






05/24/2022
Supervised By :Jagrut
Upadhyay


05/24/2022



#60
Diethylphthalate
Concen: 39.037 ng
RT: 15.495 min Scan# 2061
Delta R.T. 0.000 min
Lab File: BG053660.D
Acq: 23 May 2022 12:01

Tgt Ion:149 Resp: 285050
Ion Ratio Lower Upper
149 100
177 23.7 18.4 27.6
150 13.6 11.2 16.8






#61
4-Chlorophenyl-phenylether
Concen: 39.615 ng
RT: 15.724 min Scan# 2105
Delta R.T. 0.000 min
Lab File: BG053660.D
Acq: 23 May 2022 12:01

Instrument :

BNA_G

ClientSampleId :



SSTDCCC040



Tgt Ion:204 Resp: 150536
Ion Ratio Lower Upper
204 100
206 33.5 27.5 41.3
141 58.0 45.3 67.9


Manual Integrations
APPROVED

Reviewed By :Christian
Giraldo






05/24/2022
Supervised By :Jagrut
Upadhyay


05/24/2022



#62
4-Nitroaniline
Concen: 39.808 ng
RT: 15.759 min Scan# 2106
Delta R.T. 0.000 min
Lab File: BG053660.D
Acq: 23 May 2022 12:01

Tgt Ion:138 Resp: 64249
Ion Ratio Lower Upper
138 100
92 60.5 40.5 80.5
108 125.2 103.7 143.7






#63
Azobenzene
Concen: 39.544 ng
RT: 16.017 min Scan# 2155
Delta R.T. 0.000 min
Lab File: BG053660.D
Acq: 23 May 2022 12:01

Instrument :

BNA_G

ClientSampleId :



SSTDCCC040



Tgt Ion: 77 Resp: 295684
Ion Ratio Lower Upper
77 100
182 24.8 3.8 43.8
105 17.8 0.0 37.8
51 28.7 10.7 50.7


Manual Integrations
APPROVED

Reviewed By :Christian
Giraldo




05/24/2022
Supervised By :Jagrut
Upadhyay



05/24/2022




#64
Phenanthrene-d10
Concen: 20.000 ng
RT: 17.439 min Scan# 2392
Delta R.T. 0.000 min
Lab File: BG053660.D
Acq: 23 May 2022 12:01



Tgt Ion:188 Resp: 216891
Ion Ratio Lower Upper
188 100
94 6.4 5.4 8.0
80 8.1 6.8 10.2





#65
4,6-Dinitro-2-methylphenol
Concen: 44.450 ng
RT: 15.830 min Scan# 2118
Delta R.T. 0.000 min
Lab File: BG053660.D
Acq: 23 May 2022 12:01

Instrument :

BNA_G



ClientSampleId :

SSTDCCC040

Tgt Ion:198 Resp: 5744
Ion Ratio Lower Upper
198 100
51 41.0 28.1 68.1
105 40.4 20.3 60.3



Manual Integrations
APPROVED

Reviewed By :Christian
Giraldo






05/24/2022
Supervised By :Jagrut
Upadhyay


05/24/2022



#66
n-Nitrosodiphenylamine
Concen: 41.991 ng
RT: 15.941 min Scan# 2137
Delta R.T. 0.000 min
Lab File: BG053660.D
Acq: 23 May 2022 12:01

Tgt Ion:169 Resp: 239118
Ion Ratio Lower Upper
169 100
168 70.3 55.9 83.9
167 37.4 30.6 46.0






#67
4-Bromophenyl-phenylether
Concen: 42.359 ng
RT: 16.617 min Scan# 21
Delta R.T. 0.000 min
Lab File: BG053660.D
Acq: 23 May 2022 12:01

Instrument :

BNA_G

ClientSampleId :


SSTDCCC040



Tgt Ion:248 Resp: 10729
Ion Ratio Lower Upper
248 100
250 94.2 75.2 112.8
141 65.2 50.0 75.0


Manual Integrations
APPROVED

Reviewed By :Christian
Giraldo




05/24/2022
Supervised By :Jagrut
Upadhyay



05/24/2022




#68
Hexachlorobenzene
Concen: 41.085 ng
RT: 16.752 min Scan# 2275
Delta R.T. 0.000 min
Lab File: BG053660.D
Acq: 23 May 2022 12:01



Tgt Ion:284 Resp: 113976
Ion Ratio Lower Upper
284 100
142 27.9 23.6 35.4
249 30.6 25.8 38.6





#69
Atrazine
Concen: 38.560 ng
RT: 16.881 min Scan# 21
Delta R.T. 0.000 min
Lab File: BG053660.D
Acq: 23 May 2022 12:01

Instrument :

BNA_G



ClientSampleId :

SSTDCCC040

Tgt Ion:200 Resp: 92439
Ion Ratio Lower Upper
200 100
173 23.8 3.7 43.7
215 48.1 25.2 65.2



Manual Integrations
APPROVED

Reviewed By :Christian
Giraldo






05/24/2022
Supervised By :Jagrut
Upadhyay


05/24/2022



#70
Pentachlorophenol
Concen: 36.484 ng
RT: 17.098 min Scan# 2334
Delta R.T. 0.000 min
Lab File: BG053660.D
Acq: 23 May 2022 12:01


Tgt Ion:266 Resp: 50493
Ion Ratio Lower Upper
266 100
268 65.1 56.2 84.2
264 66.8 53.0 79.4





#71
Phenanthrene
Concen: 40.435 ng
RT: 17.480 min Scan# 21
Delta R.T. 0.000 min
Lab File: BG053660.D
Acq: 23 May 2022 12:01



Instrument :
BNA_G
ClientSampleId :
SSTDCCC040



Tgt Ion:178 Resp: 445930
Ion Ratio Lower Upper
178 100
176 18.9 14.1 21.1
179 16.1 12.1 18.1


Manual Integrations
APPROVED

Reviewed By :Christian
Giraldo




05/24/2022
Supervised By :Jagrut
Upadhyay



05/24/2022




#72
Anthracene
Concen: 40.688 ng
RT: 17.568 min Scan# 2414
Delta R.T. 0.000 min
Lab File: BG053660.D
Acq: 23 May 2022 12:01



Tgt Ion:178 Resp: 441062
Ion Ratio Lower Upper
178 100
176 18.5 15.2 22.8
179 15.9 13.0 19.4






#73
 Carbazole
 Concen: 40.082 ng
 RT: 17.839 min Scan# 2460
 Delta R.T. 0.000 min
 Lab File: BG053660.D
 Acq: 23 May 2022 12:01

Instrument :

BNA_G

ClientSampleId :



SSTDCCC040



Tgt Ion:167 Resp: 425822
 Ion Ratio Lower Upper
 167 100
 166 22.6 18.2 27.2
 139 12.5 10.1 15.1


Manual Integrations
 APPROVED

Reviewed By :Christian
 Giraldo




05/24/2022
 Supervised By :Jagrut
 Upadhyay



05/24/2022




#74
 Di-n-butylphthalate
 Concen: 40.093 ng
 RT: 18.385 min Scan# 2553
 Delta R.T. 0.000 min
 Lab File: BG053660.D
 Acq: 23 May 2022 12:01



Tgt Ion:149 Resp: 479044
 Ion Ratio Lower Upper
 149 100
 150 9.3 7.7 11.5
 104 6.2 5.4 8.0






#75
Fluoranthene
Concen: 37.853 ng
RT: 19.489 min Scan# 21
Delta R.T. 0.000 min
Lab File: BG053660.D
Acq: 23 May 2022 12:01

Instrument :

BNA_G

ClientSampleId :


SSTDCCC040



Tgt Ion:202 Resp: 54054
Ion Ratio Lower Upper
202 100
101 7.2 0.0 26.9
203 18.4 0.0 37.2


Manual Integrations
APPROVED

Reviewed By :Christian
Giraldo






05/24/2022
Supervised By :Jagrut
Upadhyay


05/24/2022



#76
Chrysene-d12
Concen: 20.000 ng
RT: 21.716 min Scan# 3120
Delta R.T. 0.000 min
Lab File: BG053660.D
Acq: 23 May 2022 12:01


Tgt Ion:240 Resp: 187606
Ion Ratio Lower Upper
240 100
120 6.3 4.6 7.0
236 24.4 20.2 30.4





#77
Benzidine
Concen: 49.958 ng
RT: 19.660 min Scan# 21
Delta R.T. 0.000 min
Lab File: BG053660.D
Acq: 23 May 2022 12:01


Instrument :
BNA_G
ClientSampleId :
SSTDCCC040



Tgt Ion:184 Resp: 202009
Ion Ratio Lower Upper
184 100
185 14.4 11.8 17.6
183 12.2 8.9 13.3


Manual Integrations
APPROVED

Reviewed By :Christian
Giraldo






05/24/2022
Supervised By :Jagrut
Upadhyay


05/24/2022



#78
Pyrene
Concen: 44.612 ng
RT: 19.853 min Scan# 2803
Delta R.T. 0.000 min
Lab File: BG053660.D
Acq: 23 May 2022 12:01

Tgt Ion:202 Resp: 541799
Ion Ratio Lower Upper
202 100
200 19.8 16.6 24.8
203 17.8 14.6 21.8






#79
Terphenyl-d14
Concen: 86.263 ng
RT: 20.041 min Scan# 21
Delta R.T. 0.000 min
Lab File: BG053660.D
Acq: 23 May 2022 12:01

Instrument :

BNA_G

ClientSampleId :



SSTDCCC040



Tgt Ion:244 Resp: 868925
Ion Ratio Lower Upper
244 100
212 8.0 6.5 9.7
122 6.6 5.0 7.4


Manual Integrations
APPROVED

Reviewed By :Christian
Giraldo




05/24/2022
Supervised By :Jagrut
Upadhyay



05/24/2022




#80
Butylbenzylphthalate
Concen: 43.475 ng
RT: 20.723 min Scan# 2951
Delta R.T. 0.000 min
Lab File: BG053660.D
Acq: 23 May 2022 12:01




Tgt Ion:149 Resp: 193613
Ion Ratio Lower Upper
149 100
91 79.1 59.6 89.4
206 23.0 18.6 28.0





#81
Benzo(a)anthracene
Concen: 40.452 ng
RT: 21.698 min Scan# 3117
Delta R.T. 0.000 min
Lab File: BG053660.D
Acq: 23 May 2022 12:01



Instrument :
BNA_G
ClientSampleId :
SSTDCCC040



Tgt Ion:228 Resp: 484044
Ion Ratio Lower Upper
228 100
226 28.0 21.6 32.4
229 20.5 16.2 24.2


Manual Integrations
APPROVED

Reviewed By :Christian
Giraldo




05/24/2022
Supervised By :Jagrut
Upadhyay



05/24/2022




#82
3,3'-Dichlorobenzidine
Concen: 43.239 ng
RT: 21.604 min Scan# 3101
Delta R.T. 0.000 min
Lab File: BG053660.D
Acq: 23 May 2022 12:01




Tgt Ion:252 Resp: 131205
Ion Ratio Lower Upper
252 100
254 65.4 49.8 74.8
126 7.8 6.0 9.0





#83
Chrysene
Concen: 41.420 ng
RT: 21.763 min Scan# 3134
Delta R.T. 0.000 min
Lab File: BG053660.D
Acq: 23 May 2022 12:01



Instrument :
BNA_G
ClientSampleId :
SSTDCCC040



Tgt Ion:228 Resp: 462895
Ion Ratio Lower Upper
228 100
226 30.3 23.4 35.2
229 20.6 15.5 23.3


Manual Integrations
APPROVED

Reviewed By :Christian
Giraldo






05/24/2022
Supervised By :Jagrut
Upadhyay


05/24/2022



#84
Bis(2-ethylhexyl)phthalate
Concen: 42.095 ng
RT: 21.581 min Scan# 3097
Delta R.T. 0.000 min
Lab File: BG053660.D
Acq: 23 May 2022 12:01


Tgt Ion:149 Resp: 258748
Ion Ratio Lower Upper
149 100
167 26.6 20.2 30.4
279 4.5 3.6 5.4





#85
Di-n-octyl phthalate
Concen: 41.447 ng
RT: 22.802 min Scan# 31
Delta R.T. 0.000 min
Lab File: BG053660.D
Acq: 23 May 2022 12:01



Instrument :
BNA_G
ClientSampleId :
SSTDCCC040



Tgt Ion:149 Resp: 43080
Ion Ratio Lower Upper
149 100
167 1.7 1.3 1.9
43 11.6 9.0 13.4


Manual Integrations
APPROVED

Reviewed By :Christian
Giraldo






05/24/2022
Supervised By :Jagrut
Upadhyay


05/24/2022



#86
Perylene-d12
Concen: 20.000 ng
RT: 24.982 min Scan# 3676
Delta R.T. 0.000 min
Lab File: BG053660.D
Acq: 23 May 2022 12:01


Tgt Ion:264 Resp: 187219
Ion Ratio Lower Upper
264 100
260 24.1 19.0 28.6
265 21.1 17.4 26.2





#87
Indeno(1,2,3-cd)pyrene
Concen: 41.401 ng
RT: 28.730 min Scan# 41
Delta R.T. 0.000 min
Lab File: BG053660.D
Acq: 23 May 2022 12:01



Instrument :
BNA_G
ClientSampleId :
SSTDCCC040



Tgt Ion:276 Resp: 53984
Ion Ratio Lower Upper
276 100
138 14.5 5.0 7.4
277 25.4 20.3 30.5


Manual Integrations
APPROVED

Reviewed By :Christian
Giraldo




05/24/2022
Supervised By :Jagrut
Upadhyay



05/24/2022




#88
Benzo(b)fluoranthene
Concen: 41.249 ng
RT: 23.942 min Scan# 3499
Delta R.T. 0.000 min
Lab File: BG053660.D
Acq: 23 May 2022 12:01



Tgt Ion:252 Resp: 457116
Ion Ratio Lower Upper
252 100
253 22.1 18.0 27.0
125 6.1 4.6 6.8





#89
Benzo(k)fluoranthene
Concen: 42.042 ng
RT: 24.013 min Scan# 31
Delta R.T. 0.000 min
Lab File: BG053660.D
Acq: 23 May 2022 12:01

Instrument :

BNA_G



ClientSampleId :

SSTDCCC040

Tgt Ion:252 Resp: 45782
Ion Ratio Lower Upper
252 100
253 20.9 18.4 27.6
125 5.8 4.8 7.2



Manual Integrations
APPROVED

Reviewed By :Christian
Giraldo






05/24/2022
Supervised By :Jagrut
Upadhyay


05/24/2022



#90
Benzo(a)pyrene
Concen: 41.287 ng
RT: 24.829 min Scan# 3650
Delta R.T. 0.000 min
Lab File: BG053660.D
Acq: 23 May 2022 12:01


Tgt Ion:252 Resp: 389430
Ion Ratio Lower Upper
252 100
253 21.9 17.8 26.8
125 6.3 4.3 6.5





#91
Dibenzo(a,h)anthracene
Concen: 41.586 ng
RT: 28.783 min Scan# 41
Delta R.T. 0.000 min
Lab File: BG053660.D
Acq: 23 May 2022 12:01


Instrument :
BNA_G
ClientSampleId :
SSTDCCC040



Tgt Ion:278 Resp: 44682
Ion Ratio Lower Upper
278 100
139 9.6 7.2 10.8
279 24.3 18.6 27.8


Manual Integrations
APPROVED

Reviewed By :Christian
Giraldo



05/24/2022
Supervised By :Jagrut
Upadhyay

05/24/2022



#92
Benzo(g,h,i)perylene
Concen: 41.240 ng
RT: 29.905 min Scan# 4514
Delta R.T. 0.000 min
Lab File: BG053660.D
Acq: 23 May 2022 12:01

Tgt Ion:276 Resp: 439220
Ion Ratio Lower Upper
276 100
277 24.3 17.3 25.9
138 13.1 9.6 14.4

