

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG052323\
 Data File : BG057512.D
 Acq On : 23 May 2023 9:45
 Operator : CG/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_G
 LabSampleId :
 SSTDCCC040

Quant Time: May 23 10:19:07 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG042723.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Sat May 20 04:19:55 2023
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	125	0.00
2	1,4-Dioxane	0.483	0.459	5.0	119	0.00
3	Pyridine	1.554	1.430	8.0	110	0.00
4	n-Nitrosodimethylamine	0.730	0.606	17.0	99	0.00
5 S	2-Fluorophenol	1.169	1.158	0.9	121	0.00
6	Aniline	2.252	2.206	2.0	119	0.00
7 S	Phenol-d6	1.775	1.829	-3.0	126	0.00
8	2-Chlorophenol	1.235	1.270	-2.8	127	0.00
9	Benzaldehyde	1.042	0.882	15.4	117	0.00
10 C	Phenol	1.731	1.770	-2.3	123	0.00
11	bis(2-Chloroethyl)ether	1.305	1.291	1.1	122	0.00
12	1,3-Dichlorobenzene	1.364	1.364	0.0	124	0.00
13 C	1,4-Dichlorobenzene	1.370	1.404	-2.5	125	0.00
14	1,2-Dichlorobenzene	1.322	1.344	-1.7	127	0.00
15	Benzyl Alcohol	1.479	1.445	2.3	115	0.00
16	2,2'-oxybis(1-Chloropropane	1.645	1.540	6.4	116	0.00
17	2-Methylphenol	1.248	1.258	-0.8	122	0.00
18	Hexachloroethane	0.494	0.494	0.0	124	0.00
19 P	n-Nitroso-di-n-propylamine	1.354	1.261	6.9	115	0.00
20	3+4-Methylphenols	1.708	1.796	-5.2	128	0.00
21 I	Naphthalene-d8	1.000	1.000	0.0	124	0.00
22	Acetophenone	0.572	0.555	3.0	119	0.00
23 S	Nitrobenzene-d5	0.477	0.460	3.6	116	0.00
24	Nitrobenzene	0.481	0.455	5.4	114	0.00
25	Isophorone	0.924	0.859	7.0	114	0.00
26 C	2-Nitrophenol	0.184	0.204	-10.9	133	0.00
27	2,4-Dimethylphenol	0.322	0.333	-3.4	122	0.00
28	bis(2-Chloroethoxy)methane	0.463	0.447	3.5	117	0.00
29 C	2,4-Dichlorophenol	0.342	0.370	-8.2	131	0.00
30	1,2,4-Trichlorobenzene	0.394	0.393	0.3	123	0.00
31	Naphthalene	1.069	1.094	-2.3	125	0.00
32	Benzoic acid	0.170	0.187	-10.0	126	0.00
33	4-Chloroaniline	0.480	0.489	-1.9	121	0.00
34 C	Hexachlorobutadiene	0.293	0.291	0.7	124	0.00
35	Caprolactam	0.117	0.119	-1.7	120	0.00
36 C	4-Chloro-3-methylphenol	0.389	0.410	-5.4	128	0.00
37	2-Methylnaphthalene	0.841	0.865	-2.9	126	0.00
38	1-Methylnaphthalene	0.792	0.804	-1.5	124	0.00
39 I	Acenaphthene-d10	1.000	1.000	0.0	122	0.00
40	1,2,4,5-Tetrachlorobenzene	0.710	0.741	-4.4	123	0.00
41 P	Hexachlorocyclopentadiene	0.287	0.281	2.1	109	0.00
42 S	2,4,6-Tribromophenol	0.284	0.307	-8.1	129	0.00
43 C	2,4,6-Trichlorophenol	0.433	0.462	-6.7	123	0.00
44	2,4,5-Trichlorophenol	0.509	0.511	-0.4	115	0.00
45 S	2-Fluorobiphenyl	1.485	1.524	-2.6	122	0.00
46	1,1'-Biphenyl	1.484	1.525	-2.8	122	0.00
47	2-Chloronaphthalene	1.104	1.136	-2.9	122	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG052323\
 Data File : BG057512.D
 Acq On : 23 May 2023 9:45
 Operator : CG/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_G
 LabSampleId :
 SSTDCCC040

Quant Time: May 23 10:19:07 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG042723.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Sat May 20 04:19:55 2023
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
48	2-Nitroaniline	0.375	0.372	0.8	115	0.00
49	Acenaphthylene	1.793	1.865	-4.0	123	0.00
50	Dimethylphthalate	1.644	1.592	3.2	118	0.00
51	2,6-Dinitrotoluene	0.336	0.349	-3.9	123	0.00
52 C	Acenaphthene	1.120	1.136	-1.4	120	0.00
53	3-Nitroaniline	0.336	0.348	-3.6	123	0.00
54 P	2,4-Dinitrophenol	0.182	0.226	-24.2	139	0.00
55	Dibenzofuran	1.882	1.916	-1.8	122	0.00
56 P	4-Nitrophenol	0.224	0.215	4.0	106	0.00
57	2,4-Dinitrotoluene	0.480	0.500	-4.2	124	0.00
58	Fluorene	1.551	1.552	-0.1	120	0.00
59	2,3,4,6-Tetrachlorophenol	0.461	0.468	-1.5	118	0.00
60	Diethylphthalate	1.669	1.611	3.5	116	0.00
61	4-Chlorophenyl-phenylether	0.934	0.924	1.1	118	0.00
62	4-Nitroaniline	0.339	0.348	-2.7	119	0.00
63	Azobenzene	1.610	1.491	7.4	111	0.00
64 I	Phenanthrene-d10	1.000	1.000	0.0	122	0.00
65	4,6-Dinitro-2-methylphenol	0.119	0.132	-10.9	121	0.00
66 c	n-Nitrosodiphenylamine	0.543	0.554	-2.0	120	0.00
67	4-Bromophenyl-phenylether	0.244	0.252	-3.3	121	0.00
68	Hexachlorobenzene	0.237	0.253	-6.8	127	0.00
69	Atrazine	0.207	0.216	-4.3	123	0.00
70 C	Pentachlorophenol	0.137	0.152	-10.9	124	0.00
71	Phenanthrene	1.025	1.036	-1.1	120	0.00
72	Anthracene	1.052	1.063	-1.0	119	0.00
73	Carbazole	0.885	0.903	-2.0	121	0.00
74	Di-n-butylphthalate	1.019	1.067	-4.7	122	0.00
75 C	Fluoranthene	1.275	1.248	2.1	116	0.00
76 I	Chrysene-d12	1.000	1.000	0.0	115	0.00
77	Benzidine	0.447	0.913	-104.3#	266#	0.00
78	Pyrene	1.453	1.486	-2.3	117	0.00
79 S	Terphenyl-d14	1.226	1.223	0.2	115	0.00
80	Butylbenzylphthalate	0.461	0.525	-13.9	126	0.00
81	Benzo(a)anthracene	1.425	1.433	-0.6	112	0.00
82	3,3'-Dichlorobenzidine	0.456	0.476	-4.4	115	0.00
83	Chrysene	1.341	1.356	-1.1	114	0.00
84	Bis(2-ethylhexyl)phthalate	0.682	0.749	-9.8	121	0.00
85 c	Di-n-octyl phthalate	1.137	1.229	-8.1	120	0.00
86 I	Perylene-d12	1.000	1.000	0.0	120	0.00
87	Indeno(1,2,3-cd)pyrene	1.458	1.432	1.8	115	0.00
88	Benzo(b)fluoranthene	1.307	1.261	3.5	113	0.00
89	Benzo(k)fluoranthene	1.329	1.261	5.1	111	0.00
90 C	Benzo(a)pyrene	1.277	1.225	4.1	113	0.00
91	Dibenzo(a,h)anthracene	1.214	1.191	1.9	115	0.00
92	Benzo(g,h,i)perylene	1.185	1.160	2.1	115	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG052323\
Data File : BG057512.D
Acq On : 23 May 2023 9:45
Operator : CG/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_G
LabSampleId :
SSTDCCC040

Quant Time: May 23 10:19:07 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG042723.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Sat May 20 04:19:55 2023
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)
----------	-------	------	-----------	------------

(#) = Out of Range

SPCC's out = 0 CCC's out = 0