

Data Path : Z:\HPCHEM1\BNA_G\DATA\BG052417\
 Data File : BG027110.D
 Acq On : 24 May 2017 1:09
 Operator : SJ/MA
 Sample : SSTDC0020
 Misc :
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 BNA_G
 LabSampleId :
 SSTD02059

Quant Time: May 24 01:48:38 2017
 Quant Method : Z:\HPCHEM1\BNA_G\METHODS\SOM-EPA-BG051817MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Wed May 24 01:27:20 2017
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	127	0.00
2 I	Naphthalene-d8	1.000	1.000	0.0	118	0.00
3	Naphthalene	1.038	1.065	-2.6	122	0.00
4	2-Methylnaphthalene	0.769	0.779	-1.3	121	0.00
5	1-Methylnaphthalene	0.828	0.837	-1.1	121	0.00
6 I	Acenaphthene-d10	1.000	1.000	0.0	116	0.00
7 S	Acenaphthylene-d8	2.051	2.098	-2.3	116	0.00
8	Acenaphthylene	2.062	2.109	-2.3	116	0.00
9 C	Acenaphthene	1.411	1.450	-2.8	117	0.00
10 S	Fluorene-d10	1.369	1.385	-1.2	116	0.00
11	Fluorene	1.504	1.541	-2.5	117	0.00
12 I	Phenanthrene-d10	1.000	1.000	0.0	112	0.00
13	Phenanthrene	1.118	1.175	-5.1	118	0.00
14 S	Anthracene-d10	0.975	1.011	-3.7	115	0.00
15	Anthracene	1.149	1.208	-5.1	117	0.00
16 C	Fluoranthene	1.143	1.216	-6.4	118	0.00
17 I	Chrysene-d12	1.000	1.000	0.0	115	0.00
18 S	Pyrene-d10	1.044	1.065	-2.0	118	0.00
19	Pyrene	1.307	1.352	-3.4	117	0.00
20	Benzo(a)anthracene	1.205	1.191	1.2	113	0.00
21	Chrysene	1.106	1.123	-1.5	116	0.00
22 I	Perylene-d12	1.000	1.000	0.0	99	0.00
23	Benzo(b)fluoranthene	1.189	1.306	-9.8	109	0.00
24	Benzo(k)fluoranthene	1.223	1.388	-13.5	110	0.00
25 S	Benzo(a)pyrene-d12	0.994	1.085	-9.2	107	0.00
26 C	Benzo(a)pyrene	1.163	1.272	-9.4	107	0.00
27	Indeno(1,2,3-cd)pyrene	1.070	0.985	7.9	90	0.00
28	Dibenzo(a,h)anthracene	0.923	0.941	-2.0	100	0.00
29	Benzo(g,h,i)perylene	0.851	0.700	17.7	77	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0