

Data Path : Z:\HPCHEM1\BNA G\DATA\BG052516\
 Data File : BG022060.D
 Acq On : 25 May 2016 15:51
 Operator : UM/SJ
 Sample : SSTDC0020
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_G
 LabSampleId :
 SSTD02016

Quant Time: May 26 03:06:18 2016
 Quant Method : Z:\HPCHEM1\BNA G\METHODS\SOM02.2-EPA-BG051816.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Fri May 20 08:11:42 2016
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	117	0.00
2 S	1,4-Dioxane-d8	0.436	0.338	22.5#	87	0.00
3 S	Phenol-d5	1.766	1.780	-0.8	120	0.00
4 S	Bis-(2-Chloroethyl)ether-d8	0.941	0.931	1.1	113	0.00
5 S	2-Chlorophenol-d4	1.450	1.482	-2.2	114	0.00
6 S	4-Methylphenol-d8	1.573	1.682	-6.9	121	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	117	0.00
8 S	Nitrobenzene-d5	0.140	0.143	-2.1	114	0.00
9 S	2-Nitrophenol-d4	0.150	0.164	-9.3	124	0.00
10 S	2,4-Dichlorophenol-d3	0.265	0.294	-10.9	126	0.00
11	Naphthalene	0.973	0.987	-1.4	116	0.00
12 S	4-Chloroaniline-d4	0.208	0.335	-61.1#	202#	0.00
13	2-Methylnaphthalene	0.705	0.737	-4.5	120	0.00
14	1-Methylnaphthalene	0.721	0.756	-4.9	118	0.00
15 I	Acenaphthene-d10	1.000	1.000	0.0	125	0.00
16 S	Dimethylphthalate-d6	1.542	1.568	-1.7	126	0.00
17 S	Acenaphthylene-d8	1.986	2.002	-0.8	122	0.00
18	Acenaphthylene	2.068	2.032	1.7	120	0.00
19 C	Acenaphthene	1.420	1.411	0.6	122	0.00
20 S	4-Nitrophenol-d4	0.271	0.240	11.4	115	0.00
21 S	Fluorene-d10	1.388	1.411	-1.7	126	0.00
22	Fluorene	1.573	1.558	1.0	121	0.00
23 I	Phenanthrene-d10	1.000	1.000	0.0	121	0.00
24 S	4,6-Dinitro-2-methylphenol-	0.099	0.112	-13.1	148	0.00
25	Phenanthrene	1.069	1.047	2.1	116	0.00
26 S	Anthracene-d10	0.921	0.912	1.0	118	0.00
27	Anthracene	1.074	1.073	0.1	117	0.00
28 C	Fluoranthene	1.111	0.960	13.6	102	0.00
29 I	Chrysene-d12	1.000	1.000	0.0	106	0.00
30 S	Pyrene-d10	1.055	1.036	1.8	103	0.00
31	Pyrene	1.377	1.284	6.8	97	0.00
32	Benzo(a)anthracene	1.109	1.094	1.4	103	0.00
33	Chrysene	1.079	1.027	4.8	101	0.00
34 I	Perylene-d12	1.000	1.000	0.0	112	0.00
35	Benzo(b)fluoranthene	1.150	1.076	6.4	103	0.00
36	Benzo(k)fluoranthene	1.131	1.053	6.9	104	0.00
37 S	Benzo(a)pyrene-d12	0.888	0.872	1.8	109	0.00
38 C	Benzo(a)pyrene	1.099	1.048	4.6	105	0.00
39	Indeno(1,2,3-cd)pyrene	1.250	1.254	-0.3	112	0.00
40	Dibenzo(a,h)anthracene	1.035	1.006	2.8	107	0.00
41	Benzo(g,h,i)perylene	1.057	1.026	2.9	107	0.00

Data Path : Z:\HPCHEM1\BNA G\DATA\BG052516\
Data File : BG022060.D
Acq On : 25 May 2016 15:51
Operator : UM/SJ
Sample : SSTDCCC020
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_G
LabSampleId :
SSTD02016

Quant Time: May 26 03:06:18 2016
Quant Method : Z:\HPCHEM1\BNA G\METHODS\SOM02.2-EPA-BG051816.M
Quant Title : SVOA CALIBRATION
QLast Update : Fri May 20 08:11:42 2016
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
Max. RRF Dev : 20% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)
----------	-------	------	-----------	------------

(#) = Out of Range

SPCC's out = 0 CCC's out = 0