


Data Path : Z:\HPCHEM1\BNA G\DATA\BG052616\
Data File : BG022071.D
Acq On : 26 May 2016 8:19
Operator : UM/SJ
Sample : DFTPP14
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
DFTPP14

Integration File: LSCINT.P

Method : Z:\HPCHEM1\BNA_G\METHODS\SOM02.2-EPA-BG052116.M
Title : SVOA CALIBRATION
Last Update : Fri May 27 05:50:42 2016



AutoFind: Scans 2568, 2569, 2570; Background Corrected with Scan 2562

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	26.5	22573	PASS
68	69	0.00	2	0.6	177	PASS
69	198	0.00	100	34.7	29586	PASS
70	69	0.00	2	0.2	50	PASS
127	198	10	80	53.5	45613	PASS
197	198	0.00	2	0.0	0	PASS
198	198	100	100	100.0	85312	PASS
199	198	5	9	7.3	6199	PASS
275	198	10	60	26.1	22301	PASS
365	198	1	100	3.2	2718	PASS
441	442	0.01	100	16.5	10973	PASS
442	198	50	100	77.8	66381	PASS
443	442	15	24	19.1	12648	PASS

Scan 2562 (17.892 min): BG022071.D
DFTPP14

m/z	Abundance
39.90	1547.0
43.90	491.0
55.00	200.0
206.90	353.0

Instrument :
BNA_G
ClientSampleId :
DFTPP14

m/z	Abundance
38.00	246.0
39.00	1558.0
39.90	1650.0
40.90	213.0
42.90	191.0
44.00	924.0
50.00	5312.0
51.00	22056.0
52.00	1444.0
54.90	227.0
55.90	715.0
57.00	1878.0
61.00	299.0
61.90	365.0
63.00	1001.0
64.90	562.0
68.00	533.0
68.90	28960.0
72.90	350.0
74.00	2636.0
75.00	4856.0
76.00	1769.0
77.00	35112.0
78.00	2581.0
79.00	1760.0
79.90	1482.0
80.90	2463.0
81.90	739.0
82.90	614.0
85.00	349.0
85.90	614.0
87.00	273.0
90.90	568.0
91.90	597.0
92.90	3471.0
94.00	248.0
95.00	162.0
95.90	182.0
97.00	163.0
97.90	2536.0
98.90	2105.0
100.90	1462.0
102.00	182.0
103.00	366.0
103.90	900.0
104.90	895.0
106.00	398.0
106.90	10109.0
107.90	1827.0
109.90	20544.0
110.90	3173.0
111.90	530.0
115.90	563.0
116.90	6776.0
117.90	489.0
121.90	816.0
122.90	1052.0
123.90	415.0
124.80	486.0
126.90	42576.0
127.90	3065.0
128.90	15821.0
129.90	1302.0
130.90	312.0
133.90	464.0
134.90	1145.0
135.90	571.0
136.90	569.0
139.90	307.0
140.90	2393.0
141.90	695.0
142.90	457.0
144.90	150.0
145.90	394.0
146.90	1042.0
147.90	2088.0
148.90	600.0
149.90	186.0
151.00	336.0
151.90	177.0
152.80	594.0
153.90	546.0
154.90	1005.0
155.90	1501.0
156.90	467.0
157.90	332.0
158.90	419.0
159.90	583.0
160.90	837.0
161.80	311.0
164.90	659.0
165.90	568.0

Instrument :
BNA_G
ClientSampleId :
DFTPP14

166.90	3734.0
167.90	1453.0
168.90	326.0
169.90	201.0
170.90	232.0
171.90	437.0
172.90	458.0
173.90	718.0
174.90	1649.0
175.90	503.0
176.90	1041.0
177.90	286.0
178.90	2903.0
179.90	1895.0
180.90	773.0
182.00	208.0
183.90	208.0
184.90	1586.0
185.90	10743.0
186.90	2909.0
187.90	438.0
188.90	645.0
189.80	170.0
190.90	393.0
191.90	1155.0
192.90	1182.0
193.80	203.0
194.80	199.0
195.90	2501.0
197.90	75944.0
198.90	5516.0
199.90	497.0
201.40	412.0
202.90	430.0
203.90	2980.0
204.90	5772.0
205.90	19928.0
206.90	3043.0
207.80	797.0
208.90	208.0
210.20	306.0
211.00	629.0
215.00	161.0
215.90	567.0
216.80	5480.0
217.80	569.0
220.90	3347.0
221.70	790.0
222.90	1149.0
223.90	11301.0
224.90	2550.0
225.80	324.0
226.90	4252.0
227.90	691.0
228.90	734.0
229.70	157.0
230.90	404.0
231.80	162.0
233.90	254.0
234.90	281.0
235.80	160.0
236.90	562.0
238.90	272.0
239.80	208.0
240.90	187.0
241.90	480.0
242.90	562.0
243.90	8624.0
244.90	1037.0
245.90	1618.0
246.90	398.0
248.80	397.0
252.70	216.0
254.90	42256.0
255.90	6694.0
256.90	502.0
257.90	2485.0
258.90	514.0
264.90	928.0
270.80	274.0
271.90	157.0
272.90	1628.0
273.90	3334.0
274.90	19072.0
275.90	2306.0
276.90	1545.0
277.90	253.0
282.90	217.0
284.90	463.0
292.90	345.0
294.90	165.0
295.90	4791.0
296.90	934.0
302.90	612.0
304.00	288.0
313.90	189.0

Instrument :
BNA_G
ClientSampleId :
DFTPP14

314.90	552.0
315.90	307.0
320.80	169.0
322.90	1490.0
323.90	365.0
326.90	340.0
333.90	1174.0
334.90	222.0
340.90	246.0
345.80	349.0
351.90	449.0
352.90	291.0
353.90	368.0
364.90	2133.0
365.90	259.0
371.90	706.0
372.90	226.0
382.80	225.0
389.80	182.0
401.90	352.0
402.80	402.0
420.80	400.0
421.90	244.0
422.90	2247.0
423.90	493.0
440.90	7493.0
441.90	47456.0
442.90	9290.0
443.90	808.0

Instrument :
BNA_G
ClientSampleId :
DFTPP14

m/z	Abundance
38.10	249.0
39.00	2019.0
39.90	1699.0
43.00	182.0
43.90	1070.0
50.00	6354.0
51.00	26360.0
52.00	1299.0
55.00	232.0
56.00	1052.0
57.00	2048.0
58.00	180.0
61.00	436.0
62.00	461.0
63.00	1426.0
65.00	676.0
67.00	195.0
68.90	34032.0
70.00	151.0
73.00	291.0
74.00	3410.0
75.00	5071.0
76.00	2038.0
77.00	41432.0
78.00	2642.0
78.90	2370.0
79.90	1972.0
80.90	3028.0
82.00	813.0
82.90	678.0
84.90	434.0
86.00	750.0
86.80	375.0
87.90	152.0
90.90	662.0
91.90	842.0
92.90	4279.0
93.90	216.0
95.00	243.0
95.90	333.0
97.90	2840.0
98.90	2619.0
99.90	201.0
100.90	1756.0
103.00	640.0
103.90	1124.0
104.90	1092.0
105.90	400.0
106.90	13313.0
107.90	2190.0
109.90	24808.0
110.90	3480.0
111.90	463.0
116.00	774.0
116.90	8722.0
118.00	571.0
119.10	154.0
119.90	312.0
121.90	955.0
122.90	1725.0
123.90	818.0
125.00	782.0
126.90	51816.0
127.90	4185.0
128.90	19952.0
129.90	1570.0
130.90	197.0
131.80	217.0
132.90	186.0
133.90	638.0
135.00	1341.0
135.90	784.0
136.90	764.0
137.90	192.0
139.90	268.0
140.90	2451.0
141.90	861.0
142.90	504.0
145.00	197.0
145.90	375.0
146.90	1432.0
147.90	2263.0
148.90	650.0
150.90	319.0
152.90	825.0
153.90	497.0
154.90	1588.0
155.90	1815.0
156.90	391.0
157.80	369.0
158.90	264.0
159.90	767.0

Instrument :
BNA_G
ClientSampleId :
DFTPP14

160.90	1329.0
161.90	311.0
162.90	218.0
163.90	180.0
164.90	1039.0
165.90	604.0
166.90	4255.0
167.90	1686.0
168.90	450.0
169.80	214.0
170.80	302.0
171.90	394.0
173.00	489.0
173.90	1164.0
174.90	1913.0
175.80	556.0
176.90	898.0
177.90	305.0
178.90	3525.0
179.90	2330.0
180.90	1489.0
181.90	245.0
182.80	190.0
183.90	327.0
184.90	1826.0
185.90	13640.0
186.90	3926.0
187.90	393.0
188.90	845.0
189.90	294.0
190.80	615.0
191.90	1198.0
193.00	1385.0
193.90	297.0
194.80	264.0
195.90	3196.0
197.90	97240.0
198.90	7416.0
199.90	547.0
201.30	320.0
202.90	748.0
203.90	3795.0
204.90	7099.0
205.90	25920.0
206.90	3547.0
207.90	994.0
208.80	267.0
210.90	1000.0
215.00	264.0
215.90	578.0
216.90	6942.0
217.90	797.0
220.90	4626.0
221.80	680.0
222.90	1528.0
223.90	14130.0
224.90	3779.0
225.90	514.0
226.90	6313.0
227.90	796.0
228.90	1149.0
229.80	184.0
230.80	531.0
231.90	176.0
233.80	453.0
234.90	509.0
236.00	264.0
236.90	505.0
238.90	352.0
239.70	212.0
240.90	413.0
241.90	631.0
242.90	894.0
243.90	11334.0
244.90	1194.0
245.90	1971.0
246.90	385.0
247.90	184.0
248.80	423.0
250.90	168.0
251.90	172.0
252.80	432.0
254.90	57368.0
255.90	7901.0
256.90	489.0
257.90	3301.0
258.90	478.0
264.90	1481.0
271.00	193.0
271.90	206.0
272.90	1768.0
273.90	4212.0
274.90	25432.0
275.90	2949.0
276.90	2112.0
277.90	286.0

Instrument :
BNA_G
ClientSampleId :
DFTPP14

279.00	154.0
280.90	168.0
282.80	300.0
283.80	188.0
284.90	399.0
292.90	553.0
293.90	181.0
294.90	211.0
295.90	6898.0
296.90	955.0
302.90	832.0
303.90	221.0
309.90	163.0
313.90	377.0
314.90	790.0
315.90	465.0
320.90	262.0
321.90	172.0
322.90	2137.0
323.90	335.0
326.90	470.0
327.90	193.0
331.80	233.0
332.80	260.0
333.90	1550.0
335.00	353.0
340.90	283.0
342.00	201.0
345.90	560.0
351.90	788.0
352.90	538.0
353.90	822.0
364.90	2842.0
365.80	394.0
370.90	270.0
371.90	1245.0
372.80	257.0
382.90	304.0
391.90	154.0
401.90	539.0
402.90	683.0
403.90	268.0
420.90	624.0
421.80	423.0
422.90	4003.0
423.90	849.0
439.20	151.0
440.90	12352.0
441.90	74400.0
442.90	13774.0
443.90	1391.0

Instrument :
BNA_G
ClientSampleId :
DFTPP14

m/z	Abundance
38.00	155.0
39.00	1306.0
39.90	1581.0
41.00	158.0
43.00	155.0
44.00	1028.0
48.90	163.0
50.00	4900.0
51.00	19304.0
52.00	1186.0
55.00	163.0
55.90	682.0
56.90	1406.0
60.90	212.0
62.00	475.0
63.00	963.0
65.00	421.0
68.90	25768.0
73.00	197.0
74.00	2242.0
74.90	3973.0
76.00	1429.0
77.00	31088.0
78.00	2056.0
78.90	1804.0
80.00	1418.0
80.90	2352.0
81.90	550.0
82.90	645.0
84.90	343.0
85.90	667.0
86.90	279.0
91.00	639.0
91.90	578.0
92.90	3281.0
93.90	279.0
97.90	2645.0
98.90	2038.0
99.90	277.0
100.90	1423.0
102.90	296.0
103.90	788.0
104.90	833.0
106.00	291.0
106.90	10053.0
107.90	1515.0
109.90	18992.0
110.90	3040.0
111.90	354.0
115.90	672.0
116.90	6259.0
117.90	572.0
119.90	177.0
121.90	660.0
122.90	1079.0
123.90	545.0
125.00	494.0
126.90	42448.0
127.90	3159.0
128.90	15761.0
129.90	1496.0
130.90	271.0
132.10	164.0
133.00	161.0
133.90	535.0
134.90	1398.0
135.90	572.0
136.90	825.0
137.90	230.0
139.90	396.0
140.90	2081.0
141.90	747.0
142.90	451.0
145.90	446.0
146.90	1164.0
147.90	1809.0
148.90	460.0
150.00	199.0
151.00	301.0
151.70	151.0
152.90	670.0
153.90	428.0
154.90	1139.0
155.90	1916.0
156.90	416.0
157.90	484.0
158.80	297.0
159.90	648.0
160.90	989.0
161.90	308.0
162.90	154.0
163.90	155.0

Instrument :
BNA_G
ClientSampleId :
DFTPP14

164.90	844.0
165.90	809.0
166.90	3295.0
167.90	1508.0
168.90	510.0
171.00	202.0
171.90	322.0
172.90	445.0
173.80	846.0
174.90	1730.0
175.90	623.0
176.90	730.0
177.90	263.0
178.90	2903.0
179.90	2078.0
180.90	1140.0
183.90	287.0
184.90	1849.0
185.90	11254.0
186.90	3261.0
187.90	269.0
188.80	661.0
190.90	345.0
191.90	1067.0
192.90	906.0
195.00	194.0
195.90	3059.0
197.90	82752.0
198.80	5666.0
200.00	323.0
201.30	402.0
201.70	267.0
202.90	875.0
203.90	3392.0
204.90	5363.0
205.90	22112.0
206.90	3121.0
207.90	792.0
208.90	258.0
210.20	376.0
210.90	920.0
214.90	317.0
215.90	513.0
216.80	5437.0
217.90	732.0
220.90	3901.0
221.60	835.0
222.90	1511.0
223.90	13089.0
224.90	3120.0
225.90	251.0
226.90	5549.0
227.90	777.0
228.80	1248.0
230.90	566.0
233.90	279.0
234.80	467.0
235.90	247.0
236.80	493.0
238.80	181.0
240.90	269.0
241.90	611.0
242.90	819.0
243.90	9757.0
244.90	1246.0
245.90	1804.0
246.80	323.0
248.80	304.0
252.00	198.0
252.90	378.0
254.90	50688.0
255.90	7165.0
256.90	672.0
257.90	2687.0
258.90	543.0
264.90	1104.0
270.80	252.0
271.80	241.0
272.90	1942.0
273.90	3648.0
274.90	22400.0
275.90	2962.0
276.90	1724.0
277.80	253.0
283.00	278.0
284.90	493.0
292.80	303.0
295.90	6208.0
296.90	931.0
301.90	167.0
302.90	805.0
303.90	163.0
313.90	476.0
314.80	704.0
315.90	453.0
320.80	247.0

Instrument :
BNA_G
ClientSampleId :
DFTPP14

321.70	155.0
322.90	2363.0
323.90	443.0
326.80	398.0
327.80	184.0
331.90	232.0
332.90	381.0
333.90	1392.0
334.90	270.0
340.80	234.0
345.90	423.0
351.90	623.0
352.90	490.0
353.80	831.0
364.90	3179.0
365.80	564.0
370.90	223.0
371.90	1297.0
372.80	317.0
382.90	295.0
389.90	286.0
401.80	551.0
402.90	758.0
404.00	273.0
420.90	627.0
421.80	646.0
422.90	3982.0
423.90	917.0
440.90	13074.0
441.90	77288.0
442.90	14880.0
443.90	1462.0

Instrument :
BNA_G
ClientSampleId :
DFTPP14