


Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG052620\
 Data File : BG045454.D
 Acq On : 26 May 2020 12:37
 Operator : CG/JU
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 DFTPP

Integration File: rteint.p

Method : Z:\SVOASRV\HPCHEM1\BNA_G\METHODS\8270-BG051520.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Wed May 27 13:40:45 2020



AutoFind: Scans 2504, 2505, 2506; Background Corrected with Scan 2497

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	26.2	15055	PASS
68	69	0.00	2	0.5	99	PASS
69	198	0.00	100	32.8	18807	PASS
70	69	0.00	2	0.0	0	PASS
127	198	10	80	41.7	23938	PASS
197	198	0.00	2	0.4	207	PASS
198	198	100	100	100.0	57402	PASS
199	198	5	9	6.5	3743	PASS
275	198	10	60	29.3	16807	PASS
365	198	1	100	4.7	2709	PASS
441	198	0.01	100	12.0	6865	PASS
442	198	50	100	86.4	49621	PASS
443	442	15	24	19.5	9664	PASS