

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG052622\
 Data File : BG053720.D
 Acq On : 26 May 2022 17:04
 Operator : CG/JU
 Sample : N3059-01
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 4922A

Quant Time: May 27 05:00:23 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG050522.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed May 25 15:11:41 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.982	152	27116	20.000	ng	0.00
21) Naphthalene-d8	10.796	136	100853	20.000	ng	0.00
39) Acenaphthene-d10	14.626	164	68780	20.000	ng	0.00
64) Phenanthrene-d10	17.369	188	147223	20.000	ng	0.00
76) Chrysene-d12	21.634	240	142433	20.000	ng	0.00
86) Perylene-d12	24.848	264	144063	20.000	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.556	112	58117	36.374	ng	0.00
7) Phenol-d6	7.172	99	53396	22.020	ng	0.00
23) Nitrobenzene-d5	9.151	82	156413	66.019	ng	0.00
42) 2,4,6-Tribromophenol	16.118	330	90789	89.498	ng	0.00
45) 2-Fluorobiphenyl	13.246	172	330786	70.278	ng	0.00
79) Terphenyl-d14	19.978	244	512096	66.962	ng	0.00
Target Compounds						
						Qvalue
2) 1,4-Dioxane	3.394	88	211	0.252	ng	# 80
3) Pyridine	3.864	79	84	0.041	ng	# 1
4) n-Nitrosodimethylamine	3.729	42	2068	2.074	ng	# 1
6) Aniline	7.313	93	2389	0.890	ng	# 85
9) Benzaldehyde	7.130	77	499	0.324	ng	# 1
10) Phenol	7.201	94	5312	2.240	ng	# 83
11) bis(2-Chloroethyl)ether	7.313	93	2389	1.351	ng	# 17
15) Benzyl Alcohol	8.235	79	1605	0.792	ng	# 82
16) 2,2'-oxybis(1-Chloropr...	8.734	45	11742	4.079	ng	# 82
17) 2-Methylphenol	8.452	107	674	0.420	ng	# 88
18) Hexachloroethane	9.087	117	726	0.961	ng	# 1
19) n-Nitroso-di-n-propyla...	8.763	70	895	0.514	ng	# 32
20) 3+4-Methylphenols	8.769	107	74130	32.745	ng	# 92
22) Acetophenone	8.810	105	3360	1.217	ng	# 49
24) Nitrobenzene	9.151	77	1310	0.571	ng	# 31
25) Isophorone	9.850	82	2210	0.554	ng	# 66
27) 2,4-Dimethylphenol	9.985	122	892	0.638	ng	# 97
28) bis(2-Chloroethoxy)met...	10.209	93	1976	0.862	ng	# 1
29) 2,4-Dichlorophenol	10.661	162	121	0.072	ng	# 1
31) Naphthalene	10.843	128	6653	1.297	ng	# 57
32) Benzoic acid	10.397	122	10529	18.407	ng	# 93
33) 4-Chloroaniline	10.914	127	292	0.136	ng	# 1
35) Caprolactam	11.742	113	246	0.401	ng	# 1
36) 4-Chloro-3-methylphenol	12.147	107	1045	0.518	ng	# 25
37) 2-Methylnaphthalene	12.458	142	2822	0.737	ng	# 1
38) 1-Methylnaphthalene	12.670	142	1882	0.504	ng	# 62
46) 1,1'-Biphenyl	13.451	154	1309	0.279	ng	# 48
48) 2-Nitroaniline	13.721	65	512	0.374	ng	# 69
49) Acenaphthylene	14.344	152	155	0.026	ng	# 23
50) Dimethylphthalate	13.968	163	2099	0.402	ng	# 88
51) 2,6-Dinitrotoluene	14.156	165	667	0.626	ng	# 16
52) Acenaphthene	14.691	154	428	0.123	ng	# 69
53) 3-Nitroaniline	14.585	138	365	0.317	ng	# 1
54) 2,4-Dinitrophenol	14.608	184	56	6.321	ng	# 34
55) Dibenzofuran	15.037	168	99	0.016	ng	# 32
56) 4-Nitrophenol	14.849	139	237	0.288	ng	# 1

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG052622\
 Data File : BG053720.D
 Acq On : 26 May 2022 17:04
 Operator : CG/JU
 Sample : N3059-01
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 4922A


Quant Time: May 27 05:00:23 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG050522.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed May 25 15:11:41 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
57) 2,4-Dinitrotoluene	14.973	165	326	0.211	ng	# 1
58) Fluorene	15.672	166	973	0.193	ng	# 28
60) Diethylphthalate	15.425	149	3636	0.673	ng	# 96
61) 4-Chlorophenyl-phenyle...	15.337	204	277	0.099	ng	# 40
62) 4-Nitroaniline	16.012	138	169	0.142	ng	# 94
63) Azobenzene	15.918	77	393	0.071	ng	# 59
65) 4,6-Dinitro-2-methylph...	16.030	198	54	0.062	ng	# 1
66) n-Nitrosodiphenylamine	16.318	169	521	0.135	ng	# 91
67) 4-Bromophenyl-phenylether	16.118	248	5787	3.366	ng	# 1
68) Hexachlorobenzene	16.888	284	75	0.040	ng	# 1
69) Atrazine	16.817	200	131	0.081	ng	# 1
71) Phenanthrene	17.416	178	2436	0.325	ng	# 92
72) Anthracene	17.481	178	564	0.077	ng	# 1
73) Carbazole	17.775	167	569	0.079	ng	# 2
74) Di-n-butylphthalate	18.327	149	3021	0.372	ng	# 85
75) Fluoranthene	19.420	202	1720	0.177	ng	# 1
77) Benzidine	19.649	184	332	0.108	ng	# 1
78) Pyrene	19.784	202	4436	0.481	ng	# 83
80) Butylbenzylphthalate	20.653	149	380	0.112	ng	# 20
81) Benzo(a)anthracene	21.617	228	1204	0.133	ng	# 42
82) 3,3'-Dichlorobenzidine	21.517	252	184	0.080	ng	# 1
83) Chrysene	21.687	228	1699	0.200	ng	# 37
84) Bis(2-ethylhexyl)phtha...	21.511	149	30389	6.512	ng	# 94
85) Di-n-octyl phthalate	22.703	149	3514	0.445	ng	# 71
87) Indeno(1,2,3-cd)pyrene	28.519	276	740	0.074	ng	# 88
88) Benzo(b)fluoranthene	23.837	252	1395	0.164	ng	# 1
89) Benzo(k)fluoranthene	23.867	252	1590	0.190	ng	# 1
90) Benzo(a)pyrene	24.683	252	448	0.062	ng	# 1
91) Dibenzo(a,h)anthracene	28.578	278	299	0.036	ng	# 1
92) Benzo(g,h,i)perylene	29.653	276	336	0.041	ng	# 1

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Quant Time: May 27 05:00:23 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG050522.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed May 25 15:11:41 2022
Response via : Initial Calibration









#1
1,4-Dichlorobenzene-d4
Concen: 20.000 ng
RT: 7.982 min Scan# 81
Delta R.T. -0.004 min
Lab File: BG053720.D
Acq: 26 May 2022 17:04




Instrument :
BNA_G
ClientSampleId :
4922A


Tgt Ion:152 Resp: 27116
Ion Ratio Lower Upper
152 100
150 149.1 131.6 197.4
115 57.2 50.9 76.3



#2
1,4-Dioxane
Concen: 0.252 ng
RT: 3.394 min Scan# 69
Delta R.T. -0.004 min
Lab File: BG053720.D
Acq: 26 May 2022 17:04

Tgt Ion: 88 Resp: 211
Ion Ratio Lower Upper
88 100
58 69.7 59.2 88.8
43 0.0 24.3 36.5








#3
Pyridine
Concen: 0.041 ng
RT: 3.864 min Scan# 14
Delta R.T. 0.055 min
Lab File: BG053720.D
Acq: 26 May 2022 17:04




Instrument :
BNA_G
ClientSampleId :
4922A


Tgt Ion: 79 Resp: 84
Ion Ratio Lower Upper
79 100
52 0.0 54.6 81.8#
51 163.4 32.5 48.7#



#4
n-Nitrosodimethylamine
Concen: 2.074 ng
RT: 3.729 min Scan# 126
Delta R.T. 0.008 min
Lab File: BG053720.D
Acq: 26 May 2022 17:04

Tgt Ion: 42 Resp: 2068
Ion Ratio Lower Upper
42 100
74 0.0 97.0 145.4#
44 35.5 7.0 10.4#









#5
2-Fluorophenol
Concen: 36.374 ng
RT: 5.556 min Scan# 41
Delta R.T. 0.002 min
Lab File: BG053720.D
Acq: 26 May 2022 17:04




Instrument :
BNA_G
ClientSampleId :
4922A


Tgt Ion:112 Resp: 58117
Ion Ratio Lower Upper
112 100
64 74.1 55.1 82.7
63 39.8 30.4 45.6



#6
Aniline
Concen: 0.890 ng
RT: 7.313 min Scan# 736
Delta R.T. 0.002 min
Lab File: BG053720.D
Acq: 26 May 2022 17:04

Tgt Ion: 93 Resp: 2389
Ion Ratio Lower Upper
93 100
66 52.8 32.9 49.3#
65 19.0 17.8 26.8









#7
Phenol-d6
Concen: 22.020 ng
RT: 7.172 min Scan# 71
Delta R.T. 0.008 min
Lab File: BG053720.D
Acq: 26 May 2022 17:04




Instrument :
BNA_G
ClientSampleId :
4922A


Tgt Ion: 99 Resp: 53396
Ion Ratio Lower Upper
99 100
42 32.5 17.0 25.6#
71 41.9 33.8 50.8



#9
Benzaldehyde
Concen: 0.324 ng
RT: 7.130 min Scan# 705
Delta R.T. 0.008 min
Lab File: BG053720.D
Acq: 26 May 2022 17:04

Tgt Ion: 77 Resp: 499
Ion Ratio Lower Upper
77 100
105 315.3 59.6 99.6#
106 44.6 56.8 96.8#









#10
Phenol
Concen: 2.240 ng
RT: 7.201 min Scan# 717
Delta R.T. 0.014 min
Lab File: BG053720.D
Acq: 26 May 2022 17:04




Instrument :
BNA_G
ClientSampleId :
4922A


Tgt Ion: 94 Resp: 5312
Ion Ratio Lower Upper
94 100
65 42.0 14.8 54.8
66 58.7 25.2 65.2



#11
bis(2-Chloroethyl)ether
Concen: 1.351 ng
RT: 7.313 min Scan# 736
Delta R.T. -0.092 min
Lab File: BG053720.D
Acq: 26 May 2022 17:04

Tgt Ion: 93 Resp: 2389
Ion Ratio Lower Upper
93 100
63 0.0 63.6 103.6#
95 0.0 10.1 50.1#









#15
Benzyl Alcohol
Concen: 0.792 ng
RT: 8.235 min Scan# 842
Delta R.T. 0.008 min
Lab File: BG053720.D
Acq: 26 May 2022 17:04




Instrument :
BNA_G
ClientSampleId :
4922A


Tgt Ion: 79 Resp: 1605
Ion Ratio Lower Upper
79 100
108 58.5 53.7 80.5
77 86.2 52.2 78.4#



#16
2,2'-oxybis(1-Chloropropane)
Concen: 4.079 ng
RT: 8.734 min Scan# 978
Delta R.T. 0.213 min
Lab File: BG053720.D
Acq: 26 May 2022 17:04

Tgt Ion: 45 Resp: 11742
Ion Ratio Lower Upper
45 100
77 8.0 0.0 32.7
79 0.0 0.0 29.7









#17
2-Methylphenol
Concen: 0.420 ng
RT: 8.452 min Scan# 91
Delta R.T. 0.008 min
Lab File: BG053720.D
Acq: 26 May 2022 17:04




Instrument :
BNA_G
ClientSampleId :
4922A


Tgt Ion:	107	Resp:	674
Ion	Ratio	Lower	Upper
107	100		
108	88.0	85.4	128.0
77	61.3	46.2	69.2
79	52.6	46.1	69.1



#18
Hexachloroethane
Concen: 0.961 ng
RT: 9.087 min Scan# 1038
Delta R.T. 0.019 min
Lab File: BG053720.D
Acq: 26 May 2022 17:04

Tgt Ion:	117	Resp:	726
Ion	Ratio	Lower	Upper
117	100		
119	369.1	75.0	112.6#
201	0.0	89.0	133.6#









#19
n-Nitroso-di-n-propylamine
Concen: 0.514 ng
RT: 8.763 min Scan# 91
Delta R.T. -0.033 min
Lab File: BG053720.D
Acq: 26 May 2022 17:04




Instrument :
BNA_G
ClientSampleId :
4922A


Tgt Ion: 70 Resp: 895
Ion Ratio Lower Upper
70 100
42 113.4 47.3 70.9#
101 44.0 7.0 10.4#
130 0.0 14.6 22.0#



#20
3+4-Methylphenols
Concen: 32.745 ng
RT: 8.769 min Scan# 984
Delta R.T. -0.004 min
Lab File: BG053720.D
Acq: 26 May 2022 17:04

Tgt Ion: 107 Resp: 74130
Ion Ratio Lower Upper
107 100
108 79.9 67.5 107.5
77 39.4 20.3 60.3
79 29.2 16.7 56.7









#21
Naphthalene-d8
Concen: 20.000 ng
RT: 10.796 min Scan# 11
Delta R.T. -0.004 min
Lab File: BG053720.D
Acq: 26 May 2022 17:04




Instrument :
BNA_G
ClientSampleId :
4922A


Tgt Ion:136	Resp:	100853
Ion Ratio	Lower	Upper
136	100	
137	9.7	8.5 12.7
54	8.4	7.4 11.2
68	7.2	4.9 7.3



#22
Acetophenone
Concen: 1.217 ng
RT: 8.810 min Scan# 991
Delta R.T. -0.004 min
Lab File: BG053720.D
Acq: 26 May 2022 17:04

Tgt Ion:105	Resp:	3360
Ion Ratio	Lower	Upper
105	100	
71	203.6	1.1 1.7#
51	62.1	25.4 38.0#
120	14.6	17.3 25.9#









#23
Nitrobenzene-d5
Concen: 66.019 ng
RT: 9.151 min Scan# 1049
Delta R.T. -0.004 min
Lab File: BG053720.D
Acq: 26 May 2022 17:04



Instrument :
BNA_G
ClientSampleId :
4922A


Tgt Ion: 82 Resp: 156413
Ion Ratio Lower Upper
82 100
128 34.4 27.9 41.9
54 53.7 41.2 61.8



#24
Nitrobenzene
Concen: 0.571 ng
RT: 9.151 min Scan# 1049
Delta R.T. -0.045 min
Lab File: BG053720.D
Acq: 26 May 2022 17:04

Tgt Ion: 77 Resp: 1310
Ion Ratio Lower Upper
77 100
123 0.0 27.4 41.0#
65 51.5 11.4 17.0#









#25
Isophorone
Concen: 0.554 ng
RT: 9.850 min Scan# 1119
Delta R.T. 0.131 min
Lab File: BG053720.D
Acq: 26 May 2022 17:04




Instrument :
BNA_G
ClientSampleId :
4922A


Tgt Ion: 82 Resp: 2210
Ion Ratio Lower Upper
82 100
95 18.9 6.2 9.4#
138 0.0 11.6 17.4#



#27
2,4-Dimethylphenol
Concen: 0.638 ng
RT: 9.985 min Scan# 1191
Delta R.T. 0.014 min
Lab File: BG053720.D
Acq: 26 May 2022 17:04

Tgt Ion: 122 Resp: 892
Ion Ratio Lower Upper
122 100
107 134.2 109.4 164.0
121 58.1 48.7 73.1





#28

bis(2-Chloroethoxy)methane

Concen: 0.862 ng

RT: 10.209 min Scan# 1177

Delta R.T. 0.008 min

Lab File: BG053720.D

Acq: 26 May 2022 17:04

Instrument :

BNA_G

ClientSampleId :

4922A





Tgt Ion: 93 Resp: 1976

Ion Ratio Lower Upper

93 100

95 659.0 26.9 40.3#

123 0.0 7.7 11.5#



#29

2,4-Dichlorophenol

Concen: 0.072 ng

RT: 10.661 min Scan# 1306

Delta R.T. 0.202 min

Lab File: BG053720.D

Acq: 26 May 2022 17:04




Tgt Ion:162 Resp: 121


Ion Ratio Lower Upper

162 100

164 0.0 46.6 86.6#

98 128.7 22.8 62.8#









#31
Naphthalene
Concen: 1.297 ng
RT: 10.843 min Scan# 11
Delta R.T. -0.004 min
Lab File: BG053720.D
Acq: 26 May 2022 17:04




Instrument :
BNA_G
ClientSampleId :
4922A


Tgt Ion:128 Resp: 6653
Ion Ratio Lower Upper
128 100
129 47.1 9.0 13.6#
127 13.4 11.1 16.7



#32
Benzoic acid
Concen: 18.407 ng
RT: 10.397 min Scan# 1261
Delta R.T. 0.237 min
Lab File: BG053720.D
Acq: 26 May 2022 17:04

Tgt Ion:122 Resp: 10529
Ion Ratio Lower Upper
122 100
105 147.1 111.8 151.8
77 131.2 109.4 149.4









#33
4-Chloroaniline
Concen: 0.136 ng
RT: 10.914 min Scan# 11
Delta R.T. -0.045 min
Lab File: BG053720.D
Acq: 26 May 2022 17:04




Instrument :
BNA_G
ClientSampleId :
4922A


Tgt Ion:127	Resp:	292
Ion	Ratio	Lower Upper
127	100	
129	166.1	26.3 39.5#
65	95.4	31.1 46.7#
92	0.0	19.1 28.7#



#35
Caprolactam
Concen: 0.401 ng
RT: 11.742 min Scan# 1490
Delta R.T. 0.002 min
Lab File: BG053720.D
Acq: 26 May 2022 17:04

Tgt Ion:113	Resp:	246
Ion	Ratio	Lower Upper
113	100	
55	1398.1	169.8 209.8#
56	438.0	141.2 181.2#





#36

4-Chloro-3-methylphenol

Concen: 0.518 ng

RT: 12.147 min Scan# 1

Delta R.T. 0.055 min

Lab File: BG053720.D

Acq: 26 May 2022 17:04





Instrument :

BNA_G

ClientSampleId :

4922A

Tgt Ion:	107	Resp:	1045
Ion Ratio	Lower	Upper	
107	100		
144	0.0	18.5	27.7#
142	0.0	55.0	82.4#



#37

2-Methylnaphthalene

Concen: 0.737 ng




RT: 12.458 min Scan# 1612


Delta R.T. 0.002 min

Lab File: BG053720.D

Acq: 26 May 2022 17:04

Tgt Ion:	142	Resp:	2822
Ion Ratio	Lower	Upper	
142	100		
141	69.2	72.3	108.5#
115	312.7	33.7	50.5#









#38
1-Methylnaphthalene
Concen: 0.504 ng
RT: 12.670 min Scan# 1
Delta R.T. -0.004 min
Lab File: BG053720.D
Acq: 26 May 2022 17:04




Instrument :
BNA_G
ClientSampleId :
4922A


Tgt Ion:142 Resp: 1882
Ion Ratio Lower Upper
142 100
141 58.5 76.4 114.6#
116 16.8 3.3 4.9#



#39
Acenaphthene-d10
Concen: 20.000 ng
RT: 14.626 min Scan# 1981
Delta R.T. 0.002 min
Lab File: BG053720.D
Acq: 26 May 2022 17:04

Tgt Ion:164 Resp: 68780
Ion Ratio Lower Upper
164 100
162 107.2 82.3 123.5
160 43.9 34.2 51.2









#42
 2,4,6-Tribromophenol
 Concen: 89.498 ng
 RT: 16.118 min Scan# 2183
 Delta R.T. 0.002 min
 Lab File: BG053720.D
 Acq: 26 May 2022 17:04




Instrument :
 BNA_G
 ClientSampleId :
 4922A


Tgt Ion:330 Resp: 90789
 Ion Ratio Lower Upper
 330 100
 332 95.6 76.0 114.0
 141 32.9 27.1 40.7



#45
 2-Fluorobiphenyl
 Concen: 70.278 ng
 RT: 13.246 min Scan# 1746
 Delta R.T. 0.002 min
 Lab File: BG053720.D
 Acq: 26 May 2022 17:04

Tgt Ion:172 Resp: 330786
 Ion Ratio Lower Upper
 172 100
 171 36.1 28.5 42.7
 170 23.4 18.9 28.3









#46
1,1'-Biphenyl
Concen: 0.279 ng
RT: 13.451 min Scan# 1
Delta R.T. -0.010 min
Lab File: BG053720.D
Acq: 26 May 2022 17:04




Instrument :
BNA_G
ClientSampleId :
4922A


Tgt Ion:154 Resp: 1309
Ion Ratio Lower Upper
154 100
153 79.9 21.6 61.6#
76 22.0 0.0 32.6



#48
2-Nitroaniline
Concen: 0.374 ng
RT: 13.721 min Scan# 1827
Delta R.T. 0.014 min
Lab File: BG053720.D
Acq: 26 May 2022 17:04

Tgt Ion: 65 Resp: 512
Ion Ratio Lower Upper
65 100
92 53.7 51.5 77.3
138 48.9 71.3 106.9#









#49
Acenaphthylene
Concen: 0.026 ng
RT: 14.344 min Scan# 1882
Delta R.T. -0.004 min
Lab File: BG053720.D
Acq: 26 May 2022 17:04




Instrument :
BNA_G
ClientSampleId :
4922A


Tgt Ion:152 Resp: 155
Ion Ratio Lower Upper
152 100
151 69.7 16.4 24.6#
153 0.0 10.8 16.2#



#50
Dimethylphthalate
Concen: 0.402 ng
RT: 13.968 min Scan# 1869
Delta R.T. -0.110 min
Lab File: BG053720.D
Acq: 26 May 2022 17:04

Tgt Ion:163 Resp: 2099
Ion Ratio Lower Upper
163 100
194 0.0 3.0 4.6#
164 14.7 8.0 12.0#









#51
2,6-Dinitrotoluene
Concen: 0.626 ng
RT: 14.156 min Scan# 1901
Delta R.T. -0.045 min
Lab File: BG053720.D
Acq: 26 May 2022 17:04




Instrument :
BNA_G
ClientSampleId :
4922A


Tgt Ion:165 Resp: 667
Ion Ratio Lower Upper
165 100
63 0.0 54.1 81.1#
89 0.0 54.8 82.2#



#52
Acenaphthene
Concen: 0.123 ng
RT: 14.691 min Scan# 1992
Delta R.T. 0.002 min
Lab File: BG053720.D
Acq: 26 May 2022 17:04

Tgt Ion:154 Resp: 428
Ion Ratio Lower Upper
154 100
153 104.1 92.7 139.1
152 0.0 42.6 64.0#









#53
3-Nitroaniline
Concen: 0.317 ng
RT: 14.585 min Scan# 1912
Delta R.T. 0.049 min
Lab File: BG053720.D
Acq: 26 May 2022 17:04




Instrument :
BNA_G
ClientSampleId :
4922A


Tgt Ion:138 Resp: 365
Ion Ratio Lower Upper
138 100
108 0.0 9.2 13.8#
92 0.0 101.9 152.9#



#54
2,4-Dinitrophenol
Concen: 6.321 ng
RT: 14.608 min Scan# 1978
Delta R.T. -0.145 min
Lab File: BG053720.D
Acq: 26 May 2022 17:04

Tgt Ion:184 Resp: 56
Ion Ratio Lower Upper
184 100
63 114.4 61.1 91.7#
154 0.0 60.6 91.0#









#55
Dibenzofuran
Concen: 0.016 ng
RT: 15.037 min Scan# 2019
Delta R.T. 0.014 min
Lab File: BG053720.D
Acq: 26 May 2022 17:04




Instrument :
BNA_G
ClientSampleId :
4922A


Tgt Ion:168	Resp:	99
Ion Ratio	Lower	Upper
168	100	
139	67.6	30.5 45.7#
169	61.6	10.3 15.5#



#56
4-Nitrophenol
Concen: 0.288 ng
RT: 14.849 min Scan# 2019
Delta R.T. -0.016 min
Lab File: BG053720.D
Acq: 26 May 2022 17:04

Tgt Ion:139	Resp:	237
Ion Ratio	Lower	Upper
139	100	
109	289.9	88.5 128.5#
65	248.0	93.9 133.9#









#57
2,4-Dinitrotoluene
Concen: 0.211 ng
RT: 14.973 min Scan# 2040
Delta R.T. -0.022 min
Lab File: BG053720.D
Acq: 26 May 2022 17:04




Instrument :
BNA_G
ClientSampleId :
4922A


Tgt Ion:165	Resp:	326
Ion Ratio	Lower	Upper
165	100	
63	135.1	34.6 52.0#
89	0.0	60.3 90.5#
182	0.0	2.2 3.4#



#58
Fluorene
Concen: 0.193 ng
RT: 15.672 min Scan# 2159
Delta R.T. -0.004 min
Lab File: BG053720.D
Acq: 26 May 2022 17:04

Tgt Ion:166	Resp:	973
Ion Ratio	Lower	Upper
166	100	
165	161.2	75.0 112.6#
167	48.8	10.8 16.2#









#60
Diethylphthalate
Concen: 0.673 ng
RT: 15.425 min Scan# 2117
Delta R.T. -0.010 min
Lab File: BG053720.D
Acq: 26 May 2022 17:04




Instrument :
BNA_G
ClientSampleId :
4922A


Tgt Ion:149 Resp: 3636
Ion Ratio Lower Upper
149 100
177 25.1 18.4 27.6
150 12.9 11.2 16.8



#61
4-Chlorophenyl-phenylether
Concen: 0.099 ng
RT: 15.337 min Scan# 2102
Delta R.T. -0.321 min
Lab File: BG053720.D
Acq: 26 May 2022 17:04

Tgt Ion:204 Resp: 277
Ion Ratio Lower Upper
204 100
206 0.0 27.5 41.3#
141 101.3 45.3 67.9#









#62
4-Nitroaniline
Concen: 0.142 ng
RT: 16.012 min Scan# 2110
Delta R.T. 0.313 min
Lab File: BG053720.D
Acq: 26 May 2022 17:04




Instrument :
BNA_G
ClientSampleId :
4922A


Tgt Ion:138 Resp: 169
Ion Ratio Lower Upper
138 100
92 52.6 40.5 80.5
108 119.9 103.7 143.7



#63
Azobenzene
Concen: 0.071 ng
RT: 15.918 min Scan# 2201
Delta R.T. -0.033 min
Lab File: BG053720.D
Acq: 26 May 2022 17:04

Tgt Ion: 77 Resp: 393
Ion Ratio Lower Upper
77 100
182 16.6 3.8 43.8
105 74.9 0.0 37.8#
51 25.6 10.7 50.7









#64
Phenanthrene-d10
Concen: 20.000 ng
RT: 17.369 min Scan# 24
Delta R.T. -0.004 min
Lab File: BG053720.D
Acq: 26 May 2022 17:04




Instrument :
BNA_G
ClientSampleId :
4922A


Tgt Ion:188 Resp: 147223
Ion Ratio Lower Upper
188 100
94 6.0 5.4 8.0
80 7.9 6.8 10.2



#65
4,6-Dinitro-2-methylphenol
Concen: 0.062 ng
RT: 16.030 min Scan# 2220
Delta R.T. 0.266 min
Lab File: BG053720.D
Acq: 26 May 2022 17:04

Tgt Ion:198 Resp: 54
Ion Ratio Lower Upper
198 100
51 161.0 28.1 68.1#
105 268.8 20.3 60.3#









#66
n-Nitrosodiphenylamine
Concen: 0.135 ng
RT: 16.318 min Scan# 2142
Delta R.T. 0.442 min
Lab File: BG053720.D
Acq: 26 May 2022 17:04




Instrument :
BNA_G
ClientSampleId :
4922A


Tgt Ion:169 Resp: 521
Ion Ratio Lower Upper
169 100
168 63.1 55.9 83.9
167 32.2 30.6 46.0



#67
4-Bromophenyl-phenylether
Concen: 3.366 ng
RT: 16.118 min Scan# 2235
Delta R.T. -0.433 min
Lab File: BG053720.D
Acq: 26 May 2022 17:04

Tgt Ion:248 Resp: 5787
Ion Ratio Lower Upper
248 100
250 216.4 75.2 112.8#
141 481.5 50.0 75.0#









#68
Hexachlorobenzene
Concen: 0.040 ng
RT: 16.888 min Scan# 21
Delta R.T. 0.202 min
Lab File: BG053720.D
Acq: 26 May 2022 17:04




Instrument :
BNA_G
ClientSampleId :
4922A


Tgt Ion:284 Resp: 75
Ion Ratio Lower Upper
284 100
142 124.9 23.6 35.4#
249 0.0 25.8 38.6#



#69
Atrazine
Concen: 0.081 ng
RT: 16.817 min Scan# 2354
Delta R.T. -0.004 min
Lab File: BG053720.D
Acq: 26 May 2022 17:04

Tgt Ion:200 Resp: 131
Ion Ratio Lower Upper
200 100
173 0.0 3.7 43.7#
215 129.1 25.2 65.2#









#71
Phenanthrene
Concen: 0.325 ng
RT: 17.416 min Scan# 2404
Delta R.T. 0.002 min
Lab File: BG053720.D
Acq: 26 May 2022 17:04




Instrument :
BNA_G
ClientSampleId :
4922A


Tgt Ion:178 Resp: 2436
Ion Ratio Lower Upper
178 100
176 19.6 14.1 21.1
179 20.0 12.1 18.1#



#72
Anthracene
Concen: 0.077 ng
RT: 17.481 min Scan# 2467
Delta R.T. -0.022 min
Lab File: BG053720.D
Acq: 26 May 2022 17:04

Tgt Ion:178 Resp: 564
Ion Ratio Lower Upper
178 100
176 53.4 15.2 22.8#
179 102.4 13.0 19.4#









#73
 Carbazole
 Concen: 0.079 ng
 RT: 17.775 min Scan# 211
 Delta R.T. -0.004 min
 Lab File: BG053720.D
 Acq: 26 May 2022 17:04




Instrument :
 BNA_G
 ClientSampleId :
 4922A


Tgt Ion:167 Resp: 569
 Ion Ratio Lower Upper
 167 100
 166 62.5 18.2 27.2#
 139 62.5 10.1 15.1#



#74
 Di-n-butylphthalate
 Concen: 0.372 ng
 RT: 18.327 min Scan# 2611
 Delta R.T. 0.002 min
 Lab File: BG053720.D
 Acq: 26 May 2022 17:04

Tgt Ion:149 Resp: 3021
 Ion Ratio Lower Upper
 149 100
 150 13.7 7.7 11.5#
 104 13.8 5.4 8.0#









#75
Fluoranthene
Concen: 0.177 ng
RT: 19.420 min Scan# 2174
Delta R.T. -0.004 min
Lab File: BG053720.D
Acq: 26 May 2022 17:04




Instrument :
BNA_G
ClientSampleId :
4922A


Tgt Ion:202 Resp: 1720
Ion Ratio Lower Upper
202 100
101 84.7 0.0 26.9#
203 49.2 0.0 37.2#



#76
Chrysene-d12
Concen: 20.000 ng
RT: 21.634 min Scan# 3174
Delta R.T. -0.004 min
Lab File: BG053720.D
Acq: 26 May 2022 17:04

Tgt Ion:240 Resp: 142433
Ion Ratio Lower Upper
240 100
120 5.8 4.6 7.0
236 26.4 20.2 30.4









#77
Benzidine
Concen: 0.108 ng
RT: 19.649 min Scan# 21
Delta R.T. 0.049 min
Lab File: BG053720.D
Acq: 26 May 2022 17:04




Instrument :
BNA_G
ClientSampleId :
4922A


Tgt Ion:184 Resp: 332
Ion Ratio Lower Upper
184 100
185 171.7 11.8 17.6#
183 199.6 8.9 13.3#



#78
Pyrene
Concen: 0.481 ng
RT: 19.784 min Scan# 2859
Delta R.T. -0.004 min
Lab File: BG053720.D
Acq: 26 May 2022 17:04

Tgt Ion:202 Resp: 4436
Ion Ratio Lower Upper
202 100
200 26.8 16.6 24.8#
203 28.1 14.6 21.8#









#79
 Terphenyl-d14
 Concen: 66.962 ng
 RT: 19.978 min Scan# 21
 Delta R.T. 0.002 min
 Lab File: BG053720.D
 Acq: 26 May 2022 17:04




Instrument :
 BNA_G
 ClientSampleId :
 4922A


Tgt Ion:244 Resp: 512096
 Ion Ratio Lower Upper
 244 100
 212 7.6 6.5 9.7
 122 6.0 5.0 7.4



#80
 Butylbenzylphthalate
 Concen: 0.112 ng
 RT: 20.653 min Scan# 3007
 Delta R.T. -0.010 min
 Lab File: BG053720.D
 Acq: 26 May 2022 17:04

Tgt Ion:149 Resp: 380
 Ion Ratio Lower Upper
 149 100
 91 163.4 59.6 89.4#
 206 24.4 18.6 28.0









#81
Benzo(a)anthracene
Concen: 0.133 ng
RT: 21.617 min Scan# 3117
Delta R.T. 0.002 min
Lab File: BG053720.D
Acq: 26 May 2022 17:04




Instrument :
BNA_G
ClientSampleId :
4922A


Tgt Ion:228 Resp: 1204
Ion Ratio Lower Upper
228 100
226 45.7 21.6 32.4#
229 60.6 16.2 24.2#



#82
3,3'-Dichlorobenzidine
Concen: 0.080 ng
RT: 21.517 min Scan# 3154
Delta R.T. -0.010 min
Lab File: BG053720.D
Acq: 26 May 2022 17:04

Tgt Ion:252 Resp: 184
Ion Ratio Lower Upper
252 100
254 0.0 49.8 74.8#
126 403.1 6.0 9.0#









#83
Chrysene
Concen: 0.200 ng
RT: 21.687 min Scan# 3134
Delta R.T. 0.002 min
Lab File: BG053720.D
Acq: 26 May 2022 17:04




Instrument :
BNA_G
ClientSampleId :
4922A


Tgt Ion:228	Resp:	1699
Ion Ratio	Lower	Upper
228	100	
226	45.2	23.4 35.2#
229	70.7	15.5 23.3#



#84
Bis(2-ethylhexyl)phthalate
Concen: 6.512 ng
RT: 21.511 min Scan# 3153
Delta R.T. 0.002 min
Lab File: BG053720.D
Acq: 26 May 2022 17:04

Tgt Ion:149	Resp:	30389
Ion Ratio	Lower	Upper
149	100	
167	28.7	20.2 30.4
279	3.6	3.6 5.4#









#85
Di-n-octyl phthalate
Concen: 0.445 ng
RT: 22.703 min Scan# 31
Delta R.T. -0.004 min
Lab File: BG053720.D
Acq: 26 May 2022 17:04




Instrument :
BNA_G
ClientSampleId :
4922A


Tgt Ion:149 Resp: 3514
Ion Ratio Lower Upper
149 100
167 11.0 1.3 1.9#
43 0.0 9.0 13.4#



#86
Perylene-d12
Concen: 20.000 ng
RT: 24.848 min Scan# 3721
Delta R.T. 0.008 min
Lab File: BG053720.D
Acq: 26 May 2022 17:04

Tgt Ion:264 Resp: 144063
Ion Ratio Lower Upper
264 100
260 22.8 19.0 28.6
265 23.0 17.4 26.2





#87

Indeno(1,2,3-cd)pyrene

Concen: 0.074 ng

RT: 28.519 min Scan# 41

Delta R.T. 0.008 min

Lab File: BG053720.D

Acq: 26 May 2022 17:04





Instrument :

BNA_G

ClientSampleId :

4922A

Tgt Ion:	276	Resp:	740
Ion Ratio	Lower	Upper	
276	100		
138	0.0	5.0	7.4#
277	30.7	20.3	30.5#



#88

Benzo(b)fluoranthene

Concen: 0.164 ng




RT: 23.837 min Scan# 3549


Delta R.T. 0.014 min

Lab File: BG053720.D

Acq: 26 May 2022 17:04

Tgt Ion:	252	Resp:	1395
Ion Ratio	Lower	Upper	
252	100		
253	86.0	18.0	27.0#
125	437.6	4.6	6.8#









#89
Benzo(k)fluoranthene
Concen: 0.190 ng
RT: 23.867 min Scan# 311
Delta R.T. -0.022 min
Lab File: BG053720.D
Acq: 26 May 2022 17:04




Instrument :
BNA_G
ClientSampleId :
4922A


Tgt Ion:252 Resp: 1590
Ion Ratio Lower Upper
252 100
253 69.4 18.4 27.6#
125 482.4 4.8 7.2#



#90
Benzo(a)pyrene
Concen: 0.062 ng
RT: 24.683 min Scan# 3693
Delta R.T. -0.004 min
Lab File: BG053720.D
Acq: 26 May 2022 17:04

Tgt Ion:252 Resp: 448
Ion Ratio Lower Upper
252 100
253 102.1 17.8 26.8#
125 615.2 4.3 6.5#









#91
Dibenzo(a,h)anthracene
Concen: 0.036 ng
RT: 28.578 min Scan# 41
Delta R.T. 0.025 min
Lab File: BG053720.D
Acq: 26 May 2022 17:04

Instrument :
BNA_G
ClientSampleId :
4922A

Tgt Ion:278 Resp: 299
Ion Ratio Lower Upper
278 100
139 340.5 7.2 10.8#
279 0.0 18.6 27.8#



#92
Benzo(g,h,i)perylene
Concen: 0.041 ng
RT: 29.653 min Scan# 4539
Delta R.T. 0.002 min
Lab File: BG053720.D
Acq: 26 May 2022 17:04

Tgt Ion:276 Resp: 336
Ion Ratio Lower Upper
276 100
277 82.8 17.3 25.9#
138 105.8 9.6 14.4#

