

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG052622\
 Data File : BG053726.D
 Acq On : 26 May 2022 21:02
 Operator : CG/JU
 Sample : N2897-08DL 10X
 Misc :
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 P042-SS001-1218-01DL

Manual Integrations
 APPROVED

Reviewed By : Christian Giraldo 05/27/2022
 Supervised By : Jagrut Upadhyay 05/27/2022

Quant Time: May 27 05:01:15 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG050522.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed May 25 15:11:41 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.987	152	36452	20.000	ng	0.00
21) Naphthalene-d8	10.795	136	145387	20.000	ng	0.00
39) Acenaphthene-d10	14.625	164	100779	20.000	ng	0.00
64) Phenanthrene-d10	17.369	188	213148	20.000	ng	0.00
76) Chrysene-d12	21.633	240	195335	20.000	ng	0.00
86) Perylene-d12	24.841	264	203699	20.000	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.561	112	19332	9.001	ng	0.00
7) Phenol-d6	7.171	99	28359	8.700	ng	0.00
23) Nitrobenzene-d5	9.150	82	19939m	5.838	ng	0.00
42) 2,4,6-Tribromophenol	16.117	330	12757	8.583	ng	0.00
45) 2-Fluorobiphenyl	13.245	172	46655	6.765	ng	0.00
79) Terphenyl-d14	19.971	244	71786	6.845	ng	0.00
Target Compounds						
84) Bis(2-ethylhexyl)phtha...	21.510	149	306974	47.965	ng	Qvalue 98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG052622\
Data File : BG053726.D
Acq On : 26 May 2022 21:02
Operator : CG/JU
Sample : N2897-08DL 10X
Misc :
ALS Vial : 16 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
P042-SS001-1218-01DL

Manual Integrations
APPROVED

Reviewed By : Christian Giraldo 05/27/2022
Supervised By : Jagrut Upadhyay 05/27/2022

Quant Time: May 27 05:01:15 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG050522.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed May 25 15:11:41 2022
Response via : Initial Calibration

