

Data Path : Z:\SVOASRV\HPCHEM1\BNA G\DATA\BG052720\
 Data File : BG045483.D
 Acq On : 27 May 2020 19:37
 Operator : CG/JU
 Sample : L2744-14
 Misc :
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 NM30-COMP-2020-0-6

Quant Time: May 28 01:26:48 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA G\METHODS\8270-BG051520.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed May 27 13:40:45 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.96	152	105987	20.00	ng	0.00
21) Naphthalene-d8	10.77	136	435518	20.00	ng	0.00
39) Acenaphthene-d10	14.60	164	296742	20.00	ng	0.00
64) Phenanthrene-d10	17.35	188	628063	20.00	ng	0.00
76) Chrysene-d12	21.61	240	557028	20.00	ng	0.00
87) Perylene-d12	24.75	264	609187	20.00	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.56	112	479385	88.39	ng	0.00
7) Phenol-d6	7.17	99	744928	96.51	ng	0.00
23) Nitrobenzene-d5	9.12	82	515921	64.60	ng	0.00
42) 2,4,6-Tribromophenol	16.09	330	368041	96.98	ng	0.00
45) 2-Fluorobiphenyl	13.22	172	1221941	69.85	ng	0.00
79) Terphenyl-d14	19.97	244	1698659	71.12	ng	0.00
Target Compounds						
50) Dimethylphthalate	14.05	163	182958	8.988	ng	Qvalue # 96
71) Phenanthrene	17.39	178	92469	3.373	ng	99
75) Fluoranthene	19.40	202	164044	4.704	ng	97
78) Pyrene	19.76	202	154848	4.877	ng	96
81) Benzo(a)anthracene	21.59	228	81674	2.632	ng	98
83) Chrysene	21.65	228	78322	2.625	ng	92
88) Benzo(b)fluoranthene	23.76	252	103354	3.163	ng	95
90) Benzo(a)pyrene	24.60	252	72689	2.389	ng	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA G\DATA\BG052720\
Data File : BG045483.D
Acq On : 27 May 2020 19:37
Operator : CG/JU
Sample : L2744-14
Misc :
ALS Vial : 12 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
NM30-COMP-2020-0-6

Quant Time: May 28 01:26:48 2020
Quant Method : Z:\SVOASRV\HPCHEM1\BNA G\METHODS\8270-BG051520.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed May 27 13:40:45 2020
Response via : Initial Calibration

