


Data Path : Z:\SVOASRV\HPCHEM1\BNA G\DATA\BG052720\
 Data File : BG045472.D
 Acq On : 27 May 2020 12:12
 Operator : CG/JU
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 DFTPP

Integration File: rteint.p

Method : Z:\SVOASRV\HPCHEM1\BNA_G\METHODS\8270-BG051520.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Wed May 27 13:40:45 2020



AutoFind: Scans 2503, 2504, 2505; Background Corrected with Scan 2497

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	27.6	18004	PASS
68	69	0.00	2	1.1	264	PASS
69	198	0.00	100	35.8	23282	PASS
70	69	0.00	2	0.3	77	PASS
127	198	10	80	46.2	30074	PASS
197	198	0.00	2	0.0	0	PASS
198	198	100	100	100.0	65122	PASS
199	198	5	9	7.8	5111	PASS
275	198	10	60	31.1	20277	PASS
365	198	1	100	5.1	3296	PASS
441	198	0.01	100	14.4	9386	PASS
442	198	50	100	93.2	60717	PASS
443	442	15	24	19.1	11586	PASS