

Data Path : Z:\HPCHEM1\BNA_G\Data\BG052815\
 Data File : BG017337.D
 Acq On : 28 May 2015 22:07
 Operator : TP/IZ
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_G
 LabSampleId :
 SSTDCCC040

Quant Time: May 29 01:15:51 2015
 Quant Method : Z:\HPCHEM1\BNA_G\METHODS\8270-BG051315.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri May 29 01:04:23 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	64	-0.05
2	1,4-Dioxane	0.445	0.505	-13.5	73	-0.06
3	Pyridine	1.346	1.418	-5.3	65	-0.06
4	n-Nitrosodimethylamine	1.021	1.014	0.7	61	-0.06
5 S	2-Fluorophenol	1.127	1.153	-2.3	62	-0.06
6	Aniline	2.189	2.170	0.9	60	-0.05
7 S	Phenol-d6	1.582	1.596	-0.9	61	-0.05
8	2-Chlorophenol	1.349	1.357	-0.6	61	-0.05
9	Benzaldehyde	1.041	1.009	3.1	61	-0.05
10 C	Phenol	1.678	1.737	-3.5	64	-0.05
11	bis(2-Chloroethyl)ether	1.258	1.291	-2.6	61	-0.05
12	1,3-Dichlorobenzene	1.506	1.501	0.3	63	-0.06
13 C	1,4-Dichlorobenzene	1.520	1.499	1.4	61	-0.05
14	1,2-Dichlorobenzene	1.452	1.417	2.4	60	-0.06
15	Benzyl Alcohol	1.537	1.476	4.0	59	-0.05
16	2,2'-oxybis(1-Chloropropane	2.040	2.158	-5.8	66	-0.05
17	2-Methylphenol	1.216	1.214	0.2	63	-0.06
18	Hexachloroethane	0.635	0.624	1.7	61	-0.06
19 P	n-Nitroso-di-n-propylamine	1.378	1.334	3.2	59	-0.05
20	3+4-Methylphenols	1.691	1.689	0.1	60	-0.06
21 I	Naphthalene-d8	1.000	1.000	0.0	59	-0.06
22	Acetophenone	0.485	0.484	0.2	59	-0.06
23 S	Nitrobenzene-d5	0.428	0.438	-2.3	59	-0.05
24	Nitrobenzene	0.419	0.432	-3.1	61	-0.05
25	Isophorone	0.790	0.801	-1.4	58	-0.06
26 C	2-Nitrophenol	0.187	0.186	0.5	58	-0.06
27	2,4-Dimethylphenol	0.308	0.310	-0.6	60	-0.06
28	bis(2-Chloroethoxy)methane	0.384	0.396	-3.1	60	-0.06
29 C	2,4-Dichlorophenol	0.291	0.300	-3.1	59	-0.06
30	1,2,4-Trichlorobenzene	0.330	0.330	0.0	58	-0.06
31	Naphthalene	0.979	0.988	-0.9	59	-0.06
32	Benzoic acid	0.209	0.160	23.4	44#	-0.04
33	4-Chloroaniline	0.463	0.451	2.6	55	-0.06
34 C	Hexachlorobutadiene	0.209	0.208	0.5	58	-0.06
35	Caprolactam	0.146	0.145	0.7	60	-0.04
36 C	4-Chloro-3-methylphenol	0.368	0.380	-3.3	60	-0.06
37	2-Methylnaphthalene	0.776	0.756	2.6	59	-0.06
38 I	Acenaphthene-d10	1.000	1.000	0.0	58	-0.05
39	1,2,4,5-Tetrachlorobenzene	0.566	0.580	-2.5	58	-0.05
40 P	Hexachlorocyclopentadiene	0.345	0.282	18.3	46#	-0.06
41 S	2,4,6-Tribromophenol	0.242	0.237	2.1	55	-0.04
42 C	2,4,6-Trichlorophenol	0.401	0.401	0.0	55	-0.05
43	2,4,5-Trichlorophenol	0.414	0.415	-0.2	56	-0.06
44 S	2-Fluorobiphenyl	1.363	1.387	-1.8	57	-0.05

Data Path : Z:\HPCHEM1\BNA_G\Data\BG052815\
 Data File : BG017337.D
 Acq On : 28 May 2015 22:07
 Operator : TP/IZ
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_G
 LabSampleId :
 SSTDCCC040

Quant Time: May 29 01:15:51 2015
 Quant Method : Z:\HPCHEM1\BNA_G\METHODS\8270-BG051315.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri May 29 01:04:23 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
45	1,1'-Biphenyl	1.441	1.450	-0.6	57	-0.05
46	2-Chloronaphthalene	1.141	1.163	-1.9	58	-0.05
47	2-Nitroaniline	0.444	0.464	-4.5	60	-0.05
48	Acenaphthylene	1.953	2.013	-3.1	57	-0.06
49	Dimethylphthalate	1.564	1.556	0.5	56	-0.05
50	2,6-Dinitrotoluene	0.347	0.356	-2.6	57	-0.04
51 C	Acenaphthene	1.154	1.164	-0.9	57	-0.05
52	3-Nitroaniline	0.384	0.390	-1.6	57	-0.05
53 P	2,4-Dinitrophenol	0.201	0.188	6.5	50#	-0.04
54	Dibenzofuran	1.715	1.730	-0.9	57	-0.05
55 P	4-Nitrophenol	0.310	0.300	3.2	56	-0.05
56	2,4-Dinitrotoluene	0.483	0.511	-5.8	59	-0.04
57	Fluorene	1.518	1.506	0.8	55	-0.05
58	2,3,4,6-Tetrachlorophenol	0.380	0.361	5.0	52	-0.05
59	Diethylphthalate	1.680	1.629	3.0	56	-0.06
60	4-Chlorophenyl-phenylether	0.780	0.781	-0.1	56	-0.05
61	4-Nitroaniline	0.429	0.457	-6.5	59	-0.04
62	Azobenzene	1.605	1.672	-4.2	59	-0.05
63 I	Phenanthrene-d10	1.000	1.000	0.0	59	-0.05
64	4,6-Dinitro-2-methylphenol	0.137	0.138	-0.7	56	-0.04
65 c	n-Nitrosodiphenylamine	0.614	0.597	2.8	56	-0.04
66	4-Bromophenyl-phenylether	0.218	0.211	3.2	56	-0.05
67	Hexachlorobenzene	0.230	0.227	1.3	57	-0.04
68	Atrazine	0.225	0.215	4.4	54	-0.05
69 C	Pentachlorophenol	0.144	0.127	11.8	50#	-0.04
70	Phenanthrene	1.031	1.033	-0.2	57	-0.05
71	Anthracene	1.045	1.054	-0.9	58	-0.05
72	Carbazole	0.960	0.996	-3.8	59	-0.05
73	Di-n-butylphthalate	1.285	1.332	-3.7	60	-0.05
74 C	Fluoranthene	1.239	1.288	-4.0	60	-0.05
75 I	Chrysene-d12	1.000	1.000	0.0	71	-0.04
76	Benzidine	0.653	0.576	11.8	59	-0.05
77	Pyrene	1.227	1.100	10.4	61	-0.04
78 S	Terphenyl-d14	0.960	0.909	5.3	65	-0.05
79	Butylbenzylphthalate	0.558	0.552	1.1	68	-0.04
80	Benzo(a)anthracene	1.146	1.161	-1.3	69	-0.05
81	3,3'-Dichlorobenzidine	0.444	0.448	-0.9	70	-0.05
82	Chrysene	1.068	1.098	-2.8	71	-0.04
83	Bis(2-ethylhexyl)phthalate	0.793	0.809	-2.0	69	-0.05
84 c	Di-n-octyl phthalate	1.320	1.366	-3.5	71	-0.06
85	Indeno(1,2,3-cd)pyrene	1.277	1.332	-4.3	71	-0.11
86 I	Perylene-d12	1.000	1.000	0.0	69	-0.07
87	Benzo(b)fluoranthene	1.166	1.208	-3.6	72	-0.06

Data Path : Z:\HPCHEM1\BNA_G\Data\BG052815\
Data File : BG017337.D
Acq On : 28 May 2015 22:07
Operator : TP/IZ
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_G
LabSampleId :
SSTDCCC040

Quant Time: May 29 01:15:51 2015
Quant Method : Z:\HPCHEM1\BNA_G\METHODS\8270-BG051315.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri May 29 01:04:23 2015
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
88	Benzo(k)fluoranthene	1.107	1.140	-3.0	69	-0.06
89 C	Benzo(a)pyrene	1.071	1.110	-3.6	71	-0.07
90	Dibenzo(a,h)anthracene	1.099	1.143	-4.0	72	-0.11
91	Benzo(g,h,i)perylene	1.035	1.075	-3.9	71	-0.12

(#) = Out of Range

SPCC's out = 0 CCC's out = 0