


Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG052824\
Data File : BG061293.D
Acq On : 28 May 2024 8:26
Operator : MA/JU
Sample : DFTPP572
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
DFTPP572

Integration File: LSCINT.P

Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\SFAM-EPA-BG052124.MA.M
Title : SVOA CALIBRATION
Last Update : Tue May 21 15:23:39 2024



AutoFind: Scans 2504, 2505, 2506; Background Corrected with Scan 2498

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	44.0	38939	PASS
68	69	0.00	2	0.7	193	PASS
69	198	0.00	100	32.5	28805	PASS
70	69	0.00	2	0.5	152	PASS
127	198	10	80	39.8	35221	PASS
197	198	0.00	2	0.3	284	PASS
198	198	100	100	100.0	88515	PASS
199	198	5	9	7.2	6376	PASS
275	198	10	60	31.9	28253	PASS
365	198	1	100	4.5	3946	PASS
441	198	0.01	100	13.6	12078	PASS
442	198	50	100	84.6	74909	PASS
443	442	15	24	19.8	14834	PASS

m/z	Abundance
44.00	1322.0

Instrument :
BNA_G
ClientSampleId :
DFTPP572

m/z	Abundance
37.10	165.0
38.10	432.0
39.00	2601.0
39.90	207.0
44.00	1441.0
49.10	252.0
50.00	9339.0
51.00	35072.0
52.00	1806.0
54.90	150.0
56.00	1067.0
57.00	2556.0
61.00	405.0
61.90	263.0
63.00	1179.0
64.00	186.0
65.00	500.0
69.00	24792.0
70.10	261.0
73.00	193.0
74.00	2725.0
75.00	4549.0
76.10	1349.0
77.00	26464.0
78.10	1800.0
79.00	2088.0
79.90	1570.0
81.00	1932.0
82.00	416.0
82.90	524.0
85.00	340.0
85.90	768.0
86.80	244.0
91.00	530.0
91.90	548.0
92.90	3396.0
94.00	205.0
96.10	179.0
98.00	2992.0
99.00	2152.0
99.70	207.0
101.00	1157.0
102.90	307.0
104.00	803.0
105.00	786.0
105.90	275.0
107.00	9631.0
108.00	1366.0
110.00	14927.0
111.00	2261.0
112.00	301.0
116.00	711.0
117.00	7603.0
118.00	585.0
122.10	522.0
123.00	946.0
123.90	401.0
125.10	637.0
127.00	30448.0
128.10	2392.0
129.00	11519.0
130.00	785.0
130.90	230.0
132.00	167.0
133.80	308.0
135.00	826.0
136.00	398.0
136.90	515.0
140.00	253.0
141.00	1729.0
142.00	719.0
143.00	360.0
146.00	288.0
147.00	856.0
148.00	1926.0
149.00	392.0
151.00	267.0
151.70	337.0
153.00	687.0
154.00	442.0
155.00	987.0
156.00	1334.0
157.00	222.0
158.00	329.0
159.00	338.0
160.20	336.0
161.10	635.0
165.00	731.0
166.00	692.0
167.00	3767.0
168.00	1556.0
169.00	312.0

Instrument :
BNA_G
ClientSampleId :
DFTPP572

171.00	161.0
171.90	374.0
173.00	408.0
173.90	715.0
175.00	1536.0
176.00	591.0
176.90	597.0
177.80	365.0
179.00	2871.0
180.00	1920.0
181.00	885.0
182.10	184.0
184.10	287.0
185.10	1646.0
186.00	9872.0
187.00	2699.0
188.00	251.0
189.00	1053.0
189.70	165.0
191.00	246.0
192.00	620.0
193.00	975.0
193.90	284.0
195.00	233.0
196.00	1800.0
196.80	853.0
198.00	67640.0
199.00	4738.0
200.10	376.0
201.20	240.0
202.90	568.0
204.10	2994.0
205.00	5018.0
206.10	21216.0
207.00	3279.0
208.10	743.0
209.00	399.0
209.70	332.0
211.10	958.0
215.00	408.0
216.10	774.0
217.00	5415.0
217.90	765.0
221.00	2492.0
222.00	904.0
223.00	1152.0
224.00	12373.0
225.00	3134.0
226.00	580.0
227.00	5520.0
228.00	825.0
229.00	1262.0
230.00	196.0
231.00	463.0
234.00	305.0
235.00	310.0
235.90	192.0
237.00	523.0
238.80	232.0
240.10	177.0
241.00	355.0
241.90	692.0
243.00	530.0
244.00	8232.0
245.00	1170.0
246.00	1779.0
246.90	370.0
249.10	352.0
253.20	318.0
255.00	48880.0
256.00	7784.0
257.00	545.0
258.00	2948.0
258.90	545.0
265.00	1102.0
266.00	203.0
272.10	195.0
273.00	1359.0
274.00	4103.0
275.00	19696.0
276.00	2553.0
277.00	1949.0
277.90	311.0
278.10	315.0
282.80	194.0
283.90	176.0
285.00	372.0
292.90	597.0
295.00	229.0
296.00	5926.0
297.10	898.0
303.00	859.0
308.00	154.0
315.00	772.0
316.10	499.0
320.90	161.0

Instrument :
BNA_G
ClientSampleId :
DFTPP572

322.20	162.0
323.00	1959.0
323.90	501.0
327.00	460.0
328.10	355.0
333.00	274.0
334.10	1275.0
335.10	321.0
341.00	274.0
346.10	427.0
352.00	546.0
352.90	312.0
354.00	534.0
365.10	2596.0
366.00	359.0
371.90	1018.0
372.80	213.0
373.30	219.0
383.10	212.0
401.90	418.0
402.90	583.0
403.10	575.0
403.80	192.0
404.20	223.0
420.90	461.0
421.90	349.0
423.00	2719.0
424.10	524.0
441.00	6837.0
442.00	42336.0
443.00	9063.0
444.00	674.0

Instrument :
BNA_G
ClientSampleId :
DFTPP572

m/z	Abundance
37.00	152.0
38.00	752.0
39.00	3483.0
40.00	216.0
44.00	1826.0
48.70	181.0
50.00	11935.0
51.00	46064.0
52.00	2301.0
54.90	350.0
55.90	1476.0
57.00	3105.0
61.00	570.0
62.00	424.0
63.00	1809.0
64.00	315.0
65.00	716.0
69.00	34840.0
74.00	3765.0
75.00	6041.0
76.10	2472.0
77.00	37880.0
78.00	2766.0
79.00	2671.0
80.00	2284.0
81.00	2950.0
82.00	676.0
83.00	552.0
85.10	556.0
85.90	1026.0
87.00	403.0
91.00	779.0
91.90	717.0
93.00	4056.0
94.00	344.0
98.00	3699.0
99.00	2781.0
100.10	245.0
101.00	1591.0
101.80	152.0
102.90	473.0
103.90	1150.0
105.00	1015.0
107.00	11941.0
108.00	1838.0
110.00	21784.0
111.00	3326.0
112.00	473.0
116.00	603.0
117.00	9674.0
118.00	751.0
119.90	241.0
121.90	690.0
123.00	1253.0
124.00	630.0
124.90	462.0
127.00	41848.0
128.00	3238.0
129.00	16355.0
130.00	1674.0
131.10	296.0
134.00	705.0
135.00	1520.0
136.00	603.0
137.10	614.0
138.10	249.0
140.10	362.0
141.00	2224.0
141.90	754.0
143.00	602.0
144.80	168.0
146.00	555.0
147.00	1195.0
148.00	2513.0
149.00	758.0
151.00	323.0
151.70	232.0
152.90	899.0
153.90	531.0
155.00	1473.0
156.00	1891.0
157.00	348.0
158.00	431.0
159.00	491.0
160.00	673.0
160.90	1033.0
162.00	201.0
163.00	181.0
165.00	1053.0
165.90	665.0
167.10	4555.0
168.00	2404.0

Instrument :
BNA_G
ClientSampleId :
DFTPP572

169.00	513.0
170.10	153.0
171.00	286.0
172.00	396.0
173.00	674.0
174.00	1324.0
175.10	2250.0
176.00	767.0
177.00	1179.0
177.90	373.0
179.00	4296.0
180.00	2998.0
181.00	1077.0
182.10	376.0
184.10	454.0
185.10	1974.0
186.00	14707.0
187.00	4324.0
188.00	615.0
188.90	1153.0
189.80	429.0
191.00	530.0
192.00	1307.0
193.10	1733.0
194.00	476.0
195.10	390.0
195.90	3037.0
198.00	105368.0
199.00	7864.0
200.00	477.0
201.50	278.0
202.00	183.0
202.90	836.0
204.00	4540.0
205.10	8171.0
206.00	31640.0
207.00	4643.0
207.90	1065.0
209.10	268.0
210.10	533.0
211.10	1232.0
215.00	517.0
216.10	1052.0
217.00	9095.0
218.00	1325.0
219.80	204.0
221.00	4302.0
222.00	1403.0
223.00	2537.0
224.00	18040.0
225.00	4538.0
226.00	605.0
227.00	9242.0
228.00	1287.0
229.10	1558.0
230.00	256.0
231.00	574.0
232.90	156.0
234.00	763.0
235.10	527.0
236.00	429.0
236.90	726.0
238.00	206.0
239.10	220.0
239.90	356.0
240.90	461.0
241.90	866.0
242.10	882.0
243.10	1290.0
244.00	13484.0
245.00	1911.0
246.00	3259.0
246.90	459.0
247.90	162.0
249.10	519.0
251.90	287.0
253.00	430.0
255.00	79448.0
256.00	11340.0
257.00	991.0
258.00	5305.0
259.00	761.0
260.90	165.0
263.80	221.0
265.00	1916.0
266.00	477.0
272.00	253.0
273.00	2123.0
274.00	6711.0
275.00	32688.0
276.00	4809.0
277.00	3145.0
277.90	568.0
279.10	196.0
280.90	158.0
283.00	392.0

Instrument :
BNA_G
ClientSampleId :
DFTPP572

283.90	219.0
285.00	493.0
291.90	153.0
293.10	583.0
293.80	274.0
296.00	9854.0
297.00	832.0
298.10	152.0
302.10	273.0
303.10	1339.0
304.00	526.0
313.90	763.0
315.00	981.0
316.10	587.0
317.00	156.0
321.10	349.0
322.00	214.0
323.00	3433.0
324.10	523.0
326.90	389.0
328.10	240.0
332.00	203.0
332.90	352.0
334.00	2258.0
334.90	455.0
335.20	494.0
341.00	301.0
346.10	559.0
352.00	1042.0
353.10	720.0
354.10	856.0
355.10	293.0
359.00	152.0
365.00	4097.0
365.80	681.0
371.00	290.0
372.00	1467.0
372.90	266.0
382.90	404.0
389.90	327.0
401.90	615.0
402.80	754.0
404.00	425.0
420.90	820.0
422.00	530.0
423.00	5060.0
424.00	811.0
424.90	186.0
441.10	13525.0
442.00	85096.0
443.00	16608.0
444.10	1811.0

Instrument :
BNA_G
ClientSampleId :
DFTPP572

m/z	Abundance
37.00	264.0
37.80	194.0
39.00	2927.0
41.00	186.0
44.00	1420.0
50.00	9036.0
51.00	35680.0
52.00	1717.0
56.00	1262.0
57.00	2206.0
58.00	184.0
60.90	374.0
62.00	582.0
62.90	1519.0
64.10	194.0
65.00	675.0
66.00	182.0
67.90	579.0
69.00	26784.0
69.90	196.0
73.10	266.0
74.00	2858.0
75.00	4473.0
76.00	1897.0
77.00	30064.0
78.00	2066.0
79.00	2491.0
80.00	1349.0
81.00	2540.0
82.00	614.0
83.00	335.0
85.00	478.0
86.00	669.0
87.00	374.0
91.00	743.0
92.10	689.0
93.00	3876.0
93.80	263.0
98.00	2939.0
99.00	2022.0
100.00	359.0
101.00	1213.0
102.20	150.0
102.90	351.0
103.90	1138.0
105.10	634.0
107.00	10434.0
108.00	1345.0
108.90	320.0
110.00	16936.0
111.00	2556.0
112.10	353.0
113.00	155.0
115.90	457.0
117.00	7902.0
117.90	813.0
120.10	157.0
122.00	900.0
123.00	1099.0
123.90	532.0
125.00	401.0
127.00	33368.0
128.10	2660.0
129.00	14413.0
130.00	1105.0
130.90	213.0
133.00	153.0
134.00	476.0
135.10	1149.0
136.00	553.0
137.00	895.0
138.10	220.0
139.90	257.0
141.00	2716.0
142.00	485.0
143.00	395.0
143.80	184.0
146.00	431.0
147.00	1134.0
148.00	2289.0
149.10	320.0
151.10	416.0
153.10	601.0
154.10	474.0
154.90	1278.0
156.00	1862.0
157.10	483.0
158.00	408.0
158.90	297.0
160.00	620.0
161.00	1120.0
162.00	243.0

Instrument :
BNA_G
ClientSampleId :
DFTPP572

163.10	256.0
165.10	740.0
166.00	753.0
167.00	4112.0
168.00	2104.0
168.90	427.0
170.10	206.0
171.10	154.0
172.10	373.0
173.00	695.0
174.00	940.0
175.00	1977.0
176.00	577.0
177.00	864.0
179.00	3845.0
180.00	2493.0
181.10	1116.0
182.00	317.0
183.50	215.0
183.80	258.0
185.10	1843.0
186.00	14235.0
187.00	4599.0
188.10	446.0
189.00	929.0
189.80	201.0
191.10	428.0
191.90	1214.0
193.10	1385.0
194.10	334.0
196.00	2864.0
198.00	92536.0
199.00	6525.0
199.90	650.0
201.60	421.0
203.00	685.0
204.10	3868.0
205.00	6111.0
206.00	29168.0
207.10	4093.0
208.00	846.0
208.90	398.0
210.00	624.0
210.80	913.0
211.10	991.0
212.10	212.0
212.90	259.0
214.90	364.0
216.00	720.0
217.00	8641.0
218.00	1119.0
219.00	211.0
221.10	2950.0
223.00	1949.0
224.00	18144.0
225.10	4462.0
226.00	649.0
227.00	8840.0
228.00	1426.0
229.00	1635.0
229.90	208.0
231.00	544.0
231.90	167.0
232.90	219.0
233.90	639.0
234.90	473.0
236.00	220.0
237.10	779.0
239.00	391.0
240.10	338.0
241.10	549.0
242.00	908.0
242.90	764.0
244.00	12657.0
245.10	1671.0
246.00	2793.0
246.90	583.0
247.90	218.0
249.10	509.0
250.10	247.0
250.70	183.0
251.90	183.0
253.00	488.0
255.00	75568.0
256.00	10767.0
257.00	1030.0
258.00	4781.0
259.00	744.0
260.10	165.0
265.00	1932.0
265.90	243.0
271.20	287.0
272.00	470.0
273.00	2587.0
274.00	5562.0
275.00	32376.0

Instrument :
BNA_G
ClientSampleId :
DFTPP572

276.00	3756.0
277.00	2990.0
278.00	387.0
283.10	389.0
284.20	176.0
285.00	417.0
286.00	151.0
289.00	206.0
293.10	717.0
294.10	157.0
294.80	191.0
296.00	9834.0
297.10	1622.0
300.90	153.0
302.00	229.0
303.00	1015.0
304.10	311.0
304.90	182.0
308.90	167.0
312.80	165.0
314.10	534.0
315.00	1169.0
316.00	559.0
321.10	293.0
322.00	205.0
323.10	3481.0
324.10	715.0
327.00	713.0
328.20	360.0
331.90	250.0
333.10	304.0
334.00	2098.0
335.10	590.0
341.00	332.0
341.90	268.0
345.90	722.0
347.10	160.0
352.00	958.0
353.00	470.0
354.10	1100.0
355.00	218.0
365.10	5145.0
365.90	757.0
371.30	262.0
372.00	1833.0
373.00	439.0
383.00	345.0
389.90	219.0
391.00	257.0
401.00	153.0
401.90	778.0
403.00	1006.0
403.90	458.0
420.90	738.0
423.00	6151.0
424.10	1255.0
427.00	161.0
441.00	15873.0
442.00	97296.0
443.00	18832.0
444.00	1830.0

Instrument :
BNA_G
ClientSampleId :
DFTPP572