


Data Path : Z:\HPCHEM1\BNA_G\DATA\BG052915\
Data File : BG017354.D
Acq On : 29 May 2015 11:43
Operator : TP/IZ
Sample : DFTPP01
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
DFTPP01

Integration File: LSCINT.P

Method : Z:\HPCHEM1\BNA_G\METHODS\SOM02.2-EPA-BG052915.M
Title : SVOA CALIBRATION
Last Update : Fri May 29 17:44:44 2015



AutoFind: Scans 2520, 2521, 2522; Background Corrected with Scan 2515

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	50.9	11497	PASS
68	69	0.00	2	1.5	166	PASS
69	198	0.00	100	47.9	10816	PASS
70	69	0.00	2	1.3	143	PASS
127	198	10	80	53.6	12105	PASS
197	198	0.00	2	0.0	0	PASS
198	198	100	100	100.0	22570	PASS
199	198	5	9	7.0	1572	PASS
275	198	10	60	29.2	6586	PASS
365	198	1	100	6.9	1551	PASS
441	443	0.01	100	72.1	2708	PASS
442	198	50	100	80.3	18130	PASS
443	442	15	24	20.7	3754	PASS

m/z	Abundance
35.00	220.0
38.00	195.0
38.90	899.0
39.90	377.0
41.00	327.0
42.90	298.0
43.80	330.0
47.00	316.0
48.90	741.0
50.00	3167.0
51.00	13366.0
52.00	749.0
55.00	288.0
56.00	632.0
57.00	1012.0
60.80	298.0
61.90	329.0
62.80	353.0
63.00	385.0
65.00	499.0
69.00	11566.0
70.10	205.0
70.90	202.0
73.00	434.0
74.00	1538.0
75.00	1982.0
76.00	840.0
77.00	11744.0
78.00	892.0
79.00	1233.0
80.00	834.0
81.00	1134.0
81.90	250.0
82.90	229.0
84.00	577.0
85.00	249.0
86.00	634.0
87.00	213.0
87.90	164.0
91.00	397.0
92.00	395.0
93.00	1613.0
95.90	180.0
97.10	188.0
98.00	1260.0
99.00	899.0
101.00	635.0
104.00	371.0
104.80	278.0
107.00	4185.0
107.90	669.0
109.00	248.0
110.00	6093.0
111.00	1049.0
111.90	225.0
117.00	4213.0
118.00	360.0
121.90	366.0
122.90	617.0
124.00	247.0
125.00	247.0
127.00	12784.0
128.00	1189.0
129.00	5832.0
130.00	598.0
133.90	275.0
135.00	593.0
135.90	214.0
137.10	416.0
138.00	278.0
141.00	771.0
141.90	264.0
143.00	209.0
146.00	225.0
146.90	349.0
147.90	811.0
149.10	401.0
153.00	331.0
155.00	380.0
156.00	786.0
156.90	186.0
159.80	152.0
161.00	426.0
162.00	232.0
164.90	338.0
166.00	375.0
167.00	1356.0
168.00	800.0
171.90	172.0
173.10	159.0
173.90	438.0

Instrument :
BNA_G
ClientSampleId :
DFTPP01

175.00	1010.0
176.00	246.0
176.70	290.0
178.90	1196.0
179.90	1047.0
181.00	423.0
183.90	172.0
185.00	376.0
186.00	4811.0
187.00	1274.0
189.00	595.0
192.00	383.0
193.00	452.0
193.80	218.0
196.00	919.0
197.90	21784.0
198.90	1762.0
199.90	188.0
203.00	185.0
204.00	1261.0
204.90	1468.0
206.00	6310.0
207.00	1409.0
208.00	367.0
208.90	154.0
211.00	585.0
216.00	288.0
216.90	1684.0
218.00	230.0
221.00	1114.0
222.90	485.0
224.00	3876.0
225.00	1106.0
226.00	273.0
227.00	2091.0
228.00	443.0
228.90	481.0
236.00	152.0
241.10	224.0
241.90	171.0
243.00	231.0
244.00	2519.0
245.10	530.0
246.00	900.0
253.00	219.0
253.70	284.0
255.00	15075.0
256.00	2553.0
257.00	212.0
258.00	1202.0
264.90	354.0
272.90	470.0
274.00	1140.0
275.00	6104.0
276.00	988.0
276.90	895.0
277.90	240.0
296.00	2994.0
297.00	481.0
303.00	330.0
315.00	238.0
323.00	614.0
326.90	160.0
334.00	436.0
346.10	192.0
351.80	199.0
353.00	203.0
353.90	260.0
365.00	1220.0
366.00	288.0
372.00	394.0
402.00	380.0
403.00	279.0
423.00	1038.0
441.00	1968.0
442.00	14144.0
443.00	2663.0
444.10	160.0

Instrument :
BNA_G
ClientSampleId :
DFTPP01

m/z	Abundance
36.90	221.0
39.00	956.0
40.00	297.0
40.90	232.0
43.10	622.0
46.90	219.0
49.00	1081.0
50.00	3408.0
51.00	13101.0
52.00	462.0
55.00	265.0
56.00	555.0
57.00	1110.0
62.00	233.0
63.00	500.0
65.10	239.0
67.90	289.0
69.00	13373.0
70.10	226.0
71.00	209.0
73.00	286.0
74.00	1597.0
75.00	2141.0
76.00	878.0
77.00	13457.0
77.90	913.0
78.90	1571.0
79.90	1002.0
81.00	1316.0
81.90	465.0
83.00	351.0
83.90	457.0
85.10	223.0
86.00	678.0
86.90	223.0
90.90	256.0
92.00	598.0
92.90	1877.0
94.00	209.0
98.00	1712.0
99.00	1168.0
100.80	344.0
103.00	160.0
104.00	594.0
105.00	524.0
106.00	183.0
107.00	4340.0
108.00	726.0
110.00	8236.0
111.00	1149.0
113.10	162.0
116.00	259.0
117.00	4672.0
117.70	164.0
122.00	414.0
123.00	407.0
123.90	360.0
124.90	284.0
127.00	13148.0
128.00	1001.0
129.00	6536.0
130.00	342.0
131.00	239.0
135.00	657.0
135.90	261.0
137.00	353.0
140.90	685.0
141.90	363.0
143.00	317.0
147.10	539.0
147.90	1205.0
149.00	417.0
151.00	200.0
151.60	182.0
153.00	323.0
153.90	348.0
155.00	374.0
156.00	706.0
158.10	241.0
159.00	270.0
159.90	322.0
161.00	560.0
164.90	311.0
166.00	323.0
167.00	2179.0
168.00	1069.0
171.80	194.0
173.00	220.0
174.00	409.0
175.00	1010.0
177.00	444.0

Instrument :
BNA_G
ClientSampleId :
DFTPP01

177.90	304.0
178.90	1265.0
180.00	909.0
181.00	484.0
183.00	167.0
185.00	922.0
186.00	4415.0
187.00	1411.0
188.00	243.0
189.00	473.0
190.90	303.0
192.00	492.0
193.00	706.0
196.10	1084.0
197.90	25304.0
198.90	1835.0
201.10	185.0
203.00	315.0
204.00	1254.0
205.00	2934.0
206.00	7960.0
207.00	1267.0
207.90	347.0
210.00	176.0
210.90	264.0
215.10	182.0
216.00	286.0
216.90	2551.0
217.90	417.0
221.00	1183.0
222.10	409.0
222.90	791.0
224.00	4197.0
225.00	1301.0
226.10	201.0
227.00	2462.0
228.00	292.0
228.90	547.0
230.10	152.0
231.00	206.0
233.80	167.0
235.10	269.0
236.90	216.0
238.10	152.0
239.00	239.0
240.10	207.0
240.90	219.0
242.00	298.0
244.00	3632.0
245.00	522.0
245.90	914.0
252.90	321.0
255.00	19720.0
256.00	2350.0
257.00	155.0
258.00	1572.0
265.00	923.0
265.90	224.0
273.00	728.0
274.00	1820.0
275.00	7799.0
275.90	971.0
277.00	800.0
278.00	380.0
282.00	160.0
285.00	262.0
292.90	232.0
296.00	2774.0
296.90	599.0
303.00	487.0
304.00	194.0
314.90	544.0
322.90	783.0
324.00	284.0
327.00	293.0
333.10	231.0
334.00	671.0
334.90	160.0
345.90	217.0
347.00	205.0
351.90	271.0
352.90	205.0
354.00	349.0
365.00	2025.0
372.00	486.0
401.80	221.0
402.90	215.0
421.00	167.0
422.00	185.0
423.00	1000.0
424.00	160.0
425.00	167.0
441.10	2939.0
442.10	21528.0
443.10	4635.0

Instrument :
BNA_G
ClientSampleId :
DFTPP01

444.00

512.0

Instrument :
BNA_G
ClientSampleId :
DFTPP01

m/z	Abundance
38.00	163.0
39.00	721.0
39.90	355.0
41.00	438.0
43.10	322.0
48.90	983.0
50.00	2069.0
51.00	8806.0
51.90	403.0
55.00	311.0
56.00	481.0
57.00	917.0
61.00	185.0
63.00	451.0
65.10	219.0
67.70	209.0
69.00	8214.0
74.00	1097.0
75.00	1872.0
76.00	511.0
77.00	8478.0
78.10	669.0
79.00	495.0
80.00	327.0
81.00	776.0
83.00	200.0
83.90	613.0
85.00	364.0
85.90	456.0
87.00	192.0
92.10	291.0
92.90	1329.0
97.90	1039.0
99.00	903.0
100.90	413.0
103.90	404.0
104.90	380.0
106.00	240.0
107.00	3392.0
108.00	564.0
109.00	218.0
110.00	5133.0
111.00	636.0
115.90	370.0
116.90	2774.0
118.00	356.0
119.10	182.0
120.00	150.0
122.00	259.0
123.00	345.0
124.90	223.0
127.00	10383.0
128.00	879.0
129.00	4616.0
129.90	568.0
133.80	180.0
135.00	365.0
136.00	266.0
136.90	177.0
140.00	230.0
140.90	545.0
142.00	178.0
147.00	291.0
148.00	1025.0
148.90	303.0
151.00	180.0
152.90	255.0
154.00	169.0
155.00	296.0
155.90	633.0
157.90	177.0
160.00	195.0
161.00	388.0
164.90	400.0
165.90	158.0
167.00	1403.0
168.00	702.0
173.90	246.0
175.10	430.0
175.90	173.0
176.90	172.0
178.90	1147.0
180.00	615.0
181.00	290.0
182.00	165.0
184.00	155.0
185.00	543.0
186.00	3551.0
187.00	946.0
188.10	190.0
188.90	388.0

Instrument :
BNA_G
ClientSampleId :
DFTPP01

191.00	237.0
191.80	429.0
192.90	378.0
195.90	506.0
197.90	20624.0
198.90	1119.0
204.00	1006.0
205.00	1380.0
206.00	6776.0
207.00	939.0
207.80	271.0
210.40	170.0
210.90	460.0
216.00	324.0
217.00	1755.0
217.80	258.0
221.00	1201.0
222.00	384.0
223.10	464.0
224.00	4284.0
225.00	1009.0
226.90	1964.0
227.90	342.0
228.90	269.0
230.90	185.0
233.90	194.0
235.00	255.0
237.00	267.0
239.00	190.0
240.80	194.0
242.00	358.0
242.80	295.0
244.00	2563.0
245.00	405.0
245.90	607.0
247.00	154.0
248.90	317.0
253.00	217.0
255.00	15167.0
256.00	2047.0
257.00	234.0
257.90	1118.0
259.00	185.0
264.90	290.0
266.10	164.0
271.00	231.0
273.00	620.0
274.00	1598.0
275.00	5855.0
276.00	764.0
277.00	785.0
283.00	169.0
296.00	2583.0
297.00	559.0
302.70	285.0
305.50	152.0
314.90	320.0
323.00	813.0
327.00	169.0
334.10	469.0
351.90	427.0
353.00	215.0
354.00	282.0
364.90	1409.0
365.70	243.0
372.00	280.0
382.90	182.0
390.00	162.0
391.00	200.0
401.90	155.0
403.00	176.0
422.00	158.0
423.00	1254.0
424.00	349.0
441.10	3218.0
442.00	18720.0
443.00	3966.0
444.10	553.0

Instrument :
BNA_G
ClientSampleId :
DFTPP01

m/z	Abundance
39.90	364.0
41.10	189.0
43.00	350.0
47.10	208.0
49.00	868.0
51.00	260.0
55.00	300.0
57.00	226.0
69.10	235.0
73.10	279.0
83.90	543.0
85.00	166.0
85.90	428.0
105.10	182.0
107.00	151.0
133.00	226.0
148.90	190.0
189.10	253.0
207.10	329.0

Instrument :
BNA_G
ClientSampleId :
DFTPP01