


Data Path : Z:\SVOASRV\HPCHEM1\BNA_G\DATA\BG052919\
 Data File : BG040934.D
 Acq On : 29 May 2019 12:54
 Operator : HP/JU
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 DFTPP

Integration File: rteint.p

Method : Z:\SVOASRV\HPCHEM1\BNA_G\METHODS\8270-BG052919.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Wed May 29 18:39:57 2019



AutoFind: Scans 2519, 2520, 2521; Background Corrected with Scan 2513

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	34.5	27349	PASS
68	69	0.00	2	0.5	139	PASS
69	198	0.00	100	36.4	28885	PASS
70	69	0.00	2	0.6	168	PASS
127	198	10	80	41.9	33248	PASS
197	198	0.00	2	0.0	0	PASS
198	198	100	100	100.0	79346	PASS
199	198	5	9	7.2	5738	PASS
275	198	10	60	27.1	21485	PASS
365	198	1	100	2.9	2315	PASS
441	198	0.01	100	8.0	6366	PASS
442	198	50	100	52.1	41333	PASS
443	442	15	24	19.6	8115	PASS