


Data Path : Z:\HPCHEM1\BNA_G\Data\BG053115\
 Data File : BG017361.D
 Acq On : 31 May 2015 9:35
 Operator : TP/IZ
 Sample : DFTPP02
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Integration File: LSCINT.P

Method : Z:\HPCHEM1\BNA_G\METHODS\SOM02.2-EPA-BG052915.M
 Title : SVOA CALIBRATION
 Last Update : Fri May 29 17:44:44 2015



AutoFind: Scans 2521, 2522, 2523; Background Corrected with Scan 2515

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	51.7	10107	PASS
68	69	0.00	2	0.0	0	PASS
69	198	0.00	100	48.8	9546	PASS
70	69	0.00	2	0.0	0	PASS
127	198	10	80	54.3	10615	PASS
197	198	0.00	2	0.0	0	PASS
198	198	100	100	100.0	19548	PASS
199	198	5	9	6.7	1318	PASS
275	198	10	60	29.4	5756	PASS
365	198	1	100	6.6	1286	PASS
441	443	0.01	100	82.7	2887	PASS
442	198	50	100	89.3	17460	PASS
443	442	15	24	20.0	3492	PASS

m/z	Abundance
38.10	163.0
39.10	962.0
39.90	278.0
41.00	164.0
43.80	234.0
50.00	3830.0
51.00	12700.0
52.00	636.0
56.00	485.0
56.90	1154.0
61.90	258.0
63.00	541.0
65.00	215.0
69.00	12710.0
74.00	1286.0
75.00	2010.0
76.10	1080.0
77.00	13687.0
78.00	763.0
79.00	953.0
79.90	849.0
81.00	921.0
82.00	222.0
82.90	192.0
84.90	229.0
85.90	301.0
90.90	186.0
91.80	291.0
93.00	1893.0
96.00	161.0
98.00	1399.0
98.90	1062.0
99.80	219.0
101.00	360.0
104.00	396.0
105.00	375.0
107.00	4271.0
107.90	721.0
108.90	155.0
110.00	6716.0
111.00	1106.0
115.80	211.0
117.00	5019.0
117.90	326.0
122.00	234.0
123.00	477.0
123.80	252.0
124.90	164.0
127.00	13639.0
128.00	1342.0
129.00	5900.0
129.90	554.0
130.90	189.0
134.00	204.0
134.70	420.0
137.00	595.0
140.90	924.0
142.00	242.0
143.00	187.0
146.00	312.0
147.00	444.0
147.90	1140.0
148.90	264.0
151.80	163.0
153.00	342.0
153.70	295.0
155.00	426.0
155.90	885.0
157.80	165.0
159.80	229.0
160.60	174.0
161.00	337.0
163.90	202.0
165.00	292.0
165.90	260.0
167.00	2290.0
167.90	726.0
169.10	166.0
172.90	174.0
173.90	224.0
174.90	646.0
176.90	403.0
178.00	181.0
178.90	1682.0
180.00	708.0
180.90	471.0
184.10	154.0
185.10	601.0
186.00	3776.0
187.00	1237.0
188.90	312.0

190.80	190.0
191.40	242.0
191.90	500.0
192.90	368.0
194.80	152.0
195.90	498.0
197.90	23192.0
199.00	1325.0
201.50	175.0
202.80	192.0
204.00	1073.0
205.00	2568.0
206.00	6949.0
207.00	1060.0
208.00	434.0
210.40	217.0
211.00	155.0
215.90	206.0
216.90	2463.0
218.00	256.0
221.00	1510.0
221.90	464.0
223.00	573.0
224.00	5113.0
225.00	1277.0
226.10	205.0
226.90	2534.0
227.90	366.0
229.00	430.0
230.70	177.0
234.00	381.0
235.80	160.0
241.90	266.0
242.90	201.0
244.00	3548.0
245.00	606.0
246.00	784.0
246.90	195.0
253.20	153.0
254.00	457.0
255.00	16354.0
256.00	2152.0
257.00	290.0
258.00	1147.0
258.90	343.0
265.00	371.0
265.80	153.0
273.00	531.0
274.00	1309.0
275.00	6324.0
276.00	882.0
277.00	692.0
283.80	186.0
293.00	178.0
294.10	151.0
296.00	3216.0
296.90	369.0
303.00	472.0
307.70	164.0
315.00	264.0
320.90	173.0
323.00	789.0
323.90	170.0
327.00	255.0
332.80	187.0
334.00	366.0
345.90	304.0
352.90	200.0
354.00	273.0
365.00	1152.0
365.90	203.0
372.10	223.0
383.00	158.0
403.00	334.0
423.00	1261.0
424.00	273.0
441.00	2506.0
442.00	17232.0
443.00	3955.0
444.00	427.0

m/z	Abundance
38.00	188.0
39.00	902.0
40.00	221.0
43.90	192.0
49.20	157.0
50.00	2758.0
51.00	12938.0
52.00	748.0
56.00	262.0
57.00	898.0
62.90	292.0
69.00	10711.0
73.10	368.0
74.00	1086.0
75.00	1816.0
76.00	503.0
77.00	12324.0
78.00	670.0
79.00	760.0
79.90	592.0
80.90	1105.0
81.90	200.0
84.80	153.0
85.90	273.0
91.90	432.0
93.00	1456.0
94.00	202.0
98.00	1288.0
98.90	916.0
101.00	623.0
104.00	426.0
105.00	365.0
106.10	173.0
107.00	3692.0
108.00	630.0
110.00	5742.0
111.00	997.0
112.00	197.0
117.00	4094.0
118.00	446.0
119.90	159.0
122.00	197.0
123.00	382.0
123.90	185.0
124.90	229.0
127.00	12353.0
128.00	1062.0
129.00	5698.0
130.00	362.0
134.00	180.0
134.90	358.0
136.00	155.0
137.00	252.0
140.90	908.0
141.90	331.0
143.10	240.0
146.90	401.0
147.90	1061.0
148.90	211.0
150.90	155.0
153.10	345.0
155.00	459.0
156.00	478.0
157.10	235.0
157.90	184.0
159.90	177.0
160.90	393.0
164.90	280.0
166.00	379.0
167.00	1863.0
168.00	856.0
168.90	331.0
172.00	243.0
172.90	175.0
173.90	344.0
175.00	673.0
175.90	227.0
176.90	237.0
178.90	1040.0
180.00	571.0
181.00	328.0
185.00	735.0
186.00	4272.0
187.00	1249.0
189.00	507.0
191.90	437.0
193.00	512.0
195.90	680.0
197.90	23400.0
198.90	1843.0
199.90	162.0

202.80	294.0
204.00	1174.0
205.00	1558.0
206.00	7215.0
207.00	909.0
208.00	479.0
211.10	433.0
211.70	192.0
217.00	2231.0
217.90	364.0
221.00	741.0
221.70	279.0
222.00	262.0
223.00	477.0
224.00	5146.0
224.90	969.0
226.90	2484.0
227.90	257.0
228.90	340.0
231.00	203.0
234.90	166.0
235.90	204.0
236.90	432.0
238.90	152.0
242.00	198.0
242.80	248.0
244.00	3086.0
245.00	489.0
246.00	900.0
247.00	285.0
252.90	152.0
255.00	15350.0
256.00	2227.0
257.90	1271.0
263.80	183.0
264.90	578.0
271.60	153.0
271.80	153.0
273.00	675.0
274.10	1453.0
275.00	7102.0
276.00	1181.0
277.00	905.0
278.00	157.0
284.00	166.0
293.00	151.0
296.00	3309.0
296.90	732.0
303.00	420.0
314.10	227.0
314.90	488.0
323.00	472.0
324.00	163.0
334.00	521.0
345.90	167.0
352.00	402.0
353.10	165.0
354.10	237.0
365.00	1273.0
366.00	196.0
372.00	554.0
403.00	167.0
421.00	211.0
422.00	182.0
423.00	1255.0
424.00	271.0
441.00	3264.0
442.00	19864.0
443.00	4053.0
444.00	459.0

m/z	Abundance
39.00	373.0
40.00	231.0
50.00	1470.0
51.00	4685.0
52.00	343.0
57.00	171.0
62.00	227.0
69.00	5217.0
73.10	175.0
74.00	548.0
75.00	794.0
76.10	318.0
77.00	4631.0
77.90	337.0
79.00	432.0
80.00	415.0
81.00	248.0
82.10	229.0
92.90	950.0
97.90	537.0
98.90	414.0
100.90	255.0
104.00	218.0
104.90	182.0
107.00	1634.0
108.00	306.0
110.00	2763.0
111.00	269.0
116.90	2456.0
121.90	174.0
123.00	254.0
127.00	5854.0
128.00	483.0
129.00	2101.0
130.00	193.0
134.00	255.0
134.90	302.0
140.80	459.0
146.90	342.0
147.80	476.0
153.00	164.0
154.80	261.0
155.90	237.0
160.00	168.0
161.10	303.0
165.90	195.0
166.90	684.0
168.00	335.0
172.80	161.0
175.00	371.0
178.90	530.0
180.00	390.0
184.90	250.0
186.00	1992.0
186.90	931.0
189.00	312.0
190.80	293.0
192.90	279.0
196.00	367.0
196.60	207.0
197.90	12052.0
198.90	788.0
202.80	295.0
204.00	902.0
205.00	927.0
206.00	3718.0
207.00	547.0
207.90	173.0
210.70	207.0
216.90	1160.0
217.90	235.0
221.00	617.0
221.80	168.0
223.00	362.0
224.00	2465.0
224.90	564.0
226.90	1270.0
228.00	333.0
229.00	191.0
242.00	165.0
244.00	2011.0
244.80	308.0
245.90	658.0
248.90	204.0
255.00	10659.0
256.00	1563.0
257.90	1278.0
259.00	151.0
265.00	585.0
273.00	508.0
273.90	968.0

275.00	3844.0
275.90	467.0
276.90	742.0
296.00	1950.0
296.90	324.0
302.90	183.0
315.00	283.0
316.00	190.0
323.00	794.0
334.10	315.0
352.00	183.0
353.80	155.0
365.00	1434.0
372.00	204.0
402.00	206.0
420.90	179.0
422.00	309.0
423.10	837.0
439.60	171.0
440.20	170.0
441.00	2892.0
442.10	15285.0
443.00	2468.0
443.90	229.0

Scan 2515 (17.464 min): BG017361.D
DFTPP02

m/z	Abundance
39.90	360.0