

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG053122\
 Data File : BG053767.D
 Acq On : 31 May 2022 19:16
 Operator : CG/JU
 Sample : SST04024
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 SST040424

Quant Time: May 31 23:15:44 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\SFAM-EPA-BG053122.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Tue May 31 23:09:35 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.105	152	19350	20.000	ng/u1	0.00
20) Naphthalene-d8	10.913	136	78615	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.726	164	55141	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.470	188	133375	20.000	ng/u1	0.00
79) Chrysene-d12	21.753	240	140873	20.000	ng/u1	0.00
88) Perylene-d12	25.026	264	145990	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.533	96	7678	14.016	ng/uL	0.00
4) Pyridine-d5	3.951	84	49041	35.859	ng/u1	0.00
7) Phenol-d5	7.247	99	63452	36.612	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.423	67	41926	39.609	ng/u1	0.00
11) 2-Chlorophenol-d4	7.635	132	50968	43.311	ng/u1	0.00
15) 4-Methylphenol-d8	8.798	113	52997	42.214	ng/u1	0.00
21) Nitrobenzene-d5	9.262	128	20688	36.558	ng/u1	0.00
24) 2-Nitrophenol-d4	9.985	143	21620	36.221	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.525	165	55691	42.888	ng/u1	0.00
31) 4-Chloroaniline-d4	11.036	131	74922	43.655	ng/u1	0.00
46) Dimethylphthalate-d6	14.127	166	178751	42.196	ng/u1	0.00
49) Acenaphthylene-d8	14.421	160	200407	38.932	ng/u1	0.00
54) 4-Nitrophenol-d4	14.897	143	25006	35.109	ng/u1	0.00
60) Fluorene-d10	15.719	176	153690	41.319	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.825	200	21552	31.753	ng/u1	0.00
73) Anthracene-d10	17.570	188	251097	40.429	ng/u1	0.00
81) Pyrene-d10	19.850	212	323748	41.081	ng/u1	0.00
92) Benzo(a)pyrene-d12	24.809	264	304767	40.402	ng/u1	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.569	88	8043	11.983	ng/uL#	87
5) Pyridine	3.968	79	52537	35.553	ng/u1	96
6) Benzaldehyde	7.235	77	38800	33.676	ng/u1	96
8) Phenol	7.276	94	68085	40.762	ng/u1	97
10) Bis(2-Chloroethyl)ether	7.517	93	52699	41.822	ng/u1	90
12) 2-Chlorophenol	7.664	128	51648	42.957	ng/u1	99
13) 2-Methylphenol	8.534	108	51851	39.941	ng/u1	97
14) 2,2'-oxybis(1-Chloropr...	8.633	45	87498	42.720	ng/u1	99
16) Acetophenone	8.927	105	83501	42.634	ng/u1	99
17) N-Nitroso-di-n-propyla...	8.915	70	44956	39.374	ng/u1	92
18) 4-Methylphenol	8.863	108	56053	41.443	ng/u1	93
19) Hexachloroethane	9.197	117	21410	40.134	ng/u1	91
22) Nitrobenzene	9.303	77	54230	32.796	ng/u1	96
23) Isophorone	9.838	82	124068	39.254	ng/u1	100
25) 2-Nitrophenol	10.020	139	23953	38.699	ng/u1	91
26) 2,4-Dimethylphenol	10.073	107	63020	41.090	ng/u1	98
27) Bis(2-Chloroethoxy)met...	10.314	93	71095	40.450	ng/u1	99
29) 2,4-Dichlorophenol	10.549	162	54870	43.461	ng/u1	97
30) Naphthalene	10.966	128	169953	41.775	ng/u1	98
32) 4-Chloroaniline	11.066	127	73781	43.751	ng/u1	100
33) Hexachlorobutadiene	11.260	225	44590	43.246	ng/u1	98
34) Caprolactam	11.824	113	7551	19.859	ng/u1	97
35) 4-Chloro-3-methylphenol	12.170	107	57865	42.252	ng/u1	94
36) 2-Methylnaphthalene	12.564	142	119446	41.603	ng/u1	96

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG053122\
 Data File : BG053767.D
 Acq On : 31 May 2022 19:16
 Operator : CG/JU
 Sample : SST04024
 Misc :
 ALS Vial : 5 Sample Multiplier: 1


Instrument :
 BNA_G
 ClientSampleId :
 SST040424


Quant Time: May 31 23:15:44 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\SFAM-EPA-BG053122.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Tue May 31 23:09:35 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.781	142	120294	41.386	ng/ul	97
39) 1,2,4,5-Tetrachloroben...	12.928	216	81484	44.708	ng/ul	97
40) Hexachlorocyclopentadiene	12.917	237	55277	43.809	ng/ul	94
41) 2,4,6-Trichlorophenol	13.158	196	47999	43.098	ng/ul	94
42) 2,4,5-Trichlorophenol	13.228	196	49332	41.017	ng/ul	98
43) 1,1'-Biphenyl	13.563	154	167583	42.530	ng/ul	99
44) 2-Chloronaphthalene	13.604	162	128985	40.880	ng/ul	97
45) 2-Nitroaniline	13.792	65	30179	32.630	ng/ul	99
47) Dimethylphthalate	14.174	163	170052	42.065	ng/ul	99
48) 2,6-Dinitrotoluene	14.291	165	27169	37.100	ng/ul	99
50) Acenaphthylene	14.450	152	210098	42.261	ng/ul	99
51) 3-Nitroaniline	14.615	138	24984	33.800	ng/ul	98
52) Acenaphthene	14.791	153	140118	43.143	ng/ul	99
53) 2,4-Dinitrophenol	14.814	184	12301	27.039	ng/ul#	87
55) 4-Nitrophenol	14.908	109	22974	30.649	ng/ul	95
56) Dibenzofuran	15.126	168	203499	42.265	ng/ul	98
57) 2,4-Dinitrotoluene	15.073	165	38473	37.027	ng/ul#	96
58) 2,3,4,6-Tetrachlorophenol	15.343	232	46736	43.110	ng/ul	97
59) Diethylphthalate	15.543	149	173348	42.340	ng/ul	99
61) Fluorene	15.772	166	166118	41.623	ng/ul	98
62) 4-Chlorophenyl-phenyle...	15.766	204	92667	43.634	ng/ul	99
63) 4-Nitroaniline	15.778	138	25671	35.316	ng/ul	95
66) 4,6-Dinitro-2-methylph...	15.837	198	20677	31.135	ng/ul	98
67) N-Nitrosodiphenylamine	15.972	169	152240	42.217	ng/ul	98
68) 4-Bromophenyl-phenylether	16.659	248	63472	48.063	ng/ul	99
69) Hexachlorobenzene	16.777	284	70307	41.214	ng/ul	98
70) Atrazine	16.924	200	68799	46.552	ng/ul	99
71) Pentachlorophenol	17.118	266	41139	39.657	ng/ul	92
72) Phenanthrene	17.511	178	288368	41.195	ng/ul	99
74) Anthracene	17.605	178	289677	41.392	ng/ul	99
75) 1,2,3,4-Tetrachloroben...	13.528	216	87749	44.174	ng/ul	96
76) Pentachlorobenzene	15.049	250	81734	42.887	ng/ul	97
77) Carbazole	17.864	167	262465	42.437	ng/ul	99
78) Di-n-butylphthalate	18.434	149	316344	43.012	ng/ul	99
80) Fluoranthene	19.521	202	370609	39.907	ng/ul	98
82) Pyrene	19.879	202	368785	39.952	ng/ul	99
83) Butylbenzylphthalate	20.772	149	134822	42.009	ng/ul	99
84) 3,3'-Dichlorobenzidine	21.642	252	139529	44.191	ng/ul	96
85) Benzo(a)anthracene	21.730	228	379952	41.789	ng/ul	99
86) Bis(2-ethylhexyl)phtha...	21.653	149	197806	44.500	ng/ul	98
87) Chrysene	21.800	228	361905	41.731	ng/ul	99
89) Di-n-octyl phthalate	22.893	149	340911	44.833	ng/ul	100
90) Benzo(b)fluoranthene	23.986	252	381742	40.495	ng/ul	98
91) Benzo(k)fluoranthene	24.057	252	360582	41.445	ng/ul	98
93) Benzo(a)pyrene	24.879	252	365276	41.042	ng/ul#	99
94) Indeno(1,2,3-cd)pyrene	28.775	276	421585	41.884	ng/ul	98
95) Dibenzo(a,h)anthracene	28.845	278	359703	42.046	ng/ul	99
96) Benzo(g,h,i)perylene	29.950	276	354546	41.669	ng/ul	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Quant Time: May 31 23:15:44 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\SFAM-EPA-BG053122.M
Quant Title : SVOA CALIBRATION
QLast Update : Tue May 31 23:09:35 2022
Response via : Initial Calibration









#2
1,4-Dioxane
Concen: 11.983 ng/uL
RT: 3.569 min Scan# 98
Delta R.T. 0.005 min
Lab File: BG053767.D
Acq: 31 May 2022 19:16




Instrument :
BNA_G
ClientSampleId :
SSTD040424


Tgt Ion: 88 Resp: 8043
Ion Ratio Lower Upper
88 100
43 46.2 48.1 72.1#
58 82.8 59.8 89.6



#5
Pyridine
Concen: 35.553 ng/uL
RT: 3.968 min Scan# 167
Delta R.T. -0.001 min
Lab File: BG053767.D
Acq: 31 May 2022 19:16

Tgt Ion: 79 Resp: 52537
Ion Ratio Lower Upper
79 100
52 90.3 75.4 113.2
51 40.8 35.2 52.8









#6
Benzaldehyde
Concen: 33.676 ng/ul
RT: 7.235 min Scan# 71
Delta R.T. -0.001 min
Lab File: BG053767.D
Acq: 31 May 2022 19:16




Instrument :
BNA_G
ClientSampleId :
SSTD040424


Tgt Ion: 77 Resp: 38800
Ion Ratio Lower Upper
77 100
105 94.7 79.7 119.5
106 92.1 71.0 106.4



#8
Phenol
Concen: 40.762 ng/ul
RT: 7.276 min Scan# 730
Delta R.T. 0.005 min
Lab File: BG053767.D
Acq: 31 May 2022 19:16

Tgt Ion: 94 Resp: 68085
Ion Ratio Lower Upper
94 100
65 32.3 24.1 36.1
66 41.9 32.5 48.7









#10
Bis(2-Chloroethyl)ether
Concen: 41.822 ng/ul
RT: 7.517 min Scan# 771
Delta R.T. -0.001 min
Lab File: BG053767.D
Acq: 31 May 2022 19:16




Instrument :
BNA_G
ClientSampleId :
SSTD040424


Tgt Ion: 93 Resp: 52699
Ion Ratio Lower Upper
93 100
63 80.4 72.3 108.5
95 31.4 29.4 44.0



#12
2-Chlorophenol
Concen: 42.957 ng/ul
RT: 7.664 min Scan# 796
Delta R.T. -0.001 min
Lab File: BG053767.D
Acq: 31 May 2022 19:16

Tgt Ion: 128 Resp: 51648
Ion Ratio Lower Upper
128 100
64 50.4 41.4 62.2
130 33.8 27.0 40.6









#13
2-Methylphenol
Concen: 39.941 ng/ul
RT: 8.534 min Scan# 944
Delta R.T. -0.001 min
Lab File: BG053767.D
Acq: 31 May 2022 19:16




Instrument :
BNA_G
ClientSampleId :
SSTD040424


Tgt Ion:108 Resp: 51851
Ion Ratio Lower Upper
108 100
107 92.0 75.8 113.8



#14
2,2'-oxybis(1-Chloropropane)
Concen: 42.720 ng/ul
RT: 8.633 min Scan# 961
Delta R.T. 0.005 min
Lab File: BG053767.D
Acq: 31 May 2022 19:16

Tgt Ion: 45 Resp: 87498
Ion Ratio Lower Upper
45 100
77 13.0 10.2 15.4
79 9.6 8.1 12.1









#16
Acetophenone
Concen: 42.634 ng/ul
RT: 8.927 min Scan# 1010
Delta R.T. 0.005 min
Lab File: BG053767.D
Acq: 31 May 2022 19:16




Instrument :
BNA_G
ClientSampleId :
SSTD040424


Tgt Ion:105 Resp: 83501
Ion Ratio Lower Upper
105 100
77 76.6 61.0 91.6
51 29.3 22.8 34.2



#17
N-Nitroso-di-n-propylamine
Concen: 39.374 ng/ul
RT: 8.915 min Scan# 1009
Delta R.T. 0.005 min
Lab File: BG053767.D
Acq: 31 May 2022 19:16

Tgt Ion: 70 Resp: 44956
Ion Ratio Lower Upper
70 100
42 70.8 50.2 75.4
101 9.2 7.1 10.7
130 18.3 15.5 23.3









#18
4-Methylphenol
Concen: 41.443 ng/ul
RT: 8.863 min Scan# 1000
Delta R.T. 0.005 min
Lab File: BG053767.D
Acq: 31 May 2022 19:16




Instrument :
BNA_G
ClientSampleId :
SSTD040424


Tgt Ion:108 Resp: 56053
Ion Ratio Lower Upper
108 100
107 106.2 90.7 136.1



#19
Hexachloroethane
Concen: 40.134 ng/ul
RT: 9.197 min Scan# 1057
Delta R.T. -0.001 min
Lab File: BG053767.D
Acq: 31 May 2022 19:16

Tgt Ion:117 Resp: 21410
Ion Ratio Lower Upper
117 100
201 130.9 117.4 176.0
199 83.0 64.0 96.0









#22
 Nitrobenzene
 Concen: 32.796 ng/ul
 RT: 9.303 min Scan# 1075
 Delta R.T. -0.001 min
 Lab File: BG053767.D
 Acq: 31 May 2022 19:16




Instrument :
 BNA_G
 ClientSampleId :
 SST040424


Tgt Ion: 77 Resp: 54230
 Ion Ratio Lower Upper
 77 100
 123 41.4 31.2 46.8
 65 14.8 10.5 15.7



#23
 Isophorone
 Concen: 39.254 ng/ul
 RT: 9.838 min Scan# 1166
 Delta R.T. 0.005 min
 Lab File: BG053767.D
 Acq: 31 May 2022 19:16

Tgt Ion: 82 Resp: 124068
 Ion Ratio Lower Upper
 82 100
 95 7.3 6.1 9.1
 138 16.9 13.6 20.4









#25
2-Nitrophenol
Concen: 38.699 ng/ul
RT: 10.020 min Scan# 1196
Delta R.T. 0.005 min
Lab File: BG053767.D
Acq: 31 May 2022 19:16




Instrument :
BNA_G
ClientSampleId :
SSTD040424


Tgt Ion:139 Resp: 23953
Ion Ratio Lower Upper
139 100
65 51.5 47.7 71.5
109 29.8 27.1 40.7



#26
2,4-Dimethylphenol
Concen: 41.090 ng/ul
RT: 10.073 min Scan# 1206
Delta R.T. -0.001 min
Lab File: BG053767.D
Acq: 31 May 2022 19:16

Tgt Ion:107 Resp: 63020
Ion Ratio Lower Upper
107 100
121 49.9 39.9 59.9
122 86.8 66.7 100.1









#27
 Bis(2-Chloroethoxy)methane
 Concen: 40.450 ng/ul
 RT: 10.314 min Scan# 1247
 Delta R.T. -0.001 min
 Lab File: BG053767.D
 Acq: 31 May 2022 19:16




Instrument :
 BNA_G
 ClientSampleId :
 SSTD040424


Tgt Ion: 93	Resp:	71095
Ion Ratio	Lower	Upper
93	100	
95	32.8	26.5 39.7
123	9.6	8.8 13.2



#29
 2,4-Dichlorophenol
 Concen: 43.461 ng/ul
 RT: 10.549 min Scan# 1287
 Delta R.T. -0.001 min
 Lab File: BG053767.D
 Acq: 31 May 2022 19:16

Tgt Ion: 162	Resp:	54870
Ion Ratio	Lower	Upper
162	100	
164	61.4	50.9 76.3
98	33.0	28.5 42.7









#30
Naphthalene
Concen: 41.775 ng/ul
RT: 10.966 min Scan# 1358
Delta R.T. -0.001 min
Lab File: BG053767.D
Acq: 31 May 2022 19:16




Instrument :
BNA_G
ClientSampleId :
SSTD040424


Tgt Ion:128 Resp: 169953
Ion Ratio Lower Upper
128 100
129 10.6 9.0 13.6
127 13.1 11.1 16.7



#32
4-Chloroaniline
Concen: 43.751 ng/ul
RT: 11.066 min Scan# 1375
Delta R.T. 0.005 min
Lab File: BG053767.D
Acq: 31 May 2022 19:16

Tgt Ion:127 Resp: 73781
Ion Ratio Lower Upper
127 100
129 31.9 25.4 38.2









#33
Hexachlorobutadiene
Concen: 43.246 ng/ul
RT: 11.260 min Scan# 1408
Delta R.T. -0.001 min
Lab File: BG053767.D
Acq: 31 May 2022 19:16




Instrument :
BNA_G
ClientSampleId :
SSTD040424


Tgt Ion:225 Resp: 44590
Ion Ratio Lower Upper
225 100
223 59.8 50.2 75.2
227 64.2 51.8 77.6



#34
Caprolactam
Concen: 19.859 ng/ul
RT: 11.824 min Scan# 1504
Delta R.T. -0.006 min
Lab File: BG053767.D
Acq: 31 May 2022 19:16

Tgt Ion:113 Resp: 7551
Ion Ratio Lower Upper
113 100
55 208.6 167.8 251.6
56 151.6 127.6 191.4









#35
4-Chloro-3-methylphenol
Concen: 42.252 ng/ul
RT: 12.170 min Scan# 1563
Delta R.T. -0.001 min
Lab File: BG053767.D
Acq: 31 May 2022 19:16




Instrument :
BNA_G
ClientSampleId :
SSTD040424


Tgt Ion:	107	Resp:	57865
Ion Ratio	Lower	Upper	
107	100		
144	25.2	21.3	31.9
142	75.1	65.0	97.4



#36
2-Methylnaphthalene
Concen: 41.603 ng/ul
RT: 12.564 min Scan# 1630
Delta R.T. -0.001 min
Lab File: BG053767.D
Acq: 31 May 2022 19:16

Tgt Ion:	142	Resp:	119446
Ion Ratio	Lower	Upper	
142	100		
141	91.1	76.3	114.5









#37
1-Methylnaphthalene
Concen: 41.386 ng/ul
RT: 12.781 min Scan# 1667
Delta R.T. -0.001 min
Lab File: BG053767.D
Acq: 31 May 2022 19:16




Instrument :
BNA_G
ClientSampleId :
SSTD040424


Tgt Ion:142 Resp: 120294
Ion Ratio Lower Upper
142 100
141 94.3 73.4 110.2
116 3.3 3.2 4.8



#39
1,2,4,5-Tetrachlorobenzene
Concen: 44.708 ng/ul
RT: 12.928 min Scan# 1692
Delta R.T. -0.001 min
Lab File: BG053767.D
Acq: 31 May 2022 19:16

Tgt Ion:216 Resp: 81484
Ion Ratio Lower Upper
216 100
214 76.0 58.1 87.1
179 18.4 14.8 22.2
108 13.6 10.4 15.6









#40
Hexachlorocyclopentadiene
Concen: 43.809 ng/ul
RT: 12.917 min Scan# 1689
Delta R.T. 0.005 min
Lab File: BG053767.D
Acq: 31 May 2022 19:16




Instrument :
BNA_G
ClientSampleId :
SSTD040424


Tgt Ion:237 Resp: 55277
Ion Ratio Lower Upper
237 100
235 63.7 47.0 70.6
272 12.4 8.6 12.8



#41
2,4,6-Trichlorophenol
Concen: 43.098 ng/ul
RT: 13.158 min Scan# 1731
Delta R.T. -0.001 min
Lab File: BG053767.D
Acq: 31 May 2022 19:16

Tgt Ion:196 Resp: 47999
Ion Ratio Lower Upper
196 100
198 94.8 81.9 122.9
200 30.8 25.7 38.5









#42
2,4,5-Trichlorophenol
Concen: 41.017 ng/ul
RT: 13.228 min Scan# 1
Delta R.T. -0.001 min
Lab File: BG053767.D
Acq: 31 May 2022 19:16




Instrument :
BNA_G
ClientSampleId :
SSTD040424


Tgt Ion:196 Resp: 49332
Ion Ratio Lower Upper
196 100
198 93.1 76.3 114.5
200 33.0 25.5 38.3



#43
1,1'-Biphenyl
Concen: 42.530 ng/ul
RT: 13.563 min Scan# 1800
Delta R.T. -0.001 min
Lab File: BG053767.D
Acq: 31 May 2022 19:16

Tgt Ion:154 Resp: 167583
Ion Ratio Lower Upper
154 100
153 40.6 32.7 49.1
76 11.5 9.0 13.6





#44

2-Chloronaphthalene

Concen: 40.880 ng/ul

RT: 13.604 min Scan# 1839

Delta R.T. -0.001 min

Lab File: BG053767.D

Acq: 31 May 2022 19:16

Instrument :

BNA_G

ClientSampleId :

SSTD040424





Tgt Ion:162 Resp: 128985

Ion Ratio Lower Upper

162 100

164 33.6 25.7 38.5

127 37.6 28.3 42.5



#45

2-Nitroaniline

Concen: 32.630 ng/ul

RT: 13.792 min Scan# 1839

Delta R.T. -0.006 min

Lab File: BG053767.D

Acq: 31 May 2022 19:16




Tgt Ion: 65 Resp: 30179


Ion Ratio Lower Upper

65 100

92 66.2 52.2 78.4

138 98.0 79.0 118.4





#47

Dimethylphthalate

Concen: 42.065 ng/ul

RT: 14.174 min Scan# 1904

Delta R.T. -0.001 min

Lab File: BG053767.D

Acq: 31 May 2022 19:16

Instrument :

BNA_G

ClientSampleId :

SSTD040424





Tgt Ion:163 Resp: 170052

Ion Ratio Lower Upper

163 100

194 3.8 3.7 5.5

164 10.2 7.9 11.9



#48

2,6-Dinitrotoluene

Concen: 37.100 ng/ul

RT: 14.291 min Scan# 1924

Delta R.T. 0.005 min

Lab File: BG053767.D

Acq: 31 May 2022 19:16




Tgt Ion:165 Resp: 27169


Ion Ratio Lower Upper

165 100

89 56.3 44.7 67.1

121 22.6 16.9 25.3









#50
Acenaphthylene
Concen: 42.261 ng/ul
RT: 14.450 min Scan# 1951
Delta R.T. -0.001 min
Lab File: BG053767.D
Acq: 31 May 2022 19:16




Instrument :
BNA_G
ClientSampleId :
SSTD040424


Tgt Ion:152 Resp: 210098
Ion Ratio Lower Upper
152 100
151 20.6 16.2 24.2
153 13.7 10.5 15.7



#51
3-Nitroaniline
Concen: 33.800 ng/ul
RT: 14.615 min Scan# 1979
Delta R.T. 0.005 min
Lab File: BG053767.D
Acq: 31 May 2022 19:16

Tgt Ion:138 Resp: 24984
Ion Ratio Lower Upper
138 100
108 22.3 17.8 26.6
92 120.8 94.2 141.4









#52
Acenaphthene
Concen: 43.143 ng/ul
RT: 14.791 min Scan# 2009
Delta R.T. -0.001 min
Lab File: BG053767.D
Acq: 31 May 2022 19:16




Instrument :
BNA_G
ClientSampleId :
SSTD040424


Tgt Ion:153 Resp: 140118
Ion Ratio Lower Upper
153 100
152 47.1 37.5 56.3
154 87.3 68.8 103.2



#53
2,4-Dinitrophenol
Concen: 27.039 ng/ul
RT: 14.814 min Scan# 2013
Delta R.T. -0.001 min
Lab File: BG053767.D
Acq: 31 May 2022 19:16

Tgt Ion:184 Resp: 12301
Ion Ratio Lower Upper
184 100
63 54.9 34.7 52.1#
154 120.8 87.3 130.9









#55
4-Nitrophenol
Concen: 30.649 ng/ul
RT: 14.908 min Scan# 2029
Delta R.T. -0.001 min
Lab File: BG053767.D
Acq: 31 May 2022 19:16




Instrument :
BNA_G
ClientSampleId :
SSTD040424


Tgt Ion:109 Resp: 22974
Ion Ratio Lower Upper
109 100
139 109.6 92.9 139.3
65 127.8 97.8 146.6



#56
Dibenzofuran
Concen: 42.265 ng/ul
RT: 15.126 min Scan# 2066
Delta R.T. 0.005 min
Lab File: BG053767.D
Acq: 31 May 2022 19:16

Tgt Ion:168 Resp: 203499
Ion Ratio Lower Upper
168 100
139 36.2 28.2 42.2









#57
2,4-Dinitrotoluene
Concen: 37.027 ng/ul
RT: 15.073 min Scan# 2056
Delta R.T. 0.005 min
Lab File: BG053767.D
Acq: 31 May 2022 19:16




Instrument :
BNA_G
ClientSampleId :
SSTD040424


Tgt Ion:165 Resp: 38473
Ion Ratio Lower Upper
165 100
63 38.7 33.0 49.4
182 2.4 2.8 4.2#



#58
2,3,4,6-Tetrachlorophenol
Concen: 43.110 ng/ul
RT: 15.343 min Scan# 2103
Delta R.T. -0.001 min
Lab File: BG053767.D
Acq: 31 May 2022 19:16

Tgt Ion:232 Resp: 46736
Ion Ratio Lower Upper
232 100
131 32.5 27.5 41.3
130 2.1 1.8 2.8
166 24.5 20.7 31.1









#59
Diethylphthalate
Concen: 42.340 ng/ul
RT: 15.543 min Scan# 2137
Delta R.T. 0.005 min
Lab File: BG053767.D
Acq: 31 May 2022 19:16




Instrument :
BNA_G
ClientSampleId :
SSTD040424


Tgt Ion:149 Resp: 173348
Ion Ratio Lower Upper
149 100
177 23.5 18.9 28.3
150 12.4 10.6 15.8



#61
Fluorene
Concen: 41.623 ng/ul
RT: 15.772 min Scan# 2176
Delta R.T. -0.001 min
Lab File: BG053767.D
Acq: 31 May 2022 19:16

Tgt Ion:166 Resp: 166118
Ion Ratio Lower Upper
166 100
165 99.7 81.7 122.5
167 13.5 11.0 16.4









#62
4-Chlorophenyl-phenylether
Concen: 43.634 ng/ul
RT: 15.766 min Scan# 2175
Delta R.T. 0.005 min
Lab File: BG053767.D
Acq: 31 May 2022 19:16




Instrument :
BNA_G
ClientSampleId :
SSTD040424


Tgt Ion:204 Resp: 92667
Ion Ratio Lower Upper
204 100
206 32.8 27.1 40.7
141 50.2 40.7 61.1



#63
4-Nitroaniline
Concen: 35.316 ng/ul
RT: 15.778 min Scan# 2177
Delta R.T. 0.005 min
Lab File: BG053767.D
Acq: 31 May 2022 19:16

Tgt Ion:138 Resp: 25671
Ion Ratio Lower Upper
138 100
92 56.5 42.0 63.0
108 93.8 79.0 118.6









#66
4,6-Dinitro-2-methylphenol
Concen: 31.135 ng/ul
RT: 15.837 min Scan# 2186
Delta R.T. 0.005 min
Lab File: BG053767.D
Acq: 31 May 2022 19:16




Instrument :
BNA_G
ClientSampleId :
SSTD040424


Tgt Ion:198 Resp: 20677
Ion Ratio Lower Upper
198 100
182 3.7 3.4 5.0
77 19.4 14.7 22.1



#67
N-Nitrosodiphenylamine
Concen: 42.217 ng/ul
RT: 15.972 min Scan# 2210
Delta R.T. -0.001 min
Lab File: BG053767.D
Acq: 31 May 2022 19:16

Tgt Ion:169 Resp: 152240
Ion Ratio Lower Upper
169 100
168 66.1 51.0 76.4
167 36.3 28.7 43.1









#68
4-Bromophenyl-phenylether
Concen: 48.063 ng/ul
RT: 16.659 min Scan# 21
Delta R.T. -0.001 min
Lab File: BG053767.D
Acq: 31 May 2022 19:16




Instrument :
BNA_G
ClientSampleId :
SSTD040424


Tgt Ion:248 Resp: 63472
Ion Ratio Lower Upper
248 100
250 98.4 78.2 117.2
141 60.3 46.7 70.1



#69
Hexachlorobenzene
Concen: 41.214 ng/ul
RT: 16.777 min Scan# 2347
Delta R.T. -0.001 min
Lab File: BG053767.D
Acq: 31 May 2022 19:16

Tgt Ion:284 Resp: 70307
Ion Ratio Lower Upper
284 100
142 26.1 20.8 31.2
249 32.2 27.4 41.0









#70
 Atrazine
 Concen: 46.552 ng/ul
 RT: 16.924 min Scan# 21
 Delta R.T. 0.005 min
 Lab File: BG053767.D
 Acq: 31 May 2022 19:16




Instrument :
 BNA_G
 ClientSampleId :
 SSTD040424


Tgt Ion:200 Resp: 68799
 Ion Ratio Lower Upper
 200 100
 173 23.8 19.8 29.6
 215 49.0 38.8 58.2



#71
 Pentachlorophenol
 Concen: 39.657 ng/ul
 RT: 17.118 min Scan# 2405
 Delta R.T. 0.005 min
 Lab File: BG053767.D
 Acq: 31 May 2022 19:16

Tgt Ion:266 Resp: 41139
 Ion Ratio Lower Upper
 266 100
 264 69.0 56.8 85.2
 268 70.5 48.0 72.0









#72
Phenanthrene
Concen: 41.195 ng/u1
RT: 17.511 min Scan# 2472
Delta R.T. -0.001 min
Lab File: BG053767.D
Acq: 31 May 2022 19:16




Instrument :
BNA_G
ClientSampleId :
SSTD040424


Tgt Ion:178 Resp: 288368
Ion Ratio Lower Upper
178 100
179 16.1 12.6 19.0
176 19.3 15.7 23.5



#74
Anthracene
Concen: 41.392 ng/u1
RT: 17.605 min Scan# 2488
Delta R.T. 0.005 min
Lab File: BG053767.D
Acq: 31 May 2022 19:16

Tgt Ion:178 Resp: 289677
Ion Ratio Lower Upper
178 100
179 15.6 12.4 18.6
176 18.7 14.6 22.0









#75
1,2,3,4-Tetrachlorobenzene
Concen: 44.174 ng/uL
RT: 13.528 min Scan# 11
Delta R.T. -0.001 min
Lab File: BG053767.D
Acq: 31 May 2022 19:16




Instrument :
BNA_G
ClientSampleId :
SSTD040424


Tgt Ion:216 Resp: 87749
Ion Ratio Lower Upper
216 100
214 75.4 64.2 96.4
218 48.3 39.0 58.4



#76
Pentachlorobenzene
Concen: 42.887 ng/uL
RT: 15.049 min Scan# 2053
Delta R.T. 0.005 min
Lab File: BG053767.D
Acq: 31 May 2022 19:16

Tgt Ion:250 Resp: 81734
Ion Ratio Lower Upper
250 100
248 60.2 47.7 71.5
252 66.9 50.0 75.0









#77
 Carbazole
 Concen: 42.437 ng/ul
 RT: 17.864 min Scan# 2532
 Delta R.T. -0.001 min
 Lab File: BG053767.D
 Acq: 31 May 2022 19:16




Instrument :
 BNA_G
 ClientSampleId :
 SSTD040424


Tgt Ion:167 Resp: 262465
 Ion Ratio Lower Upper
 167 100
 166 22.6 18.2 27.2
 139 12.0 9.0 13.6



#78
 Di-n-butylphthalate
 Concen: 43.012 ng/ul
 RT: 18.434 min Scan# 2629
 Delta R.T. -0.001 min
 Lab File: BG053767.D
 Acq: 31 May 2022 19:16

Tgt Ion:149 Resp: 316344
 Ion Ratio Lower Upper
 149 100
 150 9.2 7.1 10.7
 104 5.2 4.2 6.2





#80

Fluoranthene

Concen: 39.907 ng/u1

RT: 19.521 min Scan# 2813

Delta R.T. 0.005 min

Lab File: BG053767.D

Acq: 31 May 2022 19:16

Instrument :

BNA_G

ClientSampleId :

SSTD040424





Tgt Ion:202 Resp: 370609

Ion Ratio Lower Upper

202 100

101 8.0 6.6 9.8

100 6.1 5.8 8.8



#82

Pyrene

Concen: 39.952 ng/u1

RT: 19.879 min Scan# 2875

Delta R.T. -0.001 min

Lab File: BG053767.D

Acq: 31 May 2022 19:16




Tgt Ion:202 Resp: 368785


Ion Ratio Lower Upper

202 100

101 9.8 7.3 10.9

100 7.7 6.2 9.4









#83
 Butylbenzylphthalate
 Concen: 42.009 ng/ul
 RT: 20.772 min Scan# 30
 Delta R.T. 0.005 min
 Lab File: BG053767.D
 Acq: 31 May 2022 19:16




Instrument :
 BNA_G
 ClientSampleId :
 SSTD040424


Tgt Ion:149 Resp: 134822
 Ion Ratio Lower Upper
 149 100
 91 66.5 54.0 81.0
 206 25.4 19.5 29.3



#84
 3,3'-Dichlorobenzidine
 Concen: 44.191 ng/ul
 RT: 21.642 min Scan# 3175
 Delta R.T. 0.005 min
 Lab File: BG053767.D
 Acq: 31 May 2022 19:16

Tgt Ion:252 Resp: 139529
 Ion Ratio Lower Upper
 252 100
 254 65.5 49.6 74.4
 126 7.7 6.0 9.0









#85
Benzo(a)anthracene
Concen: 41.789 ng/ul
RT: 21.730 min Scan# 3190
Delta R.T. -0.001 min
Lab File: BG053767.D
Acq: 31 May 2022 19:16




Instrument :
BNA_G
ClientSampleId :
SSTD040424


Tgt Ion:228 Resp: 379952
Ion Ratio Lower Upper
228 100
229 19.9 16.5 24.7
226 26.8 22.1 33.1



#86
Bis(2-ethylhexyl)phthalate
Concen: 44.500 ng/ul
RT: 21.653 min Scan# 3177
Delta R.T. -0.001 min
Lab File: BG053767.D
Acq: 31 May 2022 19:16

Tgt Ion:149 Resp: 197806
Ion Ratio Lower Upper
149 100
167 28.1 21.7 32.5
279 4.7 3.9 5.9









#87
Chrysene
Concen: 41.731 ng/ul
RT: 21.800 min Scan# 31
Delta R.T. 0.005 min
Lab File: BG053767.D
Acq: 31 May 2022 19:16




Instrument :
BNA_G
ClientSampleId :
SSTD040424


Tgt Ion:228 Resp: 361905
Ion Ratio Lower Upper
228 100
226 30.6 24.2 36.4
229 20.1 16.3 24.5



#89
Di-n-octyl phthalate
Concen: 44.833 ng/ul
RT: 22.893 min Scan# 3388
Delta R.T. 0.005 min
Lab File: BG053767.D
Acq: 31 May 2022 19:16

Tgt Ion:149 Resp: 340911









#90
Benzo(b)fluoranthene
Concen: 40.495 ng/ul
RT: 23.986 min Scan# 3574
Delta R.T. -0.001 min
Lab File: BG053767.D
Acq: 31 May 2022 19:16




Instrument :
BNA_G
ClientSampleId :
SSTD040424


Tgt Ion:252 Resp: 381742
Ion Ratio Lower Upper
252 100
253 22.0 18.2 27.4
125 7.1 6.1 9.1



#91
Benzo(k)fluoranthene
Concen: 41.445 ng/ul
RT: 24.057 min Scan# 3586
Delta R.T. 0.005 min
Lab File: BG053767.D
Acq: 31 May 2022 19:16

Tgt Ion:252 Resp: 360582
Ion Ratio Lower Upper
252 100
253 22.6 17.4 26.2
125 7.2 5.4 8.0









#93
Benzo(a)pyrene
Concen: 41.042 ng/u1
RT: 24.879 min Scan# 31
Delta R.T. 0.005 min
Lab File: BG053767.D
Acq: 31 May 2022 19:16




Instrument :
BNA_G
ClientSampleId :
SSTD040424


Tgt Ion:252 Resp: 365276
Ion Ratio Lower Upper
252 100
253 22.0 17.6 26.4
125 6.8 6.8 10.2#



#94
Indeno(1,2,3-cd)pyrene
Concen: 41.884 ng/u1
RT: 28.775 min Scan# 4389
Delta R.T. 0.005 min
Lab File: BG053767.D
Acq: 31 May 2022 19:16

Tgt Ion:276 Resp: 421585
Ion Ratio Lower Upper
276 100
138 13.3 10.1 15.1
277 23.7 19.9 29.9









#95
 Dibenzo(a,h)anthracene
 Concen: 42.046 ng/ul
 RT: 28.845 min Scan# 4400
 Delta R.T. 0.005 min
 Lab File: BG053767.D
 Acq: 31 May 2022 19:16

Instrument :
 BNA_G
 ClientSampleId :
 SSTD040424

Tgt Ion:278 Resp: 359703
 Ion Ratio Lower Upper
 278 100
 139 10.1 8.6 13.0
 279 23.6 19.4 29.0



#96
 Benzo(g,h,i)perylene
 Concen: 41.669 ng/ul
 RT: 29.950 min Scan# 4589
 Delta R.T. 0.023 min
 Lab File: BG053767.D
 Acq: 31 May 2022 19:16

Tgt Ion:276 Resp: 354546
 Ion Ratio Lower Upper
 276 100
 138 13.6 11.3 16.9
 277 23.5 18.6 27.8

