


Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG053122\
 Data File : BG053763.D
 Acq On : 31 May 2022 16:35
 Operator : CG/JU
 Sample : DFTPP476
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 DFTPP476

Integration File: LSCINT.P

Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\SFAM-EPA-BG053122.M
 Title : SVOA CALIBRATION
 Last Update : Wed Jun 01 00:25:02 2022



AutoFind: Scans 2539, 2540, 2541; Background Corrected with Scan 2533

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	35.2	22692	PASS
68	69	0.00	2	0.0	0	PASS
69	198	0.00	100	32.7	21064	PASS
70	69	0.00	2	0.1	20	PASS
127	198	10	80	40.0	25744	PASS
197	198	0.00	2	0.0	0	PASS
198	198	100	100	100.0	64395	PASS
199	198	5	9	7.3	4692	PASS
275	198	10	60	30.8	19840	PASS
365	198	1	100	4.1	2665	PASS
441	198	0.01	100	13.0	8354	PASS
442	198	50	100	82.6	53195	PASS
443	442	15	24	18.7	9947	PASS

m/z	Abundance
39.10	242.0
40.00	267.0
41.10	1015.0
42.00	232.0
43.10	1517.0
44.00	1177.0
45.00	181.0
49.00	210.0
51.10	199.0
54.00	185.0
55.10	1134.0
56.10	358.0
57.10	1109.0
58.00	347.0
59.10	331.0
67.10	375.0
68.10	256.0
69.00	464.0
70.10	280.0
71.10	766.0
73.10	471.0
79.00	173.0
81.00	371.0
82.10	318.0
83.00	379.0
84.00	418.0
85.00	493.0
85.90	154.0
90.90	152.0
95.00	279.0
96.00	722.0
97.10	376.0
98.10	487.0
104.90	167.0
112.00	153.0
118.90	185.0
132.90	582.0
135.10	259.0
147.00	191.0
161.00	150.0
162.90	211.0
176.80	202.0
190.90	726.0
192.00	168.0
192.90	293.0
207.00	6738.0
207.90	1556.0
208.90	1034.0
209.80	168.0
239.10	245.0
255.20	183.0
267.10	345.0
280.90	691.0
281.90	243.0
283.10	313.0

Instrument :
BNA_G
ClientSampleId :
DFTPP476

m/z	Abundance
38.10	411.0
39.10	1860.0
40.00	296.0
41.10	906.0
42.10	315.0
43.10	1288.0
44.00	1091.0
45.00	171.0
49.00	438.0
50.10	5988.0
51.10	24456.0
52.10	1097.0
54.00	199.0
55.10	1136.0
56.00	1142.0
57.10	2422.0
58.10	510.0
59.00	423.0
61.10	289.0
62.00	371.0
63.00	799.0
65.00	450.0
67.10	319.0
69.00	23160.0
70.00	282.0
71.10	723.0
73.10	686.0
74.10	2316.0
75.00	3730.0
76.10	1430.0
77.10	24152.0
78.00	1700.0
79.00	1754.0
80.00	1277.0
81.00	2174.0
82.00	749.0
83.10	904.0
84.00	431.0
85.10	697.0
86.00	602.0
87.00	288.0
91.00	582.0
92.00	493.0
93.00	3240.0
94.00	217.0
95.00	347.0
96.00	752.0
97.00	577.0
98.00	2764.0
99.00	1757.0
99.90	199.0
101.00	1145.0
103.00	434.0
104.00	668.0
105.00	635.0
106.00	207.0
107.00	8150.0
108.00	1055.0
110.00	13105.0
111.00	2130.0
112.00	325.0
113.00	170.0
115.00	175.0
116.00	557.0
117.00	6859.0
118.00	575.0
119.00	254.0
119.90	152.0
121.90	580.0
122.90	775.0
124.00	426.0
125.00	425.0
127.00	25200.0
128.00	1913.0
129.00	10282.0
130.00	971.0
130.90	204.0
132.90	462.0
134.00	338.0
135.00	999.0
136.00	409.0
136.90	531.0
138.00	183.0
140.00	182.0
140.90	1234.0
141.90	474.0
143.00	259.0
146.00	365.0
147.00	1037.0
147.90	1580.0
149.00	435.0
151.20	214.0

Instrument :
BNA_G
ClientSampleId :
DFTPP476

152.00	165.0
152.90	510.0
153.90	410.0
155.00	877.0
156.00	1239.0
157.00	277.0
158.00	180.0
158.90	240.0
160.00	436.0
161.00	854.0
161.90	183.0
163.10	156.0
163.90	169.0
164.90	780.0
165.90	671.0
166.90	3029.0
167.90	1206.0
168.90	330.0
170.90	171.0
172.00	277.0
172.90	365.0
173.90	644.0
175.00	1282.0
175.90	497.0
176.90	765.0
178.00	223.0
178.90	2522.0
179.90	1333.0
181.00	700.0
183.90	263.0
184.90	1197.0
186.00	8397.0
186.90	2678.0
187.90	376.0
188.90	647.0
190.90	1016.0
192.00	966.0
193.00	1199.0
194.00	395.0
195.90	1701.0
197.90	61472.0
198.90	4780.0
200.00	360.0
201.40	236.0
202.80	578.0
204.00	2533.0
205.00	4217.0
206.00	18048.0
207.00	8957.0
207.90	1802.0
208.90	1339.0
210.00	457.0
211.00	818.0
215.00	166.0
215.90	630.0
216.90	5159.0
217.90	733.0
218.90	165.0
220.90	2150.0
222.00	696.0
222.90	1122.0
223.90	10456.0
224.90	2509.0
225.90	268.0
226.90	5053.0
227.90	795.0
228.90	983.0
229.90	175.0
231.00	267.0
232.90	153.0
233.90	225.0
234.90	359.0
235.90	238.0
236.90	311.0
239.00	403.0
239.90	189.0
240.90	360.0
241.90	541.0
242.90	565.0
243.90	7489.0
244.90	778.0
245.90	1568.0
246.90	339.0
248.90	416.0
252.90	360.0
254.90	41744.0
255.90	6185.0
256.80	440.0
257.90	2526.0
258.90	500.0
264.90	1055.0
265.90	208.0
267.10	317.0
271.90	201.0
272.80	1122.0
273.90	3035.0

Instrument :
BNA_G
ClientSampleId :
DFTPP476

274.90	17936.0
275.90	2400.0
276.90	1372.0
277.90	184.0
280.90	899.0
281.90	296.0
283.00	481.0
283.90	179.0
284.10	177.0
284.90	272.0
292.80	275.0
295.90	5168.0
296.90	855.0
302.90	668.0
313.80	309.0
314.80	650.0
315.90	328.0
320.90	278.0
322.90	1631.0
323.90	283.0
326.80	325.0
327.80	195.0
333.90	1085.0
334.80	272.0
340.90	275.0
345.80	396.0
351.90	563.0
353.00	288.0
353.90	408.0
354.80	224.0
364.90	2183.0
365.80	382.0
371.90	884.0
372.90	276.0
382.80	233.0
401.80	398.0
402.90	458.0
403.80	217.0
420.90	400.0
421.80	396.0
422.90	2724.0
423.90	548.0
440.80	6253.0
441.90	42544.0
442.90	7880.0
443.90	774.0

Instrument :
BNA_G
ClientSampleId :
DFTPP476

m/z	Abundance
38.10	397.0
39.10	2011.0
40.10	314.0
41.10	845.0
42.10	266.0
43.10	1374.0
44.10	1039.0
45.00	240.0
49.10	349.0
50.10	6787.0
51.10	26288.0
52.10	1302.0
53.00	194.0
55.10	1126.0
56.10	1144.0
57.10	2997.0
58.10	542.0
59.00	244.0
61.10	371.0
62.00	362.0
63.00	1065.0
64.10	188.0
65.00	620.0
67.00	409.0
68.00	562.0
69.00	24056.0
70.10	251.0
71.10	936.0
73.00	626.0
74.10	2640.0
75.00	3976.0
76.00	1418.0
77.10	27896.0
78.10	2050.0
79.00	1898.0
80.00	1505.0
81.00	2449.0
82.00	789.0
83.10	860.0
84.00	269.0
85.00	679.0
86.00	716.0
86.90	288.0
88.00	173.0
88.90	154.0
91.00	696.0
92.00	620.0
93.00	3341.0
94.00	320.0
95.10	321.0
96.10	860.0
97.10	453.0
98.00	2946.0
99.00	2063.0
100.00	266.0
101.00	1230.0
103.00	510.0
104.00	861.0
105.00	945.0
106.00	256.0
107.00	8833.0
108.00	1230.0
109.00	461.0
110.00	15161.0
111.00	2359.0
112.00	440.0
114.80	174.0
116.00	515.0
117.00	7386.0
118.00	520.0
119.00	212.0
120.00	199.0
121.90	636.0
123.00	901.0
124.00	602.0
124.90	455.0
127.00	29160.0
128.00	2247.0
129.00	12420.0
129.90	1058.0
130.90	419.0
132.00	237.0
133.00	487.0
134.00	314.0
135.00	1155.0
136.00	537.0
137.00	578.0
139.90	206.0
140.90	1664.0
142.00	580.0
143.00	420.0
146.00	357.0

Instrument :
BNA_G
ClientSampleId :
DFTPP476

147.00	1163.0
148.00	2087.0
149.00	452.0
150.90	261.0
151.90	156.0
153.00	567.0
153.90	436.0
155.00	1011.0
156.00	1472.0
157.00	274.0
158.00	274.0
159.00	243.0
160.00	669.0
160.90	879.0
162.00	315.0
163.00	190.0
165.00	765.0
166.00	704.0
167.00	3146.0
168.00	1631.0
168.80	266.0
170.00	154.0
170.90	222.0
171.90	346.0
173.00	485.0
174.00	657.0
175.00	1909.0
175.90	365.0
176.90	896.0
178.00	298.0
178.90	3192.0
180.00	2018.0
180.90	991.0
181.90	220.0
182.70	170.0
184.00	240.0
184.90	1426.0
186.00	10635.0
186.90	3429.0
187.90	326.0
188.90	879.0
190.00	175.0
190.90	1185.0
191.90	1170.0
193.00	1267.0
193.90	256.0
195.90	2305.0
197.90	73864.0
198.90	5526.0
199.90	573.0
201.40	395.0
202.90	627.0
203.90	3405.0
205.00	5320.0
206.00	21792.0
207.00	9261.0
207.90	2055.0
208.90	1086.0
210.00	540.0
210.90	1121.0
211.80	232.0
214.90	342.0
215.90	596.0
216.90	5895.0
217.90	720.0
220.90	2894.0
221.90	863.0
222.90	1427.0
223.90	12642.0
224.90	3331.0
225.90	355.0
226.90	5745.0
227.90	809.0
228.90	1336.0
229.90	174.0
231.00	436.0
233.80	503.0
234.90	411.0
235.80	213.0
236.90	564.0
239.00	547.0
240.00	194.0
241.00	280.0
241.90	629.0
242.90	649.0
243.90	9573.0
245.00	1366.0
245.90	2039.0
246.90	336.0
248.90	464.0
251.00	225.0
252.90	259.0
254.90	52928.0
255.90	7462.0
256.90	547.0
257.90	3253.0

Instrument :
BNA_G
ClientSampleId :
DFTPP476

258.90	454.0
264.90	1377.0
265.80	181.0
267.10	440.0
271.90	244.0
272.90	1798.0
273.90	4196.0
274.90	22808.0
275.90	2893.0
276.90	2109.0
277.90	331.0
281.00	682.0
281.90	256.0
283.00	620.0
284.00	194.0
284.90	406.0
285.90	160.0
291.90	161.0
292.80	450.0
295.90	6500.0
296.90	811.0
302.90	796.0
303.90	398.0
313.90	348.0
314.80	731.0
315.90	337.0
320.80	221.0
322.90	2283.0
323.90	424.0
326.80	446.0
327.90	281.0
328.90	153.0
331.80	179.0
332.90	208.0
333.90	1365.0
335.00	342.0
340.90	324.0
345.80	467.0
351.90	636.0
352.80	416.0
353.90	624.0
354.90	216.0
364.80	3177.0
365.80	359.0
370.80	199.0
371.90	978.0
372.80	264.0
382.80	360.0
389.90	197.0
401.80	562.0
402.80	602.0
403.80	187.0
420.90	561.0
421.80	564.0
422.90	3699.0
423.90	880.0
440.80	9789.0
441.90	60280.0
442.90	11449.0
443.90	1139.0

Instrument :
BNA_G
ClientSampleId :
DFTPP476

m/z	Abundance
38.00	275.0
39.10	1551.0
40.00	300.0
41.10	825.0
42.10	309.0
43.10	1446.0
44.10	991.0
45.00	247.0
49.00	407.0
50.10	4628.0
51.10	17928.0
52.10	881.0
53.10	155.0
54.00	205.0
55.10	1180.0
56.00	972.0
57.10	2428.0
58.10	640.0
59.10	375.0
61.10	274.0
62.00	181.0
63.00	654.0
63.90	182.0
65.10	463.0
66.10	176.0
67.00	317.0
69.00	17368.0
70.00	368.0
71.10	758.0
73.10	559.0
74.00	1858.0
75.00	2851.0
76.00	926.0
77.10	19520.0
78.10	1075.0
79.00	1581.0
80.00	1195.0
81.00	1717.0
82.00	673.0
83.00	766.0
84.00	337.0
85.00	706.0
86.00	563.0
87.00	351.0
88.90	174.0
91.00	531.0
92.00	494.0
93.00	2458.0
94.00	288.0
95.00	378.0
96.10	846.0
97.00	470.0
98.00	2191.0
99.00	1377.0
100.00	217.0
101.00	886.0
102.90	335.0
103.90	738.0
104.90	575.0
107.00	6499.0
108.00	902.0
110.00	10871.0
111.00	2012.0
112.00	246.0
116.00	426.0
117.00	5320.0
117.90	329.0
119.00	302.0
119.90	152.0
121.00	165.0
122.00	447.0
123.00	900.0
123.90	406.0
125.00	295.0
127.00	22872.0
128.00	1595.0
129.00	8808.0
130.00	632.0
131.00	180.0
133.00	563.0
134.00	362.0
135.00	926.0
136.00	396.0
137.00	457.0
138.00	165.0
141.00	1108.0
142.00	417.0
142.90	332.0
146.00	276.0
147.00	998.0
148.00	1387.0
149.00	253.0

Instrument :
BNA_G
ClientSampleId :
DFTPP476

151.00	201.0
151.90	154.0
153.00	491.0
153.90	305.0
155.00	692.0
156.00	1041.0
156.90	213.0
158.00	309.0
159.00	301.0
159.90	452.0
160.90	922.0
162.00	173.0
162.90	172.0
165.00	698.0
166.00	429.0
167.00	2671.0
167.90	1345.0
168.90	312.0
171.90	229.0
172.90	413.0
174.00	531.0
175.00	1094.0
175.90	395.0
176.90	901.0
177.90	306.0
178.90	2231.0
179.90	1548.0
180.90	873.0
183.90	226.0
184.90	1240.0
186.00	8801.0
187.00	2113.0
188.00	256.0
188.90	702.0
189.90	186.0
190.90	1156.0
191.90	1109.0
193.00	1278.0
194.00	324.0
194.90	185.0
195.90	1696.0
197.90	57848.0
198.90	3770.0
199.90	363.0
202.90	534.0
204.00	2354.0
205.00	3817.0
206.00	17096.0
207.00	8088.0
208.00	2026.0
208.90	1230.0
209.80	269.0
210.90	602.0
215.00	235.0
216.00	406.0
216.90	4556.0
217.90	609.0
220.00	219.0
220.90	2287.0
221.80	826.0
223.00	1269.0
223.90	9606.0
224.90	2530.0
225.90	291.0
226.90	4554.0
227.90	698.0
228.90	900.0
230.00	168.0
230.90	340.0
234.00	202.0
234.90	489.0
235.80	174.0
236.90	383.0
239.00	394.0
240.00	228.0
240.90	286.0
241.90	445.0
243.00	558.0
243.90	7358.0
245.00	974.0
245.90	1665.0
246.80	345.0
248.90	285.0
253.00	304.0
254.90	42888.0
255.90	6271.0
256.90	481.0
257.90	2411.0
258.90	524.0
264.90	1068.0
265.90	189.0
267.10	542.0
272.90	1245.0
273.90	3128.0
274.90	18776.0
275.90	2396.0

Instrument :
BNA_G
ClientSampleId :
DFTPP476

276.90	1438.0
277.80	194.0
281.00	603.0
282.00	327.0
283.00	532.0
285.00	293.0
292.90	446.0
294.90	158.0
295.90	5841.0
296.90	723.0
302.00	154.0
302.90	567.0
303.90	189.0
313.90	330.0
314.80	833.0
315.80	365.0
320.90	271.0
321.90	156.0
322.90	1796.0
323.90	321.0
326.80	411.0
327.80	246.0
332.90	190.0
333.90	1064.0
334.90	217.0
340.90	367.0
345.80	403.0
351.90	581.0
352.90	322.0
353.80	561.0
354.80	207.0
364.80	2635.0
365.90	468.0
370.80	193.0
371.90	1045.0
372.90	210.0
382.80	403.0
401.80	379.0
402.90	605.0
403.90	191.0
420.80	469.0
421.80	474.0
422.90	3394.0
423.80	647.0
440.90	9020.0
441.90	56760.0
442.90	10511.0
443.90	747.0

Instrument :
BNA_G
ClientSampleId :
DFTPP476