

Data Path : Z:\HPCHEM1\BNA_G\DATA\BG060115\
 Data File : BG017375.D
 Acq On : 1 Jun 2015 19:36
 Operator : TP/IZ
 Sample : PB83701BL
 Misc : SBLK64
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 SBLK64

Quant Time: Jun 02 03:19:35 2015
 Quant Method : Z:\HPCHEM1\BNA_G\METHODS\SOM02.2-EPA-BG052915.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Tue Jun 02 03:15:26 2015
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.65	152	24039	20.00	ng/ul	0.00
18) Naphthalene-d8	10.44	136	109443	20.00	ng/ul	0.00
35) Acenaphthene-d10	14.30	164	73634	20.00	ng/ul	0.00
61) Phenanthrene-d10	17.05	188	172565	20.00	ng/ul	0.00
75) Chrysene-d12	21.25	240	185286	20.00	ng/ul	0.00
83) Perylene-d12	23.49	264	177379	20.00	ng/ul	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
3) 1,4-Dioxane-d8	3.06	96	3172	6.27	ng/uL	0.00
5) Phenol-d5	6.85	99	56341	27.80	ng/ul	0.00
7) Bis-(2-Chloroethyl)ether-d	6.98	67	36918	30.48	ng/ul	0.00
9) 2-Chlorophenol-d4	7.18	132	51512	31.36	ng/ul	0.00
13) 4-Methylphenol-d8	8.37	113	47711	27.48	ng/ul	0.00
19) Nitrobenzene-d5	8.80	128	25175	29.70	ng/ul	0.00
22) 2-Nitrophenol-d4	9.53	143	29284	29.02	ng/ul	0.00
26) 2,4-Dichlorophenol-d3	10.08	165	46400	26.16	ng/ul	0.00
29) 4-Chloroaniline-d4	10.58	131	71159	34.94	ng/ul	0.00
43) Dimethylphthalate-d6	13.72	166	169739	32.33	ng/ul	0.00
46) Acenaphthylene-d8	14.00	160	203536	29.94	ng/ul	0.00
51) 4-Nitrophenol-d4	14.56	143	26030	25.99	ng/ul	0.00
57) Fluorene-d10	15.30	176	156246	30.88	ng/ul	0.00
62) 4,6-Dinitro-2-methylphenol	15.43	200	27806	23.02	ng/ul	0.00
70) Anthracene-d10	17.15	188	240607	29.76	ng/ul	0.00
76) Pyrene-d10	19.46	212	299061	33.33	ng/ul	0.00
87) Benzo(a)pyrene-d12	23.35	264	239119	30.19	ng/ul	0.00

Target Compounds	Qvalue
------------------	--------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA_G\DATA\BG060115\
Data File : BG017375.D
Acq On : 1 Jun 2015 19:36
Operator : TP/IZ
Sample : PB83701BL
Misc : SBLK64
ALS Vial : 10 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
SBLK64

Quant Time: Jun 02 03:19:35 2015
Quant Method : Z:\HPCHEM1\BNA_G\METHODS\SOM02.2-EPA-BG052915.M
Quant Title : SVOA CALIBRATION
QLast Update : Tue Jun 02 03:15:26 2015
Response via : Initial Calibration

