


Data Path : Z:\HPCHEM1\BNA_G\DATA\BG060215\
Data File : BG017390.D
Acq On : 2 Jun 2015 12:00
Operator : TP/IZ
Sample : DFTPP04
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
DFTPP04

Integration File: LSCINT.P

Method : Z:\HPCHEM1\BNA_G\METHODS\SOM02.2-EPA-BG052915.M
Title : SVOA CALIBRATION
Last Update : Wed Jun 03 03:45:23 2015



AutoFind: Scans 2520, 2521, 2522; Background Corrected with Scan 2515

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	45.8	10350	PASS
68	69	0.00	2	0.0	0	PASS
69	198	0.00	100	47.3	10678	PASS
70	69	0.00	2	0.0	0	PASS
127	198	10	80	49.3	11134	PASS
197	198	0.00	2	0.4	86	PASS
198	198	100	100	100.0	22584	PASS
199	198	5	9	5.9	1322	PASS
275	198	10	60	33.7	7612	PASS
365	198	1	100	7.0	1579	PASS
441	443	0.01	100	61.9	2356	PASS
442	198	50	100	78.8	17788	PASS
443	442	15	24	21.4	3809	PASS

m/z	Abundance
37.90	237.0
39.00	1127.0
40.00	380.0
40.90	369.0
43.00	187.0
46.90	338.0
49.00	1741.0
50.10	3634.0
51.00	12288.0
52.00	620.0
55.00	223.0
56.00	334.0
57.00	975.0
62.00	157.0
63.00	635.0
65.10	216.0
69.00	12814.0
73.00	212.0
74.00	1598.0
75.00	2010.0
76.00	658.0
77.00	12789.0
78.00	1085.0
79.00	843.0
79.90	877.0
81.00	1173.0
82.00	336.0
83.00	426.0
83.90	1431.0
85.00	331.0
85.90	1162.0
87.10	211.0
87.90	243.0
90.90	281.0
92.00	238.0
93.00	1949.0
94.00	153.0
97.90	1712.0
98.90	1099.0
101.00	569.0
102.00	151.0
103.80	420.0
105.00	417.0
107.00	5006.0
108.00	823.0
110.00	7241.0
111.00	1011.0
115.90	251.0
117.00	4758.0
117.90	341.0
122.00	413.0
123.00	288.0
124.00	284.0
125.00	381.0
127.00	13107.0
128.00	1337.0
129.00	6226.0
130.00	546.0
130.90	210.0
134.00	215.0
134.90	501.0
135.90	256.0
137.00	326.0
138.10	173.0
141.00	859.0
142.00	162.0
143.00	164.0
145.90	227.0
147.00	529.0
148.00	1474.0
148.90	151.0
153.10	293.0
155.00	600.0
156.00	744.0
159.80	204.0
161.00	476.0
165.00	493.0
166.00	347.0
167.00	1981.0
168.00	1238.0
171.00	225.0
171.90	169.0
173.90	320.0
174.90	949.0
176.90	247.0
178.00	240.0
178.90	1615.0
180.00	739.0
180.90	185.0
184.90	843.0
186.00	4728.0

Instrument :
BNA_G
ClientSampleId :
DFTPP04

187.00	1296.0
188.00	151.0
188.90	468.0
190.80	185.0
191.90	639.0
193.10	560.0
194.00	197.0
195.90	845.0
197.90	27000.0
198.90	1680.0
200.00	324.0
201.60	156.0
202.90	193.0
203.90	1110.0
205.00	2865.0
206.00	8620.0
207.00	1314.0
207.90	567.0
208.80	267.0
210.30	271.0
210.80	393.0
211.60	203.0
213.20	160.0
215.00	337.0
216.00	305.0
216.90	3276.0
217.90	357.0
220.90	1051.0
222.00	427.0
222.90	558.0
224.00	5580.0
225.00	1681.0
226.00	157.0
226.90	2846.0
228.00	389.0
229.00	507.0
231.00	240.0
233.90	313.0
236.80	163.0
238.90	152.0
240.00	226.0
241.40	231.0
242.00	269.0
243.00	278.0
244.00	3949.0
245.00	842.0
245.90	1113.0
247.00	242.0
252.90	170.0
255.00	20728.0
256.00	3747.0
257.00	239.0
258.00	1738.0
259.00	259.0
264.90	650.0
273.00	972.0
274.00	2114.0
275.00	8704.0
276.00	1046.0
277.00	941.0
283.00	202.0
284.90	336.0
292.90	364.0
296.00	3583.0
297.00	475.0
303.00	288.0
304.00	193.0
314.90	269.0
316.90	172.0
323.10	905.0
323.80	198.0
327.00	210.0
328.00	158.0
334.00	511.0
335.10	244.0
341.00	173.0
345.90	178.0
352.00	162.0
353.00	167.0
354.10	307.0
363.60	157.0
365.00	1898.0
372.00	523.0
382.90	258.0
402.00	180.0
403.00	339.0
420.80	314.0
423.00	1459.0
424.00	380.0
441.00	2665.0
442.00	21016.0
443.00	4865.0
444.00	395.0

Instrument :
BNA_G
ClientSampleId :
DFTPP04

m/z	Abundance
39.00	1048.0
39.90	259.0
40.90	235.0
43.10	271.0
46.90	539.0
47.90	282.0
49.00	2319.0
50.00	3184.0
51.00	12995.0
52.00	571.0
55.00	433.0
56.00	386.0
56.90	912.0
60.10	188.0
61.00	227.0
61.90	362.0
63.00	446.0
64.90	200.0
69.00	11629.0
71.00	210.0
73.00	266.0
74.00	1233.0
75.00	1853.0
76.10	689.0
77.00	13994.0
78.10	562.0
78.90	993.0
80.00	801.0
80.90	1107.0
82.00	230.0
82.90	373.0
84.00	1305.0
85.00	301.0
85.90	1214.0
90.80	253.0
91.90	325.0
93.00	1883.0
96.90	185.0
98.00	1027.0
99.00	1046.0
99.90	213.0
101.00	253.0
102.90	172.0
103.50	198.0
103.90	350.0
105.00	321.0
105.80	175.0
107.00	4459.0
108.00	504.0
110.00	7228.0
111.00	1288.0
116.00	308.0
117.00	3944.0
118.10	243.0
121.90	359.0
123.00	475.0
124.70	183.0
127.00	12275.0
128.00	906.0
129.00	6408.0
129.90	373.0
134.10	210.0
135.00	328.0
135.90	319.0
136.90	259.0
140.10	235.0
141.00	1008.0
142.00	273.0
143.00	210.0
146.90	388.0
148.00	1002.0
149.00	229.0
152.90	284.0
153.90	370.0
155.00	532.0
156.00	625.0
157.00	150.0
157.90	180.0
160.90	700.0
164.90	309.0
165.90	380.0
166.90	1590.0
167.90	687.0
171.90	207.0
172.90	165.0
173.90	404.0
174.90	804.0
177.00	279.0
178.90	1273.0
180.00	678.0
180.90	321.0

Instrument :
BNA_G
ClientSampleId :
DFTPP04

184.00	166.0
185.00	571.0
186.00	4019.0
187.00	1250.0
188.90	321.0
192.00	377.0
192.90	443.0
196.10	745.0
197.90	21848.0
198.90	991.0
201.50	210.0
202.90	225.0
203.90	1507.0
205.00	1913.0
206.00	6608.0
207.00	1003.0
207.90	269.0
209.00	216.0
211.10	229.0
216.00	222.0
216.90	2201.0
217.90	374.0
219.80	155.0
221.00	1056.0
222.00	185.0
223.00	515.0
224.00	3950.0
225.00	960.0
226.00	154.0
226.90	2344.0
227.90	244.0
228.90	158.0
231.00	373.0
233.90	154.0
236.80	162.0
240.90	155.0
242.00	273.0
243.20	272.0
244.00	3406.0
245.00	575.0
245.90	692.0
246.90	207.0
255.00	16047.0
256.00	2608.0
257.00	229.0
258.00	1300.0
258.80	208.0
264.90	549.0
265.80	249.0
266.80	167.0
272.90	556.0
274.00	1316.0
275.00	7299.0
275.90	1061.0
277.00	694.0
277.90	169.0
282.90	178.0
285.00	297.0
292.90	324.0
296.00	3020.0
296.80	360.0
303.00	254.0
314.00	166.0
315.00	268.0
315.90	156.0
322.90	529.0
324.10	173.0
334.10	346.0
346.80	178.0
352.00	166.0
352.90	203.0
354.00	238.0
365.00	1141.0
365.90	164.0
422.00	198.0
423.00	821.0
423.80	173.0
439.70	194.0
440.30	228.0
441.10	2108.0
442.00	14102.0
443.10	2895.0
444.00	266.0

Instrument :
BNA_G
ClientSampleId :
DFTPP04

m/z	Abundance
39.00	662.0
39.90	271.0
41.00	204.0
42.00	154.0
42.90	307.0
47.00	483.0
49.00	1837.0
50.00	2435.0
51.00	8022.0
52.00	381.0
55.00	332.0
56.00	309.0
57.00	587.0
60.80	176.0
61.90	162.0
63.00	362.0
65.10	173.0
69.00	8117.0
73.10	280.0
74.00	570.0
75.00	1572.0
76.10	618.0
77.00	7363.0
78.00	906.0
78.90	399.0
79.90	380.0
81.00	714.0
81.90	314.0
82.90	235.0
83.90	1198.0
84.90	155.0
86.00	557.0
88.00	219.0
91.00	268.0
92.10	177.0
93.00	1399.0
97.30	188.0
98.00	913.0
99.00	373.0
101.00	415.0
104.90	267.0
107.00	2329.0
108.00	329.0
110.00	3331.0
110.90	890.0
117.00	2893.0
117.90	269.0
122.00	219.0
122.90	293.0
123.90	203.0
127.00	8021.0
128.00	686.0
129.00	3971.0
130.10	215.0
134.80	470.0
136.00	259.0
137.00	262.0
141.00	833.0
147.00	632.0
148.00	827.0
152.90	222.0
155.00	479.0
156.00	579.0
159.90	291.0
160.80	182.0
164.80	202.0
167.00	1025.0
167.90	517.0
169.00	188.0
172.00	210.0
172.90	234.0
174.10	226.0
175.00	253.0
177.00	340.0
179.00	1037.0
179.90	611.0
180.90	384.0
185.10	655.0
186.00	3802.0
187.00	607.0
189.00	506.0
190.90	176.0
191.80	230.0
192.90	522.0
193.80	157.0
195.90	469.0
196.90	258.0
197.90	18904.0
199.00	1296.0
201.40	222.0
204.00	970.0

Instrument :
BNA_G
ClientSampleId :
DFTPP04

204.90	1638.0
206.00	6440.0
207.00	799.0
208.00	355.0
208.80	151.0
210.90	249.0
216.20	163.0
216.90	1675.0
218.00	171.0
221.00	834.0
221.70	191.0
223.00	331.0
224.00	3335.0
225.00	1021.0
226.10	165.0
227.00	1456.0
228.00	407.0
229.00	396.0
231.20	204.0
233.90	169.0
235.00	166.0
235.90	198.0
243.10	205.0
244.00	2448.0
244.90	469.0
246.00	1168.0
255.00	14862.0
256.00	1990.0
257.00	222.0
257.90	1409.0
265.00	698.0
272.90	404.0
273.90	1419.0
275.00	6834.0
276.00	1155.0
277.00	803.0
292.90	200.0
295.00	166.0
296.00	2685.0
297.00	537.0
303.00	403.0
304.00	214.0
315.00	334.0
323.00	727.0
334.00	557.0
335.10	356.0
341.00	160.0
352.00	264.0
354.00	263.0
365.00	1699.0
372.00	472.0
401.90	214.0
423.00	938.0
424.00	322.0
435.70	153.0
439.30	190.0
441.00	2297.0
442.00	18248.0
443.00	3669.0
443.90	256.0

Instrument :
BNA_G
ClientSampleId :
DFTPP04

m/z	Abundance
35.00	193.0
39.00	175.0
39.90	254.0
40.90	302.0
42.10	233.0
43.10	263.0
47.00	583.0
49.00	1793.0
50.90	751.0
55.00	321.0
57.00	352.0
69.00	175.0
72.90	290.0
83.90	1301.0
84.80	170.0
85.90	740.0
97.00	178.0
129.00	159.0
207.20	229.0

Instrument :
BNA_G
ClientSampleId :
DFTPP04