


Data Path : Z:\svoasrv\HPCHEM1\BNA G\Data\BG060220\
 Data File : BG045570.D
 Acq On : 2 Jun 2020 9:33
 Operator : CG/JU
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 DFTPP

Integration File: rteint.p

Method : Z:\SVOASRV\HPCHEM1\BNA_G\METHODS\8270-BG051520.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Mon Jun 01 15:07:00 2020



AutoFind: Scans 2502, 2503, 2504; Background Corrected with Scan 2496

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	26.3	13521	PASS
68	69	0.00	2	0.4	75	PASS
69	198	0.00	100	34.0	17449	PASS
70	69	0.00	2	1.0	171	PASS
127	198	10	80	43.6	22416	PASS
197	198	0.00	2	0.0	0	PASS
198	198	100	100	100.0	51386	PASS
199	198	5	9	6.7	3455	PASS
275	198	10	60	31.1	15981	PASS
365	198	1	100	5.1	2596	PASS
441	198	0.01	100	13.5	6948	PASS
442	198	50	100	85.3	43829	PASS
443	442	15	24	18.9	8278	PASS