

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG060222\
 Data File : BG053802.D
 Acq On : 2 Jun 2022 22:16
 Operator : CG/JU
 Sample : N3080-05RE
 Misc :
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 WC-5RE

Quant Time: Jun 03 02:28:37 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG060222.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Jun 02 23:22:27 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.090	152	11544	20.000	ng	0.00
21) Naphthalene-d8	10.898	136	50832	20.000	ng	# 0.00
39) Acenaphthene-d10	14.711	164	49136	20.000	ng	0.00
64) Phenanthrene-d10	17.455	188	182255	20.000	ng	0.00
76) Chrysene-d12	21.732	240	217246	20.000	ng	# 0.00
86) Perylene-d12	24.993	264	191480	20.000	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.657	112	100469	159.820	ng	0.00
7) Phenol-d6	7.238	99	133414	156.671	ng	0.00
23) Nitrobenzene-d5	9.247	82	90652	117.110	ng	0.00
42) 2,4,6-Tribromophenol	16.198	330	143513	240.615	ng	0.00
45) 2-Fluorobiphenyl	13.342	172	283881	86.431	ng	0.00
79) Terphenyl-d14	20.064	244	1268042	116.120	ng	0.00

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG060222\
Data File : BG053802.D
Acq On : 2 Jun 2022 22:16
Operator : CG/JU
Sample : N3080-05RE
Misc :
ALS Vial : 17 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
WC-5RE

Quant Time: Jun 03 02:28:37 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG060222.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Jun 02 23:22:27 2022
Response via : Initial Calibration

