

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG060222\
 Data File : BG053794.D
 Acq On : 2 Jun 2022 16:26
 Operator : CG/JU
 Sample : SSTDICV040
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 ICVBG060222

Quant Time: Jun 02 23:26:36 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG060222.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Jun 02 23:22:27 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	20.000	20.000	0.0	109	0.00
2	1,4-Dioxane	40.000	39.437	1.4	110	0.00
3	Pyridine	40.000	43.819	-9.5	113	0.00
4	n-Nitrosodimethylamine	40.000	43.983	-10.0	110	0.00
5 S	2-Fluorophenol	80.000	86.257	-7.8	114	0.00
6	Aniline	40.000	41.121	-2.8	107	0.00
7 S	Phenol-d6	80.000	83.502	-4.4	108	0.00
8	2-Chlorophenol	40.000	43.403	-8.5	113	0.00
9	Benzaldehyde	40.000	35.649	10.9	109	0.00
10 C	Phenol	40.000	41.804	-4.5	111	0.00
11	bis(2-Chloroethyl)ether	40.000	40.859	-2.1	108	0.00
12	1,3-Dichlorobenzene	40.000	41.128	-2.8	109	0.00
13 C	1,4-Dichlorobenzene	40.000	41.134	-2.8	109	0.00
14	1,2-Dichlorobenzene	40.000	39.846	0.4	107	0.00
15	Benzyl Alcohol	40.000	42.972	-7.4	111	0.00
16	2,2'-oxybis(1-Chloropropane	40.000	40.804	-2.0	108	0.00
17	2-Methylphenol	40.000	41.966	-4.9	110	0.00
18	Hexachloroethane	40.000	43.311	-8.3	117	0.00
19 P	n-Nitroso-di-n-propylamine	40.000	40.576	-1.4	106	0.00
20	3+4-Methylphenols	40.000	43.038	-7.6	111	0.00
21 I	Naphthalene-d8	20.000	20.000	0.0	107	0.00
22	Acetophenone	40.000	41.479	-3.7	108	0.00
23 S	Nitrobenzene-d5	80.000	93.119	-16.4	122	0.00
24	Nitrobenzene	40.000	43.603	-9.0	121	0.00
25	Isophorone	40.000	41.781	-4.5	108	0.00
26 C	2-Nitrophenol	40.000	47.035	-17.6	127	0.00
27	2,4-Dimethylphenol	40.000	42.016	-5.0	108	0.00
28	bis(2-Chloroethoxy)methane	40.000	41.625	-4.1	108	0.00
29 C	2,4-Dichlorophenol	40.000	44.865	-12.2	110	0.00
30	1,2,4-Trichlorobenzene	40.000	41.630	-4.1	108	0.00
31	Naphthalene	40.000	40.627	-1.6	106	0.00
32	Benzoic acid	40.000	42.792	-7.0	120	0.00
33	4-Chloroaniline	40.000	41.559	-3.9	107	0.00
34 C	Hexachlorobutadiene	40.000	43.066	-7.7	112	0.00
35	Caprolactam	40.000	43.092	-7.7	110	0.00
36 C	4-Chloro-3-methylphenol	40.000	43.450	-8.6	108	0.00
37	2-Methylnaphthalene	40.000	40.708	-1.8	106	0.00
38	1-Methylnaphthalene	40.000	40.274	-0.7	106	0.00
39 I	Acenaphthene-d10	20.000	20.000	0.0	106	0.00
40	1,2,4,5-Tetrachlorobenzene	40.000	41.740	-4.4	108	0.00
41 P	Hexachlorocyclopentadiene	40.000	44.819	-12.0	113	0.00
42 S	2,4,6-Tribromophenol	80.000	88.804	-11.0	112	0.00
43 C	2,4,6-Trichlorophenol	40.000	45.767	-14.4	114	0.00
44	2,4,5-Trichlorophenol	40.000	46.878	-17.2	115	0.00
45 S	2-Fluorobiphenyl	80.000	81.650	-2.1	106	0.00
46	1,1'-Biphenyl	40.000	40.534	-1.3	106	0.00
47	2-Chloronaphthalene	40.000	40.702	-1.8	106	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG060222\
 Data File : BG053794.D
 Acq On : 2 Jun 2022 16:26
 Operator : CG/JU
 Sample : SSTDICV040
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 ICVBG060222

Quant Time: Jun 02 23:26:36 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG060222.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Jun 02 23:22:27 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
48	2-Nitroaniline	40.000	43.795	-9.5	130	0.00
49	Acenaphthylene	40.000	41.119	-2.8	106	0.00
50	Dimethylphthalate	40.000	42.512	-6.3	107	0.00
51	2,6-Dinitrotoluene	40.000	43.829	-9.6	122	0.00
52 C	Acenaphthene	40.000	41.018	-2.5	106	0.00
53	3-Nitroaniline	40.000	43.951	-9.9	121	0.00
54 P	2,4-Dinitrophenol	40.000	41.139	-2.8	111	0.00
55	Dibenzofuran	40.000	40.879	-2.2	105	0.00
56 P	4-Nitrophenol	40.000	43.801	-9.5	119	0.00
57	2,4-Dinitrotoluene	40.000	43.562	-8.9	125	0.00
58	Fluorene	40.000	40.640	-1.6	105	0.00
59	2,3,4,6-Tetrachlorophenol	40.000	43.378	-8.4	114	0.00
60	Diethylphthalate	40.000	42.428	-6.1	104	0.00
61	4-Chlorophenyl-phenylether	40.000	41.371	-3.4	108	0.00
62	4-Nitroaniline	40.000	42.930	-7.3	113	0.00
63	Azobenzene	40.000	40.011	-0.0	103	0.00
64 I	Phenanthrene-d10	20.000	20.000	0.0	104	0.00
65	4,6-Dinitro-2-methylphenol	40.000	41.984	-5.0	116	0.00
66 c	n-Nitrosodiphenylamine	40.000	40.798	-2.0	105	0.00
67	4-Bromophenyl-phenylether	40.000	40.185	-0.5	105	0.00
68	Hexachlorobenzene	40.000	40.919	-2.3	105	0.00
69	Atrazine	40.000	43.203	-8.0	106	0.00
70 C	Pentachlorophenol	40.000	41.675	-4.2	112	0.00
71	Phenanthrene	40.000	40.527	-1.3	104	0.00
72	Anthracene	40.000	40.602	-1.5	103	0.00
73	Carbazole	40.000	41.236	-3.1	105	0.00
74	Di-n-butylphthalate	40.000	42.984	-7.5	104	0.00
75 C	Fluoranthene	40.000	40.761	-1.9	104	0.00
76 I	Chrysene-d12	20.000	20.000	0.0	104	0.00
77	Benzydine	40.000	39.193	2.0	100	0.00
78	Pyrene	40.000	41.617	-4.0	103	0.00
79 S	Terphenyl-d14	80.000	82.544	-3.2	105	0.00
80	Butylbenzylphthalate	40.000	44.983	-12.5	108	0.00
81	Benzo(a)anthracene	40.000	41.080	-2.7	105	0.00
82	3,3'-Dichlorobenzidine	40.000	41.981	-5.0	103	0.00
83	Chrysene	40.000	40.090	-0.2	103	0.00
84	Bis(2-ethylhexyl)phthalate	40.000	44.094	-10.2	108	0.00
85 c	Di-n-octyl phthalate	40.000	44.462	-11.2	107	0.00
86 I	Perylene-d12	20.000	20.000	0.0	105	0.00
87	Indeno(1,2,3-cd)pyrene	40.000	40.772	-1.9	104	0.00
88	Benzo(b)fluoranthene	40.000	41.279	-3.2	104	0.00
89	Benzo(k)fluoranthene	40.000	40.450	-1.1	105	0.00
90 C	Benzo(a)pyrene	40.000	40.752	-1.9	103	0.00
91	Dibenzo(a,h)anthracene	40.000	40.561	-1.4	103	0.00
92	Benzo(g,h,i)perylene	40.000	40.724	-1.8	104	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG060222\
Data File : BG053794.D
Acq On : 2 Jun 2022 16:26
Operator : CG/JU
Sample : SSTDICV040
Misc :
ALS Vial : 9 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
ICVBG060222

Quant Time: Jun 02 23:26:36 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG060222.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Jun 02 23:22:27 2022
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)
----------	--------	-------	------	-------	----------

(#) = Out of Range

SPCC's out = 0 CCC's out = 0