

Data Path : Z:\HPCHEM1\BNA_G\DATA\BG060315\
 Data File : BG017434.D
 Acq On : 4 Jun 2015 8:35
 Operator : TP/IZ
 Sample : G2429-01
 Misc :
 ALS Vial : 29 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 SEC5.1(0-4)

Quant Time: Jun 04 17:15:19 2015
 Quant Method : Z:\HPCHEM1\BNA_G\METHODS\8270-BG060315.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Jun 04 01:31:15 2015
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.62	152	15718	20.00	ng	0.00
21) Naphthalene-d8	10.42	136	64454	20.00	ng	0.00
38) Acenaphthene-d10	14.29	164	44274	20.00	ng	0.00
63) Phenanthrene-d10	17.04	188	113097	20.00	ng	0.00
75) Chrysene-d12	21.23	240	142395	20.00	ng	0.00
86) Perylene-d12	23.46	264	141287	20.00	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.22	112	108967	122.14	ng	0.00
7) Phenol-d6	6.83	99	147005	120.95	ng	0.00
23) Nitrobenzene-d5	8.79	82	95041	73.61	ng	0.00
41) 2,4,6-Tribromophenol	15.78	330	65239	106.09	ng	-0.01
44) 2-Fluorobiphenyl	12.91	172	181977	62.00	ng	0.00
78) Terphenyl-d14	19.68	244	452649	65.70	ng	0.00
Target Compounds						
49) Dimethylphthalate	13.75	163	24203	6.28	ng	Qvalue 98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA_G\DATA\BG060315\
Data File : BG017434.D
Acq On : 4 Jun 2015 8:35
Operator : TP/IZ
Sample : G2429-01
Misc :
ALS Vial : 29 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
SEC5.1(0-4)

Quant Time: Jun 04 17:15:19 2015
Quant Method : Z:\HPCHEM1\BNA_G\METHODS\8270-BG060315.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Jun 04 01:31:15 2015
Response via : Initial Calibration

