


Hexachlorocyclopentadiene

Response Ratio




Resp Ratio = $3.93 \times 10^{-1} \times \text{Amt} - 1.34 \times 10^{-1}$
Coef of Det (r^2) = 0.996 Curve Fit: Linear

Method Name: Z:\HPCHEM1\BNA_G\METHODS\8270-BG060315.M
Calibration Table Last Updated: Thu Jun 04 01:26:14 2015

2,4-Dinitrophenol

Response Ratio




Resp Ratio = $2.69 \times 10^{-1} \times \text{Amt} - 9.45 \times 10^{-2}$
Coef of Det (r^2) = 0.998 Curve Fit: Linear

Method Name: Z:\HPCHEM1\BNA_G\METHODS\8270-BG060315.M
Calibration Table Last Updated: Thu Jun 04 01:30:07 2015


4-Nitrophenol

Response Ratio



Resp Ratio = $3.41 \times 10^{-1} \times \text{Amt} - 1.36 \times 10^{-2}$
Coef of Det (r^2) = 0.997 Curve Fit: Linear

Method Name: Z:\HPCHEM1\BNA_G\METHODS\8270-BG060315.M
Calibration Table Last Updated: Thu Jun 04 01:30:31 2015



Method Name: Z:\HPCHEM1\BNA_G\METHODS\8270-BG060315.M
Calibration Table Last Updated: Thu Jun 04 01:30:47 2015