

Data Path : Z:\HPCHEM1\BNA_G\DATA\BG060315\
 Data File : BG017411.D
 Acq On : 3 Jun 2015 17:32
 Operator : TP/IZ
 Sample : SSTDICV040
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 ICVBG060315

Quant Time: Jun 04 01:37:22 2015
 Quant Method : Z:\HPCHEM1\BNA_G\METHODS\8270-BG060315.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Jun 04 01:31:15 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	20.000	20.000	0.0	93	0.00
2	1,4-Dioxane	40.000	37.121	7.2	91	0.00
3	Pyridine	40.000	42.787	-7.0	96	0.00
4	n-Nitrosodimethylamine	40.000	43.365	-8.4	96	0.00
5 S	2-Fluorophenol	80.000	82.287	-2.9	95	0.00
6	Aniline	40.000	42.487	-6.2	98	0.00
7 S	Phenol-d6	80.000	83.801	-4.8	95	0.00
8	2-Chlorophenol	40.000	41.092	-2.7	96	0.00
9	Benzaldehyde	40.000	40.743	-1.9	90	0.00
10 C	Phenol	40.000	43.788	-9.5	98	0.00
11	bis(2-Chloroethyl)ether	40.000	42.890	-7.2	98	0.00
12	1,3-Dichlorobenzene	40.000	41.782	-4.5	98	0.00
13 C	1,4-Dichlorobenzene	40.000	40.763	-1.9	95	0.00
14	1,2-Dichlorobenzene	40.000	41.298	-3.2	96	0.00
15	Benzyl Alcohol	40.000	42.623	-6.6	95	0.00
16	2,2'-oxybis(1-Chloropropane	40.000	41.541	-3.9	95	0.00
17	2-Methylphenol	40.000	42.324	-5.8	96	0.00
18	Hexachloroethane	40.000	42.610	-6.5	97	0.00
19 P	n-Nitroso-di-n-propylamine	40.000	41.383	-3.5	95	0.00
20	3+4-Methylphenols	40.000	41.293	-3.2	93	0.00
21 I	Naphthalene-d8	20.000	20.000	0.0	98	0.00
22	Acetophenone	40.000	40.610	-1.5	97	0.00
23 S	Nitrobenzene-d5	80.000	80.088	-0.1	94	0.00
24	Nitrobenzene	40.000	40.223	-0.6	96	0.00
25	Isophorone	40.000	39.036	2.4	95	0.00
26 C	2-Nitrophenol	40.000	40.147	-0.4	93	0.00
27	2,4-Dimethylphenol	40.000	39.827	0.4	95	0.00
28	bis(2-Chloroethoxy)methane	40.000	39.152	2.1	91	0.00
29 C	2,4-Dichlorophenol	40.000	40.884	-2.2	96	0.00
30	1,2,4-Trichlorobenzene	40.000	38.745	3.1	94	0.00
31	Naphthalene	40.000	39.362	1.6	94	0.00
32	Benzoic acid	40.000	37.736	5.7	88	0.00
33	4-Chloroaniline	40.000	39.836	0.4	95	0.00
34 C	Hexachlorobutadiene	40.000	40.123	-0.3	96	0.00
35	Caprolactam	40.000	40.525	-1.3	91	-0.01
36 C	4-Chloro-3-methylphenol	40.000	38.972	2.6	91	0.00
37	2-Methylnaphthalene	40.000	39.352	1.6	93	0.00
38 I	Acenaphthene-d10	20.000	20.000	0.0	95	0.00
39	1,2,4,5-Tetrachlorobenzene	40.000	40.323	-0.8	97	0.00
40 P	Hexachlorocyclopentadiene	40.000	38.145	4.6	90	0.00
41 S	2,4,6-Tribromophenol	80.000	81.240	-1.5	93	0.00
42 C	2,4,6-Trichlorophenol	40.000	40.436	-1.1	89	0.00
43	2,4,5-Trichlorophenol	40.000	39.294	1.8	93	0.00
44 S	2-Fluorobiphenyl	80.000	81.656	-2.1	96	0.00

Data Path : Z:\HPCHEM1\BNA_G\DATA\BG060315\
 Data File : BG017411.D
 Acq On : 3 Jun 2015 17:32
 Operator : TP/IZ
 Sample : SSTDICV040
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 ICSVG060315

Quant Time: Jun 04 01:37:22 2015
 Quant Method : Z:\HPCHEM1\BNA_G\METHODS\8270-BG060315.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Jun 04 01:31:15 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
45	1,1'-Biphenyl	40.000	40.994	-2.5	95	0.00
46	2-Chloronaphthalene	40.000	40.638	-1.6	92	0.00
47	2-Nitroaniline	40.000	41.943	-4.9	95	0.00
48	Acenaphthylene	40.000	40.894	-2.2	94	0.00
49	Dimethylphthalate	40.000	40.157	-0.4	92	0.00
50	2,6-Dinitrotoluene	40.000	40.417	-1.0	94	0.00
51 C	Acenaphthene	40.000	40.036	-0.1	94	0.00
52	3-Nitroaniline	40.000	40.475	-1.2	93	0.00
53 P	2,4-Dinitrophenol	40.000	39.639	0.9	95	0.00
54	Dibenzofuran	40.000	40.148	-0.4	92	0.00
55 P	4-Nitrophenol	40.000	40.108	-0.3	91	0.00
56	2,4-Dinitrotoluene	40.000	41.705	-4.3	92	0.00
57	Fluorene	40.000	40.152	-0.4	92	0.00
58	2,3,4,6-Tetrachlorophenol	40.000	43.024	-7.6	95	0.00
59	Diethylphthalate	40.000	39.699	0.8	92	0.00
60	4-Chlorophenyl-phenylether	40.000	40.345	-0.9	94	0.00
61	4-Nitroaniline	40.000	44.023	-10.1	96	0.00
62	Azobenzene	40.000	41.094	-2.7	94	0.00
63 I	Phenanthrene-d10	20.000	20.000	0.0	90	0.00
64	4,6-Dinitro-2-methylphenol	40.000	41.661	-4.2	93	0.00
65 c	n-Nitrosodiphenylamine	40.000	41.390	-3.5	92	0.00
66	4-Bromophenyl-phenylether	40.000	40.886	-2.2	94	0.00
67	Hexachlorobenzene	40.000	41.862	-4.7	95	0.00
68	Atrazine	40.000	43.106	-7.8	94	0.00
69 C	Pentachlorophenol	40.000	38.363	4.1	87	0.00
70	Phenanthrene	40.000	41.024	-2.6	93	0.00
71	Anthracene	40.000	41.286	-3.2	94	0.00
72	Carbazole	40.000	41.819	-4.5	95	0.00
73	Di-n-butylphthalate	40.000	42.197	-5.5	95	0.00
74 C	Fluoranthene	40.000	42.254	-5.6	96	0.00
75 I	Chrysene-d12	20.000	20.000	0.0	98	0.00
76	Benzidine	40.000	42.628	-6.6	97	0.00
77	Pyrene	40.000	40.128	-0.3	96	0.00
78 S	Terphenyl-d14	80.000	81.253	-1.6	98	0.00
79	Butylbenzylphthalate	40.000	41.200	-3.0	97	0.00
80	Benzo(a)anthracene	40.000	40.778	-1.9	98	0.00
81	3,3'-Dichlorobenzidine	40.000	42.254	-5.6	100	0.00
82	Chrysene	40.000	41.054	-2.6	98	0.00
83	Bis(2-ethylhexyl)phthalate	40.000	40.599	-1.5	97	0.00
84 c	Di-n-octyl phthalate	40.000	41.586	-4.0	101	0.00
85	Indeno(1,2,3-cd)pyrene	40.000	40.814	-2.0	97	0.00
86 I	Perylene-d12	20.000	20.000	0.0	98	0.00
87	Benzo(b)fluoranthene	40.000	40.164	-0.4	95	0.00

Data Path : Z:\HPCHEM1\BNA_G\DATA\BG060315\
Data File : BG017411.D
Acq On : 3 Jun 2015 17:32
Operator : TP/IZ
Sample : SSTDICV040
Misc :
ALS Vial : 9 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
ICVBG060315

Quant Time: Jun 04 01:37:22 2015
Quant Method : Z:\HPCHEM1\BNA_G\METHODS\8270-BG060315.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Jun 04 01:31:15 2015
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
88	Benzo(k)fluoranthene	40.000	41.583	-4.0	103	0.00
89 C	Benzo(a)pyrene	40.000	41.053	-2.6	99	0.00
90	Dibenzo(a,h)anthracene	40.000	40.945	-2.4	98	0.00
91	Benzo(g,h,i)perylene	40.000	40.333	-0.8	98	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0