

Data Path : Z:\HPCHEM1\BNA_G\DATA\BG060315\
 Data File : BG017420.D
 Acq On : 3 Jun 2015 23:20
 Operator : TP/IZ
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 BNA_G
 LabSampleId :
 SSTDCCC040

Quant Time: Jun 04 01:54:59 2015
 Quant Method : Z:\HPCHEM1\BNA_G\METHODS\8270-BG060315.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Jun 04 01:31:15 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	20.000	20.000	0.0	99	0.00
2	1,4-Dioxane	40.000	35.166	12.1	91	0.00
3	Pyridine	40.000	39.752	0.6	94	0.00
4	n-Nitrosodimethylamine	40.000	40.092	-0.2	94	0.00
5 S	2-Fluorophenol	80.000	80.207	-0.3	98	0.00
6	Aniline	40.000	38.516	3.7	94	0.00
7 S	Phenol-d6	80.000	78.827	1.5	95	0.00
8	2-Chlorophenol	40.000	39.029	2.4	97	0.00
9	Benzaldehyde	40.000	39.522	1.2	93	0.00
10 C	Phenol	40.000	41.238	-3.1	97	0.00
11	bis(2-Chloroethyl)ether	40.000	41.350	-3.4	100	0.00
12	1,3-Dichlorobenzene	40.000	37.493	6.3	93	0.00
13 C	1,4-Dichlorobenzene	40.000	38.028	4.9	94	0.00
14	1,2-Dichlorobenzene	40.000	38.253	4.4	94	0.00
15	Benzyl Alcohol	40.000	39.869	0.3	94	0.00
16	2,2'-oxybis(1-Chloropropane)	40.000	39.963	0.1	97	0.00
17	2-Methylphenol	40.000	39.998	0.0	96	0.00
18	Hexachloroethane	40.000	38.934	2.7	94	0.00
19 P	n-Nitroso-di-n-propylamine	40.000	38.729	3.2	94	0.00
20	3+4-Methylphenols	40.000	40.362	-0.9	96	0.00
21 I	Naphthalene-d8	20.000	20.000	0.0	98	0.00
22	Acetophenone	40.000	40.098	-0.2	96	0.00
23 S	Nitrobenzene-d5	80.000	81.074	-1.3	96	0.00
24	Nitrobenzene	40.000	40.376	-0.9	96	0.00
25	Isophorone	40.000	39.230	1.9	96	0.00
26 C	2-Nitrophenol	40.000	38.959	2.6	91	0.00
27	2,4-Dimethylphenol	40.000	39.432	1.4	94	0.00
28	bis(2-Chloroethoxy)methane	40.000	40.281	-0.7	94	0.00
29 C	2,4-Dichlorophenol	40.000	40.325	-0.8	95	0.00
30	1,2,4-Trichlorobenzene	40.000	39.979	0.1	97	0.00
31	Naphthalene	40.000	39.765	0.6	96	0.00
32	Benzoic acid	40.000	35.594	11.0	83	0.00
33	4-Chloroaniline	40.000	39.565	1.1	94	0.00
34 C	Hexachlorobutadiene	40.000	39.489	1.3	95	0.00
35	Caprolactam	40.000	38.452	3.9	87	0.00
36 C	4-Chloro-3-methylphenol	40.000	39.252	1.9	92	0.00
37	2-Methylnaphthalene	40.000	39.539	1.2	94	0.00
38 I	Acenaphthene-d10	20.000	20.000	0.0	98	0.00
39	1,2,4,5-Tetrachlorobenzene	40.000	39.340	1.6	98	0.00
40 P	Hexachlorocyclopentadiene	40.000	38.050	4.9	93	0.00
41 S	2,4,6-Tribromophenol	80.000	76.610	4.2	91	0.00
42 C	2,4,6-Trichlorophenol	40.000	38.041	4.9	87	0.00
43	2,4,5-Trichlorophenol	40.000	37.560	6.1	92	0.00
44 S	2-Fluorobiphenyl	80.000	79.009	1.2	96	0.00

Data Path : Z:\HPCHEM1\BNA_G\DATA\BG060315\
 Data File : BG017420.D
 Acq On : 3 Jun 2015 23:20
 Operator : TP/IZ
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 BNA_G
 LabSampleId :
 SSTDCCC040

Quant Time: Jun 04 01:54:59 2015
 Quant Method : Z:\HPCHEM1\BNA_G\METHODS\8270-BG060315.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Jun 04 01:31:15 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
45	1,1'-Biphenyl	40.000	40.074	-0.2	96	0.00
46	2-Chloronaphthalene	40.000	40.244	-0.6	94	0.00
47	2-Nitroaniline	40.000	39.963	0.1	94	0.00
48	Acenaphthylene	40.000	39.441	1.4	94	0.00
49	Dimethylphthalate	40.000	38.980	2.6	93	0.00
50	2,6-Dinitrotoluene	40.000	39.512	1.2	95	0.00
51 C	Acenaphthene	40.000	38.990	2.5	95	0.00
52	3-Nitroaniline	40.000	38.017	5.0	91	0.00
53 P	2,4-Dinitrophenol	40.000	37.345	6.6	92	0.00
54	Dibenzofuran	40.000	39.853	0.4	95	0.00
55 P	4-Nitrophenol	40.000	38.640	3.4	91	0.00
56	2,4-Dinitrotoluene	40.000	40.128	-0.3	92	0.00
57	Fluorene	40.000	40.107	-0.3	96	0.00
58	2,3,4,6-Tetrachlorophenol	40.000	40.108	-0.3	92	0.00
59	Diethylphthalate	40.000	38.679	3.3	93	0.00
60	4-Chlorophenyl-phenylether	40.000	39.809	0.5	96	0.00
61	4-Nitroaniline	40.000	41.871	-4.7	95	0.00
62	Azobenzene	40.000	40.443	-1.1	96	0.00
63 I	Phenanthrene-d10	20.000	20.000	0.0	93	0.00
64	4,6-Dinitro-2-methylphenol	40.000	41.264	-3.2	95	0.00
65 c	n-Nitrosodiphenylamine	40.000	40.804	-2.0	93	0.00
66	4-Bromophenyl-phenylether	40.000	40.037	-0.1	95	0.00
67	Hexachlorobenzene	40.000	39.683	0.8	93	0.00
68	Atrazine	40.000	41.052	-2.6	93	0.00
69 C	Pentachlorophenol	40.000	38.956	2.6	92	0.00
70	Phenanthrene	40.000	40.182	-0.5	95	0.00
71	Anthracene	40.000	40.755	-1.9	96	0.00
72	Carbazole	40.000	40.868	-2.2	96	0.00
73	Di-n-butylphthalate	40.000	40.847	-2.1	95	0.00
74 C	Fluoranthene	40.000	40.489	-1.2	95	0.00
75 I	Chrysene-d12	20.000	20.000	0.0	98	0.00
76	Benzidine	40.000	39.747	0.6	90	0.00
77	Pyrene	40.000	40.045	-0.1	96	0.00
78 S	Terphenyl-d14	80.000	80.805	-1.0	97	0.00
79	Butylbenzylphthalate	40.000	39.800	0.5	94	0.00
80	Benzo(a)anthracene	40.000	40.080	-0.2	96	0.00
81	3,3'-Dichlorobenzidine	40.000	40.111	-0.3	95	0.00
82	Chrysene	40.000	39.976	0.1	96	0.00
83	Bis(2-ethylhexyl)phthalate	40.000	40.307	-0.8	96	0.00
84 c	Di-n-octyl phthalate	40.000	40.457	-1.1	98	0.00
85	Indeno(1,2,3-cd)pyrene	40.000	40.688	-1.7	97	0.00
86 I	Perylene-d12	20.000	20.000	0.0	98	0.00
87	Benzo(b)fluoranthene	40.000	41.544	-3.9	98	0.00

Data Path : Z:\HPCHEM1\BNA_G\DATA\BG060315\
Data File : BG017420.D
Acq On : 3 Jun 2015 23:20
Operator : TP/IZ
Sample : SSTDCCC040
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
BNA_G
LabSampleId :
SSTDCCC040

Quant Time: Jun 04 01:54:59 2015
Quant Method : Z:\HPCHEM1\BNA_G\METHODS\8270-BG060315.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Jun 04 01:31:15 2015
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
88	Benzo(k)fluoranthene	40.000	38.876	2.8	97	0.00
89 C	Benzo(a)pyrene	40.000	40.050	-0.1	96	0.00
90	Dibenzo(a,h)anthracene	40.000	39.965	0.1	96	0.00
91	Benzo(g,h,i)perylene	40.000	39.787	0.5	97	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0