

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG060424\
 Data File : BG061486.D
 Acq On : 5 Jun 2024 14:30
 Operator : MA/JU
 Sample : PB161224BL
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 SBLK224

Quant Time: Jun 05 23:41:11 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\SFAM-EPA-BG052124.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu May 30 22:54:48 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.877	152	29769	20.000	ng/u1	0.00
20) Naphthalene-d8	10.668	136	107984	20.000	ng/u1	# 0.00
38) Acenaphthene-d10	14.511	164	77632	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.261	188	206454	20.000	ng/u1	# 0.00
79) Chrysene-d12	21.520	240	219465	20.000	ng/u1	0.00
88) Perylene-d12	24.617	264	234669	20.000	ng/u1	0.02

System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.330	96	2398	4.508	ng/uL	0.00
4) Pyridine-d5	3.735	84	40901	21.613	ng/u1	0.00
7) Phenol-d5	7.067	99	56013	22.912	ng/u1	0.01
9) Bis-(2-Chloroethyl)eth...	7.202	67	35665	23.211	ng/u1	0.00
11) 2-Chlorophenol-d4	7.413	132	47714	26.372	ng/u1	0.00
15) 4-Methylphenol-d8	8.600	113	52057	24.973	ng/u1	0.00
21) Nitrobenzene-d5	9.029	128	25205	30.316	ng/u1	0.00
24) 2-Nitrophenol-d4	9.752	143	25631	26.342	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.310	165	48564	25.601	ng/u1	0.00
31) 4-Chloroaniline-d4	10.809	131	62754	25.287	ng/u1	0.00
46) Dimethylphthalate-d6	13.917	166	157448	26.150	ng/u1	0.00
49) Acenaphthylene-d8	14.205	160	175083	27.970	ng/u1	0.00
54) 4-Nitrophenol-d4	14.758	143	15387	20.671	ng/u1	0.02
60) Fluorene-d10	15.510	176	139295	26.796	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.657	200	18352	14.697	ng/u1	0.02
73) Anthracene-d10	17.360	188	265776	29.200	ng/u1	0.00
81) Pyrene-d10	19.658	212	301296	26.738	ng/u1	0.00
92) Benzo(a)pyrene-d12	24.405	264	342459	30.355	ng/u1	0.02

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG060424\
Data File : BG061486.D
Acq On : 5 Jun 2024 14:30
Operator : MA/JU
Sample : PB161224BL
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
SBLK224

Quant Time: Jun 05 23:41:11 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\SFAM-EPA-BG052124.MA.M
Quant Title : SVOA CALIBRATION
QLast Update : Thu May 30 22:54:48 2024
Response via : Initial Calibration

