

Data Path : Z:\HPCHEM1\BNA_G\DATA\BG060615\
 Data File : BG017534.D
 Acq On : 7 Jun 2015 14:34
 Operator : TP/IZ
 Sample : G2457-05 10X
 Misc :
 ALS Vial : 29 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 SP-3

Manual Integrations
 APPROVED

apatel
 6/8/2015 6:01:10 PM

Quant Time: Jun 08 08:21:04 2015
 Quant Method : Z:\HPCHEM1\BNA_G\METHODS\8270-BG060315.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Jun 04 01:31:15 2015
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.62	152	24394	20.00	ng	0.00
21) Naphthalene-d8	10.42	136	94631	20.00	ng	0.00
38) Acenaphthene-d10	14.29	164	56702	20.00	ng	0.00
63) Phenanthrene-d10	17.05	188	116862	20.00	ng	0.00
75) Chrysene-d12	21.24	240	132387	20.00	ng	0.00
86) Perylene-d12	23.48	264	133571	20.00	ng	0.01
System Monitoring Compounds						
5) 2-Fluorophenol	5.23	112	14361	10.37	ng	0.00
7) Phenol-d6	6.84	99	16868	8.94	ng	0.00
23) Nitrobenzene-d5	8.79	82	12528	6.61	ng	0.00
41) 2,4,6-Tribromophenol	15.79	330	6250	7.94	ng	0.00
44) 2-Fluorobiphenyl	12.91	172	25905	6.89	ng	0.00
78) Terphenyl-d14	19.68	244	33212	5.19	ng	0.00
Target Compounds						Qvalue
70) Phenanthrene	17.09	178	15490	2.38	ng	95
74) Fluoranthene	19.11	202	28325	3.32	ng	92
77) Pyrene	19.47	202	29955	3.67	ng	95
80) Benzo(a)anthracene	21.23	228	16764	2.06	ng	# 87
83) Bis(2-ethylhexyl)phthalate	21.17	149	1599288	314.24	ng	# 81
87) Benzo(b)fluoranthene	22.81	252	19704m	2.33	ng	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA_G\DATA\BG060615\
Data File : BG017534.D
Acq On : 7 Jun 2015 14:34
Operator : TP/IZ
Sample : G2457-05 10X
Misc :
ALS Vial : 29 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampled :
SP-3

Manual Integrations
APPROVED

apatel
6/8/2015 6:01:10 PM

Quant Time: Jun 08 08:21:04 2015
Quant Method : Z:\HPCHEM1\BNA_G\METHODS\8270-BG060315.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Jun 04 01:31:15 2015
Response via : Initial Calibration

