

Data Path : Z:\HPCHEM1\BNA_G\Data\BG060615\
 Data File : BG017521.D
 Acq On : 7 Jun 2015 7:03
 Operator : TP/IZ
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_G
 LabSampleId :
 SSTDCCC040

Quant Time: Jun 08 09:08:27 2015
 Quant Method : Z:\HPCHEM1\BNA_G\METHODS\8270-BG060315.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Jun 04 01:31:15 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	20.000	20.000	0.0	121	0.00
2	1,4-Dioxane	40.000	32.787	18.0	104	-0.02
3	Pyridine	40.000	37.305	6.7	108	-0.02
4	n-Nitrosodimethylamine	40.000	46.827	-17.1	135	-0.02
5 S	2-Fluorophenol	80.000	76.914	3.9	115	0.00
6	Aniline	40.000	38.694	3.3	116	0.00
7 S	Phenol-d6	80.000	77.684	2.9	114	0.00
8	2-Chlorophenol	40.000	38.845	2.9	118	0.00
9	Benzaldehyde	40.000	38.506	3.7	111	0.00
10 C	Phenol	40.000	39.731	0.7	115	0.00
11	bis(2-Chloroethyl)ether	40.000	40.516	-1.3	120	0.00
12	1,3-Dichlorobenzene	40.000	39.026	2.4	119	0.00
13 C	1,4-Dichlorobenzene	40.000	37.690	5.8	114	0.00
14	1,2-Dichlorobenzene	40.000	39.802	0.5	120	0.00
15	Benzyl Alcohol	40.000	39.501	1.2	114	0.00
16	2,2'-oxybis(1-Chloropropane	40.000	39.620	1.0	118	0.00
17	2-Methylphenol	40.000	39.815	0.5	118	0.00
18	Hexachloroethane	40.000	41.320	-3.3	122	0.00
19 P	n-Nitroso-di-n-propylamine	40.000	39.313	1.7	117	0.00
20	3+4-Methylphenols	40.000	39.449	1.4	115	0.00
21 I	Naphthalene-d8	20.000	20.000	0.0	118	0.00
22	Acetophenone	40.000	40.292	-0.7	116	0.00
23 S	Nitrobenzene-d5	80.000	85.437	-6.8	122	0.00
24	Nitrobenzene	40.000	42.212	-5.5	122	0.00
25	Isophorone	40.000	38.868	2.8	114	0.00
26 C	2-Nitrophenol	40.000	41.478	-3.7	117	0.00
27	2,4-Dimethylphenol	40.000	39.219	2.0	113	0.00
28	bis(2-Chloroethoxy)methane	40.000	41.713	-4.3	118	0.00
29 C	2,4-Dichlorophenol	40.000	41.794	-4.5	119	0.00
30	1,2,4-Trichlorobenzene	40.000	40.074	-0.2	118	0.00
31	Naphthalene	40.000	41.023	-2.6	119	0.00
32	Benzoic acid	40.000	39.053	2.4	110	0.03
33	4-Chloroaniline	40.000	39.532	1.2	114	0.00
34 C	Hexachlorobutadiene	40.000	43.476	-8.7	126	0.00
35	Caprolactam	40.000	34.525	13.7	94	0.00
36 C	4-Chloro-3-methylphenol	40.000	39.902	0.2	112	0.00
37	2-Methylnaphthalene	40.000	40.435	-1.1	116	0.00
38 I	Acenaphthene-d10	20.000	20.000	0.0	119	0.00
39	1,2,4,5-Tetrachlorobenzene	40.000	41.111	-2.8	124	0.00
40 P	Hexachlorocyclopentadiene	40.000	39.014	2.5	116	0.00
41 S	2,4,6-Tribromophenol	80.000	75.274	5.9	108	0.00
42 C	2,4,6-Trichlorophenol	40.000	39.414	1.5	109	0.00
43	2,4,5-Trichlorophenol	40.000	39.688	0.8	117	0.00
44 S	2-Fluorobiphenyl	80.000	84.525	-5.7	124	0.00

Data Path : Z:\HPCHEM1\BNA_G\Data\BG060615\
 Data File : BG017521.D
 Acq On : 7 Jun 2015 7:03
 Operator : TP/IZ
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_G
 LabSampleId :
 SSTDCCC040

Quant Time: Jun 08 09:08:27 2015
 Quant Method : Z:\HPCHEM1\BNA_G\METHODS\8270-BG060315.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Jun 04 01:31:15 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
45	1,1'-Biphenyl	40.000	41.816	-4.5	121	0.00
46	2-Chloronaphthalene	40.000	41.024	-2.6	116	0.00
47	2-Nitroaniline	40.000	39.267	1.8	111	0.00
48	Acenaphthylene	40.000	40.116	-0.3	115	0.00
49	Dimethylphthalate	40.000	37.729	5.7	108	0.00
50	2,6-Dinitrotoluene	40.000	36.885	7.8	107	0.00
51 C	Acenaphthene	40.000	40.194	-0.5	118	0.00
52	3-Nitroaniline	40.000	36.328	9.2	105	0.00
53 P	2,4-Dinitrophenol	40.000	28.633	28.4#	79	0.00
54	Dibenzofuran	40.000	39.366	1.6	113	0.00
55 P	4-Nitrophenol	40.000	33.213	17.0	94	0.02
56	2,4-Dinitrotoluene	40.000	37.486	6.3	104	0.00
57	Fluorene	40.000	40.335	-0.8	116	0.00
58	2,3,4,6-Tetrachlorophenol	40.000	39.743	0.6	110	0.00
59	Diethylphthalate	40.000	37.243	6.9	108	0.00
60	4-Chlorophenyl-phenylether	40.000	41.334	-3.3	120	0.00
61	4-Nitroaniline	40.000	33.756	15.6	92	0.00
62	Azobenzene	40.000	41.897	-4.7	120	0.00
63 I	Phenanthrene-d10	20.000	20.000	0.0	107	0.00
64	4,6-Dinitro-2-methylphenol	40.000	36.260	9.4	95	0.00
65 c	n-Nitrosodiphenylamine	40.000	41.727	-4.3	109	0.00
66	4-Bromophenyl-phenylether	40.000	41.973	-4.9	114	0.00
67	Hexachlorobenzene	40.000	41.618	-4.0	112	0.00
68	Atrazine	40.000	39.390	1.5	102	0.00
69 C	Pentachlorophenol	40.000	35.549	11.1	94	0.00
70	Phenanthrene	40.000	40.225	-0.6	108	0.00
71	Anthracene	40.000	40.541	-1.4	109	0.00
72	Carbazole	40.000	37.879	5.3	102	0.00
73	Di-n-butylphthalate	40.000	39.112	2.2	103	0.00
74 C	Fluoranthene	40.000	39.274	1.8	106	0.00
75 I	Chrysene-d12	20.000	20.000	0.0	107	0.00
76	Benzidine	40.000	39.894	0.3	99	0.00
77	Pyrene	40.000	39.476	1.3	103	0.00
78 S	Terphenyl-d14	80.000	82.013	-2.5	108	0.00
79	Butylbenzylphthalate	40.000	41.051	-2.6	106	0.00
80	Benzo(a)anthracene	40.000	40.421	-1.1	106	0.00
81	3,3'-Dichlorobenzidine	40.000	39.242	1.9	101	0.00
82	Chrysene	40.000	40.449	-1.1	106	0.00
83	Bis(2-ethylhexyl)phthalate	40.000	40.311	-0.8	105	0.00
84 c	Di-n-octyl phthalate	40.000	39.395	1.5	104	0.00
85	Indeno(1,2,3-cd)pyrene	40.000	39.427	1.4	103	0.01
86 I	Perylene-d12	20.000	20.000	0.0	103	0.00
87	Benzo(b)fluoranthene	40.000	40.607	-1.5	101	0.00

Data Path : Z:\HPCHEM1\BNA_G\Data\BG060615\
Data File : BG017521.D
Acq On : 7 Jun 2015 7:03
Operator : TP/IZ
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_G
LabSampleId :
SSTDCCC040

Quant Time: Jun 08 09:08:27 2015
Quant Method : Z:\HPCHEM1\BNA_G\METHODS\8270-BG060315.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Jun 04 01:31:15 2015
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
88	Benzo(k)fluoranthene	40.000	40.840	-2.1	107	0.00
89 C	Benzo(a)pyrene	40.000	40.628	-1.6	103	0.00
90	Dibenzo(a,h)anthracene	40.000	40.979	-2.4	103	0.00
91	Benzo(g,h,i)perylene	40.000	40.572	-1.4	104	0.01

(#) = Out of Range

SPCC's out = 0 CCC's out = 0