

Data Path : Z:\HPCHEM1\BNA_G\DATA\BG060616\
 Data File : BG022174.D
 Acq On : 6 Jun 2016 18:23
 Operator : UM/SJ
 Sample : PB91111BS
 Misc :
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 PB91111BS

Quant Time: Jun 07 03:19:05 2016
 Quant Method : Z:\HPCHEM1\BNA_G\METHODS\8270-BG060616.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Jun 06 16:12:45 2016
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	8.12	152	28607	20.00	ng	0.00
21) Naphthalene-d8	10.94	136	145643	20.00	ng	0.00
38) Acenaphthene-d10	14.75	164	89904	20.00	ng	0.00
63) Phenanthrene-d10	17.49	188	210576	20.00	ng	0.00
75) Chrysene-d12	21.77	240	202626	20.00	ng	0.00
86) Perylene-d12	25.06	264	194264	20.00	ng	0.00

System Monitoring Compounds

5) 2-Fluorophenol	5.67	112	218204	147.48	ng	0.00
7) Phenol-d6	7.28	99	329161	141.50	ng	0.00
23) Nitrobenzene-d5	9.29	82	168964	80.88	ng	0.00
41) 2,4,6-Tribromophenol	16.23	330	129794	127.77	ng	0.00
44) 2-Fluorobiphenyl	13.37	172	477765	80.99	ng	0.00
78) Terphenyl-d14	20.08	244	700082	82.03	ng	0.00

Target Compounds

						Qvalue
2) 1,4-Dioxane	3.51	88	22170	31.25	ng	# 81
3) Pyridine	3.93	79	63871	33.32	ng	98
4) n-Nitrosodimethylamine	3.84	42	17840	41.62	ng	98
6) Aniline	7.44	93	88395	29.91	ng	# 79
8) 2-Chlorophenol	7.68	128	92656	45.31	ng	# 78
9) Benzaldehyde	7.27	77	5711	4.46	ng	# 88
10) Phenol	7.30	94	113524	47.33	ng	81
11) bis(2-Chloroethyl)ether	7.53	93	76677	40.10	ng	# 73
12) 1,3-Dichlorobenzene	8.01	146	86992	39.68	ng	# 89
13) 1,4-Dichlorobenzene	8.16	146	87257	39.95	ng	90
14) 1,2-Dichlorobenzene	8.47	146	87337	39.67	ng	93
15) Benzyl Alcohol	8.36	79	60956	40.56	ng	89
16) 2,2'-oxybis(1-Chloropropan	8.65	45	77253	38.94	ng	81
17) 2-Methylphenol	8.57	107	77842	43.49	ng	# 84
18) Hexachloroethane	9.21	117	31800	39.87	ng	# 81
19) n-Nitroso-di-n-propylamine	8.92	70	60227	38.25	ng	# 71
20) 3+4-Methylphenols	8.90	107	109525	42.66	ng	97
22) Acetophenone	8.95	105	123788	39.44	ng	# 81
24) Nitrobenzene	9.34	77	85495	40.70	ng	# 78
25) Isophorone	9.85	82	184594	37.77	ng	94
26) 2-Nitrophenol	10.05	139	47242	41.47	ng	# 76
27) 2,4-Dimethylphenol	10.10	122	94142	45.66	ng	90
28) bis(2-Chloroethoxy)methane	10.33	93	113432	40.51	ng	92
29) 2,4-Dichlorophenol	10.59	162	89359	46.78	ng	94
30) 1,2,4-Trichlorobenzene	10.80	180	85929	40.26	ng	95
31) Naphthalene	10.99	128	285474	39.27	ng	# 94
32) Benzoic acid	10.24	122	27746	41.93	ng	# 81
33) 4-Chloroaniline	11.10	127	63140	20.89	ng	# 86
34) Hexachlorobutadiene	11.26	225	45284	39.52	ng	95
35) Caprolactam	11.86	113	14886	14.21	ng	# 41
36) 4-Chloro-3-methylphenol	12.22	107	98854	43.66	ng	# 84
37) 2-Methylnaphthalene	12.59	142	210691	39.26	ng	94
39) 1,2,4,5-Tetrachlorobenzene	12.95	216	92892	40.14	ng	98
40) Hexachlorocyclopentadiene	12.92	237	96104	81.36	ng	99


Data Path : Z:\HPCHEM1\BNA_G\DATA\BG060616\
 Data File : BG022174.D
 Acq On : 6 Jun 2016 18:23
 Operator : UM/SJ
 Sample : PB91111BS
 Misc :
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 PB91111BS

Quant Time: Jun 07 03:19:05 2016
 Quant Method : Z:\HPCHEM1\BNA_G\METHODS\8270-BG060616.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Jun 06 16:12:45 2016
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
42) 2,4,6-Trichlorophenol	13.19	196	71550	46.09	ng	96
43) 2,4,5-Trichlorophenol	13.27	196	77760	45.33	ng #	94
45) 1,1'-Biphenyl	13.58	154	283485	41.90	ng	95
46) 2-Chloronaphthalene	13.63	162	221145	42.64	ng	95
47) 2-Nitroaniline	13.84	65	51795	41.88	ng #	48
48) Acenaphthylene	14.48	152	378595	41.60	ng	97
49) Dimethylphthalate	14.20	163	300730	40.34	ng	99
50) 2,6-Dinitrotoluene	14.33	165	67885	41.89	ng #	76
51) Acenaphthene	14.81	154	224654	41.20	ng	99
52) 3-Nitroaniline	14.67	138	41499	23.09	ng #	79
53) 2,4-Dinitrophenol	14.87	184	54353	81.84	ng #	74
54) Dibenzofuran	15.15	168	352609	41.44	ng	96
55) 4-Nitrophenol	14.98	139	109383	88.18	ng #	68
56) 2,4-Dinitrotoluene	15.11	165	94570	43.51	ng #	80
57) Fluorene	15.79	166	298809	40.77	ng	95
58) 2,3,4,6-Tetrachlorophenol	15.37	232	62430	40.58	ng	90
59) Diethylphthalate	15.55	149	318998	40.06	ng	99
60) 4-Chlorophenyl-phenylether	15.78	204	137793	41.51	ng	94
61) 4-Nitroaniline	15.82	138	77262	40.53	ng #	73
62) Azobenzene	16.07	77	251495	41.71	ng	86
64) 4,6-Dinitro-2-methylphenol	15.88	198	44592	42.48	ng #	81
65) n-Nitrosodiphenylamine	15.99	169	260443	42.93	ng	99
66) 4-Bromophenyl-phenylether	16.68	248	84347	42.75	ng	92
67) Hexachlorobenzene	16.80	284	93079	40.48	ng	91
68) Atrazine	16.93	200	91189	40.61	ng	97
69) Pentachlorophenol	17.15	266	86776	82.96	ng	99
70) Phenanthrene	17.53	178	469440	41.05	ng	98
71) Anthracene	17.63	178	466872	41.16	ng	97
72) Carbazole	17.90	167	443646	41.30	ng	97
73) Di-n-butylphthalate	18.43	149	580825	41.36	ng	99
74) Fluoranthene	19.54	202	537693	40.89	ng	97
76) Benzidine	19.72	184	186086	35.12	ng	98
77) Pyrene	19.89	202	531630	42.20	ng	100
79) Butylbenzylphthalate	20.76	149	249720	42.75	ng	90
80) Benzo(a)anthracene	21.75	228	474284	41.74	ng	100
81) 3,3'-Dichlorobenzidine	21.66	252	80736	25.31	ng	95
82) Chrysene	21.82	228	460645	41.66	ng	99
83) Bis(2-ethylhexyl)phthalate	21.62	149	364729	42.46	ng	97
84) Di-n-octyl phthalate	22.86	149	610647	42.81	ng #	85
85) Indeno(1,2,3-cd)pyrene	28.83	276	537244	43.31	ng	98
87) Benzo(b)fluoranthene	24.01	252	468823	41.38	ng	98
88) Benzo(k)fluoranthene	24.08	252	483748	43.37	ng #	98
89) Benzo(a)pyrene	24.91	252	464698	42.90	ng #	95
90) Dibenzo(a,h)anthracene	28.90	278	432268	42.37	ng #	94
91) Benzo(g,h,i)perylene	30.02	276	449076	42.31	ng #	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed



#1

1,4-Dichlorobenzene-d4

Concen: 20.00 ng

RT: 8.12 min Scan# 899

Delta R.T. -0.00 min

Lab File: BG022174.D

Acq: 6 Jun 2016 18:23

Instrument :

BNA_G

ClientSampleId :

PB91111BS


Tgt Ion:152 Resp: 28607

Ion Ratio Lower Upper

152 100

150 166.3 130.1 195.1




115 60.2 62.2 93.2#



Abundance Ion 152.00 (151.70 to 152.70): E

Ion 150.00 (149.70 to 150.70): E

Ion 115.00 (114.70 to 115.70): E



#2

1,4-Dioxane

Concen: 31.25 ng

RT: 3.51 min Scan# 115

Delta R.T. -0.01 min

Lab File: BG022174.D

Acq: 6 Jun 2016 18:23


Tgt Ion: 88 Resp: 22170

Ion Ratio Lower Upper

88 100

58 60.2 36.8 55.2#



43 17.7 10.4 15.6#




Abundance Ion 88.00 (87.70 to 88.70): BGC

Ion 58.00 (57.70 to 58.70): BGC

Ion 43.00 (42.70 to 43.70): BGC





#3

Pyridine

Concen: 33.32 ng

RT: 3.93 min Scan# 185

Delta R.T. -0.00 min

Lab File: BG022174.D

Acq: 6 Jun 2016 18:23

Instrument :

BNA_G

ClientSampleId :

PB91111BS


Tgt Ion: 79 Resp: 63871

Ion Ratio Lower Upper




79 100

52 52.8 41.6 62.4

51 26.2 19.2 28.8



Abundance Ion 79.00 (78.70 to 79.70): BGC
Ion 52.00 (51.70 to 52.70): BGC
Ion 51.00 (50.70 to 51.70): BGC



#4

n-Nitrosodimethylamine

Concen: 41.62 ng

RT: 3.84 min Scan# 171

Delta R.T. -0.00 min

Lab File: BG022174.D

Acq: 6 Jun 2016 18:23


Tgt Ion: 42 Resp: 17840

Ion Ratio Lower Upper



42 100


74 233.0 184.0 276.0

44 12.2 11.0 16.6



Abundance Ion 42.00 (41.70 to 42.70): BGC
Ion 74.00 (73.70 to 74.70): BGC
Ion 44.00 (43.70 to 44.70): BGC





#5

2-Fluorophenol

Concen: 147.48 ng

RT: 5.67 min Scan# 482

Delta R.T. 0.00 min

Lab File: BG022174.D

Acq: 6 Jun 2016 18:23

Instrument :


BNA_G

ClientSampled :

PB91111BS

Tgt Ion: 112 Resp: 218204




Ion	Ratio	Lower	Upper
112	100		
64	47.4	51.1	76.7#
63	23.3	24.6	37.0#



Abundance Ion 112.00 (111.70 to 112.70): BGC

Ion 64.00 (63.70 to 64.70): BGC

Ion 63.00 (62.70 to 63.70): BGC



#6

Aniline

Concen: 29.91 ng

RT: 7.44 min Scan# 783


Delta R.T. -0.00 min

Lab File: BG022174.D

Acq: 6 Jun 2016 18:23

Tgt Ion: 93 Resp: 88395



Ion	Ratio	Lower	Upper
93	100		
66	29.6	36.3	54.5#
65	15.8	18.4	27.6#




Abundance Ion 93.00 (92.70 to 93.70): BGC

Ion 66.00 (65.70 to 66.70): BGC

Ion 65.00 (64.70 to 65.70): BGC





#7

Phenol-d6

Concen: 141.50 ng

RT: 7.28 min Scan# 756

Delta R.T. 0.00 min

Lab File: BG022174.D

Acq: 6 Jun 2016 18:23

Instrument :

BNA_G

ClientSampleId :

PB91111BS


Tgt Ion: 99 Resp: 329161

Ion Ratio Lower Upper




99 100

42 9.6 16.4 24.6#

71 24.4 25.0 37.4#



Abundance Ion 99.00 (98.70 to 99.70): BGC
Ion 42.00 (41.70 to 42.70): BGC
Ion 71.00 (70.70 to 71.70): BGC



#8

2-Chlorophenol

Concen: 45.31 ng

RT: 7.68 min Scan# 824

Delta R.T. -0.00 min

Lab File: BG022174.D

Acq: 6 Jun 2016 18:23


Tgt Ion: 128 Resp: 92656

Ion Ratio Lower Upper



128 100


130 32.0 11.2 51.2

64 35.3 39.8 79.8#



Abundance Ion 128.00 (127.70 to 128.70): BGC
Ion 130.00 (129.70 to 130.70): BGC
Ion 64.00 (63.70 to 64.70): BGC





#9

Benzaldehyde

Concen: 4.46 ng

RT: 7.27 min Scan# 754

Delta R.T. 0.01 min

Lab File: BG022174.D

Acq: 6 Jun 2016 18:23

Instrument :

BNA_G

ClientSampleId :

PB91111BS


Tgt Ion: 77 Resp: 5711

Ion Ratio Lower Upper

77 100

105 102.7 62.1 102.1#




106 86.7 65.6 105.6



Abundance Ion 77.00 (76.70 to 77.70): BGC

Ion 105.00 (104.70 to 105.70): BGC

Ion 106.00 (105.70 to 106.70): BGC



#10

Phenol

Concen: 47.33 ng

RT: 7.30 min Scan# 760

Delta R.T. -0.00 min

Lab File: BG022174.D

Acq: 6 Jun 2016 18:23


Tgt Ion: 94 Resp: 113524

Ion Ratio Lower Upper

94 100

65 21.5 9.8 49.8



66 27.8 20.8 60.8




Abundance Ion 94.00 (93.70 to 94.70): BGC

Ion 65.00 (64.70 to 65.70): BGC

Ion 66.00 (65.70 to 66.70): BGC





#11

bis(2-Chloroethyl)ether

Concen: 40.10 ng

RT: 7.53 min Scan# 799

Delta R.T. 0.00 min

Lab File: BG022174.D

Acq: 6 Jun 2016 18:23

Instrument :

BNA_G

ClientSampleId :

PB91111BS


Tgt Ion: 93 Resp: 76677

Ion Ratio Lower Upper




93 100

63 62.8 74.9 114.9#

95 31.3 8.3 48.3



Abundance Ion 93.00 (92.70 to 93.70): BGC
Ion 63.00 (62.70 to 63.70): BGC
Ion 95.00 (94.70 to 95.70): BGC



#12

1,3-Dichlorobenzene

Concen: 39.68 ng

RT: 8.01 min Scan# 880

Delta R.T. 0.00 min

Lab File: BG022174.D

Acq: 6 Jun 2016 18:23


Tgt Ion: 146 Resp: 86992

Ion Ratio Lower Upper



146 100


148 62.9 53.7 80.5

75 21.6 27.4 41.0#



Abundance Ion 146.00 (145.70 to 146.70): BGC
Ion 148.00 (147.70 to 148.70): BGC
Ion 75.00 (74.70 to 75.70): BGC









#13
 1,4-Dichlorobenzene
 Concen: 39.95 ng
 RT: 8.16 min Scan# 905
 Delta R.T. -0.00 min
 Lab File: BG022174.D
 Acq: 6 Jun 2016 18:23

Instrument :
 BNA_G
 ClientSampleId :
 PB91111BS

Tgt Ion:146 Resp: 87257
 Ion Ratio Lower Upper
 146 100
 148 64.7 46.7 70.1
 111 34.0 33.0 49.6




Abundance Ion 146.00 (145.70 to 146.70): E
 Ion 148.00 (147.70 to 148.70): E
 Ion 111.00 (110.70 to 111.70): E






#14
 1,2-Dichlorobenzene
 Concen: 39.67 ng
 RT: 8.47 min Scan# 959
 Delta R.T. -0.00 min
 Lab File: BG022174.D
 Acq: 6 Jun 2016 18:23

Tgt Ion:146 Resp: 87337
 Ion Ratio Lower Upper
 146 100
 148 63.0 52.6 79.0
 111 37.7 36.1 54.1



Abundance Ion 146.00 (145.70 to 146.70): E
 Ion 148.00 (147.70 to 148.70): E
 Ion 111.00 (110.70 to 111.70): E









#15
Benzyl Alcohol
Concen: 40.56 ng
RT: 8.36 min Scan# 940
Delta R.T. 0.00 min
Lab File: BG022174.D
Acq: 6 Jun 2016 18:23




Instrument :
BNA_G
ClientSampleId :
PB91111BS


Tgt Ion: 79 Resp: 60956
Ion Ratio Lower Upper
79 100
108 88.4 59.4 89.0
77 62.0 52.2 78.2



#16
2,2'-oxybis(1-Chloropropane)
Concen: 38.94 ng
RT: 8.65 min Scan# 989
Delta R.T. 0.00 min
Lab File: BG022174.D
Acq: 6 Jun 2016 18:23

Tgt Ion: 45 Resp: 77253
Ion Ratio Lower Upper
45 100
77 20.0 0.0 31.2
79 14.3 0.0 29.0









#17
2-Methylphenol
Concen: 43.49 ng
RT: 8.57 min Scan# 975
Delta R.T. 0.00 min
Lab File: BG022174.D
Acq: 6 Jun 2016 18:23




Instrument :
BNA_G
ClientSampleId :
PB91111BS


Tgt Ion:107 Resp: 77842
Ion Ratio Lower Upper
107 100
108 114.6 80.2 120.4
77 39.6 40.7 61.1#
79 36.5 39.8 59.8#



#18
Hexachloroethane
Concen: 39.87 ng
RT: 9.21 min Scan# 1084
Delta R.T. -0.00 min
Lab File: BG022174.D
Acq: 6 Jun 2016 18:23

Tgt Ion:117 Resp: 31800
Ion Ratio Lower Upper
117 100
119 94.0 63.9 95.9
201 95.7 61.2 91.8#





#19

n-Nitroso-di-n-propylamine

Concen: 38.25 ng

RT: 8.92 min Scan# 1035

Delta R.T. -0.00 min

Lab File: BG022174.D

Acq: 6 Jun 2016 18:23

Instrument :

BNA_G

ClientSampleId :

PB91111BS

Tgt Ion: 70 Resp: 60227





Ion Ratio Lower Upper

70 100

42 40.1 56.7 85.1#

101 10.6 8.2 12.4

130 24.0 14.0 21.0#



#20

3+4-Methylphenols

Concen: 42.66 ng

RT: 8.90 min Scan# 1031

Delta R.T. 0.00 min

Lab File: BG022174.D

Acq: 6 Jun 2016 18:23

Tgt Ion: 107 Resp: 109525




Ion Ratio Lower Upper


107 100

108 86.2 66.2 106.2

77 26.5 11.5 51.5

79 23.7 5.6 45.6









#21
Naphthalene-d8
Concen: 20.00 ng
RT: 10.94 min Scan# 1379
Delta R.T. 0.00 min
Lab File: BG022174.D
Acq: 6 Jun 2016 18:23

Instrument :
BNA_G
ClientSampled :
PB91111BS

Tgt Ion:	136	Resp:	145643
Ion	Ratio	Lower	Upper
136	100		
137	10.5	9.0	13.4
54	5.7	9.6	14.4#
68	5.8	6.4	9.6#




Abundance Ion 136.00 (135.70 to 136.70): E
Ion 137.00 (136.70 to 137.70): E
Ion 54.00 (53.70 to 54.70): BGC
Ion 68.00 (67.70 to 68.70): BGC






#22
Acetophenone
Concen: 39.44 ng
RT: 8.95 min Scan# 1040
Delta R.T. 0.00 min
Lab File: BG022174.D
Acq: 6 Jun 2016 18:23

Tgt Ion:	105	Resp:	123788
Ion	Ratio	Lower	Upper
105	100		
71	1.3	0.0	0.0#
51	17.5	25.7	38.5#
120	24.6	17.0	25.6



Abundance Ion 105.00 (104.70 to 105.70): E
Ion 71.00 (70.70 to 71.70): BGC
Ion 51.00 (50.70 to 51.70): BGC
Ion 120.00 (119.70 to 120.70): E









#23
 Nitrobenzene-d5
 Concen: 80.88 ng
 RT: 9.29 min Scan# 1098
 Delta R.T. -0.00 min
 Lab File: BG022174.D
 Acq: 6 Jun 2016 18:23

Instrument :
 BNA_G
 ClientSampleId :
 PB91111BS

Tgt Ion: 82 Resp: 168964
 Ion Ratio Lower Upper
 82 100
 128 50.5 33.4 50.0#
 54 36.7 46.1 69.1#




Abundance Ion 82.00 (81.70 to 82.70): BGC
 Ion 128.00 (127.70 to 128.70): BGC
 Ion 54.00 (53.70 to 54.70): BGC






#24
 Nitrobenzene
 Concen: 40.70 ng
 RT: 9.34 min Scan# 1106
 Delta R.T. 0.00 min
 Lab File: BG022174.D
 Acq: 6 Jun 2016 18:23

Tgt Ion: 77 Resp: 85495
 Ion Ratio Lower Upper
 77 100
 123 57.0 32.2 48.4#
 65 10.8 10.3 15.5



Abundance Ion 77.00 (76.70 to 77.70): BGC
 Ion 123.00 (122.70 to 123.70): BGC
 Ion 65.00 (64.70 to 65.70): BGC





#25

Isophorone

Concen: 37.77 ng

RT: 9.85 min Scan# 1193

Delta R.T. -0.00 min

Lab File: BG022174.D

Acq: 6 Jun 2016 18:23

Instrument :

BNA_G

ClientSampleId :

PB91111BS


Tgt Ion: 82 Resp: 184594

Ion Ratio Lower Upper




82 100

95 6.4 6.0 9.0

138 19.6 13.4 20.0



Abundance Ion 82.00 (81.70 to 82.70): BGC
Ion 95.00 (94.70 to 95.70): BGC
Ion 138.00 (137.70 to 138.70): BGC



#26

2-Nitrophenol

Concen: 41.47 ng

RT: 10.05 min Scan# 1227

Delta R.T. -0.00 min

Lab File: BG022174.D

Acq: 6 Jun 2016 18:23


Tgt Ion: 139 Resp: 47242

Ion Ratio Lower Upper



139 100


109 23.1 26.6 40.0#

65 36.5 45.3 67.9#



Abundance Ion 139.00 (138.70 to 139.70): BGC
Ion 109.00 (108.70 to 109.70): BGC
Ion 65.00 (64.70 to 65.70): BGC









#27
 2,4-Dimethylphenol
 Concen: 45.66 ng
 RT: 10.10 min Scan# 1236
 Delta R.T. -0.00 min
 Lab File: BG022174.D
 Acq: 6 Jun 2016 18:23

Instrument :
 BNA_G
 ClientSampleId :
 PB91111BS

Tgt Ion: 122 Resp: 94142
 Ion Ratio Lower Upper
 122 100
 107 105.8 97.2 145.8
 121 58.6 49.0 73.4




Abundance Ion 122.00 (121.70 to 122.70): E
 Ion 107.00 (106.70 to 107.70): E
 Ion 121.00 (120.70 to 121.70): E






#28
 bis(2-Chloroethoxy)methane
 Concen: 40.51 ng
 RT: 10.33 min Scan# 1275
 Delta R.T. -0.00 min
 Lab File: BG022174.D
 Acq: 6 Jun 2016 18:23

Tgt Ion: 93 Resp: 113432
 Ion Ratio Lower Upper
 93 100
 95 31.2 30.1 45.1
 123 13.5 10.2 15.4



Abundance Ion 93.00 (92.70 to 93.70): BGC
 Ion 95.00 (94.70 to 95.70): BGC
 Ion 123.00 (122.70 to 123.70): E









#29
2,4-Dichlorophenol
Concen: 46.78 ng
RT: 10.59 min Scan# 1319
Delta R.T. -0.00 min
Lab File: BG022174.D
Acq: 6 Jun 2016 18:23




Instrument :
BNA_G
ClientSampleId :
PB91111BS


Tgt Ion	Ratio	Lower	Upper
162	100		
164	68.4	44.9	84.9
98	34.6	19.2	59.2



#30
1,2,4-Trichlorobenzene
Concen: 40.26 ng
RT: 10.80 min Scan# 1355
Delta R.T. 0.00 min
Lab File: BG022174.D
Acq: 6 Jun 2016 18:23

Tgt Ion	Ratio	Lower	Upper
180	100		
182	96.2	82.3	123.5
145	31.1	24.4	36.6





#31

Naphthalene

Concen: 39.27 ng

RT: 10.99 min Scan# 1387

Delta R.T. -0.00 min

Lab File: BG022174.D

Acq: 6 Jun 2016 18:23

Instrument :

BNA_G

ClientSampleId :

PB91111BS


Tgt Ion:128 Resp: 285474

Ion Ratio Lower Upper

128 100

129 11.7 7.0 10.6#




127 14.2 10.1 15.1



Abundance Ion 128.00 (127.70 to 128.70): E

Ion 129.00 (128.70 to 129.70): E

Ion 127.00 (126.70 to 127.70): E



#32

Benzoic acid

Concen: 41.93 ng

RT: 10.24 min Scan# 1259

Delta R.T. 0.00 min

Lab File: BG022174.D

Acq: 6 Jun 2016 18:23


Tgt Ion:122 Resp: 27746

Ion Ratio Lower Upper

122 100

105 105.9 104.5 144.5



77 78.0 79.9 119.9#




Abundance Ion 122.00 (121.70 to 122.70): E

Ion 105.00 (104.70 to 105.70): E

Ion 77.00 (76.70 to 77.70): BG









#33
4-Chloroaniline
Concen: 20.89 ng
RT: 11.10 min Scan# 1406
Delta R.T. -0.00 min
Lab File: BG022174.D
Acq: 6 Jun 2016 18:23

Instrument :
BNA_G
ClientSampleId :
PB91111BS

Tgt Ion:127 Resp: 63140
Ion Ratio Lower Upper
127 100
129 31.3 26.6 39.8
65 19.1 27.5 41.3#
92 14.7 14.7 22.1#




Abundance Ion 127.00 (126.70 to 127.70): E
Ion 129.00 (128.70 to 129.70): E
Ion 65.00 (64.70 to 65.70): BG
Ion 92.00 (91.70 to 92.70): BG






#34
Hexachlorobutadiene
Concen: 39.52 ng
RT: 11.26 min Scan# 1434
Delta R.T. 0.00 min
Lab File: BG022174.D
Acq: 6 Jun 2016 18:23

Tgt Ion:225 Resp: 45284
Ion Ratio Lower Upper
225 100
223 65.1 52.7 79.1
227 69.1 49.7 74.5



Abundance Ion 225.00 (224.70 to 225.70): E
Ion 223.00 (222.70 to 223.70): E
Ion 227.00 (226.70 to 227.70): E





#35

Caprolactam

Concen: 14.21 ng

RT: 11.86 min Scan# 1536

Delta R.T. -0.01 min

Lab File: BG022174.D

Acq: 6 Jun 2016 18:23

Instrument :

BNA_G

ClientSampleId :

PB91111BS





Tgt Ion:113 Resp: 14886

Ion Ratio Lower Upper

113 100

55 106.4 194.2 234.2#

56 85.5 122.6 162.6#



#36

4-Chloro-3-methylphenol

Concen: 43.66 ng

RT: 12.22 min Scan# 1597

Delta R.T. 0.00 min

Lab File: BG022174.D

Acq: 6 Jun 2016 18:23




Tgt Ion:107 Resp: 98854


Ion Ratio Lower Upper

107 100

144 27.2 19.2 28.8

142 90.1 59.1 88.7#





#37

2-Methylnaphthalene

Concen: 39.26 ng

RT: 12.59 min Scan# 1659

Delta R.T. -0.00 min

Lab File: BG022174.D

Acq: 6 Jun 2016 18:23

Instrument :

BNA_G

ClientSampleId :

PB91111BS


Tgt Ion:142 Resp: 210691

Ion Ratio Lower Upper

142 100

141 94.2 71.4 107.0




115 30.0 27.4 41.2



Abundance Ion 142.00 (141.70 to 142.70): E

Ion 141.00 (140.70 to 141.70): E

Ion 115.00 (114.70 to 115.70): E



#38

Acenaphthene-d10

Concen: 20.00 ng

RT: 14.75 min Scan# 2027

Delta R.T. -0.00 min

Lab File: BG022174.D

Acq: 6 Jun 2016 18:23


Tgt Ion:164 Resp: 89904

Ion Ratio Lower Upper

164 100

162 108.0 78.0 117.0



160 48.7 32.9 49.3




Abundance Ion 164.00 (163.70 to 164.70): E

Ion 162.00 (161.70 to 162.70): E

Ion 160.00 (159.70 to 160.70): E









#39
 1,2,4,5-Tetrachlorobenzene
 Concen: 40.14 ng
 RT: 12.95 min Scan# 1721
 Delta R.T. -0.00 min
 Lab File: BG022174.D
 Acq: 6 Jun 2016 18:23

Instrument :
 BNA_G
 ClientSampleId :
 PB91111BS

Tgt Ion:	216	Resp:	92892
Ion	Ratio	Lower	Upper
216	100		
214	79.3	62.2	93.2
179	21.8	16.7	25.1
108	19.3	14.0	21.0




Abundance Ion 216.00 (215.70 to 216.70): E
 Ion 214.00 (213.70 to 214.70): E
 Ion 179.00 (178.70 to 179.70): E
 Ion 108.00 (107.70 to 108.70): E






#40
 Hexachlorocyclopentadiene
 Concen: 81.36 ng
 RT: 12.92 min Scan# 1716
 Delta R.T. -0.00 min
 Lab File: BG022174.D
 Acq: 6 Jun 2016 18:23

Tgt Ion:	237	Resp:	96104
Ion	Ratio	Lower	Upper
237	100		
235	62.4	43.2	83.2
272	12.8	0.0	33.4



Abundance Ion 236.80 (236.50 to 237.50): E
 Ion 234.90 (234.60 to 235.60): E
 Ion 271.75 (271.45 to 272.45): E









#41
 2,4,6-Tribromophenol
 Concen: 127.77 ng
 RT: 16.23 min Scan# 2280
 Delta R.T. -0.00 min
 Lab File: BG022174.D
 Acq: 6 Jun 2016 18:23

Instrument :
 BNA_G
 ClientSampleId :
 PB91111BS

Tgt Ion:330 Resp: 129794
 Ion Ratio Lower Upper
 330 100
 332 95.1 78.0 117.0
 141 37.3 33.8 50.8




Abundance Ion 329.80 (329.50 to 330.50): E
 Ion 331.80 (331.50 to 332.50): E
 Ion 141.00 (140.70 to 141.70): E






#42
 2,4,6-Trichlorophenol
 Concen: 46.09 ng
 RT: 13.19 min Scan# 1762
 Delta R.T. -0.00 min
 Lab File: BG022174.D
 Acq: 6 Jun 2016 18:23

Tgt Ion:196 Resp: 71550
 Ion Ratio Lower Upper
 196 100
 198 94.3 80.1 120.1
 200 30.1 24.1 36.1



Abundance Ion 196.00 (195.70 to 196.70): E
 Ion 198.00 (197.70 to 198.70): E
 Ion 200.00 (199.70 to 200.70): E









#43
 2,4,5-Trichlorophenol
 Concen: 45.33 ng
 RT: 13.27 min Scan# 1775
 Delta R.T. -0.00 min
 Lab File: BG022174.D
 Acq: 6 Jun 2016 18:23




Instrument :
 BNA_G
 ClientSampleId :
 PB91111BS


Tgt Ion:196 Resp: 77760
 Ion Ratio Lower Upper
 196 100
 198 96.7 78.0 117.0
 97 42.9 43.0 64.6#
 132 26.0 23.0 34.4



#44
 2-Fluorobiphenyl
 Concen: 80.99 ng
 RT: 13.37 min Scan# 1793
 Delta R.T. 0.00 min
 Lab File: BG022174.D
 Acq: 6 Jun 2016 18:23

Tgt Ion:172 Resp: 477765
 Ion Ratio Lower Upper
 172 100
 171 39.1 27.8 41.6
 170 25.7 19.6 29.4









#45
 1,1'-Biphenyl
 Concen: 41.90 ng
 RT: 13.58 min Scan# 1829
 Delta R.T. 0.00 min
 Lab File: BG022174.D
 Acq: 6 Jun 2016 18:23

Instrument :
 BNA_G
 ClientSampleId :
 PB91111BS

Tgt Ion:154 Resp: 283485
 Ion Ratio Lower Upper
 154 100
 153 43.1 19.7 59.7
 76 12.1 0.0 33.0




Abundance Ion 154.00 (153.70 to 154.70): E
 Ion 153.00 (152.70 to 153.70): E
 Ion 76.00 (75.70 to 76.70): BGC






#46
 2-Chloronaphthalene
 Concen: 42.64 ng
 RT: 13.63 min Scan# 1837
 Delta R.T. -0.00 min
 Lab File: BG022174.D
 Acq: 6 Jun 2016 18:23

Tgt Ion:162 Resp: 221145
 Ion Ratio Lower Upper
 162 100
 127 35.4 32.0 48.0
 164 32.3 27.0 40.4



Abundance Ion 162.00 (161.70 to 162.70): E
 Ion 127.00 (126.70 to 127.70): E
 Ion 164.00 (163.70 to 164.70): E





#47

2-Nitroaniline

Concen: 41.88 ng

RT: 13.84 min Scan# 1872

Delta R.T. -0.00 min

Lab File: BG022174.D

Acq: 6 Jun 2016 18:23

Instrument :

BNA_G

ClientSampleId :

PB91111BS





Tgt Ion: 65 Resp: 51795

Ion Ratio Lower Upper

65 100

92 86.7 49.2 73.8#

138 155.8 75.0 112.6#



#48

Acenaphthylene

Concen: 41.60 ng

RT: 14.48 min Scan# 1981

Delta R.T. 0.00 min

Lab File: BG022174.D

Acq: 6 Jun 2016 18:23




Tgt Ion: 152 Resp: 378595


Ion Ratio Lower Upper

152 100

151 22.0 16.6 24.8

153 13.5 10.2 15.2





#49

Dimethylphthalate

Concen: 40.34 ng

RT: 14.20 min Scan# 1933

Delta R.T. -0.00 min

Lab File: BG022174.D

Acq: 6 Jun 2016 18:23





Instrument :

BNA_G

ClientSampleId :

PB91111BS

Tgt Ion:	163	Resp:	300730
Ion Ratio	Lower	Upper	
163	100		
194	4.3	3.0	4.6
164	10.3	8.1	12.1



#50

2,6-Dinitrotoluene

Concen: 41.89 ng




RT: 14.33 min Scan# 1955


Delta R.T. 0.00 min

Lab File: BG022174.D

Acq: 6 Jun 2016 18:23

Tgt Ion:	165	Resp:	67885
Ion Ratio	Lower	Upper	
165	100		
63	33.3	48.2	72.4#
89	48.4	45.3	67.9





#51

Acenaphthene

Concen: 41.20 ng

RT: 14.81 min Scan# 2038

Delta R.T. -0.00 min

Lab File: BG022174.D

Acq: 6 Jun 2016 18:23

Instrument :

BNA_G

ClientSampleId :

PB91111BS


Tgt Ion:154 Resp: 224654

Ion Ratio Lower Upper

154 100

153 114.3 91.9 137.9




152 53.8 42.0 63.0



Abundance Ion 154.00 (153.70 to 154.70): E

Ion 153.00 (152.70 to 153.70): E

Ion 152.00 (151.70 to 152.70): E



#52

3-Nitroaniline

Concen: 23.09 ng

RT: 14.67 min Scan# 2013

Delta R.T. 0.00 min

Lab File: BG022174.D

Acq: 6 Jun 2016 18:23


Tgt Ion:138 Resp: 41499

Ion Ratio Lower Upper

138 100

108 7.8 9.1 13.7#



92 102.5 102.6 154.0#




Abundance Ion 138.00 (137.70 to 138.70): E

Ion 108.00 (107.70 to 108.70): E

Ion 92.00 (91.70 to 92.70): BGC









#53
 2,4-Dinitrophenol
 Concen: 81.84 ng
 RT: 14.87 min Scan# 2048
 Delta R.T. -0.00 min
 Lab File: BG022174.D
 Acq: 6 Jun 2016 18:23




Instrument :
 BNA_G
 ClientSampleId :
 PB91111BS


Tgt Ion:184 Resp: 54353
 Ion Ratio Lower Upper
 184 100
 63 42.1 57.3 85.9#
 154 55.8 55.2 82.8



#54
 Dibenzofuran
 Concen: 41.44 ng
 RT: 15.15 min Scan# 2095
 Delta R.T. 0.00 min
 Lab File: BG022174.D
 Acq: 6 Jun 2016 18:23

Tgt Ion:168 Resp: 352609
 Ion Ratio Lower Upper
 168 100
 139 35.4 31.1 46.7
 169 13.8 11.0 16.6









#55
4-Nitrophenol
Concen: 88.18 ng
RT: 14.98 min Scan# 2066
Delta R.T. -0.00 min
Lab File: BG022174.D
Acq: 6 Jun 2016 18:23




Instrument :
BNA_G
ClientSampleId :
PB91111BS


Tgt Ion:139 Resp: 109383
Ion Ratio Lower Upper
139 100
109 52.7 49.2 89.2
65 63.4 83.2 123.2#



#56
2,4-Dinitrotoluene
Concen: 43.51 ng
RT: 15.11 min Scan# 2089
Delta R.T. -0.00 min
Lab File: BG022174.D
Acq: 6 Jun 2016 18:23

Tgt Ion:165 Resp: 94570
Ion Ratio Lower Upper
165 100
63 29.0 37.4 56.0#
89 60.6 59.4 89.0
182 3.0 2.1 3.1









#57
 Fluorene
 Concen: 40.77 ng
 RT: 15.79 min Scan# 2205
 Delta R.T. -0.00 min
 Lab File: BG022174.D
 Acq: 6 Jun 2016 18:23




Instrument :
 BNA_G
 ClientSampleId :
 PB91111BS


Tgt Ion:166 Resp: 298809
 Ion Ratio Lower Upper
 166 100
 165 104.0 79.0 118.4
 167 13.5 10.7 16.1



#58
 2,3,4,6-Tetrachlorophenol
 Concen: 40.58 ng
 RT: 15.37 min Scan# 2133
 Delta R.T. -0.00 min
 Lab File: BG022174.D
 Acq: 6 Jun 2016 18:23

Tgt Ion:232 Resp: 62430
 Ion Ratio Lower Upper
 232 100
 131 52.5 36.3 54.5
 130 2.7 1.9 2.9
 166 36.3 25.0 37.4









#59
Diethylphthalate
Concen: 40.06 ng
RT: 15.55 min Scan# 2163
Delta R.T. -0.00 min
Lab File: BG022174.D
Acq: 6 Jun 2016 18:23




Instrument :
BNA_G
ClientSampleId :
PB91111BS


Tgt Ion:149 Resp: 318998
Ion Ratio Lower Upper
149 100
177 21.5 17.3 25.9
150 13.6 9.7 14.5



#60
4-Chlorophenyl-phenylether
Concen: 41.51 ng
RT: 15.78 min Scan# 2202
Delta R.T. -0.00 min
Lab File: BG022174.D
Acq: 6 Jun 2016 18:23

Tgt Ion:204 Resp: 137793
Ion Ratio Lower Upper
204 100
206 32.9 27.8 41.8
141 57.6 51.0 76.4









#61
4-Nitroaniline
Concen: 40.53 ng
RT: 15.82 min Scan# 2210
Delta R.T. -0.00 min
Lab File: BG022174.D
Acq: 6 Jun 2016 18:23




Instrument :
BNA_G
ClientSampleId :
PB91111BS


Tgt Ion:138 Resp: 77262
Ion Ratio Lower Upper
138 100
92 42.4 37.7 77.7
108 51.8 59.4 99.4#



#62
Azobenzene
Concen: 41.71 ng
RT: 16.07 min Scan# 2252
Delta R.T. -0.00 min
Lab File: BG022174.D
Acq: 6 Jun 2016 18:23

Tgt Ion: 77 Resp: 251495
Ion Ratio Lower Upper
77 100
182 28.8 4.1 44.1
105 21.8 0.0 39.8
51 18.6 11.1 51.1





#63

Phenanthrene-d10

Concen: 20.00 ng

RT: 17.49 min Scan# 2494

Delta R.T. -0.00 min

Lab File: BG022174.D

Acq: 6 Jun 2016 18:23

Instrument :

BNA_G

ClientSampleId :

PB91111BS


Tgt Ion:188 Resp: 210576

Ion Ratio Lower Upper

188 100

94 10.6 7.5 11.3




80 9.9 8.6 13.0



Abundance Ion 188.00 (187.70 to 188.70): E

Ion 94.00 (93.70 to 94.70): BGC

Ion 80.00 (79.70 to 80.70): BGC



#64

4,6-Dinitro-2-methylphenol

Concen: 42.48 ng

RT: 15.88 min Scan# 2219

Delta R.T. -0.00 min

Lab File: BG022174.D

Acq: 6 Jun 2016 18:23


Tgt Ion:198 Resp: 44592

Ion Ratio Lower Upper

198 100

51 25.7 29.2 69.2#



105 42.2 24.0 64.0




Abundance Ion 198.00 (197.70 to 198.70): E

Ion 51.00 (50.70 to 51.70): BGC

Ion 105.00 (104.70 to 105.70): E









#65
n-Nitrosodiphenylamine
Concen: 42.93 ng
RT: 15.99 min Scan# 2239
Delta R.T. -0.00 min
Lab File: BG022174.D
Acq: 6 Jun 2016 18:23




Instrument :
BNA_G
ClientSampleId :
PB91111BS


Tgt Ion:169 Resp: 260443
Ion Ratio Lower Upper
169 100
168 73.3 59.6 89.4
167 38.5 31.2 46.8



#66
4-Bromophenyl-phenylether
Concen: 42.75 ng
RT: 16.68 min Scan# 2355
Delta R.T. -0.00 min
Lab File: BG022174.D
Acq: 6 Jun 2016 18:23

Tgt Ion:248 Resp: 84347
Ion Ratio Lower Upper
248 100
250 103.1 76.6 114.8
141 72.1 63.8 95.6





#67

Hexachlorobenzene

Concen: 40.48 ng

RT: 16.80 min Scan# 2376

Delta R.T. 0.00 min

Lab File: BG022174.D

Acq: 6 Jun 2016 18:23

Instrument :

BNA_G

ClientSampleId :

PB91111BS


Tgt Ion:284 Resp: 93079

Ion Ratio Lower Upper

284 100

142 32.8 31.8 47.6




249 30.2 27.3 40.9



Abundance Ion 283.80 (283.50 to 284.50): E

Ion 142.00 (141.70 to 142.70): E

Ion 249.00 (248.70 to 249.70): E



#68

Atrazine

Concen: 40.61 ng

RT: 16.93 min Scan# 2399

Delta R.T. -0.00 min

Lab File: BG022174.D

Acq: 6 Jun 2016 18:23


Tgt Ion:200 Resp: 91189

Ion Ratio Lower Upper

200 100

173 26.3 5.1 45.1



215 45.8 28.0 68.0




Abundance Ion 200.00 (199.70 to 200.70): E

Ion 173.00 (172.70 to 173.70): E

Ion 215.00 (214.70 to 215.70): E





#69

Pentachlorophenol

Concen: 82.96 ng

RT: 17.15 min Scan# 2435

Delta R.T. -0.00 min

Lab File: BG022174.D

Acq: 6 Jun 2016 18:23

Instrument :

BNA_G

ClientSampleId :

PB91111BS


Tgt Ion:266 Resp: 86776

Ion Ratio Lower Upper

266 100

268 69.5 54.5 81.7




264 67.0 53.8 80.6



Abundance Ion 265.70 (265.40 to 266.40): E

Ion 268.00 (267.70 to 268.70): E

Ion 264.00 (263.70 to 264.70): E



#70

Phenanthrene

Concen: 41.05 ng

RT: 17.53 min Scan# 2501

Delta R.T. -0.00 min

Lab File: BG022174.D

Acq: 6 Jun 2016 18:23


Tgt Ion:178 Resp: 469440

Ion Ratio Lower Upper

178 100

176 20.7 16.2 24.4



179 16.0 12.0 18.0




Abundance Ion 178.00 (177.70 to 178.70): E

Ion 176.00 (175.70 to 176.70): E

Ion 179.00 (178.70 to 179.70): E









#71
 Anthracene
 Concen: 41.16 ng
 RT: 17.63 min Scan# 2517
 Delta R.T. -0.00 min
 Lab File: BG022174.D
 Acq: 6 Jun 2016 18:23




Instrument :
 BNA_G
 ClientSampleId :
 PB91111BS


Tgt Ion:178 Resp: 466872
 Ion Ratio Lower Upper
 178 100
 176 20.2 15.0 22.4
 179 16.1 12.2 18.4



#72
 Carbazole
 Concen: 41.30 ng
 RT: 17.90 min Scan# 2563
 Delta R.T. -0.00 min
 Lab File: BG022174.D
 Acq: 6 Jun 2016 18:23

Tgt Ion:167 Resp: 443646
 Ion Ratio Lower Upper
 167 100
 166 25.0 18.6 27.8
 139 12.6 10.6 15.8









#73
Di-n-butylphthalate
Concen: 41.36 ng
RT: 18.43 min Scan# 2654
Delta R.T. -0.00 min
Lab File: BG022174.D
Acq: 6 Jun 2016 18:23




Instrument :
BNA_G
ClientSampleId :
PB91111BS


Tgt Ion:149 Resp: 580825
Ion Ratio Lower Upper
149 100
150 9.7 7.4 11.0
104 4.5 4.0 6.0



#74
Fluoranthene
Concen: 40.89 ng
RT: 19.54 min Scan# 2842
Delta R.T. -0.00 min
Lab File: BG022174.D
Acq: 6 Jun 2016 18:23

Tgt Ion:202 Resp: 537693
Ion Ratio Lower Upper
202 100
101 13.1 0.0 31.0
203 18.5 0.0 37.9





#75

Chrysene-d12

Concen: 20.00 ng

RT: 21.77 min Scan# 3222

Delta R.T. -0.00 min

Lab File: BG022174.D

Acq: 6 Jun 2016 18:23

Instrument :

BNA_G

ClientSampleId :

PB91111BS


Tgt Ion:240 Resp: 202626

Ion Ratio Lower Upper

240 100

120 12.6 8.4 12.6




236 24.8 19.6 29.4



Abundance Ion 240.00 (239.70 to 240.70): E

Ion 120.00 (119.70 to 120.70): E

Ion 236.00 (235.70 to 236.70): E



#76

Benzidine

Concen: 35.12 ng

RT: 19.72 min Scan# 2873

Delta R.T. 0.00 min

Lab File: BG022174.D

Acq: 6 Jun 2016 18:23


Tgt Ion:184 Resp: 186086

Ion Ratio Lower Upper

184 100

185 15.6 11.5 17.3



183 12.7 10.2 15.4




Abundance Ion 184.00 (183.70 to 184.70): E

Ion 185.00 (184.70 to 185.70): E

Ion 183.00 (182.70 to 183.70): E









#77
 Pyrene
 Concen: 42.20 ng
 RT: 19.89 min Scan# 2903
 Delta R.T. -0.00 min
 Lab File: BG022174.D
 Acq: 6 Jun 2016 18:23




Instrument :
 BNA_G
 ClientSampleId :
 PB91111BS


Tgt Ion: 202 Resp: 531630
 Ion Ratio Lower Upper
 202 100
 200 22.8 18.2 27.2
 203 18.9 15.0 22.4



#78
 Terphenyl-d14
 Concen: 82.03 ng
 RT: 20.08 min Scan# 2935
 Delta R.T. -0.00 min
 Lab File: BG022174.D
 Acq: 6 Jun 2016 18:23

Tgt Ion: 244 Resp: 700082
 Ion Ratio Lower Upper
 244 100
 212 7.7 6.5 9.7
 122 13.4 8.6 12.8#









#79
Butylbenzylphthalate
Concen: 42.75 ng
RT: 20.76 min Scan# 3051
Delta R.T. -0.00 min
Lab File: BG022174.D
Acq: 6 Jun 2016 18:23




Instrument :
BNA_G
ClientSampleId :
PB91111BS


Tgt Ion	Resp	Lower	Upper
149	249720		
91	62.2	57.8	86.8
206	19.2	17.0	25.4



#80
Benzo(a)anthracene
Concen: 41.74 ng
RT: 21.75 min Scan# 3218
Delta R.T. -0.00 min
Lab File: BG022174.D
Acq: 6 Jun 2016 18:23

Tgt Ion	Resp	Lower	Upper
228	474284		
226	28.0	22.6	34.0
229	20.1	16.1	24.1









#81
 3,3'-Dichlorobenzidine
 Concen: 25.31 ng
 RT: 21.66 min Scan# 3203
 Delta R.T. -0.00 min
 Lab File: BG022174.D
 Acq: 6 Jun 2016 18:23




Instrument :
 BNA_G
 ClientSampleId :
 PB91111BS


Tgt Ion:252 Resp: 80736
 Ion Ratio Lower Upper
 252 100
 254 60.9 51.8 77.6
 126 12.2 8.8 13.2



#82
 Chrysene
 Concen: 41.66 ng
 RT: 21.82 min Scan# 3230
 Delta R.T. -0.00 min
 Lab File: BG022174.D
 Acq: 6 Jun 2016 18:23

Tgt Ion:228 Resp: 460645
 Ion Ratio Lower Upper
 228 100
 226 31.0 25.4 38.0
 229 20.6 15.9 23.9









#83
 Bis(2-ethylhexyl)phthalate
 Concen: 42.46 ng
 RT: 21.62 min Scan# 3197
 Delta R.T. -0.00 min
 Lab File: BG022174.D
 Acq: 6 Jun 2016 18:23




Instrument :
 BNA_G
 ClientSampleId :
 PB91111BS


Tgt Ion:149 Resp: 364729
 Ion Ratio Lower Upper
 149 100
 167 26.1 22.4 33.6
 279 3.9 3.2 4.8



#84
 Di-n-octyl phthalate
 Concen: 42.81 ng
 RT: 22.86 min Scan# 3407
 Delta R.T. 0.00 min
 Lab File: BG022174.D
 Acq: 6 Jun 2016 18:23

Tgt Ion:149 Resp: 610647
 Ion Ratio Lower Upper
 149 100
 167 1.6 1.3 1.9
 43 5.7 9.8 14.6#









#85
 Indeno(1,2,3-cd)pyrene
 Concen: 43.31 ng
 RT: 28.83 min Scan# 4424
 Delta R.T. -0.02 min
 Lab File: BG022174.D
 Acq: 6 Jun 2016 18:23




Instrument :
 BNA_G
 ClientSampled :
 PB91111BS


Tgt Ion: 276 Resp: 537244
 Ion Ratio Lower Upper
 276 100
 138 16.0 14.0 21.0
 277 25.6 20.6 30.8



#86
 Perylene-d12
 Concen: 20.00 ng
 RT: 25.06 min Scan# 3782
 Delta R.T. -0.00 min
 Lab File: BG022174.D
 Acq: 6 Jun 2016 18:23

Tgt Ion: 264 Resp: 194264
 Ion Ratio Lower Upper
 264 100
 260 22.9 20.2 30.4
 265 22.3 17.8 26.8









#87
Benzo(b)fluoranthene
Concen: 41.38 ng
RT: 24.01 min Scan# 3604
Delta R.T. 0.00 min
Lab File: BG022174.D
Acq: 6 Jun 2016 18:23

Instrument :
BNA_G
ClientSampleId :
PB91111BS

Tgt Ion:252 Resp: 468823
Ion Ratio Lower Upper
252 100
253 21.9 17.5 26.3
125 12.8 8.6 13.0




Abundance Ion 252.00 (251.70 to 252.70): E
Ion 253.00 (252.70 to 253.70): E
Ion 125.00 (124.70 to 125.70): E






#88
Benzo(k)fluoranthene
Concen: 43.37 ng
RT: 24.08 min Scan# 3615
Delta R.T. -0.00 min
Lab File: BG022174.D
Acq: 6 Jun 2016 18:23

Tgt Ion:252 Resp: 483748
Ion Ratio Lower Upper
252 100
253 22.0 17.8 26.8
125 12.6 8.3 12.5#



Abundance Ion 252.00 (251.70 to 252.70): E
Ion 253.00 (252.70 to 253.70): E
Ion 125.00 (124.70 to 125.70): E





#89

Benzo(a)pyrene

Concen: 42.90 ng

RT: 24.91 min Scan# 3756

Delta R.T. -0.00 min

Lab File: BG022174.D

Acq: 6 Jun 2016 18:23

Instrument :


BNA_G

ClientSampleId :

PB91111BS

Tgt Ion:252 Resp: 464698




Ion	Ratio	Lower	Upper
252	100		
253	23.1	17.1	25.7
125	14.5	9.6	14.4#



Abundance Ion 252.00 (251.70 to 252.70): E

Ion 253.00 (252.70 to 253.70): E

Ion 125.00 (124.70 to 125.70): E



#90

Dibenzo(a,h)anthracene

Concen: 42.37 ng

RT: 28.90 min Scan# 4435


Delta R.T. -0.01 min

Lab File: BG022174.D

Acq: 6 Jun 2016 18:23

Tgt Ion:278 Resp: 432268



Ion	Ratio	Lower	Upper
278	100		
139	18.9	11.9	17.9#
279	24.2	18.2	27.2




Abundance Ion 278.00 (277.70 to 278.70): E

Ion 139.00 (138.70 to 139.70): E

Ion 279.00 (278.70 to 279.70): E





#91

Benzo(g,h,i)perylene

Concen: 42.31 ng

RT: 30.02 min Scan# 4627

Delta R.T. -0.00 min

Lab File: BG022174.D

Acq: 6 Jun 2016 18:23

Instrument :

BNA_G

ClientSampleId :

PB91111BS


Tgt Ion: 276 Resp: 449076

Ion Ratio Lower Upper

276 100

277 24.0 19.4 29.2

138 26.9 17.4 26.0#



Abundance Ion 276.00 (275.70 to 276.70): E

Ion 277.00 (276.70 to 277.70): E

Ion 138.00 (137.70 to 138.70): E

