

Data Path : Z:\HPCHEM1\BNA_G\DATA\BG060616\
 Data File : BG022179.D
 Acq On : 6 Jun 2016 21:46
 Operator : UM/SJ
 Sample : H3404-01
 Misc :
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 B1-19-20

Quant Time: Jun 07 03:22:01 2016
 Quant Method : Z:\HPCHEM1\BNA_G\METHODS\8270-BG060616.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Jun 06 16:12:45 2016
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	8.12	152	27195	20.00	ng	0.00
21) Naphthalene-d8	10.94	136	118309	20.00	ng	0.00
38) Acenaphthene-d10	14.74	164	75764	20.00	ng	0.00
63) Phenanthrene-d10	17.49	188	181113	20.00	ng	0.00
75) Chrysene-d12	21.77	240	195253	20.00	ng	0.00
86) Perylene-d12	25.06	264	187270	20.00	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.67	112	168449	119.76	ng	0.00
7) Phenol-d6	7.28	99	237811	107.54	ng	0.00
23) Nitrobenzene-d5	9.29	82	117887	69.47	ng	0.00
41) 2,4,6-Tribromophenol	16.24	330	89807	104.90	ng	0.00
44) 2-Fluorobiphenyl	13.37	172	347105	69.82	ng	0.00
78) Terphenyl-d14	20.09	244	597033	72.59	ng	0.00
Target Compounds						
49) Dimethylphthalate	14.19	163	40653	6.47	ng	Qvalue 99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA_G\DATA\BG060616\
Data File : BG022179.D
Acq On : 6 Jun 2016 21:46
Operator : UM/SJ
Sample : H3404-01
Misc :
ALS Vial : 16 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
B1-19-20

Quant Time: Jun 07 03:22:01 2016
Quant Method : Z:\HPCHEM1\BNA_G\METHODS\8270-BG060616.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Jun 06 16:12:45 2016
Response via : Initial Calibration

