


Data Path : Z:\HPCHEM1\BNA\_G\Data\BG060616\  
 Data File : BG022171.D  
 Acq On : 6 Jun 2016 16:22  
 Operator : UM/SJ  
 Sample : DFTPP  
 Misc :  
 ALS Vial : 1 Sample Multiplier: 1

Instrument :  
 BNA\_G  
 ClientSampleId :  
 DFTPP

Integration File: rteint.p

Method : Z:\HPCHEM1\BNA\_G\METHODS\8270-BG060616.M  
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
 Last Update : Mon Jun 06 16:12:45 2016



AutoFind: Scans 2563, 2564, 2565; Background Corrected with Scan 2556

| Target Mass | Rel. to Mass | Lower Limit% | Upper Limit% | Rel. Abn% | Raw Abn | Result Pass/Fail |
|-------------|--------------|--------------|--------------|-----------|---------|------------------|
| 51          | 198          | 10           | 80           | 24.0      | 6174    | PASS             |
| 68          | 69           | 0.00         | 2            | 1.6       | 127     | PASS             |
| 69          | 198          | 0.00         | 100          | 31.5      | 8079    | PASS             |
| 70          | 69           | 0.00         | 2            | 0.0       | 0       | PASS             |
| 127         | 198          | 10           | 80           | 50.7      | 13031   | PASS             |
| 197         | 198          | 0.00         | 2            | 0.0       | 0       | PASS             |
| 198         | 198          | 100          | 100          | 100.0     | 25682   | PASS             |
| 199         | 198          | 5            | 9            | 7.6       | 1943    | PASS             |
| 275         | 198          | 10           | 60           | 27.0      | 6929    | PASS             |
| 365         | 198          | 1            | 100          | 3.4       | 867     | PASS             |
| 441         | 442          | 0.01         | 24           | 17.4      | 3813    | PASS             |
| 442         | 198          | 50           | 100          | 85.5      | 21957   | PASS             |
| 443         | 442          | 15           | 24           | 19.6      | 4295    | PASS             |