

Data Path : Z:\HPCHEM1\BNA G\DATA\BG060716\  
 Data File : BG022222.D  
 Acq On : 8 Jun 2016 2:47  
 Operator : UM/SJ  
 Sample : H3424-05  
 Misc :  
 ALS Vial : 16 Sample Multiplier: 1

Instrument :  
 BNA\_G  
 ClientSampleId :  
 CTM-5(7.5-10)

Quant Time: Jun 08 03:49:50 2016  
 Quant Method : Z:\HPCHEM1\BNA G\METHODS\8270-BG060616.M  
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
 QLast Update : Mon Jun 06 16:12:45 2016  
 Response via : Initial Calibration

| Internal Standards          | R.T.  | QIon | Response | Conc   | Units | Dev(Min)       |
|-----------------------------|-------|------|----------|--------|-------|----------------|
| 1) 1,4-Dichlorobenzene-d4   | 8.12  | 152  | 22043    | 20.00  | ng    | 0.00           |
| 21) Naphthalene-d8          | 10.94 | 136  | 107860   | 20.00  | ng    | 0.00           |
| 38) Acenaphthene-d10        | 14.75 | 164  | 72475    | 20.00  | ng    | 0.00           |
| 63) Phenanthrene-d10        | 17.49 | 188  | 171913   | 20.00  | ng    | 0.00           |
| 75) Chrysene-d12            | 21.76 | 240  | 169395   | 20.00  | ng    | 0.00           |
| 86) Perylene-d12            | 25.06 | 264  | 159772   | 20.00  | ng    | 0.00           |
| System Monitoring Compounds |       |      |          |        |       |                |
| 5) 2-Fluorophenol           | 5.66  | 112  | 153897   | 134.99 | ng    | 0.00           |
| 7) Phenol-d6                | 7.27  | 99   | 229351   | 127.95 | ng    | 0.00           |
| 23) Nitrobenzene-d5         | 9.29  | 82   | 116551   | 75.34  | ng    | 0.00           |
| 41) 2,4,6-Tribromophenol    | 16.23 | 330  | 89247    | 108.98 | ng    | 0.00           |
| 44) 2-Fluorobiphenyl        | 13.37 | 172  | 340276   | 71.55  | ng    | 0.00           |
| 78) Terphenyl-d14           | 20.08 | 244  | 551881   | 77.35  | ng    | 0.00           |
| Target Compounds            |       |      |          |        |       |                |
| 49) Dimethylphthalate       | 14.19 | 163  | 37697    | 6.27   | ng    | Qvalue<br># 99 |

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA G\DATA\BG060716\  
Data File : BG022222.D  
Acq On : 8 Jun 2016 2:47  
Operator : UM/SJ  
Sample : H3424-05  
Misc :  
ALS Vial : 16 Sample Multiplier: 1

Instrument :  
BNA\_G  
ClientSampleId :  
CTM-5(7.5-10)

Quant Time: Jun 08 03:49:50 2016  
Quant Method : Z:\HPCHEM1\BNA G\METHODS\8270-BG060616.M  
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
QLast Update : Mon Jun 06 16:12:45 2016  
Response via : Initial Calibration

