


Data Path : Z:\HPCHEM1\BNA G\Data\BG060717\
 Data File : BG027286.D
 Acq On : 7 Jun 2017 16:41
 Operator : SJ/MA
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 DFTPP

Integration File: rteint.p

Method : Z:\HPCHEM1\BNA_G\METHODS\8270-BG060517.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Mon Jun 05 16:00:35 2017



AutoFind: Scans 2484, 2485, 2486; Background Corrected with Scan 2477

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	36.8	18443	PASS
68	69	0.00	2	0.0	0	PASS
69	198	0.00	100	35.3	17678	PASS
70	69	0.00	2	0.0	0	PASS
127	198	10	80	45.1	22581	PASS
197	198	0.00	2	0.0	0	PASS
198	198	100	100	100.0	50088	PASS
199	198	5	9	7.3	3681	PASS
275	198	10	60	26.7	13366	PASS
365	198	1	100	3.4	1694	PASS
441	442	0.01	24	15.6	6904	PASS
442	198	50	100	88.2	44160	PASS
443	442	15	24	20.1	8896	PASS