

Data Path : Z:\SVOASRV\HPCHEM1\BNA_G\DATA\BG060920\
 Data File : BG045655.D
 Acq On : 9 Jun 2020 9:02
 Operator : CG/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_G
 LabSampleId :
 SSTDCCC040

Quant Time: Jun 09 10:47:35 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_G\METHODS\8270-BG051520.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Jun 01 15:07:00 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	89	0.00
2	1,4-Dioxane	0.507	0.531	-4.7	93	0.00
3	Pyridine	1.413	1.582	-12.0	95	0.00
4	n-Nitrosodimethylamine	0.462	0.522	-13.0	99	0.00
5 S	2-Fluorophenol	1.023	1.152	-12.6	97	0.00
6	Aniline	1.901	2.168	-14.0	100	0.00
7 S	Phenol-d6	1.457	1.689	-15.9	100	0.00
8	2-Chlorophenol	1.122	1.259	-12.2	98	-0.01
9	Benzaldehyde	0.949	0.931	1.9	83	0.00
10 C	Phenol	1.501	1.751	-16.7	100	-0.01
11	bis(2-Chloroethyl)ether	1.171	1.313	-12.1	96	0.00
12	1,3-Dichlorobenzene	1.349	1.477	-9.5	96	0.00
13 C	1,4-Dichlorobenzene	1.331	1.496	-12.4	98	0.00
14	1,2-Dichlorobenzene	1.280	1.415	-10.5	98	0.00
15	Benzyl Alcohol	1.203	1.439	-19.6	104	0.00
16	2,2'-oxybis(1-Chloropropane	1.111	1.294	-16.5	103	0.00
17	2-Methylphenol	1.124	1.310	-16.5	100	-0.01
18	Hexachloroethane	0.510	0.558	-9.4	95	0.00
19 P	n-Nitroso-di-n-propylamine	1.007	1.210	-20.2	104	0.00
20	3+4-Methylphenols	1.530	1.825	-19.3	103	-0.01
21 I	Naphthalene-d8	1.000	1.000	0.0	91	0.00
22	Acetophenone	0.474	0.526	-11.0	101	0.00
23 S	Nitrobenzene-d5	0.367	0.417	-13.6	103	0.00
24	Nitrobenzene	0.393	0.445	-13.2	105	0.00
25	Isophorone	0.722	0.835	-15.7	104	0.00
26 C	2-Nitrophenol	0.168	0.192	-14.3	104	-0.01
27	2,4-Dimethylphenol	0.249	0.288	-15.7	104	0.00
28	bis(2-Chloroethoxy)methane	0.379	0.426	-12.4	103	0.00
29 C	2,4-Dichlorophenol	0.294	0.336	-14.3	105	0.00
30	1,2,4-Trichlorobenzene	0.348	0.394	-13.2	104	0.00
31	Naphthalene	0.932	1.047	-12.3	102	0.00
32	Benzoic acid	0.150	0.167	-11.3	108	0.00
33	4-Chloroaniline	0.390	0.455	-16.7	104	0.00
34 C	Hexachlorobutadiene	0.246	0.274	-11.4	102	0.00
35	Caprolactam	0.108	0.119	-10.2	97	0.00
36 C	4-Chloro-3-methylphenol	0.332	0.393	-18.4	107	-0.01
37	2-Methylnaphthalene	0.695	0.796	-14.5	104	-0.01
38	1-Methylnaphthalene	0.656	0.753	-14.8	103	0.00
39 I	Acenaphthene-d10	1.000	1.000	0.0	97	0.00
40	1,2,4,5-Tetrachlorobenzene	0.559	0.612	-9.5	107	0.00
41 P	Hexachlorocyclopentadiene	0.322	0.331	-2.8	99	0.00
42 S	2,4,6-Tribromophenol	0.256	0.275	-7.4	101	0.00
43 C	2,4,6-Trichlorophenol	0.388	0.426	-9.8	105	0.00
44	2,4,5-Trichlorophenol	0.443	0.468	-5.6	100	0.00

Data Path : Z:\SVOASRV\HPCHEM1\BNA_G\DATA\BG060920\
 Data File : BG045655.D
 Acq On : 9 Jun 2020 9:02
 Operator : CG/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_G
 LabSampleId :
 SSTDCCC040

Quant Time: Jun 09 10:47:35 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_G\METHODS\8270-BG051520.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Jun 01 15:07:00 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
45 S	2-Fluorobiphenyl	1.179	1.318	-11.8	105	0.00
46	1,1'-Biphenyl	1.229	1.374	-11.8	107	0.00
47	2-Chloronaphthalene	0.998	1.101	-10.3	107	0.00
48	2-Nitroaniline	0.328	0.376	-14.6	106	0.00
49	Acenaphthylene	1.603	1.796	-12.0	107	0.00
50	Dimethylphthalate	1.372	1.522	-10.9	103	0.00
51	2,6-Dinitrotoluene	0.310	0.340	-9.7	102	0.00
52 C	Acenaphthene	1.073	1.064	0.8	93	0.00
53	3-Nitroaniline	0.315	0.344	-9.2	104	0.00
54 P	2,4-Dinitrophenol	0.180	0.181	-0.6	92	0.00
55	Dibenzofuran	1.593	1.763	-10.7	104	0.00
56 P	4-Nitrophenol	0.238	0.225	5.5	86	0.00
57	2,4-Dinitrotoluene	0.448	0.491	-9.6	104	0.00
58	Fluorene	1.309	1.456	-11.2	105	0.00
59	2,3,4,6-Tetrachlorophenol	0.390	0.411	-5.4	99	-0.01
60	Diethylphthalate	1.428	1.579	-10.6	103	0.00
61	4-Chlorophenyl-phenylether	0.749	0.842	-12.4	106	0.00
62	4-Nitroaniline	0.329	0.348	-5.8	98	0.00
63	Azobenzene	1.332	1.498	-12.5	106	0.00
64 I	Phenanthrene-d10	1.000	1.000	0.0	92	0.00
65	4,6-Dinitro-2-methylphenol	0.116	0.126	-8.6	96	0.00
66 c	n-Nitrosodiphenylamine	0.480	0.534	-11.3	103	0.00
67	4-Bromophenyl-phenylether	0.217	0.240	-10.6	102	0.00
68	Hexachlorobenzene	0.227	0.255	-12.3	103	0.00
69	Atrazine	0.198	0.196	1.0	90	0.00
70 C	Pentachlorophenol	0.118	0.108	8.5	82	0.00
71	Phenanthrene	0.873	0.957	-9.6	99	0.00
72	Anthracene	0.877	0.965	-10.0	100	0.00
73	Carbazole	0.855	0.917	-7.3	96	0.00
74	Di-n-butylphthalate	1.026	1.104	-7.6	95	0.00
75 C	Fluoranthene	1.111	1.164	-4.8	93	0.00
76 I	Chrysene-d12	1.000	1.000	0.0	81	0.00
77	Benzidine	0.562	0.513	8.7	71	0.00
78	Pyrene	1.140	1.328	-16.5	92	0.00
79 S	Terphenyl-d14	0.858	0.994	-15.9	91	0.00
80	Butylbenzylphthalate	0.492	0.542	-10.2	87	0.00
81	Benzo(a)anthracene	1.114	1.245	-11.8	90	0.00
82	3,3'-Dichlorobenzidine	0.437	0.477	-9.2	87	0.00
83	Chrysene	1.071	1.191	-11.2	88	0.00
84	Bis(2-ethylhexyl)phthalate	0.676	0.760	-12.4	90	0.00
85 c	Di-n-octyl phthalate	1.162	1.288	-10.8	89	0.00
86	Indeno(1,2,3-cd)pyrene	1.401	1.503	-7.3	88	0.00
87 I	Perylene-d12	1.000	1.000	0.0	81	0.00

Data Path : Z:\SVOASRV\HPCHEM1\BNA_G\DATA\BG060920\
Data File : BG045655.D
Acq On : 9 Jun 2020 9:02
Operator : CG/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_G
LabSampleId :
SSTDCCC040

Quant Time: Jun 09 10:47:35 2020
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_G\METHODS\8270-BG051520.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Jun 01 15:07:00 2020
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
88	Benzo(b)fluoranthene	1.073	1.196	-11.5	89	0.00
89	Benzo(k)fluoranthene	1.008	1.150	-14.1	92	0.00
90 C	Benzo(a)pyrene	0.999	1.125	-12.6	91	0.00
91	Dibenzo(a,h)anthracene	1.004	1.120	-11.6	91	-0.02
92	Benzo(g,h,i)perylene	1.004	1.096	-9.2	89	-0.02

(#) = Out of Range

SPCC's out = 0 CCC's out = 0