

Method Path : Z:\HPCHEM1\BNA_G\METHODS\
 Method File : 625-TERPINEOL-BG040915.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Thu Jun 11 04:12:20 2015
 Response Via : Initial Calibration

Calibration Files

2.5 =BG017601.D 10 =BG017602.D 25 =BG017603.D 40 =BG017604.D 50 =BG017605.D 60 =BG017606.D 80 =BG017607.D

	Compound	2.5	10	25	40	50	60	80	Avg	%RSD
1) I	1,4-Dichlorobenzen...	-----ISTD-----								
2) S	2-Fluorophenol	0.981	0.963	1.027	1.030	0.970	1.053	1.041	1.009	3.65
3) S	Phenol-d6	1.101	1.210	1.290	1.302	1.320	1.337	1.282	1.263	6.50
4) I	Naphthalene-d8	-----ISTD-----								
5) S	Nitrobenzene-d5	0.334	0.340	0.379	0.368	0.368	0.379	0.379	0.364	5.28
6)	Terpineol	0.035	0.036	0.035	0.037	0.036	0.036	0.036	0.036	2.00
7) I	Acenaphthene-d10	-----ISTD-----								
8) S	2,4,6-Tribromo...	0.185	0.240	0.231	0.258	0.256	0.264	0.264	0.243	11.65
9) S	2-Fluorobiphenyl	1.071	1.131	1.075	1.089	1.141	1.136	1.161	1.115	3.20
10) I	Phenanthrene-d10	-----ISTD-----								
11) I	Chrysene-d12	-----ISTD-----								
12) S	Terphenyl-d14	0.946	0.904	0.972	1.016	1.019	1.031	1.077	0.995	5.83
13) I	Perylene-d12	-----ISTD-----								

(#) = Out of Range