

Data Path : Z:\HPCHEM1\BNA_G\DATA\BG061015\
Data File : BG017608.D
Acq On : 10 Jun 2015 19:35
Operator : TP/IZ
Sample : SSTDICV040
Misc :
ALS Vial : 9 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
ICVBG061015

Quant Time: Jun 11 04:20:19 2015
Quant Method : Z:\HPCHEM1\BNA_G\METHODS\625-TERPINEOL-BG040915.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Jun 11 04:12:20 2015
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	20.000	20.000	0.0	102	0.00
2 S	2-Fluorophenol	80.000	80.563	-0.7	100	0.00
3 S	Phenol-d6	80.000	83.804	-4.8	103	0.00
4 I	Naphthalene-d8	20.000	20.000	0.0	106	0.00
5 S	Nitrobenzene-d5	80.000	83.059	-3.8	108	0.00
6	Terpineol	40.000	41.886	-4.7	112	0.00
7 I	Acenaphthene-d10	20.000	20.000	0.0	106	0.00
8 S	2,4,6-Tribromophenol	80.000	75.285	5.9	94	0.00
9 S	2-Fluorobiphenyl	80.000	80.389	-0.5	110	0.00
10 I	Phenanthrene-d10	20.000	20.000	0.0	104	0.00
11 I	Chrysene-d12	20.000	20.000	0.0	96	0.00
12 S	Terphenyl-d14	80.000	82.201	-2.8	97	0.00
13 I	Perylene-d12	20.000	20.000	0.0	97	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0