

Data Path : Z:\SVOASRV\HPCHEM1\BNA G\DATA\BG061020\
 Data File : BG045717.D
 Acq On : 11 Jun 2020 10:46
 Operator : CG/JU
 Sample : L2965-06
 Misc :
 ALS Vial : 30 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 BASMENT-GRAY-WALL

Quant Time: Jun 11 11:46:49 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA G\METHODS\8270-BG051520.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Jun 01 15:07:00 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.95	152	66326	20.00	ng	0.00
21) Naphthalene-d8	10.75	136	250366	20.00	ng	0.00
39) Acenaphthene-d10	14.59	164	164360	20.00	ng	0.00
64) Phenanthrene-d10	17.34	188	332689	20.00	ng	0.00
76) Chrysene-d12	21.59	240	304330	20.00	ng	0.00
87) Perylene-d12	24.73	264	350560	20.00	ng	-0.02
System Monitoring Compounds						
5) 2-Fluorophenol	5.53	112	474976	139.95	ng	0.00
7) Phenol-d6	7.14	99	705631	146.08	ng	0.00
23) Nitrobenzene-d5	9.11	82	472875	103.00	ng	0.00
42) 2,4,6-Tribromophenol	16.08	330	292656	139.23	ng	0.00
45) 2-Fluorobiphenyl	13.21	172	987513	101.91	ng	0.00
79) Terphenyl-d14	19.96	244	1301373	99.73	ng	0.00
Target Compounds						
35) Caprolactam	11.67	113	10732	7.933	ng	94
50) Dimethylphthalate	14.04	163	62421	5.536	ng	98
60) Diethylphthalate	15.40	149	53107	4.525	ng	99
74) Di-n-butylphthalate	18.30	149	203278	11.913	ng	98
80) Butylbenzylphthalate	20.64	149	71141	9.501	ng	94
84) Bis(2-ethylhexyl)phthalate	21.49	149	131816	12.806	ng	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA G\DATA\BG061020\
Data File : BG045717.D
Acq On : 11 Jun 2020 10:46
Operator : CG/JU
Sample : L2965-06
Misc :
ALS Vial : 30 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
BASMENT-GRAY-WALL

Quant Time: Jun 11 11:46:49 2020
Quant Method : Z:\SVOASRV\HPCHEM1\BNA G\METHODS\8270-BG051520.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Jun 01 15:07:00 2020
Response via : Initial Calibration

